
Einthoven's Triangle *Prodigy Anesthesia*

The placement of the leads for a three-lead electrocardiogram can be envisioned through the diagram known as Einthoven's triangle. The lines in the triangle represent the placement of the positive and negative electrodes for each of the bipolar limb leads. For example, to create Lead I, the negative electrode is placed on the right arm and the positive electrode is placed on the left arm (the left leg becomes the neutral lead). Likewise, for lead II, the left leg becomes the positive lead, the right arm is negative, and the left arm is the neutral lead. For lead III, the left leg is positive, the left arm is negative, and the right arm is neutral.



Limb leads

Kossick MA. EKG Interpretation: Simple-Thorough-Practical. 2nd ed. Parkridge, IL: AANA Publishing; 2000.

Copyright 2015 Prodigy Anesthesia, Inc. It is unlawful to copy, distribute, and/or sell this material.