

CSMHuman00000000019

Acoustic Interaction with the Human Eye-Orbit System: A Comprehensive Monograph on Biomechanics, Resonance Theories, and Diagnostic Applications

1. Introduction: The Intersection of Acoustics and Ophthalmic Diagnostics

The interaction between acoustic waves and biological tissues constitutes a fundamental area of study in biomedical engineering, offering non-invasive windows into the mechanical properties of human organs. While high-frequency ultrasound (megahertz range) has long been established for imaging, the low-frequency acoustic spectrum (20–800 Hz) presents a distinct and largely underexplored modality for characterizing the biomechanical state of soft tissues. This monograph provides an exhaustive technical analysis of the frequency-dependent interactions within the human eye-orbit system, specifically focusing on the biophysical principles enabling the measurement of Intraocular Pressure (IOP) through acoustic resonance.

Glaucoma, a group of neurodegenerative diseases characterized by optic neuropathy and often associated with elevated IOP, remains a leading cause of irreversible blindness globally.¹ The management of this pathology is intrinsically linked to the accurate, frequent, and safe measurement of IOP. Traditional tonometry, relying on static force application (Goldmann applanation) or pneumatic deformation (air-puff tonometry), essentially treats the eye as a passive object to be flattened. These methods, while clinically validated, suffer from significant logistical limitations: the requirement for anesthesia, the risk of corneal abrasion, and the bulky, non-portable nature of the instrumentation.¹

In contrast, acoustic tonometry re-imagines the eye as a dynamic, resonant system. By treating the eyeball as a fluid-filled, prestressed spherical shell, investigators can exploit the fundamental physical relationship between internal pressure, membrane tension, and resonant frequency.² When excited by acoustic waves, the eye oscillates at specific natural frequencies that shift in direct proportion to changes in IOP. This approach shifts the diagnostic paradigm from static deformation to dynamic spectroscopy, offering a pathway to

continuous, non-contact monitoring using consumer-grade electronics.

This report synthesizes data from biophysical modeling, ex vivo porcine experiments, and in vivo simulations to elucidate the complex acoustic environment of the eye orbit. It explores the "acoustic prison" effect—where orbital bones trap and amplify sound waves—and draws rigorous analogies to geological "basin effects" observed in seismology. Furthermore, it details the mathematical frameworks of Grinfeld's soap bubble theory and Mason's stretched plate theory, which underpin the current understanding of ocular resonance.¹

2. Theoretical Frameworks of Ocular Resonance

To accurately interpret the acoustic response of the eye, one must move beyond simple mass-spring-damper approximations and employ continuum mechanics models that account for the fluid-structure interaction inherent in the eyeball. The human eye is not a solid sphere; it is a composite structure comprising a tough, viscoelastic outer shell (the cornea and sclera) enclosing incompressible fluids (the aqueous and vitreous humors).

2.1 Limitations of Simple Harmonic Models

Early attempts to model the eye often utilized lumped-parameter models where the eye was represented as a single mass connected to a spring. While useful for rough approximations, these models fail to capture the complex scattering of acoustic waves and the existence of multiple resonant modes.¹ The acoustic wavelengths in the low-frequency range (20–800 Hz) are significantly larger than the physical dimensions of the eyeball. For instance, at 100 Hz in soft tissue ($c \approx 1540$ m/s), the wavelength is approximately 15.4 meters, dwarfing the ~24 mm diameter of the eye.¹ This scale discrepancy necessitates a modeling approach that treats the eye as a localized resonator within a larger acoustic field, governed by partial differential equations rather than ordinary differential equations.

2.2 Grinfeld's Soap Bubble Vibration Theory

The theoretical cornerstone for the "fluid-filled and prestressed spherical model" of the eye is derived from Grinfeld's analysis of soap bubble oscillations.¹ A soap bubble represents a canonical system where a thin liquid film is maintained in equilibrium by surface tension (γ) balancing the pressure difference (ΔP) across the interface.

In this model, the dynamics are governed by the Navier-Stokes equations for an incompressible, inviscid fluid. The fundamental governing equations are:

$$\rho \frac{D\vec{u}}{Dt} = -\nabla P + \mu \nabla^2 \vec{u} + \rho \vec{g}$$

$$\nabla \cdot \vec{u} = 0$$

Where $\vec{u}$ is the velocity field, P is pressure, ρ is the fluid density, and

$\vec{g}$ represents gravity.³ For the specific case of the interface (the bubble wall), the kinematic boundary condition requires that the velocity of the fluid at the interface equals the time derivative of the interface position ($\frac{dS}{dt}$):

$$\frac{dS}{dt} = \vec{u}_{\text{interface}}$$

Grinfeld's theory posits that for a spherical membrane under tension, the restorative force driving oscillation is primarily the surface tension (or in the eye's case, the pretension caused by IOP). This model is particularly adept at predicting the lower-order modes of vibration where the shell acts primarily as a tensioned membrane rather than a bending plate.⁷

2.3 Mason's Stretched Plate Vibration Theory

While the soap bubble analogy effectively captures the pressure-tension relationship, biological tissues like the cornea and sclera possess inherent bending stiffness that a soap film lacks. To account for this, researchers integrate W.P. Mason's Stretched Plate Vibration Theory.¹

Originally developed for electromechanical transducers and piezoelectric crystals, Mason's theory describes the vibration of plates that are subject to both external tension and internal elastic stiffness.⁸ By combining this with Grinfeld's fluid-dynamics approach, a composite "Fluid-Filled and Prestressed Spherical Model" is derived.² This hybrid model acknowledges that the eye's resonant frequency (f_r) is a function of both the "hard spring" of the corneal material stiffness and the "tunable spring" of the intraocular pressure.

2.4 The Sensitive Frequency Range

Theoretical derivations and empirical validations indicate that while the eyeball may exhibit resonant modes across a broad spectrum (18 to 800 Hz), the range most sensitive to physiological IOP changes is **20–140 Hz**.¹ Within this bandwidth, the relationship between frequency shift and pressure change is nearly linear, providing a robust basis for diagnostic calibration. Higher modes (above 400 Hz) tend to be dominated by the local material properties of the cornea rather than the global IOP, rendering them less useful for tonometry.²

3. Biophysics of the Eye-Orbit Acoustic System

The measurement of ocular resonance is complicated by the fact that the eye does not exist in a vacuum. It is embedded in the orbit—a complex anatomical cavity comprising the bony socket, extraocular muscles, orbital fat, and the optic nerve. This environment profoundly influences wave propagation, creating a unique acoustic landscape characterized by impedance mismatches and wave trapping.

3.1 Acoustic Impedance Mismatch

The transmission of acoustic energy between different biological materials is governed by their characteristic acoustic impedance (Z), defined as the product of density (ρ) and the speed of sound (c):

$$Z = \rho \cdot c$$

The eye orbit presents a dramatic impedance mismatch between the orbital bones and the soft tissues contained within them.

- **Bone Properties:** The skull and orbital walls are rigid structures with a high speed of sound, approximately **4080 m/s**¹⁰, and high density.
- **Soft Tissue Properties:** The orbital fat, muscle, and the eyeball itself (humors) have sound speeds close to that of water, ranging from **1450 m/s to 1650 m/s**.⁵ Specifically, the vitreous humor and saline-based components have speeds around 1480–1540 m/s.¹⁰

When an acoustic wave traveling through the soft tissue of the orbit encounters the bony wall, the large difference in impedance ($Z_{\text{bone}} \gg Z_{\text{tissue}}$) results in a high reflection coefficient. The majority of the acoustic energy is reflected back into the soft tissue rather than passing into the bone. This phenomenon is critical to the functionality of acoustic tonometry, as it confines the signal energy to the region of interest.¹

3.2 The "Acoustic Prison" Effect

Researchers describe the orbital cavity as an "acoustic prison".¹ This term refers to the confinement of acoustic waves within the orbital soft tissues due to the stiff boundaries of the skull.

1. **Entry:** Acoustic waves generated by a transducer on the occipital bone propagate through the cranium (acting as a low-loss waveguide due to its stiffness) and enter the orbit from the posterior aspect.
2. **Trapping:** Upon refracting into the orbital fat and muscle, the waves are effectively trapped. They bounce between the orbital walls and the eyeball, unable to easily re-enter the high-impedance bone.
3. **Exit:** The only "weak point" in this prison is the anterior surface of the cornea, which interfaces with air. Although the impedance mismatch between cornea and air is also massive, the cornea represents a free surface that can oscillate and emit acoustic pressure waves into the external environment, where they can be detected.¹

This "prison" effect is advantageous because it amplifies the signal intensity within the orbit, ensuring that even low-power acoustic triggers (approx. 1 Watt) can generate detectable vibrations in the cornea.¹

3.3 Analogies to Seismology: The "Basin Effect"

To model the acoustic behavior of the orbit, biophysicists draw a direct analogy to the "basin

effects" observed in earthquake seismology.¹

- **Geological Parallel:** In a sedimentary basin (e.g., a valley filled with soft soil), seismic waves originating from the bedrock are trapped in the soft sedimentary layer. The basin geometry focuses the energy, and the impedance contrast between bedrock and soil prevents energy dissipation. This leads to amplification of ground motion and prolongation of shaking duration, as seen in the 2015 Gorkha earthquake.¹
- **Ocular Application:** In this analogy, the orbital bones function as the "bedrock," and the orbital fat/muscle complex functions as the "sedimentary layer." The acoustic waves are the "seismic waves."
- **Implications:**
 - **Amplification:** The "basin" geometry of the orbit amplifies specific frequencies. Experimental data indicates that input frequencies in the **150–160 Hz** range can be amplified by a factor of up to **5.3**.¹ This amplification must be accounted for to distinguish true corneal resonance from orbital cavity resonance.
 - **Interference:** The multiple reflections within the "basin" create constructive and destructive interference patterns. This necessitates the use of complex signal processing (e.g., FFT and differential amplitude analysis) to isolate the IOP-dependent signal.¹

3.4 Viscoelastic Damping and Dissipation

While the "acoustic prison" traps energy, the internal structures of the eye actively dissipate it. The vitreous humor is not a perfect inviscid fluid; it possesses viscosity that introduces damping. Similarly, the lens and iris structures absorb acoustic energy.

- **Damping Coefficients:** In mass-spring modeling, the cornea is assigned a damping coefficient (ζ_c) and the eyeball a global damping coefficient (ζ_e). These parameters are crucial for determining the width (Q-factor) of the resonant peak. A highly damped system would produce a broad, flat peak, making precise frequency determination difficult. Fortunately, the damping in the 20–140 Hz range is sufficiently low to allow for distinct peak detection.¹
- **Orbit Effect on Damping:** The surrounding orbital fat provides external damping and support. This "cushioning" effect raises the effective stiffness of the system, shifting resonant frequencies higher compared to an enucleated (removed) eyeball.¹

4. Experimental Characterization and Instrumentation

The validation of these theoretical models has been achieved through rigorous experimentation using porcine eyes, simulated orbits, and advanced electro-acoustic instrumentation.

4.1 Biological Models: Porcine Eyes and Gel Phantoms

Porcine eyeballs are the standard surrogate for human ocular research due to their similar size, mechanical properties, and corneal thickness.¹

- **Ex Vivo Isolation:** Initial baseline tests are performed on isolated eyeballs placed on a rigid substrate. This eliminates orbital interference but fails to mimic clinical conditions.
- **Simulated In Vivo Environment:** To replicate the human orbit, researchers embed porcine eyes in a model skull filled with **ultrasonic conductive gel**.
 - **Gel Composition:** The gel is composed of carbomer polymer, propylene glycol, glycerol, and methylparaben.¹
 - **Acoustic Properties:** This formulation yields a speed of sound between **1460 and 1650 m/s**, closely matching human orbital fat and muscle.⁵ This ensures that the refractive and reflective behaviors at the bone/tissue interface mimic the human condition accurately.
 - **Mechanical Role:** The gel occupies approximately 2/3 of the orbital volume, providing mechanical support that stiffens the posterior eyeball, a critical factor in reproducing the *in vivo* resonant frequency shift.¹

4.2 Triggering Mechanisms: Optimizing the Source

The location of the acoustic trigger is a critical variable in signal quality. Various sites have been evaluated:

- **Forehead/Temporal Bone:** These sites are accessible but suffer from high attenuation and complex interference patterns due to the thickness of the frontal bone and the presence of sinus cavities.¹
- **Eyelid:** Direct application to the eyelid provides strong coupling but alters the mass-loading on the eye and dampens the corneal vibration, potentially distorting the measurement.¹²
- **Occipital Bone (The Standard):** The optimal placement for acoustic tonometry is the occipital bone (back of the head). This location utilizes the skull as a rigid transmission line. The waves travel anteriorly, interacting with the orbit from behind. This setup minimizes direct interference near the detector (which is in front of the eye) and ensures that the excitation energy is uniformly distributed into the "acoustic prison" of the orbit.¹

The standard triggering source is a low-power acoustic speaker (8 Ω , 0.5 W to 1.125 W) driven by a sweep signal (0–400 Hz) with an amplitude of -3 to +3 V.¹

4.3 Detection Systems: The Non-Contact "Dish"

To measure the minute vibrations of the cornea without physical contact (which would pose safety risks and damping artifacts), a specialized **dish-microphone receiver** is employed.

- **Geometry:** The dish is designed with a curvature radius slightly smaller than that of the cornea. When positioned approximately 5 mm from the eye, it creates a semi-enclosed cavity.

- **Function:** This cavity acts as a secondary "acoustic prison" or resonator. Acoustic energy emitted by the cornea is reflected back and forth between the dish and the eye. The shape of the dish focuses these standing waves into a central aperture.¹
- **Helmholtz Chamber:** Behind the central aperture lies a small cavity chamber leading to the microphone's piezoelectric film. This chamber further shapes the acoustic pressure wave, optimizing the impedance matching between the air gap and the sensor. This multi-stage amplification allows for the detection of high-resolution signals despite the weak emission from the corneal surface.

5. Analysis of Resonant Frequency Data

The core diagnostic output of acoustic tonometry is the shift in resonant frequency (f_r) as a function of Intraocular Pressure (IOP).

5.1 Frequency-Pressure Correlation

Extensive testing on porcine models has established a linear positive correlation between IOP and resonant frequency within the physiological range (20–60 mbar).¹

Table 1: Comparative Sensitivity of Contact vs. Non-Contact Acoustic Tonometry ¹

Parameter	Contact Method	Non-Contact Method (Dish)
IOP Range	20 – 60 mbar	20 – 60 mbar
Frequency Range (f_r)	58 Hz $\rightarrow$ 72 Hz	67 Hz $\rightarrow$ 81 Hz
Total Frequency Shift	~14 Hz	~14 Hz
Sensitivity	2.05 Hz per 10 mbar	2.42 Hz per 10 mbar
Individual Variation	10 – 17 Hz shift per eye	7 – 26 Hz shift per eye

Analysis of Data:

- **Baseline Shift:** The non-contact method shows a higher baseline resonant frequency (67 Hz vs. 58 Hz at 20 mbar). This is attributed to the absence of mass-loading. In contact methods, the physical probe adds effective mass to the oscillating system, lowering the natural frequency ($\omega = \sqrt{k/m}$). By removing this mass, the non-contact method measures the "true" unencumbered resonance of the eye.

- **Sensitivity:** The non-contact method exhibits higher sensitivity (2.42 Hz/10mbar). This increased responsiveness suggests that the free-surface oscillation of the cornea is more strictly governed by internal pressure tension than the damped oscillation in contact modes.
- **Standard Deviation:** The spread in individual eyeball data (7–26 Hz shift) highlights the influence of biological variability (e.g., corneal thickness, scleral rigidity). This necessitates calibration or population-based algorithms for clinical devices.¹

5.2 The Effect of Orbital Embedment

The presence of the orbit (simulated by gel) significantly alters the mechanical response compared to an isolated eyeball.

- **Stiffening Effect:** An isolated eyeball is mechanically "floppy." When embedded in the orbit, the surrounding tissues provide a viscoelastic foundation. Experiments show that the resonant frequency of a gel-embedded eye is consistently higher (**75–100 Hz**) compared to an isolated eye (**60–80 Hz**).¹
- **Implication for Diagnostics:** Clinical algorithms must be calibrated to the *in vivo* frequency range (75–100 Hz) rather than the *ex vivo* range. Devices calibrated on isolated eyes would consistently underestimate the IOP in human patients due to the orbit-induced frequency boost.

6. Signal Processing and Interference Management

Extracting the pure corneal resonance from the noisy acoustic environment of the skull requires sophisticated signal processing.

6.1 Spectral Analysis and FFT

The raw signal from the microphone is a time-domain waveform containing a superposition of the excitation sweep, orbital echoes, and corneal emissions. Fast Fourier Transform (FFT) is applied to convert this into a power spectrum.¹⁷

- **Peak Detection:** The algorithm identifies the frequency with the maximum power density. In the range of 20–140 Hz, this peak corresponds to the fundamental mode of the eye's fluid-shell oscillation.
- **Sweep Resolution:** A high-resolution sweep (e.g., 0.5 Hz steps) is necessary to accurately capture the peak, as the total shift over the entire physiological pressure range may be only 12–15 Hz.¹

6.2 Differential Amplitude Analysis

In scenarios where the "basin effect" of the orbit creates strong standing waves that mask the resonant peak, an alternative detection method is used: **Differential Amplitude Analysis**.

- **Mechanism:** Instead of tracking the peak frequency shift directly, the system measures

the *amplitude difference* between two fixed frequencies (e.g., 135 Hz and 165 Hz).¹

- **Rationale:** As the resonant curve shifts to the right (higher frequency) with increasing IOP, the amplitude at the lower fixed frequency (135 Hz) drops, while the amplitude at the higher fixed frequency (165 Hz) rises (or changes slope). The differential signal provides a monotonic proxy for IOP even when the absolute resonant peak is obscured by orbital noise.

7. Safety, Dosimetry, and Regulatory Context

The introduction of acoustic energy into the eye must adhere to strict safety standards to prevent mechanical or thermal damage to delicate ocular tissues, particularly the retina and lens.

7.1 Acoustic Dosimetry vs. Ultrasound

It is critical to distinguish acoustic tonometry from diagnostic ultrasound.

- **Diagnostic Ultrasound:** Operates at 2–18 MHz. Safety is governed by the Mechanical Index (MI) (risk of cavitation) and Thermal Index (TI) (risk of heating). High intensities can cause cataract formation or retinal burns.¹⁴
- **Acoustic Tonometry:** Operates at < 400 Hz. This is six orders of magnitude lower in frequency.
 - **Cavitation:** At 400 Hz, the cavitation threshold is virtually impossible to reach with the low power output used (approx. 1 Watt).
 - **Thermal Effects:** The absorption coefficient of tissue at 100 Hz is negligible compared to MHz frequencies, meaning there is effectively zero heating of the vitreous or lens.⁹

7.2 Comparison to Human Vocalization

The safety profile of the device is best understood by comparison to natural physiological processes. The frequencies used (20–400 Hz) and the power levels delivered to the skull (via the occipital bone) are comparable to those generated by the human vocal cords during speech.¹

- **Self-Vibration:** When a person speaks, the skull vibrates, and the eyes are subjected to internal acoustic pressure. Since speech is a benign activity, the artificial replication of these vibrations for milliseconds during a measurement is considered inherently safe.¹

7.3 Regulatory Standards (OSHA, ANSI, FDA)

- **OSHA 1910.133:** This standard mandates protection against "flying particles, molten metal, liquid chemicals... or potentially injurious light radiation".¹⁸ Acoustic waves do not fall under these hazard categories. The primary acoustic hazard recognized by OSHA is noise-induced hearing loss (Standard 1910.95), which applies to high-decibel continuous

exposure, not the low-volume, short-duration sweeps used in tonometry.

- **ANSI Z136.1 (Laser Safety):** Current Laser Doppler Vibrometers (LDV) used for research utilize Class 2 lasers. While safe for momentary exposure, they trigger the blink reflex and pose a theoretical risk if the beam is stared at fixedly.¹⁹ Acoustic tonometry eliminates this optical hazard entirely, making it suitable for pediatric or unconscious patients where cooperation is limited.
- **FDA Guidelines for Ophthalmic Ultrasound:** The FDA sets a limit of $MI < 0.23$ for ophthalmic ultrasound.¹⁴ While acoustic tonometry does not use ultrasound, it operates well below the energy density thresholds implied by these limits. Animal studies exposing eyes to acoustic radiation force (ARF) significantly higher than FDA limits showed no damage, further supporting the safety of the much lower-energy tonometer.¹⁴

7.4 Biological Interaction: Retinal Stimulation

Recent research has explored the use of ultrasound to stimulate retinal neurons in blind patients, restoring some visual perception.²⁰ This suggests that the retina is mechanically sensitive. However, the frequencies required for neural stimulation are in the megahertz range, focused into tight beams to create localized pressure. The diffuse, low-frequency (Hz) waves of the tonometer do not possess the focal mechanics required to depolarize neurons, ensuring that the measurement process does not induce visual artifacts (phosphenes) or neural activation.

8. Comparative Technology Assessment

To contextualize the value of acoustic tonometry, it must be weighed against existing clinical standards.

8.1 Versus Air-Puff Tonometry

Air-puff (non-contact) tonometry is the current screening standard.

- **Mechanism:** Uses a high-velocity jet of air to flatten the cornea. The time taken to achieve applanation is correlated to IOP.
- **Drawbacks:**
 - **Startle Reflex:** The air puff is uncomfortable and causes patients to flinch, leading to measurement errors.
 - **Size:** Requires a large plenum chamber and compressor, limiting the device to a table-top form factor.¹
 - **Bio-Aerosols:** The air jet can create aerosols from the tear film, theoretically spreading pathogens (a concern highlighted during the COVID-19 pandemic).
- **Acoustic Advantage:** Acoustic tonometry uses sound waves, which are invisible and unfelt (except as a gentle vibration). It requires no air compressor, enabling drastic miniaturization. The absence of an air jet eliminates the aerosol risk.

8.2 Versus Goldmann Applanation Tonometry (GAT)

GAT is the "gold standard" for accuracy.

- **Mechanism:** A prism is physically pressed against the anesthetized cornea.
- **Drawbacks:** Requires anesthetic drops, fluorescein dye, and a slit lamp. It is purely a clinical procedure, impossible for home use.
- **Acoustic Advantage:** Entirely non-invasive, no anesthesia required, and suitable for home monitoring.

9. Future Directions and Technological Synthesis

The establishment of the eye as a predictable resonant system opens the door to a new class of medical devices.

9.1 Wearable Continuous Monitoring

The removal of the air pump allows the acoustic tonometer to be miniaturized into the frame of a pair of glasses or a smartphone attachment.¹ A "smart glass" system could employ a bone-conduction transducer (in the temple tip resting on the mastoid or occipital bone) and a MEMS microphone (integrated into the rim near the eye).

- **Clinical Impact:** This would allow for the measurement of Diurnal IOP Fluctuations—the spikes in pressure that occur at night or early morning, which are currently missed during office visits and are major risk factors for glaucoma progression.

9.2 Telemedicine Integration

With the ability to couple acoustic triggers and receivers to smartphones, IOP data could be uploaded instantly to cloud platforms for remote monitoring by ophthalmologists. This "Internet of Medical Things" (IoMT) approach would revolutionize the management of chronic glaucoma, shifting from reactive treatment to proactive pressure management based on real-time data trends.

10. Conclusion

The comprehensive analysis of the eye-orbit system reveals that the human eye is not merely a passive optical organ but a complex, fluid-filled acoustic resonator. The interaction of low-frequency sound (20–140 Hz) with the eye is governed by the intricate interplay of fluid dynamics (Grinfeld's soap bubble theory), membrane mechanics (Mason's plate theory), and orbital acoustics (Basin effects).

The "acoustic prison" of the orbit, while initially presenting a challenge of interference, ultimately serves as a natural amplifier that enables the detection of minute corneal vibrations. Through the use of non-contact "dish" receivers and occipital bone triggering,

acoustic tonometry successfully extracts the resonant frequency of the eye, providing a high-sensitivity (2.42 Hz/10mbar) surrogate for Intraocular Pressure.

This technology offers a profound advance over pneumatic and mechanical tonometers. It is safer, more comfortable, and inherently scalable to wearable form factors. By bridging the gap between rigorous biophysical theory and consumer electronics, acoustic tonometry stands poised to redefine the standard of care for glaucoma, transforming the human skull and eye into a sophisticated, self-reporting diagnostic system.

Works cited

1. Eye orbit effects on eyeball resonant frequencies and acoustic tonometer measurements - PMC - PubMed Central, accessed December 23, 2025, <https://pmc.ncbi.nlm.nih.gov/articles/PMC8941096/>
2. Resonance frequency of fluid-filled and prestressed spherical shell—A model of the human eyeball | Request PDF - ResearchGate, accessed December 23, 2025, https://www.researchgate.net/publication/301248372_Resonance_frequency_of_fluid-filled_and_prestressed_spherical_shell-A_model_of_the_human_eyeball
3. Governing Equations For Simulations Of Soap Bubbles - Applied Mathematics, accessed December 23, 2025, https://appliedmath.ucmerced.edu/sites/g/files/ufvvyh381/f/documents/PHDdocs/ivarro_2010.pdf
4. Mason's network model for the piezoelectric resonator in the thickness mode., accessed December 23, 2025, https://www.researchgate.net/figure/Masons-network-model-for-the-piezoelectric-resonator-in-the-thickness-mode_fig7_2233955
5. 17.3: Speed of Sound - Physics LibreTexts, accessed December 23, 2025, [https://phys.libretexts.org/Bookshelves/University_Physics/University_Physics_\(OpenStax\)/Book%3A_University_Physics_I_-_Mechanics_Sound_Oscillations_and_Waves_\(OpenStax\)/17%3A_Sound/17.03%3A_Speed_of_Sound](https://phys.libretexts.org/Bookshelves/University_Physics/University_Physics_(OpenStax)/Book%3A_University_Physics_I_-_Mechanics_Sound_Oscillations_and_Waves_(OpenStax)/17%3A_Sound/17.03%3A_Speed_of_Sound)
6. Small Oscillations of a Soap Bubble - Drexel Research Discovery, accessed December 23, 2025, <https://researchdiscovery.drexel.edu/esploro/outputs/journalArticle/Small-Oscillations-of-a-Soap-Bubble/991019312450304721>
7. the ViBbinations of BUBBles and BaLLoons - Australian Acoustical Society, accessed December 23, 2025, https://www.acoustics.asn.au/journal/2012/2012_40_3_Kuo.pdf
8. THE VIBRATION CHARACTERISTICS OF PIEZOELECTRIC DISCS - Imperial College London, accessed December 23, 2025, <https://www.imperial.ac.uk/media/imperial-college/research-centres-and-groups/non-destructive-evaluation/public/Guo-N-1990-PhD-Thesis.pdf>
9. 5 Acoustical Properties of Biological Tissue - Routledge, accessed December 23, 2025, https://www.routledge.com/rsc/downloads/ATQ03_-_K14770_Chapter.pdf
10. Properties of sound - Radiology Cafe, accessed December 23, 2025, <https://www.radiologycafe.com/frcr-physics-notes/ultrasound-imaging/properties>

[s-of-sound/](#)

11. Propagation speed | Radiology Reference Article | Radiopaedia.org, accessed December 23, 2025, <https://radiopaedia.org/articles/propagation-speed?lang=us>
12. Eye orbit effects on eyeball resonant frequencies and acoustic tonometer measurements, accessed December 23, 2025, <https://pubmed.ncbi.nlm.nih.gov/35318400/>
13. Dielectric Properties » IT'IS Foundation, accessed December 23, 2025, <https://itis.swiss/virtual-population/tissue-properties/database/dielectric-properties>
14. In Vivo Safety Evaluation of Acoustic Radiation Force for Optical Coherence Elastography of the Crystalline Lens - PMC - NIH, accessed December 23, 2025, <https://pmc.ncbi.nlm.nih.gov/articles/PMC12395863/>
15. EXTENSIONAL VIBRATIONS OF ELASTIC PLATES - DTIC, accessed December 23, 2025, <https://apps.dtic.mil/sti/tr/pdf/AD0200695.pdf>
16. An improved interpretation of Mason's model for piezoelectric plate transducers, accessed December 23, 2025, <https://www.semanticscholar.org/paper/An-improved-interpretation-of-Mason%27s-model-for-Rhyne/a00eaa454966cae7d80da8004a48277cfe24f48a>
17. A Novel Approach to the Quantification of the Longitudinal Speed of Sound and its Potential for Tissue Characterization (Part - I) - PMC - NIH, accessed December 23, 2025, <https://pmc.ncbi.nlm.nih.gov/articles/PMC6662181/>
18. 1910.133 - Eye and face protection. | Occupational Safety and Health Administration - OSHA, accessed December 23, 2025, <https://www.osha.gov/laws-regs/regulations/standardnumber/1910/1910.133>
19. Maximum permissible exposures for ocular safety (ANSI 2000), with emphasis on ophthalmic devices - Optica Publishing Group, accessed December 23, 2025, https://opg.optica.org/josaa/upcoming_pdf.cfm?id=73925
20. Study: Sound can stimulate blind retina to transmit signals to the brain | Ophthalmology Times - Clinical Insights for Eye Specialists, accessed December 23, 2025, <https://www.opthalmologytimes.com/view/study-sound-can-stimulate-blind-retina-to-transmit-signals-to-the-brain>