Table 5. Minimum Disclosure and Consent Checklist for Connected Implantable Neuromodulation Systems.

Domain	Minimum disclosure element	Patient-facing question	Responsible party
Data categories	List neural, physiologic, device-use, programming, patient-generated, and app-derived data collected	“What data does my device collect?”	Manufacturer
Data flow	Map data movement among implant, programmer, app, cloud, clinician portal, manufacturer, and third parties	“Where does my data go?”	Manufacturer
Care-essential vs optional use	Clearly distinguish therapy/safety uses from non-essential uses, including research, analytics, product development, or marketing	“Which data uses are required for my care?”	Manufacturer/clinician
Cloud and remote monitoring	State whether data are uploaded, where they are uploaded, and whether the data are remotely monitored	“Does my data leave the device or clinic? What happens with my data after it leaves?”	Manufacturer/clinician
Third-party sharing	Identify recipients of data and provide examples	“Who else can receive my data?”	Manufacturer
Retention	Define retention window for all data types	“How long is my data kept?”	Manufacturer
Access/deletion	Describe process, timeline, data scope, and exceptions for access and deletion	“Can I access, correct, export, or delete my data?”	Manufacturer/institution
Opt-out consequences	Explain essential and non-essential consequences of data opt-out	“Can I opt out without losing therapy?”	Manufacturer/clinician
Security lifecycle	Describe patching, vulnerability handling, and security support duration	“How is this system protected over time?”	Manufacturer
Breach notification	Explain notification, security process, and timeline during a data breach	“How will I be told if my data is compromised?”	Manufacturer/institution
End-of-life support	Describe manufacturer support obligations of software, app, or device after discontinuation	“What happens if the platform is retired?”	Manufacturer