

Sort Report

System

Sort Start:	9/8/2021 12:18:49 PM	Server:	Utopex
Application:	BD FACS™ Software	Build:	1.2.0.108
Version:	1.2.0.142	Cytometer Model:	BD Influx System (USB)
ValComp:	7.5.1.3.16	Cytometer Serial #:	X646500R5001

Details

Data Source:	Cytometer	Sort Mode:	1.0 Drop Pure
Nozzle Diameter (µm):	0.00	Drop Envelope:	1.0 Drop
Sheath Pressure (PSI):	0.00	Sort Objective:	Purify
Sort Device:	2 Tube Holder - 2 Way Sort	Phase Mask:	16/16
Piezo Amplitude:	1.27	Extra Coincidence Bits:	4
Drop Delay:	34.0015	Drop Frequency (kHz):	39.00

Sort Details

Name	Population	Event Limit	Event Count	Sort Count	Sort Rate	Aborts	Abort Rate	Efficiency	Time (sec)
Left	-	-	-	-	-	-	-	-	-
Right	live	Unlimited	416,435	200,705	1,026	37,069	189	84.4%	195