

Quantitation Report

7700S

7700s

Sample Name 0.1CK
Acq Time 12/12/2015 9:22
Operator Name
File Name S3.D
File Path C:\ICPMH\1\DATA\15Z1201.B
Method File leon-16.m
Method Path C:\ICPMH\1\METHODS\
Last Calib 12/12/2015 9:13
Total Dilution 1.000

Tune Step **Tune File**
1 copfapt.u
2 nogas.u

Element	Mass	Tune	CPS	Conc	Units	RSD(%)	Time(sec)	Det	Rep
Na	23	1	28,742.46	0.0975	ppb	0.66	0.66	Pulse	3
Mg	24	1	12,554.48	0.0982	ppb	1.37	0.66	Pulse	3
Al	27	1	12,361.98	0.1028	ppb	1.01	3.00	Pulse	3
K	39	1	13,555.38	0.0985	ppb	2.33	0.66	Pulse	3
Ca	40	1	8,918.91	0.0986	ppb	2.18	3.00	Pulse	3
Cr	52	1	5,680.19	0.0998	ppb	1.25	0.99	Pulse	3
Mn	55	1	8,838.91	0.0999	ppb	1.89	0.99	Pulse	3
Fe	56	1	5,629.34	0.1039	ppb	1.80	0.99	Pulse	3
Ni	58	1	4,070.94	0.1008	ppb	2.77	0.99	Pulse	3
Co	59	1	5,981.67	0.1018	ppb	3.14	0.99	Pulse	3
Cu	63	1	4,493.20	0.1002	ppb	1.02	3.00	Pulse	3
Zn	66	2	2,978.40	0.1005	ppb	2.14	3.00	Pulse	3
In	115	1	3,205.77	0.0974	ppb	1.60	0.99	Pulse	3
Ce	140	2	23,818.81	0.0995	ppb	1.08	3.00	Pulse	3
W	182	2	6,515.91	0.0971	ppb	0.83	3.00	Pulse	3
Pb	208	2	10,723.10	0.0980	ppb	0.37	3.00	Pulse	3