

FECHA: 14/06/26
HORA INICIO: 01:21:01
SN: 150P10FEB048RV
CICLO N: 470
<<INSTRUMENTAL>>
TEMP: 121øC EST: 20 min SEC: 20 min
TEMP: 121 øC P: 300 kpa
TEMP: 121 øC P: 300 kpa
TEMP: 121 øC P: 300 kpa
TEMP: 121 øC P: 300 kpa
TEMP: 121 øC P: 300 kpa
TEMP: 121 øC P: 300 kpa
TEMP: 121 øC P: 300 kpa
TEMP: 121 øC P: 300 kpa
TEMP: 121 øC P: 300 kpa
TEMP: 100 øC P: 110 kpa
TEMP: 100 øC P: 110 kpa
TEMP: 100 øC P: 110 kpa
TEMP: 100 øC P: 110 kpa
TEMP: 100 øC P: 110 kpa
TEMP: 100 øC P: 110 kpa
TEMP: 100 øC P: 110 kpa
TEMP: 100 øC P: 110 kpa
HORA FIN: 13:23:23
CICLO COMPLETO
Duraci?n: 38 min
DESCARGA: _____
FECHA: 14/06/26
HORA INICIO: 01:21:22
SN: 150P10FEB048RV
CICLO N: 471
<<INSTRUMENTAL>>
TEMP: 121øC EST: 20 min SEC: 20 min