



GENSAN MEDICAL CENTER

Purok Veterans, Brgy Calumpang, National Highway, General Santos City, 9500

REQUEST FOR REPAIR OF ACCOUNTABILITY

Control No. : 719

Date : 01-Oct-2025

To: Inventory Section

The following items are request list for repair.

HISTOPATH OVEN
LABORATORY

Property No. : 20251000785

NO HEATING. REPAIR BLOWER & TROUBLESHOOT.

Findings : DEFECTIVE BEARING & HEATER.

I certify that item(s) mentioned above reasonable requires
repair of accountability in view of reason(s) stated above.

MAGLASANG SHEENALYN TAPANGAN

Name of Employee
(Sign over printed name)

Checked by:

Section/Division/Dept. Head