



## GENSAN MEDICAL CENTER

Purok Veterans, Brgy Calumpang, National Highway, General Santos City, 9500

### REQUEST FOR REPAIR OF ACCOUNTABILITY

Control No. : 719

Date : 01-Oct-2025

To: Inventory Section

The following items are request list for repair.

**HISTOPATH OVEN**  
LABORATORY

Property No. : 20251000785

NO HEATING. REPAIR BLOWER & TROUBLESHOOT.

Findings : DEFECTIVE BEARING & HEATER.

I certify that item(s) mentioned above reasonable requires repair of accountability in view of reason(s) stated above.

**MAGLASANG SHEENALYN TAPANGAN**

Name of Employee  
(Sign over printed name)

Checked by:

MAE QUIAMCO - BATAN, RSW

10/02/2025

Section/Division/Dept. Head