



GENSAN MEDICAL CENTER

Purok Veterans, Brgy Calumpang, National Highway, General Santos City, 9500

REQUEST FOR REPAIR OF ACCOUNTABILITY

Control No. : 749

Date : 23-Oct-2025

To: Inventory Section

The following items are request list for repair.

NEGATOSCOPE

OPERATING ROOM (OR)

Property No. : 20251000997

FOR REPAIR / DAMAGED PCB.

Findings :

FOR REPAIR / DAMAGED PCB.

I certify that item(s) mentioned above reasonable requires repair of accountability in view of reason(s) stated above.

APROSTA FRITZY JOY CASI

Name of Employee

(Sign over printed name)

Checked by:

10/23/2025

MAE QUIAMCO – BATAN, RSW

Section/Division/Dept. Head