



GENSAN MEDICAL CENTER

Purok Veterans, Brgy Calumpang, National Highway, General Santos City, 9500

REQUEST FOR REPAIR OF ACCOUNTABILITY

Control No. : 793
Date : 28-Nov-2025

To: Inventory Section

The following items are request list for repair.

LABOR

OPERATING ROOM (OR)

Property No. : 20251101125

LABOR PACKAGE, 6500 X 2 = 13000


Findings :

REPROCESS, REPLACED PCB BOARD & RECHARGING OF FREON.

I certify that item(s) mentioned above reasonable requires repair of accountability in view of reason(s) stated above.

APROSTA FRITZY JOY CASI

Name of Employee
(Sign over printed name)

Checked by:  12/1/2025
MAE QUIAMCO - BATAN, RSW
Section/Division/Dept. Head

Friday, 28 - Nov - 2025

Page 1 of 1