Sony Plate Setup

1) Set up the plate adapter in the sort chamber
    a. Locate the desired plate base. On the plate base it will an etch on the metal surface stating either 96 or 364. Locate the L arm that the plate base sits upon. To attach to the sort chamber platform you will need to locate the small screw found on the bottom right side of the platform. Unscrew the screw line up the arm to the hole and screw the arm in place. Once it is set place the base platform on with A1 to the left side. It fits one way.
2) Once you have set your target population with the appropriate gating strategy, you can start setting up your sort parameters.
    a. In the bottom left center of the screen, there will be the sort control section
        i. Select the method of collection (96 v bottom well plate,364 bottom well plate...etc)
    b. Open Sorting settings
        i. This will open a wizard stating Sort Setting- ex.96 well plate. Click on the tab labeled plate adjustment
        ii. Place your plate with the cover on so we can visually check the position of the droplets on the cover.
        iii. In the Sort Test section of the window make sure you have selected the option perform sort test with sheath
        iv. Start at position A1, in the sort test type select the option four corners and center wells.
            i. Click start, there is a status bar that will inform you the test was completed when the bar fills, and your plate will be in the unload position.
            ii. Check the drop landing and adjust the plate as needed.
                a. To adjust select the well and using the arrows in the droplet position to where you need the drop to be center. Do this for all or any that needs readjusting. Recheck each time.
                b. Once you are satisfied you can save this plate custom position for future use or Close the window.
3) Setting up Index sorting
    a. When you can associate the cells sorted in each well with the events for those cells on plots. This function allows you to select a specific well and view the data for cells collected in that well.
        i. In the Sort Setting window have the box check marked< add index sort information> in the plate sort setting tab. Once your done configuring the plate and setting close the window.
        ii. Once you start your sample in the sort control window click index sort and record start.
        iii. To review or analyze the data. In the worksheet tab select analyze index data. The dialog box of the plate will appear with a plot of FSC-A vs BSC-A in the

analysis mode well select mode. If you want to view your cells in a gate region just click the drop down window in the analysis mode to fate select mode. Then click on the desired well to show the location of the sorted item.

iv. You are able to export this data in CSV file in the worksheet tab under the tools group.