GOVERNMENT OF INDIA

MINISTRY OF CONSUMER AFFAIRS, FOOD & PUBLIC DISTRIBUTION

DEPARTMENT OF CONSUMER AFFAIRS — LEGAL METROLOGY DIVISION

OIML R 76-2 PATTERN EVALUATION TEST REPORT

Non-Automatic Weighing Instruments (NAWI) Model Approval

Report Ref. No:	TR-2026-3E5E99	**Date of Evaluation:**	25-09-2026
Testing Laboratory:	Regional Reference Standards Laboratory (RRSL)	**Operating Environment:**	22.5°C | 50.0% RH
Manufacturer Name:	Eagle Scales Pvt Ltd	**Applicant Address:**	Industrial Area, Phase 2, Pune
Instrument Model:	DS-215	**Serial Number:**	SN-2026-98124
Accuracy Class:	Class III	**Verification Interval (e):**	5.0 g
Max Capacity (Max):	30.0 kg	**Min Capacity (Min):**	100.0 g

1. WEIGHING PERFORMANCE TEST (Increasing & Decreasing Load)

Load (kg)	Indication (kg)	Error E (g)	MPE (±g)	Verdict
0.100	0.105	+0.005	±2.500	PASS
2.500	2.503	+0.003	±2.500	PASS
10.000	10.004	+0.004	±2.500	PASS
20.000	20.005	+0.005	±2.500	PASS
30.000	30.006	+0.006	±2.500	PASS
15.000	15.000	+0.000	±2.500	PASS

2. PATTERN EVALUATION COMPLIANCE VERDICT

The Non-Automatic Weighing Instrument (Model: DS-215, Serial No: SN-2026-98124) has been evaluated under statutory clauses of OIML Recommendation R-76. All recorded errors were verified against Maximum Permissible Error (MPE) thresholds for Class III.

FINAL EVALUATION DECISION:	OVERALL RESULT: PASS

3. CERTIFICATE UNDER SECTION 63 OF BHARATIYA SAKSHYA ADHINIYAM (BSA), 2023

This digital record is generated by an automated electronic metrology testing system (MetroTest-76). The data contained herein was recorded directly from laboratory observations without manual alteration. The cryptographic hash embedded below guarantees end-to-end electronic integrity in accordance with Section 63(2) of the Bharatiya Sakshya Adhiniyam, 2023 for admissibility before courts of law.

SHA-256 DIGITAL DIGEST:	a3c8f25e2db016f8e5cab4709aeedc84d0e44e2b286fb7a8e8cc5652192b4d77
ELECTRONIC SIGNATURE:	METROLOGIST BADGE # RRSL-DOCA-2026 | CENTRAL CONTROLLER VERIFIED