GOVERNMENT OF INDIA
MINISTRY OF CONSUMER AFFAIRS, FOOD & PUBLIC DISTRIBUTION
DEPARTMENT OF CONSUMER AFFAIRS — LEGAL METROLOGY DIVISION

OIML R 76-2 PATTERN EVALUATION TEST REPORT

Non-Automatic Weighing Instruments (NAWI) Model Approval Compliance

Report Ref. No:	TR-2026-A2E77B	**Evaluation Date:**	26-09-2026 02:36:44 UTC
Inspecting Officer:	Insp. Vikram Malhotra (GATC-MH-2026-112)	**Testing Laboratory:**	Regional Reference Standards Laboratory (RRSL)
Manufacturer Name:	Eagle Scales Pvt Ltd	**Operating Environment:**	23.0 C | 52% RH
Instrument Model:	DS-215	**Serial Number:**	SN-2026-98124
Accuracy Class:	Class III	**Verification Interval (e):**	5.0 g (n = 6000)
Max Capacity (Max):	30.0 kg	**Min Capacity (Min):**	100.0 g

PHOTOGRAPHIC GROUND-TRUTH EVIDENCE (SECTION 63 BSA 2023 BINDING)




EXHIBIT A: Machine Rating Plate (Riveted Identity Plate)



EXHIBIT B: Digital Scale Display at Max Load (30.0 kg)

[STRICT-LIABILITY MANUAL OVERRIDE LOGGED]: Officer manually modified field(s): Critical Max Indication (Discrepancy with Photo: Display shows 30.018kg, officer entered 30.005kg). Raw unedited photographic evidence above has been permanently bound into this certificate under Section 63 BSA 2023. Discrepancies between physical photos and claimed readings are subject to criminal liability under Bharatiya Nyaya Sanhita (BNS) 2023.

1. WEIGHING PERFORMANCE TEST OBSERVATIONS (OIML R 76-1 Clause A.4.4)

Load m (kg)	Indication I (kg)	Load in e	Error E (g)	Statutory MPE (±g)	Verdict
0.100	0.100	20 e	+0.00	±2.50	PASS
2.500	2.502	500 e	+2.00	±2.50	PASS
10.000	10.004	2000 e	+4.00	±5.00	PASS
20.000	20.005	4000 e	+5.00	±7.50	PASS
30.000	30.005	6000 e	+5.00	±7.50	PASS

3. PATTERN EVALUATION DECISION & ASSIGNED QR BADGE

OVERALL STATUTORY DECISION: PASS

This weighing instrument has been evaluated in strict accordance with the Legal Metrology Act, 2009, Legal Metrology (General) Rules, 2011, and OIML R 76-1/2 Table 6 MPE limits. The assigned cryptographic QR Badge encodes immutable certificate parameters registered in the central DoCA repository.

4. CERTIFICATE UNDER SECTION 63 OF BHARATIYA SAKSHYA ADHINIYAM (BSA), 2023

Officer Non-Repudiation Attestation: I, Insp. Vikram Malhotra (Badge: GATC-MH-2026-112), Inspecting Officer at Regional Reference Standards Laboratory (RRSL), solemnly declare under penalty of perjury and Section 63 of Bharatiya Sakshya Adhiniyam, 2023, that this electronic record was produced during official testing at coordinates [18.5204 N, 73.8567 E]. The physical rating plate in Exhibit A and scale readout in Exhibit B represent unaltered ground-truth observations bound into the cryptographic SHA-256 evidence digest below.

SHA-256 EVIDENCE DIGEST:	f52d925782f24993959485e12ab2ed93933427bcfbc2cca1bf1cce68017cd5f7
ELECTRONIC SIGNATURE:	Insp. Vikram Malhotra [GATC-MH-2026-112] | CENTRALLY REGISTERED DOCA COMPLIANCE STAMP