# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 1 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| Permanent Facility | | | | | |
| 1 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | (Digital/Mechanical) Bevel Protractor / Level and Slope Meter L.C.: 1 min | Using Angle Gauge By Comparison Method | 0° to 90° | 3 min of Arc |
| 2 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Air Gauge Unit L.C:. 0.001 mm (Digital / Dial Type) | Using Plain Ring Gauge set By Comparison Method | Up to 0.1 mm | 2.84 µm |
| 3 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Bore Gauge (Transmission Error) | Using Dial Calibration Tester & Master Plunger Dial By Comparison Method | 0 to 1 mm | 4.7 µm |
| 4 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | CD & PCD Gauge | Using Electronic 2D Height Gauge/Surface plate By Comparison Method | 50 mm to 500 mm | 51.08 µm |
| 5 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Coating Thickness Foils | Using LVDT probe with DRO, Slip Gauge Set By Comparison Method | 20 µm to 2 mm | 3.32 µm |
| 6 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Coating Thickness Gauge / DFT Meter L.C.: 0.1 µm & 1 µm | Using Thickness Foils By Comparison Method | 25 µm to 1880 µm | 3.14 µm |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 2 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 7 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Combination Set / Angle Protractor L.C.: 1° | Using Angle Gauge by Comparison Method | 0°- 90°- 0° | 34.7 min of Arc |
| 8 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Comparator Stand (Flatness of Working Surface) | Using Digital Dial, Comparator Stand By Comparison Method | 0 to 300 mm | 4.96 µm |
| 9 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Cylindrical Setting Master (Concentricity / Runout) | Using FCDM By Comparison method | 3 mm to 100 mm | 4.0 µm |
| 10 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Cylindrical Setting Master (Diameter) | Using Slip Gauge and Comparator stand with Digital Dia By Comparison Method | Up to 100 mm | 4.3 µm |
| 11 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Depth Gauge (Dial / Digital / Vernier ) L.C.: 0.01 mm | Using Slip Gauge / Long Slip Gauge Block, V-Block, Dial Indicator By Comparison Method | 0 to 300 mm | 13.27 µm |
| 12 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Depth Gauge (Dial / Digital / Vernier ) L.C.: 0.02 mm | Using Slip Gauge / Long Slip Gauge Block, V-Block, Dial Indicator By Comparison Method | 0 to 600 mm | 19.19 µm |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 3 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 13 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Depth Micrometer L.C.:0.001 mm | Using Slip Gauge / Long Slip Gauge Block, V-Block, Dial Indicator By Comparison Method | 0 to 300 mm | 6.7 µm |
| 14 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Dial Snap Gauge (Parallesm) L.C.:0.001 mm | Using Slip Gauge Set By Direct Measurement | 5 mm to 200 mm | 3.83 µm |
| 15 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Dial Snap Gauge (Parallesm) L. C: 0.001 mm | Using Gauge Blocks by comparison method | 200 mm to 500 | 5 µm |
| 16 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Dial Thickness Gauge L.C.: 0.001 mm | Using Slip Gauge Set By Comparison Method | 0 to 25 mm | 1.44 µm |
| 17 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Engineers Square (Squareness) | Using Comparator & 2D Height Gauge By Comparison Method | 50 mm to 600 mm | 13.54 µm |
| 18 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Engineers Square (Straighness) | Using Comparator & 2D Height Gauge By Comparison Method | 50 mm to 600 mm | 13.54 µm |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 4 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 19 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | External Micrometer L.C.: 0.001 mm | Using Slip Gauge Block / Long Slip Gauge Blocks By Comparison Method | 0 to 300 mm | 3.23 µm |
| 20 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | External Micrometer L.C.: 0.001 mm | Using Slip Gauges Block , Long Slip Gauge Block By Comparison Method | 300 mm to 1000 mm | 10.42 µm |
| 21 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | External Micrometer (L.C: 0.01 mm) | Using Slip gauge Block, Long Slip Gauge Block By comparison method | 1000 mm to 2000 mm | 16.6 µm |
| 22 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Feeler Gauge Set | Using LVDT probe with DRO, Slip Gauge Set By Comparison Method | 0.02 mm to 1 mm | 3.4 µm |
| 23 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Flush Pin Gauge Height / Depth Checking Gauge | Using Electronic 2D Height Gauge By comparison Method | 0.1 mm to 500 mm | 8 µm |
| 24 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Height Gauge (Dial / Digital / Vernier ) L.C.: 0.01 mm | Using Long Slip Gauge Block, Surface Plate By Comparison Method | 0 to 1000 mm | 15.93 µm |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 5 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 25 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Height Gauge (Dial / Digital / Vernier) L.C.:0.01 mm | Using Caliper Checker, Gr.Surface Plate By Comparison Method | Up to 600 mm | 15.25 µm |
| 26 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Inclinometer L.C: 0.01° | Using Angle Gauge Blocks By Comparison Method | 0° to 90° | 6.9 min of Arc |
| 27 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Inside / Out side Dial Caliper (Two Pin Dial)- Traverse Error L.C.: 0.001 mm | Using Digital Micrometer / Gauge blocs By Comparison Method | Up to 200 mm | 7.4 µm |
| 28 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Internal Micrometer L.C.: 0.01 mm | Using Slip Gauge Set, Long Slip Gauge Block By Comparison Method | 50 mm to 350 mm | 8.14 µm |
| 29 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Jig & Fixtures Outer Dia, Inner Dia, Length , Width, Depth, Height, Concentricity | Using Electronic 2D Height Gauge with Gr Surface Plate by Comparison method | 10 mm to 600 mm | 22 µm |
| 30 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Lever Dial Gauge L.C.: 0.001 mm | Using Dial Calibration Tester By Comparison Method | 0 to 0.14 mm | 3.21 µm |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 6 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 31 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Lever Dial Gauge L.C.: 0.01 mm | Using Dial Calibration Tester By Comparison Method | 0 to 1 mm | 6.34 µm |
| 32 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Measuring Pin | Using LVDT probe with DRO, Slip Gauge Set By Comparison Method | 0.170 mm to 20 mm | 3.48 µm |
| 33 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Measuring Scale L.C: 0.5 mm | Using Scale Tape & Scale Calibrator by Direct Method | 0 to 1000 mm | 150 µm |
| 34 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Measuring Tape L.C.: 1 mm | Using Tape & Scale Calibrator By Comparison Method | Up to 30 m | 328.30 µm |
| 35 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Micrometer Setting Stick/Height Block | Using Digital Dial, Slip Gauge Set,Comparator stand By Comparison Method | 100 mm to 300 mm | 5.65 µm |
| 36 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Micrometer Setting Stick/Height Block | Using Digital Dial, Slip Gauge Set, Comparator Stand With Digital Dial By Comparison Method | 25 mm to 100mm | 4.56 µm |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 7 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 37 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Micrometer Setting Stick/Height Block | Using Digital Dial, Slip Gauge Set, Comparator stand By Comparison Method | 300 mm to 1000 mm | 11.38 µm |
| 38 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Mould (Cube, Beam, Flakiness & Elongation Gauge, Slump Cone, Core Cutter, Width Gauge, Industrial Templates(ID, OD, Height, Thickness) | Using Digital Vernier Caliper by Comparison method | Up to 600 mm | 39.3 µm |
| 39 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | O D Master | Using Digital Dial, Slip Gauge Set By Comparison Method | 1 mm to 100 mm | 2.5 µm |
| 40 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | O D Master | Using Digital Dial, Slip Gauge Set, Long Gauge Block By Comparison Method | 100 mm to 300 mm | 4.58 µm |
| 41 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Paddle Gauge Diameter | Using Digital Dial, Slip Gauge Set, Long Gauge Block By Comparison Method | 100 mm to 300 mm | 4.58 µm |
| 42 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Paddle Gauge Diameter | Using Digital Dial, Slip Gauge Set, Long Slip Gauge Block By Comparison Method | 300 mm to 500 mm | 5.75 µm |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 8 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 43 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Paddle Gauge Diameter | Using Digital Dial, Slip Gauge Set By Comparison Method | 5 mm to 100 mm | 2.5 µm |
| 44 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Paddle Gauge Diameter | Using Digital Dial , Comparator stand, Slip Gauge, Long Slip Gauge By comparison method | 500 mm to 1000 mm | 9.7 µm |
| 45 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Pie Tape (0.5 mm, 1 mm) | Using Tape & Scale Calibrator by comparison method | 0 to 6000 mm | 78.2 x sqrt (L) µm, where L in meter. |
| 46 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Pistol Caliper L.C.: 0.01 mm | Using Slip Gauge Set By Comparison Method | 0 to 50 mm | 32.46 µm |
| 47 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Plain Mandrel (Variation in dia) | Using Digital Dial with Comparator stand by comparison method | 5 mm to 100 mm | 3.4 µm |
| 48 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Plain Plug Gauge | Using Digital Dial, Slip Gauge Set, Long Gauge Block By Comparison Method | 100 mm to 300 mm | 4.58 µm |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 9 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 49 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Plain Plug Gauge | Using Digital Dial, Slip Gauge Set, Long Slip Gauge Block By Comparison Method | 300 mm to 500 mm | 5.75 µm |
| 50 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Plain Plug Gauge | Using Digital Dial gauge , Comparator stand, Slip Gauge, long slip gauge by comparison method | 500 mm to 1000 mm | 10.2 µm |
| 51 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Plain Plug Gauge / Width Gauge / Plain Mandrill | Using Digital Dial, Slip Gauge Set By Comparison Method | 0.5 mm to 100 mm | 3.1 µm |
| 52 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Plain Ring Gauge / Setting Ring Gauge | Using Length Measuring Machine & Setting Ring Gauge by Comparison method | 150 mm to 300 mm | 4.3 µm |
| 53 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Plain Ring Gauge / Setting Ring Gauge | Using Length Measuring Machine & Master setting Ring Gauge by Comparison Method | 100 mm to 160 mm | 3.23 µm |
| 54 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Plain Ring Gauge / Setting Ring Gauge | Using Length Measuring Machine & Master setting Ring Gauge by Comparison Method | 3 mm to 100 mm | 1.5 µm |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 10 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 55 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Plunger Dial Gauge L.C.: 0.001 mm | Using Dial Calibration Tester By Comparison Method | 0 to 12.7 mm | 2.9 µm |
| 56 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Plunger Dial Gauge L.C.: 0.01 mm | Using Dial Calibration Tester By Comparison Method | 0 to 50 mm | 6.37 µm |
| 57 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Setting Gauge (Height Master) | Using Digital Dial, Slip Gauge Set, Long Gauge Block By Comparison Method | 100 mm to 300 mm | 4.58 µm |
| 58 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Setting Gauge (Height Master) | Using Plunger Dial, Comparator stand, Slip Gauge, long slip gauge by Comparison method | 500 mm to 1000 mm | 11.6 µm |
| 59 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Setting Gauge(Height Master) | Using Digital Dial, Slip Gauge Set, Long Slip Gauge Block By Comparison Method | 300 mm to 500 mm | 5.75 µm |
| 60 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Setting Gauge(Hight Master) | Using Digital Dial, Slip Gauge Set By Comparison Method | 1 mm to 100 mm | 2.5 µm |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 11 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 61 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Slump Cone (Diameter & Height) | Using Digital Vernier Caliper by Comparison method | 0 to 300 mm | 39.1 µm |
| 62 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Snap Gauge | Using Slip Gauge blocks, Long slip gauge By comparison method | 500 mm to 1000 mm | 10.33 µm |
| 63 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Snap Gauge / Gap Gauge | Using Slip Gauge Block By Comparison Method | 100 mm to 200 mm | 2.80 µm |
| 64 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Snap Gauge / Gap Gauge | Using Slip Gauge Blocks By Comparison Method | 2 mm to 100 mm | 2.46 µm |
| 65 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Snap Gauge / Gap Gauge | Using Slip Gauge Block, Long Slip Gauge Block By Comparison Method | 200 mm to 500 mm | 5.37 µm |
| 66 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Sphere / Spherical Ball Diameter | Using Electronic Probe With DRO, Slip Gauge set By Comparison Method | Up to 100 mm | 3.63 µm |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 12 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 67 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Spline Plug Gauge ( Diameter over Pin & Major Diameter) | Using Slip Gauge, Floating Carriage Diameter Measuring Machine & Cylindrical Setting Master by Comparison Method | 10 mm to 100 mm | 4.0 µm |
| 68 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Spline Ring Gauge | Using Grade '0' Grade Slip Gauge by Comparison Method | 10 mm to 100 mm | 2.1 µm |
| 69 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Taper Thread Plug Gauge Effective Diameter | Using FCDM With Dial, Cylindrical Setting Master, Thread Measuring Wire By Comparison Method | 2 mm to 100 mm | 4.90 µm |
| 70 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Test Sieves | Using Digital Caliper By Comparison Method | 1 mm to 100 mm | 41.18 µm |
| 71 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Thread Measuring Wire | Using LVDT probe with DRO, Slip Gauge Set By Comparison Method | 0.17 mm to 6.35 mm | 0.93 µm |
| 72 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Thread Plug Gauge Effective Diameter & Major Diameter | Using Length Measuring Machine , Cylindrical Setting Master& Measuring wire / Long Slip Gauge Block by Comparison method | 100 mm to 300 mm | 5.2 µm |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 13 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 73 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Thread Plug Gauge Major Diameter and Effective Diameter | Using FCDM With Dial , Cylindrical Setting Master, Thread Measuring Wire By Comparison Method | 2 mm to 100 mm | 4.91 µm |
| 74 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Thread Ring Gauge Major Diameter and Effective Diameter | Using Length Measuring Machine & Master Setting Ring Gauge by Comparison Method | 3 mm to 100 mm | 3.23 µm |
| 75 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Three Point Micrometer (L.C.: 0.001 mm) | Using Setting Ring Gauge by Comparison Method | 100 mm to 200 mm | 7.5 µm |
| 76 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Three Point Micrometer LC:.0.001 mm | Using Setting Ring Gauge By Comparison Method | 5 mm to 100 mm | 3.27 µm |
| 77 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Ultrasonic Thickness Gauge L.C.: 0.01 mm | Using Slip Gauge Set, Long Slip Gauge By Comparison | 0 to 200 mm | 57.95 µm |
| 78 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | V Block Perpendicularity | Using Surface Plate, Engineering Square, Slip gauge set By Comparison Method | Up to 300 mm | 15.2 µm |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 14 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 79 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | V Block Parallelism | Using Surface Plate, Digital Plunger dial By Comparison Method: | upto 300mm | 9.58 µm |
| 80 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | V Block Symmetricity | Using Surface Plate, Digital Plunger dial , Plain Mandril By Comparison Method: | upto 300 mm | 11.4 µm |
| 81 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Vernier Caliper (Dial / Digital / Vernier ) L.C.: 0.01 mm | Using Long Slip Gauge Block By Comparison method | 0 to 1000 mm | 17.63 µm |
| 82 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Vernier Caliper (Dial / Digital / Vernier ) L.C.:0.01 mm | Using Caliper Checker By Comparison Method | Up to 600 mm | 16.10 µm |
| 83 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Vernier Caliper (Vernier / Dial / Digital) | Using Length bar Slip Gauge & Digital Micrometer By comparison method | 0 to 2000 mm | 30 µm |
| 84 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Welding Gauge, Weld Fillet Gauge | Using Gauge Blocks, Electronic 2D Height Gauge by comparison method | 1 mm to 100 mm | 120.8 µm |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 15 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 85 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Width Gauge | Using Digital Dial, Slip Gauge Set By Comparison Method | 0.5 mm to 100 mm | 3.6 µm |
| 86 | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Caliper Checker / Check Master | Using Electronic 2D Height Gauge By comparison Method | 0 to 600 mm | 12.4 µm |
| 87 | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Electronic 2D Height Gauge 0.0001 mm (Squareness) | Using Granite Square by Comparison Method | 0 to 600 mm | 10.3 µm |
| 88 | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Electronic 2D Height Gauge 0.0001mm Linear Measurement | Using Long Slip Gauge Block By Comparison method | 0 to 600 mm | 8.63 µm |
| 89 | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Surface Roughness Specimen (Master) | Using Portable Surface Roughness Tester by Comparison method | 0.1 µm to 6.2 µm | 7.4 %rdg |
| 90 | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Surface Roughness Tester (Ra & Rz) | Using Three different values of Roughness Masters By Direct Comparison Method | up to 6.2 µm | 8 % |
| 91 | MECHANICAL-PRESSURE INDICATING DEVICES | Pneumatic Pressure gauge (Analog/ Digital) Differential Pressure Gauge, Pressure Transmitter with Indicator, Pressure Switch, Pressure Sensor with Indiactor | Using Pneumatic Comparator and Digital Pressure Gauge, Digital Multimeter By Comparison Method as per DKD R6-1 | 0 to 40 bar | 0.025 bar |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 16 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 92 | MECHANICAL-PRESSURE INDICATING DEVICES | Pressure Gauge (Dial / Digital Pressure Gauges)-Hydraulic Pressure Transducer/Transmitter / Pressure Switch/ Pressure Gauge/ Pressure Sensor with Indicator | Using Digital Pressure Gauge with Hydraulic Pump Comparator, Digital Multimeter by Comparison Method as per DKD R6-1 | 0 to 1000 bar | 3.36 bar |
| 93 | MECHANICAL-PRESSURE INDICATING DEVICES | Pressure Gauges-Hydraulic ( Analog/ Digital ) , Transmitter with indicator , Pressure Transducer/Pressure Calibrator/Pressure Gauge, Pressure Switch | Using Digital Pressure Gauge with Hydraulic Pump, Digital Multimeter By Comparison Method as per DKD R6-1 | 0 to 700 bar | 0.95 bar |
| 94 | MECHANICAL-PRESSURE INDICATING DEVICES | Vacuum Gauges (Analog/ Digital) Vacuum Transmitter, Vacuum Switch) | Using Pneumatic hand pump and Digital Pressure Gauge, Digital Multimeter By Comparison Method as per DKD R6-1 | (-) 0.80 bar to 0 | 0.009 bar |

# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 17 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| Site Facility | | | | | |
| 1 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Air Gauge Unit L.C:. 0.001 mm (Digital / Dial Type) | Using Plain Ring Gauge set By Comparison Method | Up to 0.1 mm | 2.84 µm |
| 2 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Bench Center Co-Axiality | Using Plunger Dial & Plain Mandrel by Comparison method | 0 to 600 mm | 8.74 µm |
| 3 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Bench Center Parallelism | Using Plunger Dial & Plain, Taper Mandrel by Comparison method | 0 to 600 mm | 9.3 µm |
| 4 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | External Micrometer (L.C: 0.01 mm) | Using Slip gauge Block, Long Slip Gauge Block By comparison method | 1000 mm to 2000 mm | 16.6 µm |
| 5 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Slump Cone (Diameter & Height) | Using Digital Vernier Caliper by Comparison method | 0 to 300 mm | 39.1 µm |
| 6 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Surface Plate (Flatness) | Using Sprit Level by Comparison Method | 3000 mm to 3000 mm | 2.6 sqrt {(L+W)/150} µm, 'L'& 'W' in mm |

# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 18 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 7 | MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.) | Vernier Caliper (Vernier / Dial / Digital) | Using Length bar Slip Gauge & Digital Micrometer By comparison method | 0 to 2000 mm | 30 µm |
| 8 | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Electronic 2D Height Gauge 0.0001 mm (Squareness) | Using Granite Square by Comparison Method | 0 to 600 mm | 10.3 µm |
| 9 | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Electronic 2D Height Gauge 0.0001mm Linear Measurement | Using Long Slip Gauge Block By Comparison method | 0 to 600 mm | 8.63 µm |
| 10 | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Profile Profile Projector / Video Measuring Machine (Angular Scale ) L.C: 1 min | Using Angle Gauge by Comparison Method | 0º to 360º | 1.8 min |
| 11 | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Profile Projector / Microscope ( Magnification ) | Using Glass Scale, Slip Gauge , Digital Caliper by Comparison Method | 5 X to 100 X | 0.24 % |
| 12 | MECHANICAL-DIMENSION (PRECISION INSTRUMENTS) | Profile Projector / Video Measuring Machine Linear Scale L. C.: 0.0001 mm | Using Slip Gauge by Comparison Method | 0 to 200 mm | 6.4 µm |
| 13 | MECHANICAL-HARDNESS TESTING MACHINES | Brinell Hardness Testing Machine Indirect method | Using Standard Hardness Blocks by Indirect Method as per IS 1500(2)-2021 | 10/ 3000 HBW | 3.01 % |
| 14 | MECHANICAL-HARDNESS TESTING MACHINES | Rockwell Hardness tester & Portable Hardness Tester Indirect method | Using Standard Hardness Blocks Part-2: 2018 by Indirect Method as per IS 1586 | HRB | 1.6 HRB |

# SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 19 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 15 | MECHANICAL-HARDNESS TESTING MACHINES | Rockwell Hardness Tester / Portable Hardness tester Indirect method | Using Hardness Blocks by Indirect Method as per IS 1586 Part-2: 2018 | HRC | 1.31 HRC |
| 16 | MECHANICAL-HARDNESS TESTING MACHINES | Vickers Hardness Tester by Indirect method | Using Standard Hardness Block as per ISO 6507-2-2018, IS 1501(2)-2020 | HV30 | 1.99 % |
| 17 | MECHANICAL-PRESSURE INDICATING DEVICES | Pneumatic Pressure gauge (Analog/ Digital) Differential Pressure Gauge, Pressure Transmitter with Indicator, Pressure Switch, Pressure Sensor with Indiactor | Using Pneumatic Comparator and Digital Pressure Gauge, Digital Multimeter By Comparison Method as per DKD R6-1 | 0 to 40 bar | 0.025 bar |
| 18 | MECHANICAL-PRESSURE INDICATING DEVICES | Pressure Gauge (Dial / Digital Pressure Gauges)-Hydraulic Pressure Transducer/Transmitter / Pressure Switch/ Pressure Gauge/ Pressure Sensor with Indicator | Using Digital Pressure Gauge with Hydraulic Pump Comparator, Digital Multimeter by Comparison Method as per DKD R6-1 | 0 to 1000 bar | 3.36 bar |
| 19 | MECHANICAL-PRESSURE INDICATING DEVICES | Pressure Gauges-Hydraulic ( Analog/ Digital ) , Transmitter with indicator , Pressure Transducer/Pressure Calibrator/Pressure Gauge, Pressure Switch | Using Digital Pressure Gauge with Hydraulic Pump, Digital Multimeter By Comparison Method as per DKD R6-1 | 0 to 700 bar | 0.95 bar |

# National Accreditation Board for Testing and Calibration Laboratories

## SCOPE OF ACCREDITATION

**Laboratory Name :** JYOTIBA INSTRUMENTS ENTERPRISES, SHREE COMPLEX, BLOCK NO. 7, 1ST FLOOR, PLOT NO. 182/A & 186, SR. NO. 347/1B, UDYAMBAG, BELGAUM, BELAGAVI, KARNATAKA, INDIA

**Accreditation Standard** ISO/IEC 17025:2017

**Certificate Number** CC-3243 **Page No** 20 of 20

**Validity** 11/05/2025 to 10/05/2029 **Last Amended on** 30/05/2025

| S.No | Discipline / Group | Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument | Calibration or Measurement Method or procedure | Measurement range and additional parameters where applicable(Range and Frequency) | * Calibration and Measurement Capability(CMC)(±) |
|---|---|---|---|---|---|
| 20 | MECHANICAL-PRESSURE INDICATING DEVICES | Vacuum Gauges (Analog/ Digital) Vacuum Transmitter, Vacuum Switch) | Using Pneumatic hand pump and Digital Pressure Gauge, Digital Multimeter By Comparison Method as per DKD R6-1 | (-) 0.80 bar to 0 | 0.009 bar |

*** CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of k = 2.**