

Certificate of Analysis

Product Name			
Cat. No.			
LOT No.		MFG Date	
		EXP Date	
Quantity		Date of Release	
Testing categories	Specification		Result
Appearance Inspection	Test strip size · W: 5.5 ± 0.05mm · L: 30 ± 0.05mm		Pass
	Quantity per one container: 25 T/vial		Pass
Repeatability	· Low: CV ≤ 5% · Mid/High: CV ≤ 4%		Pass
Coding test	· Testing Temperature Range: 10-40°C 50-104°F · Testing Range: 10-600mg/dL 0.6-33.3mmol/L		Pass
Packaging Inspection	Packing list	Test strip (25T) X 2 containers / Code chip / Test strip package insert	Pass
	Expiration date between test strip label and package		Pass
	LOT No. between test strip label and package		Pass
	Code chip number between container label and code chip		Pass

According to SD Biosensor quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.

* Date issued :

SunYoung Jeong
Manager of Regulatory Affairs & QA Dept.
SD Biosensor, Inc.