

## Certificate of Analysis

Product Name			
Cat. No.			
LOT No.		MFG Date	
		EXP Date	
Quantity		Date of Release	
Testing categories	Specification		Result
Appearance Inspection	Test strip size · W: 5.5 ± 0.05mm · L: 30 ± 0.05mm		Pass
	Quantity per one container: 25 T/vial		Pass
Repeatability	· Low: CV ≤ 5% · Mid/High: CV ≤ 4%		Pass
Coding test	· Testing Temperature Range: 10-45°C   50-113°F · Testing Range: 10-600mg/dL   0.6-33.3mmol/L		Pass
Packaging Inspection	Packing list	Test strip (25T) X 2 containers / Test strip package insert	Pass
	Expiration date between test strip label and package		Pass
	LOT No. between test strip label and package		Pass

*According to SD Biosensor quality assurance system, the above QC tests have been completed and passed the specification. We hereby approve this product release.*

\* Date issued :

SunYoung Jeong  
Manager of Regulatory Affairs & QA Dept.  
SD Biosensor, Inc.