
cvdentus

TIPS

ULTRASONIC & ROTARY



Experience Innovation

CATALOG 2025

Clinical PLUS | DentSurg PRO

CONTENTS

01 CVD DIAMOND 04

02 DLC DIAMOND 05

03 ENDODONTICS 07

04 PERIODONTICS 17

05 RESTORATIVE DENTISTRY 25

06 PROSTHODONTICS 33


07 SURGERY 41

08 INDEX 48


cvdentus

CVD Diamond

CVDentus® CVD Diamond (the process known as Chemical Vapor Deposition) is a 100% Brazilian technology, patented worldwide. Its hardness is close to that of natural diamond and it is 30x harder than tooth enamel (95 GPa vs. 3.3 GPa for dental enamel). Being a unique stone with controlled roughness, in addition to its high durability, its characteristics also enable exclusive procedures and differentiated preparation quality.



Unused



After 5 min. of use

CONVENTIONAL DIAMOND-COATED TIP

· Natural or synthetic diamond powder


· Lower durability

· Surface with irregular edges


· Uneven finish

· Leaves residues

· Direct contact with metal



Unused



After 5 min. of use

CVDENTUS® DIAMOND TIP

· Single diamond stone


· High durability (30x greater)

· Homogeneous surface, free of edges

· Uniform finish

· Contact only with diamond (biologically compatible)

Beyond the **homogeneous surface**, CVDentus® technology allows controlled grain size in the diamond production process, enabling the professional to perform procedures from wear to pre-polishing. There are **five different roughness types** to meet the widest range of needs: Extreme Ultra Fine, Ultra Fine, Fine, Coarse and Extreme Coarse.




EXTREME ULTRA FINE - UX

Ideal for finishing or the stage prior to polishing.




ULTRA FINE - U

Performs the refinement of the preparation




FINE - F

Ideal for fine wear.



COARSE

Performs controlled wear.




EXTREME COARSE - XR

Ideal for heavy wear.


DLC Diamond

The coating known as **Diamond Like Carbon** (DLC) by CVDentus®, is a patented technology applied as a film on the tips of the surgical line, providing **wear resistance, a reduced friction coefficient** and also a **bactericidal effect**. Applied over a titanium alloy, a noble material used in aircraft and satellites, it has characteristics that make it ideal for manufacturing ultrasonic surgical tips. It has **low weight and excellent physical-chemical properties, such as high mechanical strength, resistance to high temperatures and also to chemical corrosion.**






Tips with varied shapes and active-part thicknesses developed from 0.28 mm, with diameters that ensure precision without unnecessary wear, enabling high-performance, safe surgical procedures.




DLC HAS AN EXTREMELY LOW FRICTION COEFFICIENT, CLOSE TO TEFLON, ENSURING LESS TISSUE DAMAGE AND LESS HEAT GENERATION

Comparative study of the tribological behavior under hybrid lubrication of diamond-like carbon films with different adhesion interfaces
R.P.C Costa . D.A. Lima-Oliveira , F.R. Marciano , A.O. Lobo , E.J. Corat . V.J. Trava- Airoldi



DLC INCREASES THE SERVICE LIFE OF THE PART BY PROTECTING IT FROM ELECTROCHEMICAL CORROSION.

Crystalline diamond particles into diamond-like carbon films: The influence of the particle sizes on the electrochemical corrosion resistance
F.R. Marciano. E.C Almeida, D.A Lima-Oliveira, E.J. Corat , V.J. Trava-Airoldi



DLC "DOPED" WITH TITANIUM DIOXIDE (TiO_2) NANOPARTICLES REDUCES BACTERIAL ACTIVITY BY UP TO 75%

Anticabterial activity of DLS films containing Tio nanoparticles
F.R Marciano, D.A. Lima-Oliveira, N.S. da-SIlva, A.V. Diniz , E.J. Corat , V.J. Trava - Airoldi

ENDODONTICS



cvdentus

PULP CHAMBER

Access








Code **TR18-H**
Ø 0.5mm

· Rotary bur. High wear for access to the pulp chamber









Code **TR38-H**
Ø 1.0mm

· Rotary bur. High wear for access to the pulp chamber









Code **R2**
Ø 1.2mm

· Wear for access to the pulp chamber









Code **R4**
Ø 1.6mm

· Wear for access to the pulp chamber

Pulp chamber preparation








Code **R1-E**
Ø 1.0mm

· Performs wear and enables removal of the chamber roof and dentin









Code **C22-E**
Ø 1.0mm

· Performs wear and enables removal of the chamber roof and dentin

Pulp stone removal








Code **R1-E**
Ø 1.0mm










Code **T1F-E**
Ø 0.5mm

Fine wear of the walls









Code **CR4F-ST**
Ø 1.6mm

· Does not wear the floor, since it has no diamond on the tip

Canal location








Code **R1-E**
Ø 1.0mm








Code **T0T-E**
Ø 0.4mm

· Including MB2







Code **T1T-E**
Ø 0.65mm


> These 2 inserts are very interesting: the TR1F-ST works in the isthmus region and refines the canal entrance, and the CR4F-ST finishes and smooths the pulp chamber walls, **ensuring the safety of the procedures and extreme quality in the preparations.**

PROF. **RODRIGO VILLAGRA**
Endodontics specialist.

Dentin triangle removal








Code **T0T-E**
Ø 0.4mm








Code **T1T-E**
Ø 0.65mm

Canal clearing








Code **T1F-E**
Ø 0.5mm

Refinement of the canal entrance







Code **TR1F-ST**
Ø 1.0mm


· Has no diamond on the tip

MIDDLE THIRD

Isthmus








Code **T1F-E**
Ø 0.5mm

· Cleaning of the isthmus area








Code **TR1F-ST**
Ø 1.0mm

· Has no diamond on the tip, cleaning only the lateral part of the isthmus region

Gutta-percha








Code **T0T-E1**
Ø 0.3mm

· Cutting and thermoplasticization of gutta-percha








Code **T0T-E2**
Ø 0.3mm

· Cutting and thermoplasticization of gutta-percha








Code **CR2T-E**
Ø 1.2mm

· Cutting and compaction of gutta-percha

Cleaning of flattened canals









Code **T0T-E3**
Ø 0.45mm

· Cleaning of hard-to-reach areas






Photos: Prof. Rodrigo Villagra -
Cutting and thermoplasticization of gutta-percha
Code C2T-E and R1-E

Fractured instrument removal








Code **TOT-E**
Ø 0.4mm









Code **T0T-E1**
Ø 0.3mm







Code **T0T-E2**
Ø 0.3mm






Photos: Prof. Rodrigo Villagra -
Pulp chamber cleaning and cutting and thermoplasticization of gutta-percha
Code TOT-E1, TOT-E2 and TOT-E3

> Together with CVDentus tips, ultrasound is a tool of great importance for those seeking **excellence in endodontic treatment**, as it helps in several steps, such as: conservative access surgery, cavity cleaning, PUI, post removal, retreatment, among others.
>
> DR. RODRIGO VILLAGRA
> Endodontics specialist

ACTIVATION






Code **T0T-E1**
Ø 0,3mm

· Passive irrigation (PUI and CUI)
Agitation of cement, acids and cleaning liquids






Code **T0T-E2**
Ø 0.3mm

· Passive irrigation (PUI and CUI)
Agitation of cement, acids and cleaning liquids






Code **T0T-E5**
Ø 0,3mm

· Passive irrigation (PUI and CUI)
Agitation of cement, acids and cleaning liquids

CANAL CLEANING IN RETREATMENT






Code **T0T-E**
Ø 0.4mm

· Removal of gutta-percha from the cervical half at the start of endodontic retreatment






Code **T1T-E**
Ø 0.65mm

· Removal of gutta-percha from the cervical half at the start of endodontic retreatment






Code **T0T-E3**
Ø 0.45mm

· Cleaning of hard-to-reach areas, removing remaining gutta-percha and cement

POST REMOVAL








Code **T0T-E**
Ø 0.4mm

· Removal at the cement line around the metal post and inside the canal up to the middle third / Completion of fiberglass post removal








Code **T1T-E**
Ø 0.65mm

· Removal at the cement line around the metal post / Completion of fiberglass post removal








Code **C2T**
Ø 1.6mm

· Removal of cast or threaded metal posts

FIBERGLASS POST REMOVAL








Code **TR1-E**
Ø 0.8mm

· Performs wear of the post, aiding its removal







Photos: Prof. Ellen Mangabeira -
Wear and removal of fiberglass post
Code TR1-E

> The TR1-E insert is a market differentiator, as it provides wear of the fiberglass post while preserving the maximum amount of dentin compared to traditional methods, thanks to the shape, usability and differentiated diamond surface treatment technology that only CVDENTUS has.
>
> PROF. ELLEN MANGABEIRA
>
> Specialist in Prosthodontics (APCD SP) and in Endodontics (USP); Master's in Prosthodontics from Faculdade São Leopoldo Mandic; Professor of Prosthodontics; Partner in the development of the TR1-E insert

Bone window opening








Code **VR04XR**
1.6mm x 0.4mm

Apicoectomy








Code **VR04XR**
1.6mm x 0.4mm








Code **W1**
0.5mm x 2.5mm








Code **W1-O**
0.5mm x 2.5mm

Root-end preparation








Code **T0U-2**
Ø 0.25mm








Code **T1F-2**
Ø 0.5mm

Root-end preparation in posterior teeth








Code **T0U-L**
Ø 0.25mm








Code **T0U-R**
Ø 0.25mm








Code **T1F-L**
Ø 0.5mm








Code **T1F-R**
Ø 1.0mm

PERIODONTICS




CLINICAL CROWN LENGTHENING

With flap opening








Code **L3S-O**
0.5mm x 3.5mm

· Performs detachment of the gingiva








Code **FS-O**
0.4mm x 2.5mm

· Performs scraping of thick bone








Code **FS2-O**
3.5mm x 0.4mm

· Performs wider scraping of thick bone








Code **C4T**
Ø 1.6mm

· Indicated for thicker bone. Has diamond only on the tip and performs bone wear through an opening in the gingiva (mini flap) for tip insertion
· Has depth marking at 4 and 5 mm








Code **CH4T-M**
Ø 1.6mm

· Indicated for thicker bone. Has diamond only on the tip and performs bone wear through an opening in the gingiva (mini flap) for tip insertion
· Has depth marking at 4 and 5 mm








Code **CH4T-L**
Ø 1.6mm

· Indicated for thicker bone. Has diamond only on the tip and performs bone wear through an opening in the gingiva (mini flap) for tip insertion
· Has depth marking at 4 and 5 mm








Code **CH4T-R**
Ø 1.6mm

· Indicated for thicker bone. Has diamond only on the tip and performs bone wear through an opening in the gingiva (mini flap) for tip insertion
· Has depth marking at 4 and 5 mm

Without flap opening








Code **TR1-PK**
Ø 0.5mm

· Performs bone wear without injuring the gingiva (flapless)
· Has 3 mm of diamond and depth marking at 4 and 5 mm
Code








Code **VR1-RPK**
Ø 1,2mm

· Performs bone wear without injuring the gingiva (flapless)
· Has 3 mm of diamond and depth marking at 4 and 5 mm
Code







Code **VR1-LPK**
Ø 1,2mm

· Performs bone wear without injuring the gingiva (flapless)
· Has 3 mm of diamond and depth marking at 4 and 5 mm
Code





Code **TR12-PK**
Ø 0.5mm

· Performs bone wear without injuring the gingiva
· Has 2 mm of diamond and depth marking at 3 and 4 mm
Code



Photo: Prof. Paulo Kano -
Detail of excess gingiva removed
Code VR1-LPK



Photo: Prof. Marcelo de Faveri -
Flapless surgery
Code TR1-PK



Photo: Prof. Paulo Kano -
Postoperative (1 day)

> I use CVDentus diamond inserts for crown lengthening without flap opening, without damage to the gingival tissue, with better precision, **providing more comfort for the patient** with a faster postoperative period.

PROF. **PAULO KANO**


Master's in Implantology, Director of the Paulo Kano Institute, Creator and Developer of the Clones method, Creator of the AST (Atomic Shell Technique) / SKIN Concept and dental prosthetics technician.

SCALING WITHOUT DIAMOND

Supragingival








Code **T0T**
Ø 0.30mm

· Delicate interproximal scaling (without diamond)








Code **T0TC**
Ø 0.60mm








Code **T1T**
Ø 0.6mm

Code **T1T-P**
Ø 0.8mm

· For cleaning implants and ceramic parts








Code **W1T-O**
0.5mm x 2.5mm

Subgingival








Code **T0T**
Ø 0.3mm

· Delicate interproximal scaling (without diamond)








Code **T1T-P**
Ø 0.8mm

· For cleaning implants and ceramic parts

SCALING WITH DIAMOND

Supragingival






Code **T0UX**
Ø 0.25mm

· Delicate supra- and subgingival interproximal scaling






Code **T0U**
Ø 0.25mm

· Delicate supra- and subgingival interproximal scaling






Code **T1U**
Ø 0.5mm

· Indicated for all teeth






Code **T1UX**
Ø 0.5mm

· Indicated for all teeth






Code **TR1U-SD**
Ø 1.0mm

· Has diamond only on the lateral working surfaces






Code **FR4U**
Ø 1.6mm

· Cleaning of the lingual surface of anterior teeth

Subgingival






Code **T0UX**
Ø 0.25mm

· Delicate supra- and subgingival interproximal scaling








Code **T0U**
Ø 0.25mm

· Delicate supra- and subgingival interproximal scaling








Code **T1UX**
Ø 0.50mm

· Indicated for all teeth








Code **T1U-E**
Ø 0.5mm








Code **TR1U-SD**
Ø 1.0mm

· Has diamond only on the lateral working surfaces








Code **T0U-2**
Ø 0.25mm

· Delicate scaling in the furcation area








Code **T1U-2**
Ø 0.5mm

· Scaling in the furcation area

ROOT PLANING








Code **T1F-E**
Ø 0.5mm

PERIOTOMY








Code **VR1T**
1.2mm x 0.4mm

· In all teeth

GINGIVAL PEELING









Code **CR1**
Ø 1.2mm

· For narrow areas







Code **CR4**
Ø 1.6mm

· For wide areas

> Among its many applications, Cvdentus tips also allow melanoplasty to be performed with great safety and comfort for the patient.
>
> PROF. **DIEGO BAZAN**
> Dentist, specialist in Oral and Maxillofacial Surgery, Periodontics and Implantology.






Photos: Diego Bazan -
Gingival peeling and postoperative
Code CR1 and CR4

> I use diamond tips because they are better, effectively proven. Better than laser peeling surgery or with an Orban scalpel. The pigments take a long time to reappear and, in many cases, never appear again.
>
> DR. DÉBORA AYALA
> Master's in Laser, Specialist in Periodontics, Specialist in Restorative Dentistry and Specialist in Implantology.

RESTORATIVE DENTISTRY




cvdentus

CAVITY ACCESS

High wear








Code **TR18-H**
Ø 0.5mm

· Rotary bur








Code **TR38-H**
Ø 1.0mm

· Rotary bur








Code **CR1XR**
Ø 1.2mm








Code **CR4XR**
Ø 1.6mm

Wear









Code **R0,5**
Ø 0.5mm

· Small cavities








Code **R1**
Ø 1.0mm

· Small cavities









Code **R1-A**
Ø 1.0mm

· Small hard-to-reach cavities, such as tunnel cavities








Code **R1-B**
Ø 1.0mm

· Small hard-to-reach cavities, such as on the lingual surface of anterior teeth and tunnel-type access









Code **R2**
Ø 1.2mm

· Medium cavities









Code **R4**
Ø 1.5mm

· Large cavities








Code **T1**
Ø 0.5mm

· Small cavities








Code **C1**
Ø 1.2mm

· Angled cavity preparation








Code **CH1-M**
1.0mm x 0.5mm

· Small interproximal cavities (mesial)








Code **CH1-D**
1.0mm x 0.5mm

· Small interproximal cavities (distal)









Code **CH4-M**
1.6mm x 0.8mm

· Large interproximal cavities (mesial)









Code **CH4-D**
1.6mm x 0.8mm

· Large interproximal cavities (distal)

Safety and precision are all I need in my daily practice as a general practitioner. Being able to work with CVDentus makes every procedure I perform easier, from the simplest to the most complex, in addition to being completely portable and having **an innovative and bold design.**

PROF. **GUSTAVO PETRIS**
Master's in Clinical Dentistry - Biomaterials and Specialist in Implantology.





Photos: Gustavo Petris -
Molar wear in a hard-to-reach area while preserving the adjacent tooth
Code CH4-M and CH4-D


In Pediatric Dentistry I recommend the use of the CVDentus R0,5 diamond tip: developed with high technology, it has the perfect size and angulation for small cavity preparations, enabling minimally invasive procedures.

DR. LIETE YASSUDA
Specialist in Pediatric Dentistry and Periodontics

GROOVE PREPARATION WITH MORE INCIPIENT LESIONS








Code **T1**
Ø 0.5mm

ANGLE PREPARATION








Code **C22-E**
Ø 1.2mm









Code **C1**
Ø 1.2mm

BEVELING







Code **C1**
Ø 1.2mm


> With CVDentus diamond, it is possible to perform **treatments and preparations that no other tip can do.** It is spectacular!
>
> PROF. **DIRCEU VIEIRA**
>
> PhD and Master's in Restorative Dentistry, Coordinator of the Specialization courses at ABO Osasco, Santos and ABS - São Paulo.

REMOVAL OF RESTORATIVE MATERIAL

High wear








Code **FR4-H**
Ø 1.6mm

· Rotary bur. Indicated for the lingual surface of anterior teeth and occlusal adjustment







Code **TR18-H**
Ø 0.5mm

· Rotary bur








Code **TR38-H**
Ø 1.0mm

· Rotary bur

Wear








Code **CR4**
Ø 1.6mm

· Finishing the removal








Code **F4**
Ø 1.6mm

· Removal on the lingual surface of anterior teeth or the occlusal surface

Fine wear








Code **T1F**
Ø 0.5mm

· Fine wear in the interpapillary region, supra- and subgingival








Code **T1F-2**
Ø 0.5mm

· Fine wear in the interpapillary region, supra- and subgingival on the lingual surface









Code **T1F-L**
Ø 0.5mm

· Fine wear in the interpapillary region, supra- and subgingival on the buccal surface








Code **T1F-R**
Ø 0.5mm

· Fine wear in the interpapillary region, supra- and subgingival on the buccal surface

Wear refinement








Code **T0U**
Ø 0.25mm

· Delicate refinement in interproximal areas








Code **FR4U**
Ø 1.6mm

· Refinement on the lingual surface of anterior teeth or the occlusal surface

Amalgam removal








Code **T1**
Ø 0.5mm

· Small area








Code **W1**
0.5mm x 2.5mm

· Large area








Code **W1-O**
0.5mm x 2.5mm

· Large area

PAPILLARY CONTOURING









Code **R1**
Ø 1.0mm








Code **R1-E**
Ø 1.0mm

TEXTURING OF RESTORATIVE MATERIAL








Code **T1F**
Ø 0.5mm

TOOTH SECTIONING







Code **VR04XR**
1.6mm x 0.4mm

PROSTHODONTICS




cvdentus

PREPARATION FOR VENEERS AND CONTACT LENSES

High wear








Code **TR18-H**
Ø 0.5mm

· Rotary bur








Code **TR38-H**
Ø 1.0mm


· Rotary bur













Code **C10.5G-H**
0.5mm



Code **C11.0G-H**
1.0mm


Code **C11.5G-H**
1.5mm



Code **C12.0G-H**
2.0mm


· Rotary bur.
Has a standardized active-area depth for wear, ensuring precision in the preparation and fit of the restoration.








Photos: Prof. Celso Sakakura






Photos: Prof. Celso Sakakura








Code **CR1XR**
Ø 1.2mm








Code **CR4XR**
Ø 1.6mm

Preparation finish line






Code **CH1-M**
1.0mm x 0.5mm

· Interproximal wear (mesial). Narrow margin





Code **CH1-D**
1.0mm x 0.5mm

· Interproximal wear (distal). Narrow margin




Photo: Prof. Renato Voss
Preparation finish line
Code CR1



Photo: Prof. Thiago Ottoboni -
Preparation finish line
Code CR4






Code **CR1**
Ø 1.2mm


· Wear. Narrow margin






Code **CH4-M**
1.6mm x 0.8mm


· Interproximal wear (mesial). Wide margin






Code **CH4-D**
1.6mm x 0.8mm

· Interproximal wear (distal). Wide margin






Code **CR4**
Ø 1.6mm

· Wear. Wide margin

Preparation refinement








Code **CR1U**
Ø 1.2mm

· Refinement and aid for hybridization of the immediate dentin sealing. Narrow margin








Code **CR4U**
Ø 1.6mm

· Refinement and aid for hybridization of the immediate dentin sealing. Wide margin








Code **T0U**
Ø 0.25mm

· Delicate refinement in interproximal areas








Code **T1U**
Ø 0.5mm

BEVELING








Code **C1**
Ø 1.2mm

TEXTURING OF RESTORATIVE MATERIAL








Code **T1F**
Ø 0.5mm

VENEER REMOVAL







Code **W1**
0.5mm x 2.5mm



Photo: Prof. Gustavo Greco -
Veneer removal




Photo: Prof. Clébio Domingues -
Veneer removal

> Using the piezo and the CVDentus W1 tip, I can completely remove veneers and contact lenses. I make a section and, with dilation, I perform the debonding. **A very fast, simple procedure that does not damage the tooth.**
>
> PROF. **GUSTAVO GRECO**
> Doctorate in Dentistry, Master's, Specialist in Prosthodontics, Coordinator of the Specialization course in Prosthodontics and Implantology - ABO-MG.

POST REMOVAL








Code **C2T**
Ø 1.6mm

· Removal of cast or threaded metal posts








Code **TOT-E**
Ø 0.4mm

· Removal at the cement line around the metal post and inside the canal up to the middle third / Completion of fiberglass post removal








Code **T1T-E**
Ø 0.65mm

· Removal at the cement line around the metal post / Completion of fiberglass post removal








Code **TR1-E**
Ø 0.8mm

· Performs wear of the fiberglass post, aiding its removal

CEMENTATION OF CONTACT LENSES AND PROSTHESES







Code **CM4**
Ø 3.7mm



Photo: Prof. Ricardo Bertholdi -
Cementation
Code CM4



Photo: Prof. Rogério Marcondes -
Cementation
Code CM4


> The CVDentus CM4 Cementation tip is **a unique product in the world market, providing perfect seating of ceramic restorations,** regardless of the viscosity of the cementing material.
>
> PROF. **ROGÉRIO MARCONDES**
> Specialist in Restorative Dentistry, Founding Member of Paraná Dental Team, Clinical Director of Dental Esthetic Center and Professor at Dental Esthetic Center Training.

REMOVAL OF EXCESS MATERIAL

High wear








Code **FR4-H**
Ø 1.6mm

· Rotary bur. Indicated for the lingual surface of anterior teeth and occlusal adjustment








Code **TR18-H**
Ø 0.5mm

· Rotary bur








Code **TR38-H**
Ø 1.0mm

· Rotary bur

Wear








Code **F4**
Ø 1.6mm

· Indicated for the lingual surface of anterior teeth or the occlusal surface

Fine wear








Code **T1F**
Ø 0.5mm

· Fine wear in the interpapillary region, supra- and subgingival








Code **TR1F-RV**
0.5mm x 2.5mm

· Has no diamond on the tip, performing cervical finishing in composite resin restorations

Wear refinement








Code **FR4U**
Ø 1.6mm

· Indicated for the lingual surface of anterior teeth and the occlusal surface








Code **CR1U**
Ø 1.0mm








Code **CR4U**
Ø 1.6mm








Code **T0U**
Ø 0.25mm

· Only for hard-to-reach areas








Code **T1U**
Ø 1.0mm

SLICE / INTERPROXIMAL WEAR








Code **FR02XR**
1.6mm x 0.2mm








Code **FR04XR**
1.6mm x 0.4mm

TOOTH SECTIONING







Code **TR18-H**
Ø 0.5mm

· Rotary bur








Code **TR38-H**
Ø 1.0mm

· Rotary bur








Code **VR04XR**
1.6mm x 0.4mm








Code **W1**
0.5mm x 2.5mm







Code **W1-O**
0.5mm x 2.5mm.

SURGERY




cvdentus

OSTEOTOMY AND OSTEOECTOMY








Code **SF11**
0.5mm

· High bone-cutting efficiency








Code **SF3**
0.28mm

· High bone-cutting efficiency








Code **SF38**
0.28mm

· High bone-cutting efficiency








Code **SF32**
0,3mm

· High bone-cutting efficiency in hard-to-reach locations








Code **SFR3**
0.5mm

· High bone-cutting efficiency.
· Has laser marking for depth control








Code **SFR4**
0.45mm

· High bone-cutting efficiency








Code **SFR5**
0.35mm

· High bone-cutting efficiency.
· Has laser marking for depth control







Code **SFR42**
0.60mm

· High bone-cutting efficiency in hard-to-reach locations








Code **SFR43**
0.40mm

· High bone-cutting efficiency in hard-to-reach locations








Code **SF3-R**
0.40mm

· Lateral bone cutting, for hard-to-reach areas








Code **SF3-L**
0.40mm

· Lateral bone cutting, for hard-to-reach areas








Code **SF32-O**
0.40mm

· Orthogonal bone cutting, for hard-to-reach areas








Code **SF38-O**
0.28mm

· Orthogonal bone cutting, for hard-to-reach areas








Code **SFTV2-O**
0.40mm

· Osteotomy around apices or fractured implants
· Opening of surgical window

For delicate areas








Code **SFTV-O**
0.40mm








Code **SF3F**
0.40mm








Code **SF3F-O**
0.40mm

· Orthogonal bone cutting








Code **SL1**
0.28mm







Code **SL2**
0.30mm




Photo: Profs. Alan Costa and Shaban Burgoa -
Bone-guided regularization
Code SF3 and SF38





Photo: Prof. Ricardo Scotton -
Lateralization of the Inferior Alveolar Nerve
Code SF3 and SF38

> I use the SF3 and SF38 tips because of their **high cutting power,** always preserving the greatest amount of material with great predictability and safety.
>
> PROF. **RICARDO SCOTTON**
> Master's and Specialist in Implantology, Coordinator of the Advanced Surgery course In Surgery.

SCRAPING / SURFACE REGULARIZATION AND AUTOGENOUS BONE HARVESTING









Code **SFS3-O**
0.4mm

· Scraping and autogenous bone harvesting








Code **SR11**
Ø 3.4mm

· Scraping and surface regularization








Code **SE9F**
Ø 3.9mm

· Scraping for finishing and regularization of the bone surface








Code **SE9G**
Ø 3.9mm

· Scraping and regularization of the bone surface








Code **SL3G**
1.3mm

· Scraping and regularization of the bone surface








Code **SCP**
Ø 3.0mm
3.5mm
3.8mm
4.1mm
4.8mm

· Block-type harvesting
· Has mm markings to aid depth measurement








Code **SCPF**
Ø 3.0mm
3.5mm
3.8mm
4.1mm
4.8mm

· Block-type harvesting.
· More delicate bone drilling
· Has mm markings to aid depth measurement

BONE CAVITY CLEANING







Code **ST10G**
Ø 3.5mm

I found the ST10G insert spectacular due to the size and anatomy of its active part. It allows excellent and quick cleaning of bone cavitations/NICOs... In addition, the internal irrigation with double perforation (at the apex and on the side of the active part) ensures great cooling during the procedure. Highly recommended!!

PROF. **CÉSAR PANIGHEL**

Specialist and Master's in Implantology; Postgraduate in Nutritional Adequacy and Maintenance of Homeostasis - Prevention and Treatment of Age-Related Diseases; Clinical Hypnosis; Biological Dentistry; Partner in the development of the ST10G insert.






Photo: Prof. César Panighel

SOCKET CLEANING







Code **SCR7**
Ø 2,5mm

The SCR7 performs an **excellent decortication and removal of the periodontal ligament from extraction sockets,** facilitating the osseointegration of ceramic implants.

DR. **FERNANDO AVELAR**

Ceramic implantologist - Swiss BioHealth Academy





Photo: Dr. Fernando Avelar

TREPHINE









Code **SCP**
Ø 3.0mm
3.5mm
3.8mm
4.1mm
4.8mm

· Bone drilling
· Has mm markings to aid depth measurement








Code **SCPF**
Ø 3.0mm
3.5mm
3.8mm
4.1mm
4.8mm

· More delicate bone drilling
· Has mm markings to aid depth measurement

SINUS MEMBRANE ELEVATION








Code **SE1**
Ø 3.9mm








Code **SLO-2**
0.70mm x 3.5mm

FLAP ELEVATION








Code **SLTV-O**
0,40mm

BONE DRILLING / IMPLANT SITE PREPARATION








Code **SR2**
Ø 1.3mm

· Has mm markings to aid depth measurement









Code **SR4**
Ø 1.5mm

· Has mm markings to aid depth measurement









Code **SR9**
Ø 2.5mm








Code **SR11**
Ø 3.4mm








Code **STR1**
Ø 0.6mm

· Has mm markings to aid depth measurement








Code **STR4**
Ø 1.5mm

· Has mm markings to aid depth measurement








Code **SI**
Ø 3.0mm
3.5mm
3.8mm
4.1mm
4.8mm

· Guide for SCP








Code **SCP**
Ø 3.0mm
3.5mm
3.8mm
4.1mm
4.8mm

· Has mm markings to aid depth measurement








Code **SCPF**
Ø 3.0mm
3.5mm
3.8mm
4.1mm
4.8mm

· More delicate bone drilling
· Has mm markings to aid depth measurement








Code **SF2**
0.6 /1.0mm

· Has 3.5 mm of active-area depth






Code **ST23**
Ø 2.45mm

· Long tip for drilling in hard-to-reach locations. Has depth marking








Code **ST33**
Ø 2.6mm

· Long tip for drilling in hard-to-reach locations. Has depth marking

PERIOTOMY








Code **SFTV-O**
0.40mm

· In all teeth








Code **SLTV-O**
Ø 0.40mm

· In molars

ROOT SECTIONING








Code **SF3**
0.28mm

· Sectioning of ankylosed roots and apicoectomy








Code **SF38**
0.28mm

· Sectioning of ankylosed roots and apicoectomy








Code **SF38-O**
0.28mm

· Sectioning of ankylosed roots and apicoectomy








Code **SFR4**
0.45mm

· Sectioning of ankylosed roots and apicoectomy







Code **SFTV2-O**
0.40mm

· Apicoectomy



Photo: Prof. Túlio Valcanaia -
Detachment and extraction of the premolar
Code SFTV2-O




Photo: Prof. Túlio Valcanaia -
Detachment and extraction of the premolar
Code SFTV2-O

> The SFTV2-O tip can be used both for opening a surgical window, apicoectomy and osteotomy around apices and fractured implants, as well as for delicate osteotomies, **allowing complicated situations to be resolved with minimal bone loss.**
>
> PROF. **TÚLIO VALCANAIA**
>
> Specialist, Master's and Doctorate in Oral and Maxillofacial Surgery and Traumatology, Partner and Director of IOA - Instituto de Odontologia das Américas and Partner and Director of IOA Miami - USA

INDEX

Rotary Burs

TR18-H	ENDO	- High wear for access to the pulp chamber	08
	RESTORATIVE	- Cavity access	26
		- Removal of restorative material	30
	PROSTHO	- Preparation for veneers and contact lenses	34
		- Removal of excess material	38
		- Tooth sectioning	40
TR38-H	ENDO	- High wear for access to the pulp chamber	08
	RESTORATIVE	- Cavity access	26
		- Removal of restorative material	30
	PROSTHO	- Preparation for veneers and contact lenses	34
		- Removal of excess material	38
		- Tooth sectioning	40
FR4-H	RESTORATIVE	- Removal of restorative material on the lingual surface of anterior teeth and occlusal adjustment	30
	PROSTHO	- Removal of excess material on the lingual surface of anterior teeth and occlusal adjustment	38
C10.5G-H **C11.0G-H** **C11.5G-H** **C12.0G-H**	PROSTHO	- Has a depth "stopper", ensuring standardization and control during wear.	34

Clinical Inserts

C1	RESTORATIVE	- Angled cavity preparation	27
		- Angle preparation	29
	PROSTHO/ RESTORATIVE	- Beveling	29/36
C22-E	ENDO	- Pulp chamber preparation	08
	RESTORATIVE	- Angle preparation	29
C2T	ENDO/ PROSTHO	- Removal of cast or threaded metal posts	14/ 37
CR2T-E **C4T** **CH4T-M** **CH4T-L** **CH4T-R**	ENDO	- Cutting and compaction of gutta-percha	11
	PERIO	- Clinical crown lengthening without flap opening. (mini flap) Indicated for wear of thicker bone.	18
CH1-M (mesial) **CH1-D** (distal)	RESTORATIVE	- Opening of small interproximal cavities (mesial or distal)	27
	PROSTHO	- Wear for the preparation finish line with a narrow margin or in the interproximal region (mesial or distal).	35
CH4-M (mesial) **CH4-D** (distal)	RESTORATIVE	- Opening of large interproximal cavities (mesial or distal)	28
	PROSTHO	- Wear for the preparation finish line with a wide margin or in the interproximal region (mesial or distal).	35
CM4	PROSTHO	- Cementation of contact lenses and prostheses	38
CR1XR	RESTORATIVE	- High wear for cavity access and removal of resin and orthodontic bracket residue	26
	PROSTHO	- High wear for preparation of veneers and contact lenses	34
CR1	PERIO	- Gingival peeling. Narrow area	23
	PROSTHO	- Wear for the preparation finish line. Narrow margin	35
CR1U	PROSTHO	- Preparation refinement and aid for hybridization of the immediate dentin sealing. Narrow margin	36
		- Wear refinement when removing excess material	39
CR4XR	RESTORATIVE	- High wear for cavity access and removal of resin and orthodontic bracket residue	26
	PROSTHO	- High wear for preparation of veneers and contact lenses	34
CR4	PERIO	- Gingival peeling. Wide area	23
	PROSTHO	- Preparation finish line (wide margin)	35
	RESTORATIVE	- Wear. Completion of restorative material removal	30
CR4F-ST	ENDO	- Fine wear of the pulp chamber walls without wearing the floor	09
CR4U	PROSTHO	- Preparation refinement and aid for hybridization of the immediate dentin sealing. Wide margin	36
		- Wear refinement for removal of excess material	39
F4	RESTORATIVE/ PROSTHO	- Removal of excess restorative material on the lingual surface of anterior teeth or the occlusal surface	30/ 38
FR4U	RESTORATIVE/ PROSTHO	- Wear refinement of restorative material on the lingual surface of anterior teeth and the occlusal surface	31/39
	PERIO	- Supragingival scaling with diamond on the lingual surface of anterior teeth	21
FR02-XR **FR04-XR**	PROSTHO	- Slice/ Interproximal wear	40
FS-O	PERIO	- Crown lengthening with flap opening (scraping of thick bone)	18
FS2-O	PERIO	- Crown lengthening with flap opening (wider scraping of thick bone)	18
L3S-O	PERIO	- Detachment of the gingiva	18
R0,5	RESTORATIVE	- Small cavity access	26
R1	RESTORATIVE	- Small cavity access	26
		- Papillary contouring	32
R1-A	RESTORATIVE	- Small hard-to-reach cavity access, such as tunnel	26
R1-B	RESTORATIVE	- Small hard-to-reach cavity access, such as the lingual surface of anterior teeth and tunnel-type access	27
R1-E	ENDO	- Wear for pulp chamber preparation, allowing removal of the roof and dentin	08
		- Pulp stone removal	09
		- Canal location	09
	RESTORATIVE	- Papillary contouring	32
R2	ENDO	- Wear for access to the pulp chamber	08
	RESTORATIVE	- Wear. Medium cavities	27
R4	ENDO	- Wear for access to the pulp chamber	08
	RESTORATIVE	- Wear. Large cavities	27
T0T	PERIO	- Delicate conventional supra- and subgingival interproximal scaling (without diamond)	20
T0TC	PERIO	- Supragingival scaling	20
T0T-E **T1T-E**	ENDO	- Canal location	09
		- Dentin triangle removal	10
		- Cleaning of canals in retreatment	13
		- Fractured instrument removal - (TOT-E)	12
	ENDO/ PROSTHO	- Removal at the cement line around the metal post and - inside the canal - (TOT-E). Completion of fiberglass post removal	14/37
T0T-E1 **T0T-E2**	ENDO	- Passive irrigation (PUI and CUI)	13
		- Agitation of cement, acids and cleaning liquids	13
		- Cutting and thermoplasticization of gutta-percha	11
		- Fractured instrument removal	12
T0T-E3	ENDO	- Cleaning of flattened canals	11
		- Cleaning of canals in retreatment	13
T0T-E5	ENDO	- Passive irrigation (PUI and CUI)	13
		- Agitation of cement, acids and cleaning liquids	13
T0U	RESTORATIVE/ PROSTHO	- Delicate refinement in interproximal areas	36/39
	PERIO	- Delicate supra- and subgingival interproximal scaling (with diamond)	21/22
T0U-2	PERIO	- Delicate scaling in the furcation area (with diamond)	22
	ENDO	- Paraendodontic surgery - Root-end preparation in anterior teeth	15
T0U-L **T0U-R**	ENDO	- Paraendodontic surgery - Root-end preparation in posterior teeth	16
T0UX	PERIO	- Delicate supra- and subgingival interproximal scaling (with diamond)	21
T1	RESTORATIVE	- Wear for small cavity	27
		- Removal of small amalgam or restorative material	31
		- Groove preparation with more incipient lesions	29
T1F	PROSTHO/ RESTORATIVE	- Texturing of restorative material	32/36
		- Fine wear for removal of restorative material in the interpapillary region, supra- and subgingival	30/39

T1F-2	ENDO	- Paraendodontic surgery - Root-end preparation in anterior teeth	15
	RESTORATIVE	- Fine wear in the interpapillary region, supra- and subgingival on the lingual surface	30
T1F-E	ENDO	- Canal location - Canal clearing - Cleaning of the isthmus area	9 10 10
	PERIO	- Pulp stone removal - Root planing	22
T1F-L (distal) **T1F-R** (right)	ENDO	- Paraendodontic surgery - Root-end preparation in posterior teeth	16
	RESTORATIVE	- Fine wear in the interpapillary region, supra- and subgingival on the buccal surface	31
T1T	PERIO	- Conventional supragingival scaling (without diamond)	20
T1T-P	PERIO	- Conventional supra- and subgingival scaling (without diamond) for cleaning implants and ceramic parts	20
T1U	PERIO	- Supragingival scaling (with diamond)	21
	PROSTHO	- Wear refinement for removal of excess restorative material and for preparation for veneers and contact lenses	36/39
T1UX	PERIO	- Sub- and supragingival scaling (with diamond)	21/22
T1U-2	PERIO	- Scaling in the furcation area (with diamond)	22
T1U-E	PERIO	- Subgingival scaling (with diamond)	22
TR1-E	ENDO	- Fiberglass post removal	14/37
TR1-PK **VR1-RPK** **VR1-LPK**	PERIO	- Clinical crown lengthening without flap opening (flapless). Performs bone wear without injuring the gingiva.	19
TR12-PK	PERIO	- Clinical crown lengthening without flap opening (flapless). Performs bone wear without injuring the gingiva.	19
TR1F-RV	PROSTHO	- Fine wear at the cervical margin of composite resin restorations	39
TR1F-ST	ENDO	- Lateral cleaning of the isthmus region - Fine wear at the canal entrance	10 10
TR1U-SD	PERIO	- Supra- and subgingival scaling with diamond only on the lateral surfaces	21/22
VR1T	PERIO	- Periotomy in all teeth	23
VR04XR	ENDO	- Bone window opening - Apicoectomy/ root sectioning	15 15
	RESTORATIVE/ PROSTHO	- Tooth sectioning	32/40
W1	ENDO	- Apicoectomy	15
	RESTORATIVE	- Removal of restorative material (large amalgam)	31
	PROSTHO	- Tooth sectioning - Veneer removal	40 37
W1-O	ENDO	- Apicoectomy	15
	RESTORATIVE	- Removal of restorative material (large amalgam)	31
	PROSTHO	- Tooth sectioning	40
W1T-O	PERIO	- Conventional supragingival scaling (without diamond)	20

Surgical Inserts

SCP	- Preparation/ drilling for implant - Autogenous bone harvesting - Trephine	49 45 47
SCPF	- Preparation/ drilling for implant (near soft tissue) - Autogenous bone harvesting (near soft tissue) - Trephine with more delicate action	49 45 47
SCR7	- Socket cleaning	46
SI	- Guide for SCP	48
SE1	- Sinus membrane elevation	47
SE9F	- Scraping and regularization of the bone surface	45
SE9G	- Scraping and regularization of the bone surface	45
SF2	- Bone drilling	49
SF11	- Osteotomy and osteoectomy	42
SF32	- Osteotomy and osteoectomy (hard-to-reach area)	42
SF3 **SF38**	- Osteotomy and osteoectomy - Root cutting. Sectioning of ankylosed root	42 50
SF32-O	- Osteotomy and ostectomy on the orthogonal (hard-to-reach area)	43
SF38-O	- Osteotomy and ostectomy on the orthogonal (hard-to-reach area) - Root cutting. Sectioning of ankylosed root	42 49
SF3F	- Delicate osteotomy and osteoectomy (near soft tissue)	44
SF3F-O	- Delicate osteotomy and ostectomy on the orthogonal (hard-to-reach area near soft tissue)	44
SF3-L (left)	- Lateral osteotomy and osteoectomy (hard-to-reach areas)	43
SF3-R (right)	- Lateral osteotomy and osteoectomy (hard-to-reach areas)	42
SFR3	- Osteotomy and osteoectomy	42
SFR4	- Osteotomy and osteoectomy - Root cutting. Sectioning of ankylosed root	42 50
SFR5	- Osteotomy and osteoectomy	42
SFR42 **SFR43**	- Osteotomy and osteoectomy	43
SFS3-O	- Scraping and autogenous bone harvesting	45
SFTV-O	- Periotomy in all teeth - Delicate corticotomy	49 44
SFTV2-O	- Apicoectomy - Osteotomy around apices or fractured implants/ Opening of surgical window	50 43
SL1	- Delicate osteotomy (near soft tissue)	44
SL2	- Osteotomy and osteoectomy near soft tissue	44
SL3G	- Scraping and regularization of the bone surface	45
SLO-2	- Sinus membrane elevation	47
SLTV-O	- Flap elevation - Periotomy in molars	47 49
SR2	- Bone drilling/ initial preparation for implant placement	48
SR4	- Bone drilling/ initial preparation for implant placement	48
SR9	- Bone drilling/ Bone window opening	48
SR11	- Bone wear/ Bone window opening	45/48
ST23 **ST33**	- Bone drilling/ initial preparation for implant placement	49
STR1 **STR4**	- Bone drilling/ initial preparation for implant placement	48
ST10G	- Bone cavity cleaning	46

cvdentus

@cvdentus

/cvdentus

CVDentus Diamantes

cvdentus.com.br

Estrada Principal Torrão de Ouro, 500 – Torrão de Ouro
12229-840 – São José dos Campos, SP
Phone: +55 (12) 3944-1126

*Images are for illustrative purposes only; dimensions are approximate.

V.01_2026