cvdentus

TIPS

ULTRASONIC & ROTARY



Experience Innovation

CATALOG 2025

OPTYMUS

CONTENTS

01 CVD DIAMOND 05

02 ENDODONTICS 07


03 PERIODONTICS 17

04 INDEX 21


cvdentus

CVD Diamond

CVDentus® CVD Diamond (the process known as Chemical Vapor Deposition) is a 100% Brazilian technology, patented worldwide. Its hardness is close to that of natural diamond and it is 30x harder than tooth enamel (95 GPa vs. 3.3 GPa for dental enamel). Being a unique stone with controlled roughness, in addition to its high durability, its characteristics also enable exclusive procedures and differentiated preparation quality.



Unused



After 5 min. of use

CONVENTIONAL DIAMOND-COATED TIP

· Natural or synthetic diamond powder


· Lower durability

· Surface with irregular edges


· Uneven finish

· Leaves residues

· Direct contact with metal



Unused



After 5 min. of use

CVDENTUS® DIAMOND TIP

· Single diamond stone

· High durability (30x greater)

· Homogeneous surface, free of edges

· Uniform finish

· Contact only with diamond (biologically compatible)

Beyond the **homogeneous surface**, CVDentus® technology allows controlled grain size in the diamond production process, enabling the professional to perform procedures from wear to pre-polishing. There are **five different roughness types** to meet the widest range of needs: Extreme Ultra Fine, Ultra Fine, Fine, Coarse and Extreme Coarse.




EXTREME ULTRA FINE - UX

Ideal for finishing or the stage prior to polishing.




ULTRA FINE - U

Performs the refinement of the preparation




FINE - F

Ideal for fine wear.



COARSE

Performs controlled wear.



EXTREME COARSE - XR

Ideal for heavy wear.

ENDODONTICS




cvdentus

PULP CHAMBER

Access








Code **TR18-H**
Ø 0.5mm

· Rotary bur. High wear for access to the pulp chamber









Code **TR38-H**
Ø 1.0mm

· Rotary bur. High wear for access to the pulp chamber

Pulp chamber preparation








Code **R1-E**
Ø 1.0mm

· Performs wear and enables removal of the chamber roof and dentin









Code **C22-E**
Ø 1.0mm

· Performs wear and enables removal of the chamber roof and dentin

Pulp stone removal








Code **R1-E**
Ø 1.0mm








Code **T1F-E**
Ø 0.5mm


Fine wear of the walls




Code **CR4F-ST**
Ø 1.6mm

· Does not wear the floor, since it has no diamond on the tip

Canal location




Code **R1-E**
Ø 1.0mm




Code **TOT-E**
Ø 0.4mm

· Including MB2




Code **T1T-E**
Ø 0.65mm

Dentin triangle removal



Code **TOT-E**
Ø 0.4mm




Code **T1T-E**
Ø 0.65mm

Canal clearing








Code **T1F-E**
Ø 0.5mm

Refinement of the canal entrance







Code **TR1F-ST**
Ø 1.0mm


· Has no diamond on the tip

MIDDLE THIRD

Isthmus








Code **T1F-E**
Ø 0.5mm

· Cleaning of the isthmus area








Code **TR1F-ST**
Ø 1.0mm

· Has no diamond on the tip, cleaning only the lateral part of the isthmus region

> These 2 inserts are very interesting: the TR1F-ST works in the isthmus region and refines the canal entrance, and the CR4F-ST finishes and smooths the pulp chamber walls, **ensuring the safety of the procedures and extreme quality in the preparations.**
>
> PROF. **RODRIGO VILLAGRA**
> Endodontics specialist

Gutta-percha






Code **T0T-E1**
Ø 0.3mm

· Cutting and thermoplasticization of gutta-percha






Code **T0T-E2**
Ø 0.3mm

· Cutting and thermoplasticization of gutta-percha






Code **CR2T-E**
Ø 1.2mm

· Cutting and compaction of gutta-percha


Cleaning of flattened canals






Code **T0T-E3**
Ø 0.45mm

· Cleaning of hard-to-reach areas






Photos: Prof. Rodrigo Villagra -
Cutting and thermoplasticization of gutta-percha
Code C2T-E and R1-E

Fractured instrument removal






Code **TOT-E**
Ø 0.4mm








Code **T0T-E1**
Ø 0.3mm







Code **T0T-E2**
Ø 0.3mm






Photos: Prof. Rodrigo Villagra -
Pulp chamber cleaning and cutting and thermoplasticization of gutta-percha
Code TOT-E1, TOT-E2 and TOT-E3

> Together with CVDentus tips, ultrasound is a tool of great importance for those seeking **excellence in endodontic treatment**, as it helps in several steps, such as: conservative access surgery, cavity cleaning, PUI, post removal, retreatment, among others.
>
> DR. **RODRIGO VILLAGRA**
> Endodontics specialist

ACTIVATION



Code **T0T-E5**
Ø 0.3mm


· Passive irrigation (PUI and CUI) Agitation of cement, acids and cleaning liquids

Code **T0T-E1**
Ø 0.3mm

· Passive irrigation (PUI) and cement agitation

Code **T0T-E2**
Ø 0.3mm
· Passive irrigation (PUI) and cement agitation

CANAL CLEANING IN RETREATMENT



Code **T0T-E**
Ø 0.4mm

· Removal of gutta-percha from the cervical half at the start of endodontic retreatment


Code **T1T-E**
Ø 0.65mm

· Removal of gutta-percha from the cervical half at the start of endodontic retreatment

Code **T0T-E3**
Ø 0.45mm


· Cleaning of hard-to-reach areas, removing remaining gutta-percha and cement

POST REMOVAL




Code **TOT-E**
Ø 0.4mm

· Removal at the cement line around the metal post and inside the canal up to the middle third / Completion of fiberglass post removal



Code **T1T-E**
Ø 0.65mm


· Removal at the cement line around the metal post / Completion of fiberglass post removal



Code **C2T**
Ø 1.6mm


· Removal of cast or threaded metal posts

Bone window opening



Code **VR04XR**
1.6mm x 0.4mm

Apicoectomy




Code **VR04XR**
1.6mm x 0.4mm



Code **W1**
0.5mm x 2.5mm








Code **W1-O**
0.5mm x 2.5m

Root-end preparation








Code **T0U-2**
Ø 0.25mm








Code **T1F-2**
Ø 0.5mm

Root-end preparation in posterior teeth








Code **T0U-L**
Ø 0.25mm








Code **T0U-R**
Ø 0.25mm








Code **T1F-L**
Ø 0,5mm








Code **T1F-R**
Ø 1,0mm


PERIODONTICS



SCALING WITHOUT DIAMOND

Supragingival






Code **T0T**
Ø 0.3mm








Code **T0TC**
Ø 0.60mm








Code **T1T**
Ø 0.6mm








Code **W1T-O**
0.5mm x 2.5mm

Subgingival







Code **T0T**
Ø 0.3mm


SCALING WITH DIAMOND

Supragingival




Code **TR1U-SD**
Ø 1.0mm

· Has diamond only on the lateral working surfaces



Code **T0UX**
Ø 0.25mm


· Delicate supra- and subgingival interproximal scaling



Code **T1UX**
Ø 0.50mm


· Indicated for all teeth

Subgingival




Code **TR1U-SD**
Ø 1.0mm

· Has diamond only on the lateral working surfaces




Code **T0UX**
Ø 0.25mm

· Delicate supra- and subgingival interproximal scaling



Code **T1UX**
Ø 0.50mm

· Indicated for all teeth




Code **T0U-2**
Ø 0.25mm

· Scaling in the furcation area







Code **T1U-2**
Ø 0.5mm

· Scaling in the furcation area

ROOT PLANING

Code **T1F-E**
Ø 0.5mm

INDEX

Rotary Burs

TR18-H	ENDO	- High wear for access to the pulp chamber	08
	RESTORATIVE	- Cavity access - Removal of restorative material	
	PROSTHO	- Preparation for veneers and contact lenses - Removal of excess material - Tooth sectioning	
TR38-H	ENDO	- High wear for access to the pulp chamber	08
	RESTORATIVE	- Cavity access - Removal of restorative material	
	PROSTHO	- Preparation for veneers and contact lenses - Removal of excess material - Tooth sectioning	
FR4-H	RESTORATIVE	- Removal of restorative material on the lingual surface of anterior teeth and occlusal adjustment	08
	PROSTHO	- Removal of excess material on the lingual surface of anterior teeth and occlusal adjustment	

Clinical Inserts

C22-E	ENDO	- Pulp chamber preparation	08
C2T	ENDO	- Removal of cast or threaded metal posts	14
CR2T-E	ENDO	- Cutting and compaction of gutta-percha	11
CR4F-ST	ENDO	- Fine wear of the pulp chamber walls without wearing the floor	09
R1-E	ENDO	- Wear for pulp chamber preparation, allowing removal of the roof and dentin - Pulp stone removal - Canal location	08 08 09
T0T	PERIO	- Conventional supragingival scaling (without diamond) - Conventional supragingival scaling (without diamond)	18 18
T0TC	PERIO	- Conventional supragingival scaling (without diamond)	18
T0T-E **T1T-E**	ENDO	- Canal location - Dentin triangle removal - Cleaning of canals in retreatment - Fractured instrument removal - (TOT-E) - Removal at the cement line around the metal post and - inside the canal - (TOT-E). Completion of fiberglass post removal	09 09 13 12 14
T0T-E1 **T0T-E2**	ENDO	- Passive irrigation (PUI and CUI) - Cutting and thermoplasticization of gutta-percha - Fractured instrument removal	13 13 11 12
T0T-E3	ENDO	- Cleaning of flattened canals - Cleaning of canals in retreatment	11 13
T0T-E5	ENDO	- Passive irrigation (PUI and CUI) - Agitation of cement, acids and cleaning liquids	13
T0U-2	PERIO	- Scaling in the furcation area (with diamond)	19
	ENDO	- Paraendodontic surgery - Root-end preparation in anterior teeth	15
T0U-L **T0U-R**	ENDO	- Paraendodontic surgery - Root-end preparation in posterior teeth	15
T0UX	PERIO	- Delicate supra- and subgingival interproximal scaling (with diamond)	19
T1UX	PERIO	- Delicate supra- and subgingival interproximal scaling (with diamond)	19
T1F-2	ENDO	- Paraendodontic surgery - Root-end preparation in anterior teeth	15
T1F-E	ENDO	- Canal location and clearing - Cleaning of the isthmus area - Pulp stone removal	10 10 08
	PERIO	- Root planing	20
T1F-L (distal) **T1F-R** (right)	ENDO	- Paraendodontic surgery - Root-end preparation in posterior teeth	15
T1T	PERIO	- Conventional supragingival scaling (without diamond)	18
T1U-2	PERIO	- Scaling in the furcation area (with diamond)	20
TR1F-ST	ENDO	- Lateral cleaning of the isthmus region - Fine wear at the canal entrance	10 10
TR1U-SD	PERIO	- Supragingival scaling (with diamond) - Supra- and subgingival scaling with diamond only on the lateral surfaces	19 19
VR04XR	ENDO	- Bone window opening - Apicoectomy/ root sectioning	14 14
W1	ENDO	- Apicoectomy	14
W1-O	ENDO	- Apicoectomy	15
W1T-O	PERIO	- Conventional supragingival scaling (without diamond)	18

cvdentus

@cvdentus

/cvdentus

CVDentus Diamantes

cvdentus.com.br

Estrada Principal Torrão de Ouro, 500 – Torrão de Ouro
12229-840 – São José dos Campos, SP
Phone: +55 (12) 3944-1126

*Images are for illustrative purposes only; dimensions are approximate.

V.01_2025