cvdentus

ULTRASONIC

TIPS AND BURS



Try the innovation

2023 CATALOG

Clinical PLUS | DentSurg PRO

SUMMARY

01 CVD DIAMOND 04

02 DLC DIAMOND 05

03 ENDODONTICS 07

04 PERIODONTICS 17

05 DENTISTRY 25

06 PROSTHESIS 33


07 SURGERY 41

08 INDEX 48

cvdentus

CVD Diamond


The CVD Diamond (process known as Chemical Vapor Deposition) from CVDentus® is a 100% Brazilian technology, which is patented all over the world. Ihe hardness is close to natural diamond and it is 30x harder than the tooth enamel (95GPa VS. tooth enamel 3,3GPa). Being a single stone with controlled roughness, besides its high durability, its characteristics also allow exclusive procedures and unique prepping.



Unused | After 5 min. of usage

CONVENTIONAL DIAMOND TIP/BUR

- Natural or synthetic diamond powder
- Less durability
- Surface with sharp edges
- Not uniform finishing
- Generates waste (metal and diamond powder)
- It has direct contact with metal of the tip/bur



Unused | After 5 min. of usage

CVDENTUS® DIAMOND TIP/BUR

- Single Diamond stone
- High durability (30x higher)
- Surface controled with no sharp edges
- Uniform finishing
- Contact only with diamond (biologically compatible)

Besides the **uniform surface**, CVDentus® technology allows controlled granulation during the diamond's production process, which allows the professional to perform procedures since abrasion to pre-polishing. There are **five types of roughness** to meet the most diverse needs: Extreme Ultra Thin, Ultra Thin, Thin, Rough and Extreme Rough.



EXTREME ULTRA THIN - UX

Ideal for finishing or a step before polishing



ULTRA THIN - U

Performs the finishing preparation



THIN - F

Ideal for small abrasions



ROUGH

Performs midium abrasion



EXTREME ROUGH - XR

Ideal for high abrasions


DLC Diamond

The coating known as **Diamond Like Carbon** - DLC - from CVDentus®, is a patented technology and it is applied in the form of film in the surgical tips, promoting wear **resistance, reduction of the friction coefficient** and it also has a **bactericidal effect**.
Applied to titanium, a noble material used in aircrafts and satellites, it has characteristics that makes it ideal for manufacturing surgical ultrasonic tips.
It has **low weight and excellent physical and chemical properties, such as high mechanical strength, resistance to high temperature and also to chemical corrosion.**






Tips with different shapes, thicknesses of the active part developed from 0.28mm. diameters that guarantee precision without unnecessary wear, allowing high performance and safe surgical procedures.




DLC HAS AN EXTREMELY LOW FRICTION COEFFICIENT, CLOSE TO TEFLON, ALLOWING LESS DAMAGE TO THE TISSUE AND LESS HEAT GENERATION.

Comparative study of the tribological behavior under hybrid lubrication of diamond-like carbon films with different adhesion interfaces
R.P.C Costa , D.A. Lima-Oliveira , F.R. Marciano , A.O. Lobo ,
E.J. Corat , V.J. Trava- Airoldi



DLC INCREASES THE DURABILITY PART'S BY PROTECTING IT FROM ELECTROCHEMICAL CORROSION.

Crystalline diamond particles into diamond-like carbon films: The influence of the particle sizes on the electrochemical corrosion resistance
F.R. Marciano[a], E.C Almeida[a], D.A Lima-Oliveira[a], E.J. Corat , V.J. Trava-Airoldi[a]



"ACTIVATED" DLC WITH TITANIUM DIOXIDE NANOPARTICLES (TIO2) REDUCES BACTERIAL ACTIVITY BY UP TO 75%

Anticabterial activity of DLS films containing Tio nanoparticles
F.R Marciano[a], D.A. Lima-Oliveira[a], N.S. da-Silva[a], A.V. Diniz , E.J. Corat , V.J. Trava - Airoldi

ENDODONTICS



cvdentus

PULP CHAMBER

Acess



Cod. **TR18-H**
Ø 0,5mm

· Rotation bur. High abrasion to access pulp chamber



Cod. **TR38-H**
Ø 1,0mm

· Rotation bur. High abrasion to access pulp chamber



Cod. **R2**
Ø 1,2mm


· Abrasion to access pulp chamber



Cod. **R4**
Ø 1,6mm

· Abrasion to access pulp chamber

Pulp chamber preparation



Cod. **R1-E**
Ø 1,0mm

· Performs abrasion and allows roof and dentin removal.



Cod. **C22-E**
Ø 1,0mm

· Performs abrasion and allows roof and dentin removal.

Pulp nodule removal



Cod. **R1-E**
Ø 1.0mm



Cod. **T1F-E**
Ø 0,5mm

Thin wall abrasion




Cod. **CR4F-ST**
Ø 1,6mm

· Does not wear out the floor since it does not have diamond on the top.

Canal location




Cod. **R1-E**
Ø 1.0mm



Cod. **TOT-E**
Ø 0,4mm


· Including MV2



Cod. **T1T-E**
Ø 0,65mm

> These 2 inserts are very interesting, the TR1F-ST works in the isthmus area and prepare the canal's entrance, and the CR4F-ST performs finishing and smoothing of the pulp chamber walls, ensuring safe procedure and extreme quality preparing.
>
> PROFESSOR **RODRIGO VILLAGRA**
> Specialist in Endodontics

Dentin trigone removal



Cod. **TOT-E**
Ø 0,4mm



Cod. **T1T-E**
Ø 0,65mm

Canal cleaning



Cod. **T1F-E**
Ø 0,5mm

Canal entrance refinement



Cod. **TR1F-ST**
Ø 1,0mm

· Does not have diamond on the top

MIDDLE THIRD

Isthmus



Cod. **T1F-E**
Ø 0,5mm

· Isthmus area clean



Cod. **TR1F-ST**
Ø 1,0mm


· Does not have diamond on the top, cleaning only the side of the isthmus area.

Gutta-percha



Cod. **T0T-E1**
Ø 0,3mm

· Cutting and thermoplasticizing gutta-percha



Cod. **T0T-E2**
Ø 0,3mm

· Cutting and thermoplasticizing gutta-percha



Cod. **CRZT-E**
Ø 1.2mm


· Cutting and compacting gutta-percha

Flat canal cleaning



Cod. **T0T-E3**
Ø 0,45mm

· Cleaning hard-to-reach areas





Photos: Professor Rodrigo Villagra ·
Cutting and thermoplasticizing gutta-percha
Cod. C2T and R1-EZ


Fractured instrument removal



Cod. **T0T-E**
Ø 0,4mm



Cod. **T0T-E1**
Ø 0,3mm



Cod. **T0T-E2**
Ø 0,3mm






Photos: Professor Rodrigo Villagra
Cleaning the pulp chamber, cutting and thermoplasticizing gutta-percha
Cod. T0T-E1, T0T-E2 and T0T-E3

> Along with CVDentus tips, ultrasound isa very important tool for those seeking **excellence in endodontic treatment,** as it helps in several steps, such as: conservative access surgery, cavity cleaning, PUI, nucleus removal, retreatment, amongothers.
>
> DR. **RODRIGO VILLAGRA**
> Specialist in Endodontics

ACTIVATION



Cod. **T0T-E1**
Ø 0,3mm

· Passive irrigation (P.U.I e C.U.I), agitation of the sealer, acids and cleaning liquids.


Cod. **T0T-E2**
Ø 0,3mm

· IPassive irrigation (P.U.I e C.U.I), agitation of the sealer, acids and cleaning liquids.

Cod. **T0T-E5**
Ø 0,3mm

· Passive irrigation (P.U.I e C.U.I), agitation of the sealer, acids and cleaning liquids.

CANAL CLEANING IN RETREATMENT



Cod. **TOT-E**
Ø 0,4mm

· Gutta-percha removal from the cervical 1/2 at the beginning of endodontic retreatment.

Cod. **T1T-E**
Ø 0,65mm

· Gutta-percha removal from the cervical 1/2 at the beginning of endodontic retreatment.

Cod. **T0T-E3**
Ø 0,45mm

· Cleaning hard-to-reach areas by removing gutta-percha and the remaining sealer.

PIN REMOVAL



Cod. **TOT-E**
Ø 0,4mm

· Removal in the sealer line around the metal pin and inside the canal up to 1/3 average / Finishing the removal of the fiberglass pin



Cod. **T1T-E**
Ø 0,65mm

· Removal in the sealer line around the metal pin / Finishing the removal of the fiberglass pin



Cod. **C2T**
Ø 1,6mm

· Removal of cast threaded metal pins

FIBERGLASS PIN REMOVAL



Cod. **TR1-E**
Ø 0,8mm

· Makes the wear of the pin assisting in its removal





Photos: . Prof. Ellen Mangabeira
FIBERGLASS PIN REMOVAL
Cod. TR1-E

"The TR1-E insert is a differentiator in the market, as it provides fiber post wear while preserving the dentin as much as possible compared to traditional methods, the format, usability and differentiated technology of diamond surface treatment that only CVDENTUS has."

PROFESSOR. **ELLEN MANGABEIRA**

Specialist in Dental Prosthesis APCD SP and Endodontics by USP; Master in Dental Prosthesis by Faculdade São Leopoldo Mandic; Prof. Dental Prosthesis; Partner in the development of the Insert TR1-E

PARENDODONTIC SURGERY

Bone window opening



Cod. **VR04XR**
1.6mm x 0,4mm

Apicoectomy




Cod. **VR04XR**
1.6mm x 0,4mm



Cod. **W1-O**
0.5mm x 2.5mm

Retropreparation




Cod. **T0U-2**
Ø 0,25mm



Cod. **T1F-2**
Ø 0,5mm


Retropreparation in posteriors






Cod. **T0U-L**
Ø 0,25mm






Cod. **T0U-R**
Ø 0,25mm





Cod. **T1F-L**
Ø 0,5mm






Cod. **T1F-R**
Ø 1,0mm

PERIODONTICS



CLINICAL CROWN LENGTHENING (OPENFLAP)

Crown lengthening (Openflap)




Cod. **L3S-O**
0.5mm x 3.5mm

· Flap detachment



Cod. **FS-O**
0.4mm x 2.5mm

· Strait bone volume reduction (Scraping)



Cod. **FS2-O**
3.5mm x 0,4mm

· Bone volume reduction (Scraping)

Crown lengthening (Flapless)



Cod. **TR1-PK**
Ø 0,5mm

· Performs the bone wear ness without injuring the gingiva (Flapless)
· It has 3mm of diamond and a depth marking at 4 and 5mm



Cod. **VR1-RPK**
Ø 1.2mm

· Performs the bone wear ness without injuring the gingiva (Flapless)
· It has 3mm of diamond and a depth marking at 4 and 5mm



Cod. **VR1-LPK**
Ø 1.2mm

· Performs the bone wear ness without injuring the gingiva (Flapless)
· It has 3mm of diamond and a depth marking at 4 and 5mm



Cod. **TR12-PK**
Ø 0,5mm

· Performs the bone wear ness without injuring the gingiva (Flapless)
· It has 2mm of diamond and a depth marking at 3 and 4mm



Cod. **C4T**
Ø 1,6mm

- Indicated for thicker bone. It has diamond only at the top and performs bone abrasion through an opening in the gingiva (mini flap) for the tip insertion.
- It has a depth marking at 4 and 5mm



Cod. **CH4T-M**
Ø 1,6mm

- Indicated for thicker bone. It has diamond only at the top and performs bone abrasion through an opening in the gingiva (mini flap) for the tip insertion.
- It has a depth marking at 4 and 5mm



Cod. **CH4T-L**
Ø 1,6mm

- Indicated for thicker bone. It has diamond only at the top and performs bone abrasion through an opening in the gingiva (mini flap) for the tip insertion.
- It has a depth marking at 4 and 5mm



Cod. **CH4T-R**
Ø 1,6mm

- Indicated for thicker bone. It has diamond only at the top and performs bone abrasion through an opening in the gingiva (mini flap) for the tip insertion.
- It has a depth marking at 4 and 5mm



Photo: Professor Paulo Kano -
Detail of the excess gum removed
Cod. TR1-PK



Photo: Professor Marcelo de Faveri -
Flapless surgery
Cod. VR1-LPK



Photo: Professor Paulo Kano -
Postoperative (Day 1)
Cod. VR1-LPK

"I use **CVDentus diamond inserts** to increase crown without flap opening, without damage to gingival tissue, better precision, **providing more comfort to the patient** with a faster postoperative period."

PROFESSOR **PAULO KANO**
Master in Implantology, Director of the Paulo Kano Institute, Creator

SCALING WITHOUT DIAMOND

Supragingival




Cod. **T0T**
Ø 0,3mm

· Delicate conventional interproximal scaling (no diamond)



Cod. **T1T**
Ø 0,6mm



Cod. **T1T-P**
Ø 0,8mm

· For cleaning implants and ceramic parts




Cod. **W1T-O**
0.5mm x 2.5mm

Subgingival



Cod. **T0T**
Ø 0,3mm

· Delicate conventional interproximal scaling (no diamond)



Cod. **T1T-P**
Ø 0,8mm

· For cleaning implants and ceramic parts

Supragingival



Cod. **T0UX**
Ø 0.25mm

· Delicate interproximal scraping (with diamond)



Cod. **T0U**
Ø 0.25mm

· Delicate interproximal scraping (with diamond)



Cod. **T1U**
Ø 0.5mm

· Suitable for all elements



Cod. **TR1U-SD**
Ø 1.0mm

· It had diamond only at the sides



Cod. **FR4U**
Ø 1.6mm

· Cleaning on the lingual surface of anterior teeth

Subgingival



Cod. **T0UX**
Ø 0.25mm

· Delicate interproximal scraping (with diamond)



Cod. **T0U**
Ø 0.25mm

· Delicate interproximal scraping (with diamond)




Cod. **T1U-E**
Ø 0,5mm



Cod. **TR1U-SD**
Ø 1,0mm

· It has diamond only on the action sides



Cod. **T0U-2**
Ø 0,25mm

· Delicate scaling on furc (with Diamond)



Cod. **T1U-2**
Ø 0,5mm

· Scraping in the furcation region

ROOT PLANING



Cod. **T1F-E**
Ø 0,5mm

PERITOMY




Cod. **VR1T**
1,2mm x 0,4mm

· In all elements



Cod. **CR1**
Ø 1.2mm

· For restricted areas



Cod. **CR4**
Ø 1.6mm

· For wide areas

Among the most diverse applications, CVDentus tips also allow melanoplasty to be performed with **great safety and comfort for the patient.**

PROFESSOR **DIEGO BAZAN**
Dental surgeon, specialist in Oral and Maxillofacial Surgery, Periodontics and Implantology.





Photos: Diego Bazan -
Gingival peeling and postoperative
Cod. CR1 and CR4

I use diamond tips because they are better, effectively proven. Better than laser peeling surgery or an orban scalpel. The pigments are slow to reappear and many times they never appear again.

DRA. **DÉBORA AYALA**
Master in Laser, Specialist in Periodontics, Specialist in Restorative Dentistry and Specialist in Implantology

DENTISTRY



cvdentus


CAVITY OPENING

High abrasion





Cod. **TR18-H**
Ø 0,5mm

· Rotation bur





Cod. **TR38-H**
Ø 1,0mm

· Rotation bur





Cod. **CR1XR**
Ø 1,2mm


Cod. **CR4XR**
Ø 1,6mm

Abrasion


Cod. **R0,5**
Ø 0,5mm

· Small cavities


Cod. **R1**
Ø 1,0mm

· Small cavities

Cod. **R1-A**
Ø 1,0mm

· Small cavities that are difficult to access as in tunnel



Cod. **R1-B**
Ø 1,0mm

· Small, difficult-to-access cavities such as the lingual surface of anterior teeth and tunnel-type access



Cod. **R2**
Ø 1,2mm

· Medium cavities



Cod. **R4**
Ø 1,5mm

· Large cavities



Cod. **T1**
Ø 0,5mm

· Small cavities



Cod. **C1**
Ø 1,2mm

· Angled cavity preparation



Cod. **CH1-M**
1.0mm x 0.5mm

· Small interproximal cavities (mesial)



Cod. **CH1-D**
1.0mm x 0.5mm

· Small interproximal cavities (mesial)



Cod. **CH4-M**
1.6mm x 0.8mm

· Large interproximal cavities (mesial)




Cod. **CH4-D**
1.6mm x 0.8mm

· Large interproximal cavities (distal)

Safety and precision is all I need in my daily practice, as a general practitioner. Being able to work with CVDentus brings ease in all procedures that I perform, from the simplest to the most complex, in addition to being completely portable and having an **innovative and bold design.**

PROFESSOR **GUSTAVO PETRIS**
Master in Dental Clinic - Biomaterials and Specialist in Implantology.





Photos: Gustavo Petris -
Molar abrasion in difficult-to-access area with adjacent tooth conservation
Cod. CH4-M and CH4-D

In Pediatric Dentistry, I recommend using the CVDentus R0.5 diamond tip: developed with high technology, it has the perfect size and angulation for small cavity preparations, allowing minimally invasive procedures.

DRA. **LIETE YASSUDA**
Specialist in Pediatric Dentistry and Periodontics

PREPARATION IN GROOVES WITH MORE INCIPIENT LESIONS



Cod. **T1**
Ø 0,5mm

PREPARATION IN ANGLES



Cod. **C22-E**
Ø 1,2mm



Cod. **C1**
Ø 1,2mm

BEVELLING




Cod. **C1**
Ø 1,2mm

> With CVDentus diamond, it is possible to make **treatments and preparations that no other tip can do**. It's spectacular!
>
> PROFESSOR **DIRCEU VIEIRA**
> Doctor and Master of Dentistry, Coordinator

REMOVAL OF RESTORING MATERIAL

Amalgam removal



Cod. **T1**
Ø 0,5mm

· Small area


Cod. **W1**
0.5mm x 2.5mm

· Large area

Cod. **W1-O**
0.5mm x 2.5mm

· Large area

High abrasion



Cod. **FR4-H**
Ø 1.6mm

· Rotation bur. Indicated for lingual aspect of teeth and occlusal adjustment


Cod. **TR18-H**
Ø 0,5mm

· Rotation bur

Cod. **TR38-H**
Ø 1,0mm

· Rotation bur

Abrasion



Cod. **CR4**
Ø 1.6mm

· Finishing the removal

Cod. **F4**
Ø 1.6mm

· Lingual removal of anterior or occlusal teeth

Thin abrasion



Cod. **T1F**
Ø 0,5mm

· Fine abrasion in the intrapapillary, supra and subgingival region



Cod. **T1F-2**
Ø 0,5mm

· Fine abrasion in the intrapapillary, supra and subgingival region on the lingual surface



Cod. **T1F-L**
Ø 0,5mm

· Fine abrasion in the intrapapillary, supra and subgingival region on the buccal surface



Cod. **T1F-R**
Ø 0,5mm

· Fine abrasion in the intrapapillary, supra and subgingival region on the buccal surface

Finishing the abrasion



Cod. **T0U**
Ø 0,25mm

· Delicate refinement in interproximal areas




Cod. **FR4U**
Ø 1,6mm

· Finishing the lingual surface of anterior or occlusal teeth

PAPILLARY CONTOUR



Cod. **R1**
Ø 1,0mm



Cod. **R1-E**
Ø 1,0mm

TEXTURING IN RESTORING MATERIAL



Cod. **T1F**
Ø 0,5mm

ODONTOSECTION



Cod. **VR04XR**
1,6mm x 0,4mm

PROSTHESIS



cvdentus

PREPARATION FOR LAMINATED AND VENEERS

High abrasion



Cod. **TR18-H**
Ø 0.5mm

· Rotation bur



Cod. **TR38-H**
Ø 1.0mm

· Rotation bur



Cod. **CR1XR**
Ø 1.2mm



Cod. **CR4XR**
Ø 1.6mm

Finishing preparation



Cod. **CH1-M**
1.0mm x 0.5mm

· Interproximal (mesial) abrasion. Narrow margin



Cod. **CH1-D**
1.0mm x 0.5mm

· Interproximal (distal) abrasion. Narrow margin



Photo: Professor Renato Voss -
Finishing preparation
Cod. CR1



Photo: Professor Thiago Ottoboni -
Finishing preparation
Cod. CR4



Cod. **CR1**
Ø 1.2mm

· Abrasion: narrow margin

Cod. **CH4-M**
⫞⫟ 1.6mm x 0.8mm

· Interproximal (mesial) abrasion. Wide margin


Cod. **CH4-D**
⫞⫟ 1.6mm x 0.8mm

· Interproximal (distal) abrasion. Wide margin

Cod. **CR4**
Ø 1.6mm

· Abrasion: large margin

Pre-polishing



Cod. **CR1U**
Ø 1.2mm

· Refinement and aid for the hybridization of the immediate dentin seal. Narrow margin

Cod. **CR4U**
Ø 1.6mm

· Refinement and aid for the hybridization of the immediate dentin seal. Narrow margin

Cod. **T0U**
Ø 0.25mm

· Delicate refinement in interproximal areas

Cod. **T1U**
Ø 0.5mm

BEVELLING



Cod. **C1**
Ø 1,2mm

TEXTURING IN RESTORING MATERIAL



Cod. **T1F**
Ø 0,5mm

VENEER REMOVAL



Cod. **W1**
0.5mm x 2.5mm



Photo: Professor Gustavo Greco -
Veneer removal
Cod. W1




Photo: Professor Clébio Domingues -
Veneer removal
Cod. W1

> With the use of the CVDentus piezo and W1 tip, I can completely remove the contact lenses. I make a section, and with the dilation I make the scaling. **A very quick, simple procedure, without damaging the tooth.**

PROFESSOR **GUSTAVO GRECO**
PhD in Dentistry, Master in Dental Prosthesis Specialist, Coordinator of the Specialization Course in Dental Prosthesis and Implantology - ABO-MG.

PINS REMOVAL



Cod. **C2T**
Ø 1,6mm

· Cast or threaded metal pins removal

Cod. **TOT-E**
Ø 0,4mm

· Removal the cement line around the metal pin and inside the canal up to 1/3avg. Finalize the fiberglass pin removall

Cod. **T1T-E**
Ø 0,65mm

· Removal the cement line around a metal post / Finishing the removal of the fiberglass pin

Cod. **TR1-E**
Ø 0,8mm

· Makes the wear of the fiberglass pin helping to remove it

VENEERS AND PROSTHESIS CEMENTATION



Cod. **CM4**
Ø 3,7mm



Photo: Professor Ricardo Bertholdi - **Cementation**
Cod. CM4



Photo: Professor Rogério Marcondes - **Cementation.**
Cod. CM4

CVDentus' CM4 Cementing tip is a **unique product on the world market, which provides a perfect settlement of ceramic artifacts,** egardless of the viscosity of the cementing material.


PROFESSOR **ROGÉRIO MARCONDES**

Specialist in Restorative Dentistry. Founding Member of the Paraná Dental Team and Clinical Director of the Dental Esthetic Center and Professor of the Dental Esthetic Center Training.

EXCESS MATERIAL REMOVAL

High abrasion






Cod. **FR4-H**
Ø 1.6mm

· Rotation bur. Indicated for lingual aspect of anterior teeth and occlusal adjustment






Cod. **TR18-H**
Ø 0,5mm

· Rotation bur






Cod. **TR38-H**
Ø 1,0mm

· Rotation bur

Abrasion






Cod. **F4**
Ø 1,6mm

· Indicated for lingual aspect of anterior teeth and occlusal adjustment

Thin abrasion






Cod. **T1F**
Ø 0,5mm

· Thin abrasion in the intra-papillary, supra and subgin-gival region






Cod. **TR1F-RV**
0.5mm x 2.5mm

· It does not have diamond on the top, performing the cervical finishing in composi-te resin restorationsPros

Abrasion finishing






Cod. **FR4U**
Ø 1,6mm


· Indicated for lingual aspect of anterior teeth and occlusal adjustment






Cod. **CR1U**
Ø 1,0mm





Cod. **CR4U**
Ø 1,6mm






Cod. **T0U**
Ø 0,25mm

· Only for difficult-to-access areas





Cod. **T1U**
Ø 1,0mm

SLICE / INTERPROXIMAL ABRASION



Cod. **FR02XR**
1,6mm x 0,2mm



Cod. **FR04XR**
1,6mm x 0,4mm

ODONTOSECTION



Cod. **TR18-H**
Ø 0,5mm

· Rotation bur



Cod. **TR38-H**
Ø 1,0mm


· Rotation bur



Cod. **VR04XR**
1,6mm x 0,4mm



Cod. **W1**
0.5mm x 2.5mm



Cod. **W1-O**
0.5mm x 2.5mm.

SURGERY




cvdentus

OSTEOTOMY AND OSTEOCTOMY








Cod. **SF11**
0,5mm

· High efficiency in bone cutting








Cod. **SF3**
0.28mm

· High efficiency in bone cutting







Cod. **SF32**
0,30mm

· High efficiency in bone cutting for hard acess areas








Cod. **SF32-O**
0.4mm

· Orthogonal bone cut for hard acess areas







Cod. **SF38**
0.28mm

· High efficiency in bone cutting








Cod. **SF38-O**
0.28mm

· Orthogonal bone cut for hard acess areas







Cod. **SF3-R**
0.40mm

· Lateral bone cutting, for hard acess areas








Cod. **SF3-L**
0,40mm

· Lateral bone cutting. for hard acess areas








Cod. **SFR4**
0,45mm

· High efficiency in bone cutting








Cod. **SFR42**
0,60mm

· High efficiency in bone cutting for hard acess areas







Cod. **SFTV2-O**
0,40mm

· Osteotomy around apexes or fractured implants/ Opening a surgical pocket




Photo: Professors Alan Costa e Shaban Burgoa -
Bone-Guided Regularization
Cod. SF3 and SF38




Photo: Professor Ricardo Scotton -
Lateralization of the inferior alveolar nerve
Cod. SF3 and SF38


I use the tips SF3 and SF38 due to their **high cutting power**, always preserving the largest amount of material with great **predictability and safety.**

PROFESSOR **RICARDO SCOTTON**

Master and Specialist in Implantology. Coordinator of the Advanced Surgery In Surgery course.


For delicate areas






Cod. **SFTV-O**
0.40mm





Cod. **SF3F**
0.40mm






Cod. **SF3F-O**
0.40mm

· Orthogonal bone cut






Cod. **SL1**
0.28mm

SURFACE SCRAPING / REGULATION AND AUTOGENOUS BONE COLLECTION






Cod. **SFS3-O**
0.40mm

·Collect autogenous bone (scraping)






Cod. **SCP**
Ø 3.0mm
3.5mm
3.8mm
4.1mm
4.8mm

· Block type collection
· It has a mm mark to help in depth measurement






Cod. **SCPF**
Ø 3.0mm
3.5mm
3.8mm
4.1mm
4.8mm

· Block type collection
· Delicate bone perforation
· It has a mm mark to help in depth measurement






Cod. **SE9F**
Ø 3.9mm

· Scraping for finishing / bone surface smoothing






Cod. **SE9G**
Ø 3.9mm

· Scraping and smoothing bone surface

THEPHINE






Cod. **SCP**
Ø 3.0mm
3.5mm
3.8mm
4.1mm
4.8mm

· Bone perforation
· It has a mm mark to help in depth measurement






Cod. **SCPF**
Ø 3.0mm
3.5mm
3.8mm
4.1mm
4.8mm

· Delicate bone perforation
· It has a mm mark to help in depth measurement


SINUS MEMBRANE DETACHMENT






Cod. **SE1**
Ø 3,9mm





Cod. **SLO-2**
0,70mm x 3,5mm

FLAP DETACHMENT



Cod. **SLTV-O**
0,40mm

BONE PERFORATION / IMPLANT PREPARATION




Cod. **SR2**
Ø 1,3mm

· It has a mm mark to help in depth measurement



Cod. **SR4**
Ø 1,5mm

·It has a mm mark to help in depth measurement



Cod. **SR9**
Ø 2.5mm



Cod. **STR1**
Ø 0,6mm

· It has a mm mark to help in depth measurement



Cod. **STR4**
Ø 1,5mm

· It has a mm mark to help in depth measurement



Cod. **SI**
Ø 3.0mm
3.5mm
3.8mm
4.1mm
4.8mm

· Guide for SCP



Cod. **SCP**
Ø 3.0mm
3.5mm
3.8mm
4.1mm
4.8mm
· It has a mm mark to help in depth measurement



Cod. **SCPF**
Ø 3.0mm
3.5mm
3.8mm
4.1mm
4.8mm
· Delicate bone perforation
· It has a mm mark to help in depth measurement

PERITOMY



Cod. **SFTV-O**
⫞⫟ 0,40mm

· For all elements



Cod. **SLTV-O**
Ø 0,40mm

· In molars

ROOT CUT



Cod. **SF3**
0.28mm

· Ankylosed root fractionation and apicectomy



Cod. **SF38**
0.28mm

· Ankylosed root fractionation and apicectomy



Cod. **SF38-O**
0.28mm

· Ankylosed root fractionation and apicectomy



Cod. **SFR4**
0.45mm

· Ankylosed root fractionation and apicectomy



Cod. **SFTV2-O**
0.40mm

· Apicectomy



Photo: Professor Túlio Valcanaia -
Premolar detachment and extraction
Cod. SFTV2-O



Photo: Professor Túlio Valcanaia -
Premolar detachment and extraction
Cod. SFTV2-O

"The SFTV2-O tip can be used both for opening a surgical store, apicectomy and osteotomy around fractured apexes and implants, as well as for delicate osteotomies **allowing to solve complicated situations with a minimum of bone loss.**"

PROFESSOR **TÚLIO VALCANAIA**

Specialist, Master and Doctor in Oral and Maxillofacial Surgery and Traumatology. Partner and Director of IOA - Institute of Dentistry of the Americas and Partner and Director of IOA Miami-USA

INDEX

Rotation Bur

TR18-H	ENDO	- High abrasion to access pulp chamber	08
	DENTISTRY	- Cavity opening - Removal of restoring material	26 30
	PROSTHESIS	- Preparation for laminated and veneers - Excess material removal - Odontosection	34 38 40
TR38-H	ENDO	- High abrasion to access pulp chamber	08
	DENTISTRY	- Cavity opening - Removal of restoring material	26 30
	PROSTHESIS	- Preparation for laminated and veneers - Excess material removal - Odontosection	34 38 40
FR4-H	DENTISTRY	- Material removal for lingual aspect of teeth and occlusal adjustment	30
	PROSTHESIS	- Excess material removal for lingual aspect of teeth and occlusal adjustment	38

Clinical Inserts


C1	DENTISTRY	- Angled cavity preparation - Angled preparation	27 29
	PROSTHESIS/ DENTISTRY	- Bevelling	36
C22-E	ENDO/ DENTISTRY	- Pulp chamber preparation - Angled preparation	08 29
C2T	ENDO/ PROSTHESIS	- Removal of cast threaded metal pins	14/ 37
CR2T-E	ENDO	- Cutting and compacting gutta-percha	11
C4T **CH4T-M** **CH4T-L** **CH4T-R**	PERIO	- Flapless crown lengthening. (Mini Flap) Indicated for thicker bone abrasion	19
CH1-M (mesial) **CH1-D** (distal)	DENTISTRY	- Opening of large interproximal cavities (mesial or distal)	27
	PROSTHESIS	- Abrasion for finishing preparation in the wide margin or in the interproximal area (mesial or distal)	34
CH4-M (mesial) **CH4-D** (distal)	DENTISTRY	- Opening of large interproximal cavities (mesial or distal)	28
	PROSTHESIS	- Abrasion for finishing preparation in the wide margin or in the interproximal area (mesial or distal)	35
CM4	DENTISTRY	- Veneers and prosthesis cementation	37
CR1XR	DENTISTRY	-High abrasion for cavity opening and resign removal and post bracket	26
	PROSTHESIS	- High abrasion for preparation of laminated and veneers	34
CR1	PERIO	- Gingival peeling. Narrow area	23
	PROSTHESIS	- Abrasion for finishing preparation. Narrow margin.	34
CR1U	PROSTHESIS	- Preparation refinement and aid for the hybridization of the immediate dentin seal. Narrow margin	35
		- Abrasion refinement in excess material removal	39
CR4XR	DENTISTRY	- High abrasion for cavity opening and resign removal and post bracket	26
	PROSTHESIS	- High abrasion for preparation of laminated and veneers	34
CR4	PERIO	- Gingival peeling. Wide area	23
	PROSTHESIS	- Finishing preparation (wide margin)	35
	DENTISTRY	- Abrasion. Finishing the removal of restoring material	30
CR4F-ST	ENDO	- Thin wall abrasion for pulp chamber with no floor abrasion	09
CR4U	PROSTHESIS	- Refinement and aid for the hybridization of the immediate dentin seal. Narrow margin	35
		- Abrasion refinement for excess of restoring material removal	39
F4	DENTISTRY/ PROSTHESIS	- Excess restaurative material removal in the lingual aspect of teeth or occlusal	30/ 38
FR4U	DENTISTRY/ PROSTHESIS	- Excess restoring material removal in the lingual aspect of teeth or occlusal	31/39
	PERIODONTICS	- Supragingival diamond scaling in the lingual aspect of teeth or occlusal	21
FR02-XR **FR04-XR**	PROSTHESIS	- Slice/Interproximal abrasion	40
FS-O	PERIO	- Clinical crown lengthening with openflap (bone volume reduction)	18
FS2-O	PERIO	- Clinical crown lengthening with openflap (wider bone volume reduction)	18
L3S-O	PERIO	- Gingiva detachment	18
R0,5	DENTISTRY	- Small cavity opening	26
R1	DENTISTRY	- Small cavity opening - Papillary contour	26 32
R1-A	DENTISTRY	- Small cavities that are difficult to access as in tunnel	26
R1-B	DENTISTRY	- Small cavities that are difficult to access as in n the lingual aspect of teeth and in tunnel	27
R1-E	ENDO	- Abrasion for pulp chamber preparation, allowing roof and dentin removal. - Pulp nodule removal - Canal location	08 09 09
	DENTISTRY	- Papillary contour	32
R2	ENDO	- Abrasion to access pulp chamber	08
	DENTISTRY	- Abrasion. Medium cavities	27
R4	ENDO	- Abrasion to access pulp chamber	08
	DENTISTRY	- Abrasion. Large cavities	27
T0T	PERIO	- Delicate conventional supra and subgingival interproximal scaling (no diamond)	20
T0T-E **T1T-E**	ENDO	- Canal location - Dentin trigone removal - Canal cleaning in retreatment - Fractured instrument removal (T0T-E)	09 10 13 13
	ENDO/ PROSTHESIS	- Removal in the sealer line around the metal pin and - inside the canal - (T0T-E). Finishing the removal of the fiberglass pin	14/37
T0T-E1 **T0T-E2**	ENDO	- Passive irrigation (P.U.I e C.U.I) - Agitation of the sealer, acids and cleaning liquids. - Cutting and thermoplasticizing gutta-percha - Fractured instrument removal	13 13 11 12
T0T-E3	ENDO	- Flat canal cleaning and retreatment canals	13
T0T-E5	ENDO	- Passive irrigation (P.U.I e C.U.I) - Agitation of the sealer, acids and cleaning liquids.	13 13
T0U	DENTISTRY/ PROSTHESIS	- Delicate refinement in interproximal areas	31/35
	PERIO	- Delicate supra and subgingival interproximal scraping (with diamond)	21
T0U-2	PERIO	- Delicate scaling on furc (with Diamond)	22
	ENDO	- Parendodontic surgery - Retropreparation in anteriors	15
T0U-L **T0U-R**	ENDO	- Parendodontic surgery - Retropreparation in posteriors	16
T0UX	PERIO	- Delicate supra and subgingival interproximal scraping (with diamond)	21
T1	DENTISTRY	- Abrasion. Small cavity. - Small amalgam removal or restoring material - Preparation in grooves with more incipient lesions	27 30 29
T1F	DENTISTRY/ PROSTHESIS	- Texturing in restoring material - Thin abrasion for restoring material removal in the intrapapillary, supra and subgingival region.	32/36 31/38
T1F-2	ENDO	- Parendodontic surgery - Retropreparation in posteriors	15
	DENTISTRY	- Thin abrasion in the intrapapillary, supra and subgingival in the lingual surface.	31

T1F-E	ENDO	- Location and canal cleaning - Isthmus area clean - Pulp nodule removal	10 11 09
	PERIO	- Root planing	22
T1F-L (distal) **T1F-R** (right)	ENDO	- Parendodontic surgery. Retropreparation in posteriors.	16
	DENTISTRY	- Thin wall intrapapillary, supra and subgingival abrasion on the buccal surface	31
T1T	PERIO	- Conventional supragingival scaling (without diamond)	20
T1T-P	PERIO	- Conventional supragingival scaling (without diamond)	20
T1U	PERIO	- Supragingival scaling (with diamond)	21
	PROSHTESIS	- Finishing abrasion for excess restauring material removal end preparationg for laminated and veeners.	35
T1U-2	PERIO	- Scraping in the furcation region (with diamond)	22
T1U-E	PERIO	- Subgingival scraping (with diamond)	22
TR1-E	ENDO	- Fiberglass pin removal	14/37
TR1-PK **VR1-RPK** **VR1-LPK**	PERIO	- Clinical crown lenghening (Flapless). Performs the bone wear ness without injuring the gingiva	18
TR12-PK	PERIO	- Clinical crown lenghening (Flapless). Performs the bone wear ness without injuring the gingiva	18
TR1F-RV	PROSTHESIS	- Thin abrasion in the cervical finishing in composite resin restorations	38
TR1F-ST	ENDO	- Isthmus area clean - Thin abrasion in the canal entrance	10 10
TR1U-SD	PERIO	- Supra and subgingival scaling with diamond only at the sides	21/22
VR1T	PERIO	- Peritomy in all elements	22
VR04XR	ENDO	- Bone window opening - Apicoectomy/root sectioning	15 15
	DENTISTRY PROSTHESIS	- Odontosection	32/40
W1	DENTÍSTICA	- Removal of restoring material (large amalgam)	30
	PRÓTESE	- Odontosecção - Remoção de facetas	40 36
W1-O	ENDO	- Apicoectomy - Removal of restoring material (large amalgam)	15
	DENTISTRY		30
	PROSTHESIS	- Odontosection	40
W1T-O	PERIO	- Conventional upragingival diamond scaling (withouth diamond	20

Surgical inserts

SCP	- Preparation/perforartion for implants - Autogenous bone collection - Thephine	47 44 45
SCPF	- Preparation/perforartion for implants (near soft tissue) - Autogenous bone collection (near soft tissue - Thephine with more delicate action	47 44 45
SI	- Guide for SCP	47
SE1	- Sinus membrane detachment	45
SE9F	- Scraping and smoothing bone surface	45
SE9G	- Scraping and smoothing bone surface	45
SF11	- Osteotomy and Osteoctomy	42
SF32	- Osteotomy and Osteoctomy for hard acess areas	42
SF3 **SF38**	- Osteotomy and Osteoctomy - Root cut. Ankylosed root fractionation and apicectomy	42 48
SF32-O	- Osteotomy and Osteoctomy in the orthogonal for hard acess areas	42
SF38-O	- Osteotomy and Osteoctomy - Root cut. Ankylosed root fractionation and apicectomy	42 48
SF3F	- Delicate osteotomy and Osteoctomy (near soft tissue)	44
SF3F-O	- Osteotomy and Osteoctomy in the orthogonal for hard acess areas near soft tissues	43
SF3-L (left)	- Lateral osteotomy and osteoctomy for hard acess areas	42
SF3-R (right)	- Lateral osteotomy and osteoctomy for hard acess areas	42
SFR4	- Osteotomy and osteoctomy - Root cut. Ankylosed root fractionation and apicectomy	43 48
SFR42	- Osteotomy and Osteoctomy for hard acess areas	44
SFS3-O	- Autogenous bone collection and reduction	44
SFTV-O	- Peritomy in all elements - Delicate corticotomy	47 44
SFTV2-O	- Apicoectomy - Osteotomy around apexes or fractured implants/ Opening a surgical pocket	48 43
SL1	- Delicate osteotomy (near soft tissues)	44
SLO-2	- Sinus membrane detachment	45
SLTV-O	- Flap detachment - Peritomy in molars	46 47
SR2	- Bone perforation/ Initial preparation for implant placement	45
SR9	- Bone perforation/ Bone window opening	46
SR4	- Bone perforation/ Initial preparation for implant placement	46
STR1	- Bone perforation/ Initial preparation for implant placement	46
STR4	- Bone perforation/ Initial preparation for implant placement	47







Estrada Torrão de Ouro, 500 / Torrão de Ouro / Postal Code:
12229-840 - São José dos Campos, SP - Brazil
Phone: +55 (12) 3944-1126 / Whatsapp: +55 (12) 99142-7233

*Images are purely illustrative and approximate dimensions

V.06_2023