



PET (Positron Emission Tomography)

Infographic Challenge

Andrea Bertola
Prateek Gupta
Nadine La Salvia
Mariasole Pasinato

History

We can trace back the origins of PET to **1932** when **Anderson** proved the existence of **positrons**, the 'anti-electrons'. After only two decades, **Brownell** and **Sweet** were the first to use positron annihilation radiation for **medical imaging (1951)**.

Physical principles

The basic principle of PET is the detection of photons emitted from the body after the administration of a **radiotracer**, which undergoes radioactive decay and emits **positrons**. The positron travels a short distance before interacting with an **electron** in the surrounding tissue. This results in the **annihilation** of both particles, producing two **photons** traveling in opposite directions.

A ring of **detectors** surrounds the patient: the detectors are typically **scintillation crystals** coupled to **photomultiplier tubes** (PMTs). The photon-crystal interaction produces a burst of light, detected by the PMTs. This **coincidence detection** of opposite photons allows the localization of the annihilation event along the line connecting the two detectors involved.

Applications

Fluorodeoxyglucose (^{18}F -FDG), the most used PET tracer, allows the assessment of **glucose metabolism**. Glucose exhibits different behavior in the presence of cancerous cells, which explains the importance of FDG in **oncology**. Glucose investigation plays a key role in **cardiology** and neurology as well: it helps diagnose **Alzheimer's disease** and locate seizure foci in **epileptic patients**.

Other commonly used types of tracers are Amyloid tracers and Tau tracers, especially in the early detection of **Alzheimer's disease**.

Resources

Anderson, C.D. (1933) The Positive Electron. *Physical Review*, 43, 491-494.
<https://doi.org/10.1103/PhysRev.43.491>

Bloomfield, P. S., Brigadoi, S., Rizzo, G., & Veronese, M. "Basic Neuroimaging – a guide to the methods and their applications" (2021). ISBN: 979-8721692239

Wagner H. N., Jr (1998). A brief history of positron emission tomography (PET). *Seminars in nuclear medicine*, 28(3), 213–220. [https://doi.org/10.1016/s0001-2998\(98\)80027-5](https://doi.org/10.1016/s0001-2998(98)80027-5)

Sweet, W. H. (1951). The uses of nuclear disintegration in the diagnosis and treatment of brain tumor. *The New England journal of medicine*, 245(23), 875–878.
<https://doi.org/10.1056/NEJM195112062452301>

Brownell, G. L. (1999). A history of positron imaging. Physics Research Laboratory, Massachusetts General Hospital, MIT, 1.

Anand, S. S., Singh, H., & Dash, A. K. (2009). Clinical Applications of PET and PET-CT. *Medical journal, Armed Forces India*, 65(4), 353–358.
[https://doi.org/10.1016/S0377-1237\(09\)80099-3](https://doi.org/10.1016/S0377-1237(09)80099-3)

Hoh C. K. (2007). Clinical use of FDG PET. *Nuclear medicine and biology*, 34(7), 737–742.
<https://doi.org/10.1016/j.nucmedbio.2007.07.001>

Cherry, S.R., Dahlbom, M. (2006). PET: Physics, Instrumentation, and Scanners. In: Phelps, M.E. (eds) PET. Springer, New York, NY. https://doi.org/10.1007/0-387-34946-4_1

Schwarz, C.G. (2021) Uses of Human MR and PET Imaging in Research of Neurodegenerative Brain Diseases. *Neurotherapeutics* 18, 661–672.
<https://doi.org/10.1007/s13311-021-01030-9>