EU MDR POST-MARKET SURVEILLANCE (MAUDE + CLINICALTRIALS.GOV)

Device Scope: Ambulatory Insulin Infusion Pumps — OpenFDA MAUDE & ClinicalTrials.gov v2 | Window: Live API Snapshot (2026-11-12 UTC)

OPENFDA MAUDE EVENT SUMMARY

n = 2,056,124

Total Matched Device Reports

- **Malfunction:** 1,786,860 reports
- **Injury:** 266,177 reports
- **Death:** 2,191 reports

TOP PROBLEM MODES (MAUDE)

Trend Baseline Stable (12 mo)

Pumping Stopped	421,798 (20.5%)
Obstruction of Flow	134,822 (6.6%)
Break	127,253 (6.2%)
Adverse Event Without Identified Device or Use Problem	122,700 (6.0%)
Charging Problem	120,157 (5.8%)

CLINICALTRIALS.GOV PMCF & COMPETITOR TRIALS (n = 15)

Completed: 12 | Recruiting: 2 | Terminated: 1

[NCT01742741 | Completed | n=5] Early Feasibility Study 2 of Outpatient Control-to-Range - Testing System...

[NCT04761094 | Completed | n=2,977] Clinical Efficacy of Insulin Pumps in Type 1 Diabetes Mellitus Patients in...

[NCT00891995 | Terminated | n=71] Effect of Metabolic Control at Onset of Diabetes on Progression of Type 1...

[NCT03838900 | Completed | n=1,212] The Loop Observational Study (Jaeb Center for Health Research)

[NCT03815487 | Completed | n=40] Comparison of Two Management Systems in Patients With Type 1 Diabetes (Pediatric...