

Display Time	User	Description
3/24/09 4:05:44 PM	ROYAL,JOANN	Patient admitted to 'TRG-D', Reason: Antepartum testing
3/24/09 4:05:47 PM		Monitor status: On
3/24/09 4:05:47 PM		Monitor ID: M1350B, protocol Revision: A20, software revision: C.01.03, serial number: 3650G17149
3/24/09 4:05:47 PM		Recorder status: Off
3/24/09 4:05:47 PM		FHR1 measurement method: No transducer
3/24/09 4:05:47 PM		FHR2 measurement method: No transducer
3/24/09 4:05:47 PM		FHR3 measurement method: No transducer
3/24/09 4:05:47 PM		MHR measurement method: No transducer
3/24/09 4:05:47 PM		Toco measurement method: No transducer
3/24/09 4:06:05 PM		ALERT SETTINGS Basic Alerting Severe Tachy 180 bpm for 300 sec Tachy 160 bpm for 600 sec Brady 110 bpm for 240 sec Severe Brady 90 bpm for 60 sec Signal Loss 25 % within 3 min NST Off
3/24/09 4:06:43 PM		Toco measurement method: external
3/24/09 4:06:45 PM		ALERT SETTINGS Basic Alerting Severe Tachy 180 bpm for 300 sec Tachy 160 bpm for 600 sec Brady 110 bpm for 240 sec Severe Brady 90 bpm for 60 sec Signal Loss 25 % within 3 min NST Off
3/24/09 4:06:45 PM		FHR1 measurement method: US
3/24/09 4:06:45 PM		FMP Status: On
3/24/09 4:07:33 PM		FHR1: SIGNAL LOSS Bad FHR signal. Please adjust FHR sensor.
3/24/09 4:07:48 PM	ROYAL,JOANN	Alarm acknowledged
3/24/09 4:09:53 PM		Maternal blood pressure: 112 / 72 mmHg, Mean value: 86 mmHg
3/24/09 4:09:53 PM		Maternal heart rate (NIBP): 65 BPM
3/24/09 4:09:54 PM		Recorder status: On
3/24/09 4:10:37 PM		Recorder status: Off
3/24/09 4:10:52 PM	PROPER, YLONDA	Maternal respiration: 20 / min
3/24/09 4:10:56 PM	PROPER, YLONDA	Maternal temperature: 98.0°F
3/24/09 4:20:35 PM	System	Acknowledged: Fetal Monitor on
3/24/09 4:20:35 PM	ROYAL,JOANN	Patient discharged to 'UNDEL'
3/24/09 4:20:35 PM	System	Monitor status: Off
3/24/09 4:20:35 PM	ROYAL,JOANN	Episode closed