



Composite multi-view echocardiographic reference rules for tricuspid regurgitant jet span: simulation results


a Maximum-type rules stay within about ± 1 mm across unknown view accuracy; mean-type rules reach -3 mm



b A percentage beat window adds error to the anchor-view mean



c Fixed-millimetre cross-view triggers detect few true errors and fire on measurement noise



d Protocol-built references inflate independent AI error, hide shared error

