DEPARTMENT OF ENERGY SCIENCE
ALAGAPPA UNIVERSITY
(A State University Accredited with A+ Grade by NAAC (CGPA: 3.64) in the Third Cycle)
KARAIKUDI - 630 003

PEROVSKITE SOLAR CELL FABRICATION LOW HUMIDITY GLOVE BOX

Specifications:

Perovskite solar cell fabrication low humidity glove box designed for clean and sealed working environment suitable for assembling perovskite solar cell. The glove box should have large effective space approximately 1500 x 750 x 900 mm to place assembling equipment and should have three gloves. Perovskite solar cell assembly equipment like Laboratory Hot Plate with aluminium/silicon alloy top plate for heat transmission, plate dimension of 150 x 150 mm should be given.

Vacuum Chuck Spin Coater with substrate size up to 25 x 15 mm, maximum 8000 RPM, Max Rotation time of 3000 s, 0.06 to 0.09MPa vacuum input, > 15l/min pump speed.

Pipette with one hand volume adjustment, volume 10-100µl, subdivision 0.1 µl and tip size of 2-200 µl, 5-300µl, filter tips of 2-20 µl and 5-100µl.

Analytical balance with scale range from 0.001g to 200g and 0.0001g readability. Max capacity: 200 g, min capacity 10 mg, setting time < 8 s, operating temperature of 17.5 to 22.50C.

Ultrasonic Bath with stainless steel tank and basket. Tank size 150 x 135 x 99 mm, Ultrasonic power 60 W, heating power 100 W, Tank capacity of 2 L usable capacity of 1.6 L and temperature setting up to 800°C and a Glass Cutter to be provided.