

Compliance

1. Relevancy

1.1 Is the dimension "Compliance" relevant for your application? ☐ Yes [-] ☐ No [-]

2. Documentation

2.1 Is there documentation about **relevant laws and regulations** with respect to the intended use of the data? ☐ Yes [1] ☐ No [0]

2.1.1 If yes, were they implemented? ☐ Yes [1] ☐ No [0]

2.2 Is there documentation about **relevant standards** with respect to the intended use of the data? ☐ Yes [1] ☐ No [0]

2.2.1 If yes, were they implemented? ☐ Yes [1] ☐ No [0]

2.3 Were **permissions** (including ethical, legal, and institutional) obtained to enable dataset creation? [Alderman et al. 2025] ☐ Yes [1] ☐ No [0]

3. Methods and measures

3.1 Is the dataset **standardized to ensure compliance with external reporting requirements**? ☐ Yes [1] ☐ No [0]

3.2 Are processes integrated for handling **personal or sensitive data according to data protection laws** (e.g. GDPR)? [ALTAI, Alderman et al. 2025] ☐ Yes [1] ☐ No [0]

3.3 Is the dataset regularly reviewed for compliance with **changing regulations**? ☐ Yes [1] ☐ No [0]

3.4 Does the manufacturer establish procedures to collect patient data based on well know practices such as **Good Clinical Practice (GCP)**? (e.g. consent of participants is collected and handled accordingly, etc.) [EN ISO 14155, IG_NB] ☐ Yes [1] ☐ No [0]

3.5 Does the manufacturer describe the collected data using **descriptive statistics**? [BS/AAMI 34971, clauses 5.3.2, 5.3.3 ISO/IEC TS 4213, clause 6 [IG_NB]] ☐ Yes [1] ☐ No [0]

4. Definition-specifics

4.1 Anonymization

4.1.1 If patient data is included, was ensured that **patient data is adequately protected** (incl. adequate protection of training data or in case of model extraction)? [BS/AAMI 34971, clause 5.3.4, Regulation (EU) 2016/679 [IG_NB], Alderman et al. 2025] ☐ Yes [1] ☐ No [0]

4.1.2 Were the datasets **de-identified or anonymised** so that individuals cannot be identified? [Healtsheets] ☐ Yes [1] ☐ No [0]

4.2 Consent

4.2.1 Did individuals who are represented in this data **give consent** for using their information developing this use case? [Fehr et al. 2022, Healthsheets] ☐ Yes [1] ☐ No [0]

4.2.2 Were individuals provided with any mechanism to **revoke their consent** in the future or for specific uses? [Fehr et al. 2022, Healthsheets] ☐ Yes [1] ☐ No [0]

4.3 Ethics

4.3.1 Were any **ethical review processes** conducted (e.g., by an institutional review board)? [Datasheets] ☐ Yes [1] ☐ No [0]

4.4 Risk management

4.4.1 Does the manufacturer assess the risk arising from **imbalanced datasets** based on the intended target patient population of the device? [Regulation (EU) 2017/745, Annex I 1.,3.,5. , (EU) 2017/746, Annex I 1.,3.,5. • EN ISO 14971, clauses 5.3, 5.4, 5.5 [IG_NB]] ☐ Yes [1] ☐ No [0]

4.4.2 Does the manufacturer consider risks arising from differences in demographic, anthropometric, anatomical and physiological differences among **target patient population** when establishing datasets? [Artificial Intelligence (AI) in medical devices Page 7 of 28, ISO/IEC TR 24027, Regulation (EU) 2017/745, Annex I 1., 3., 5., Regulation (EU) 2017/746, Annex I 1., 3., 5. • EN ISO 14971, clause 5.3 [IG_NB]] ☐ Yes [1] ☐ No [0]

4.5 Data storage

4.5.1 Does the manufacturer establish procedures for **data storage and retention according to the applicable regulatory requirements**? [EN ISO 13485, clauses 4.2.4, 4.2.5 [IG_NB]] ☐ Yes [1] ☐ No [0]