




Figure 7: DBS Suppresses Beta Oscillations in Parkinsonian Network



DBS MECHANISM (Simplified Model)

DBS Parameters:

- Target: STN
- Frequency: 130.0 Hz
- Effect: Disrupts STN-GPe loop

Key Results:

- GPe Beta: 6.1% → 5.3%
- Beta suppression: 14%

Interpretation:
DBS disrupts pathological synchrony in the STN-GPe feedback loop, reducing beta oscillations.