EMERGENCY ACTION CARD

bBraun Noradrenaline Pump Failure

 Issue can occur during Noradrenaline pump startup (within first 120mins) leading to unexpected occlusion alarm and/or shutdown.

•Set up Noradrenaline infusion with NO SMARTSITE and 960mmHg pressure limit

• If "double-pumping" from an old infusion, retain the old one for a few hours as backup.

•In case of pump failure - restart old infusion.

• If no infusion available - call for help

Move new infusion to a new pump.

• Medical /ACCP team to assist if required with boluses of vasopressors

Check patient

3

4

5

6.

8.

 Send bBraun to medical physics for check (fill in as much detail as possible about what happened.

• Complete DATIX form

Title: Emergency Action Card in case of bBraun Spaceplus Infusomat Failure	
	Authors: David Briody
Status Draft/Final: DRAFT	Approved by:
Reviewed by:	Written: 14 th December 2021
Reviewed on:	Next review: 14 th January 2022