



MUNSON HOME MEDICAL TRAVERSE CITY
3816 W. FRONT STREET
TRAVERSE CITY
Michigan, 49684

Phone: 231-935-2100
Email: MHH-MHMEsleep@mhc.net

06/08/2023 - 07/07/2023

Compliance Report

Initial compliance period

06/08/2023 - 07/07/2023

Compliance met Yes
Compliance percentage 70%
Payor Standard

Usage

06/08/2023 - 07/07/2023

Usage days 22/30 days (73%)
 >= 4 hours 21 days (70%)
 < 4 hours 1 day (3%)

Usage hours 149 hours 48 minutes
 Average usage (total days) 5 hours 0 minutes
 Average usage (days used) 6 hours 49 minutes
 Median usage (days used) 6 hours 55 minutes
Total used hours (value since last reset - 07/07/2023) 276 hours

AirSense 11 AutoSet

Serial number 23231350306
Mode CPAP
Set pressure 15 cmH2O
EPR Fulltime
EPR level 3

Therapy

Leaks - L/min Median: 34.5 95th percentile: 62.6 Maximum: 72.1
Events per hour AI: 4.0 HI: 2.1 AHI: 6.1
Apnea Index Central: 1.8 Obstructive: 0.4 Unknown: 1.7
RERA Index 1.3
Cheyne-Stokes respiration (average duration per night) 7 minutes (2%)

Usage - hours

