




CHARLESTON 2090 EXECUTIVE HALL RD STE 100 CHARLESTON South Carolina, 29407

Phone: 843-554-4000 Fax: 843-769-6849

Email: tj@atlanticrespiratory.com



Compliance Report

_			
Com	nl	ıan	CO
COILL	\mathcal{O} I	ıaıı	cc

Payor Standard


Usage	04/01/2023 - 10/29/2023
Usage days	207/212 days (98%)
>= 4 hours	200 days (94%)
< 4 hours	7 days (3%)
Usage hours	1,458 hours 37 minutes
Average usage (total days)	6 hours 53 minutes
Average usage (days used)	7 hours 3 minutes
Median usage (days used)	7 hours 9 minutes
Total used hours (value since last reset - 10/29/2023)	1,449 hours

AirSense 10 AutoSet

Serial number	23223420062
Mode	AutoSet
Min Pressure	5 cmH2O
Max Pressure	20 cmH2O
EPR	Fulltime
EPR level	3
Response	Standard

Therapy						
Pressure - cmH2O	Median:	6.4	95th percentile:	9.6	Maximum:	11.1
Leaks - L/min	Median:	7.5	95th percentile: 3	33.2	Maximum:	54.8
Events per hour	AI:	1.3	HI:	8.0	AHI:	2.1
Apnea Index	Central:	0.5	Obstructive:	0.6	Unknown:	0.2
RERA Index						0.1
Chevne-Stokes respira	tion (averag	ne dur	ation per night)		1 minutes	: (0%)

Usage - hours







CHARLESTON 2090 EXECUTIVE HALL RD STE 100 CHARLESTON South Carolina, 29407

Phone: 843-554-4000


Fax: 843-769-6849 Email: tj@atlanticrespiratory.com








Usage (hours)



Leak (L/min)



Pressure (cmH2O)

Mode Set EPR	AutoSet Fulltime, 3.0
Set Max Pressure	20.0
Set Min Pressure	5.0
Maximum (avg)	11.1
95th % (avg)	9.6
Median (avg)	6.4

AHI (events/hour)

