

Equipment Specifications

Scanning Systems: Ultrasonic (XR US)

V19 ★ DENVER CO

554BXXXXX

A. REQUIREMENT OVERVIEW

Facility Name	Rocky Mountain Regional VA Medical Center
Delivery Address	1700 N Wheeling St, Aurora, CO 80045
Requesting Department(s)	Radiology
Quantity	2

B. TECHNICAL REQUIREMENTS

1. Unit physical specifications

a. Minimum screen size [in]	23.5
b. Minimum active screen size [in]	21.5
c. Minimum number of split screens	2
d. Minimum monitor rotation [degrees]	45
e. Monitor height range from floor [in]	
f. Minimum number of transducer ports	10
g. Minimum image storage [days or TB]	
h. Minimum battery life [min]	30
i. Maximum equipment dimensions (HxWxD) [in]	67.5in H x 23.9in W x 43in D
j. Maximum equipment weight [lbs]	350

2. Scanning modes

☒	a. Two-dimensional (2D)
☒	b. Three-dimensional (3D) – freehand
☒	c. Navigation Software
☒	d. Image Fusion Software
☒	e. Tissue Harmonic Imaging
☒	f. M-Mode
☒	g. Simultaneous M-Mode
☒	h. Pulsed Wave Doppler
☒	i. Continuous Wave Doppler
☒	j. Velocity Color Doppler



3. Doppler displays

☒	a. Frequency
☒	b. Velocity
☒	c. Duplex

4. Control panel specifications

☒	a. Physical keyboard
☒	b. Touchscreen monitor
☒	c. Keyboard on touchscreen
☒	d. Articulating support arm to allow for vertical and horizontal adjustment of the monitors for viewing from anywhere in the room

5. Additional specifications

☒	a. Panoramic mode
☒	b. Multi-modality display
☒	c. Image annotation
☒	d. One-button image optimization
☒	e. One-button Doppler optimization
☒	f. Programmable protocols
☒	g. Built-in, customizable PACS-compatible form
☐	h. Image editing – easy video editing and annotation
☐	i. Phantom to be used for quality control testing
☒	j. Digital calipers
☒	k. Other specifications

6. Analysis packages

☒	a. Procedural
☒	b. Abdominal
☒	c. Musculoskeletal Clinical
☒	d. Obstetric
☒	e. Cardiovascular Clinical
☒	f. Cardiac 3D Quantification
☒	g. Other analysis packages

Vendors must include in their offers analysis packages that meet the criteria above. Please include all other analysis packages offered by your company in the optional section on the quotes.



7. Transducers

Description	Qty	Frequency Range [MHz]
a. C5-1	1	5 to 1 MHz
b. C6-2	1	6 to 2 MHz
c. L12-3	1	12 to 3 MHz

Vendors must include in their offers transducers that meet the criteria above. Please include all other transducers offered by your company in the optional section on the quotes.

8. Advanced features

☒	Fusion-Targeted Biopsy System
☒	a. Image storage for any captured images

9. Security/Connectivity requirements

☒	a. OEM-supported operating system
☒	b. Latest DICOM print, store, commit, structured report (SR) and modality worklist
☒	c. Encrypted hard drive
☒	d. PACS compatibility – VA VistA Imaging
☒	e. Wireless connectivity to VA network – Compatible with 802.11b/g/n and FIPS 140-2 or FIPS 140-3 compliant
☒	f. PACS compatibility – [specify vendor]

C. TRAINING REQUIREMENTS

1. Clinical Training

☒	a. On-site clinical applications training for [##] technologists during go-live
☒	b. On-site clinical applications training for [##] physicians during go-live
☒	c. Vendors shall be responsible for accommodating different personnel shifts for clinical applications training during go-live.

2. Biomedical Technician Training

Please reference the “Instructions to Offers” for further information about the type of information to provide by equipment type not by specific request. Please also reference the “Instructions to Offers” for response format.

Technical training information to include detailed information about the curriculum and length of the biomedical technical training required for each equipment type.

Although the NAC will not award this training along with the equipment, it is imperative that the customer is informed that this training is available. Vendors must demonstrate that they can provide any required off-site training, therefore off-site training should be quoted as an optional item. Off-site training will be



purchased at the time of need via a modification (if the original order remains open) or via a separate order. No travel expenses for any VA employees will be included in any HTME equipment or training order.

D. SERVICE REQUIREMENTS

1. VPN/Remote Access – The vendor shall provide any and all equipment service programs, such as remote diagnostics, during the warranty period. The vendor shall provide post-warranty remote diagnostic service program as an “Add Option” with the offer. The system shall provide vendor remote diagnostics via VPN. The vendor shall either utilize the VA national site-to-site VPN or work with the Office of Cyber and Information Security and the VAMC Information Systems Security Officer to establish a client-based VPN.
2. Service and Operator Manuals – The vendor shall provide the following documentation for the proposed systems:
 - a. Two (2) copies of operator instruction manuals (one (1) electronic and one (1) physical copy)
 - b. Two (2) copies of a service manuals (one (1) electronic and one (1) physical copy)

*Vendors can include the physical copy as a priced line item in their quote as applicable.

3. Minimum Warranty – The system and accessories shall be covered under the manufacturer’s warranty and shall include all parts and labor for one year following acceptance by the VAMC. This warranty must include PMs as required by the manufacturer. The manufacturer’s factory-trained field service representatives shall perform installation and maintenance during the warranty period.

Vendors are encouraged to include any offerings for service, warranty, and training that may exceed the minimum requirements, to include information on their service support structure during and after the warranty period. Vendors who do not include any added value offerings for service, warranty, and training will not be docked or excluded from consideration. However, any such offerings will be evaluated based on added value.

E. ADDED VALUE

This section can be used to list preferred system specifications. Vendors who cannot meet the specifications listed in this section will not be excluded from consideration; however, vendors who can meet the specifications listed in this section can be evaluated based on the added value these characteristics provide. In compiling this document, feel free to move specifications listed elsewhere to this section and specifications listed in this section to other sections as appropriate for the present requirement.

Specifications listed below are not required, but preferred. Vendors who do not include the below specifications in the submitted offer will not be docked or excluded from consideration. Specifications listed below will be evaluated based on added value.

☒	a. Additional year(s) of warranty
☒	b. Version/platform long-range plan

F. OTHER INFORMATION/DOCUMENTATION REQUIRED



Please reference the “Instructions to Offers” for further information about the type of information to provide by equipment type not by specific request. Please also reference the “Instructions to Offers” for response format.

1. Completed pre-procurement assessment form (6550 Appendix A)
2. Completed Manufacture Disclosure Statement for Medical Device Security (MDS2) form
3. Federal Information Processing Standard (FIPS) 140-2 or (FIPS) 140-3 certification, if applicable
4. Product brochures
5. Technical specification sheets, to include dimensions and weight of the system
6. Typical drawings (pdf version of the CAD drawings)
7. Technical training- Biomedical: information to include detailed information about the curriculum and length of the biomedical technical training required for each equipment type.
 - Although the NAC will not award this training along with the equipment, it is imperative that the customer is informed that this training is available. Vendors must demonstrate that they can provide any required off-site training, therefore off-site training should be quoted as an optional item. Off-site training will be purchased at the time of need via a modification (if the original order remains open) or via a separate order. No travel expenses for any VA employees will be included in any HTME equipment or training order.
8. Support information to include your company’s support structure during and after the warranty period
 - On-line or telephonic applications support and availability (include third party coverage)
 - A listing of field service engineer locations and availability
 - A listing of part depots

G. TRADE-IN

☒	a. In instances where sanitization of ePHI compromises the OS and/or application software, the operating system and application software will be reloaded by VA or a vendor contracted by VA on the native system post drive sanitization. Verification of system operation is the responsibility of the vendor.
-------------------------------------	--

