

User:	5	Supervisor
Product:	4	Ampoules. 10ml
Batch:	PQ1_TS03	
Notes:	10ML PLAIN GLASS AMPOULE	

17 November 2015 09:41:31

### Date Time

Starting date:	<b>17/11/2015</b>	9:17:05 AM
Ending date:	<b>17/11/2015</b>	9:41:31 AM

### Counters:

INSPECTED	<b>35</b>		
ACCEPTED	<b>22</b>	62.86	%
REJECTED	<b>13</b>	37.14	%
TECNICAL REJECT	<b>0</b>	0.00	%

### Reject details:

ST0	AVAILABLE		ON
<i>Totals:</i>	1	2.86	%
Height	0	0.00	%
Shape	0	0.00	%
Black Spot	1	2.86	%
Colour Rings	0	0.00	%
ST1	AVAILABLE		ON
<i>Totals:</i>	6	17.14	%
Single Particles	3	8.57	%
Total Particles	6	17.14	%
	0	0.00	%
Fill Level	0	0.00	%
ST2	AVAILABLE		ON
<i>Totals:</i>	9	25.71	%
Single Particles	4	11.43	%
Total Particles	9	25.71	%
	0	0.00	%
	0	0.00	%
ST3	AVAILABLE		ON
<i>Totals:</i>	0	0.00	%
Still Particles	0	0.00	%
Tot. Still Particles	0	0.00	%
Charrings	0	0.00	%
Tot. Charrings	0	0.00	%

ST4	AVAILABLE		ON
<i>Totals:</i>	11	31.43	%
Single Particles	7	20.00	%
Total Particles	11	31.43	%
	0	0.00	%
	0	0.00	%
ST5	NOT AVAILABLE		OFF
<i>Totals:</i>	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%

LKD counters			
LKD1			ON
<i>Totals:</i>	0	0.00	%
LKD2			ON
<i>Totals:</i>	0	0.00	%

Approved by:

Verified by: