

User:	1	adamb
Product:	1	Production 1ml
Batch:	challenge 1705165	
Notes:		

21 June 2017 10:32:50

Date Time

Starting date:	21/06/2017	10:19:52 AM
Ending date:	21/06/2017	10:32:50 AM

Counters:

INSPECTED	19		
ACCEPTED	6	31.58	%
REJECTED	13	68.42	%
TECNICAL REJECT	0	0.00	%

Reject details:

ST0	AVAILABLE		ON
<i>Totals:</i>	2	10.53	%
Height	0	0.00	%
Shape	0	0.00	%
Black Spot	2	10.53	%
Colour Rings	0	0.00	%
ST1	AVAILABLE		ON
<i>Totals:</i>	10	52.63	%
Single Particles	6	31.58	%
Total Particles	6	31.58	%
	0	0.00	%
Fill Level	9	47.37	%
ST2	AVAILABLE		ON
<i>Totals:</i>	8	42.11	%
Single Particles	7	36.84	%
Total Particles	8	42.11	%
	0	0.00	%
	0	0.00	%
ST3	AVAILABLE		ON
<i>Totals:</i>	9	47.37	%
Still Particles	9	47.37	%
Tot. Still Particles	6	31.58	%
Charrings	0	0.00	%
Tot. Charrings	0	0.00	%

ST4	AVAILABLE		ON
<i>Totals:</i>	11	57.89	%
Single Particles	11	57.89	%
Total Particles	11	57.89	%
	0	0.00	%
	0	0.00	%
ST5	NOT AVAILABLE		OFF
<i>Totals:</i>	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%

LKD counters			
LKD1			ON
<i>Totals:</i>	1	5.26	%
LKD2			ON
<i>Totals:</i>	0	0.00	%

Approved by:

Verified by: