

User:	5	Supervisor
Product:	2	Ampoules. 2ml
Batch:	PQ1_TS03	
Notes:	2ML PLAIN GLASS AMPOULE	

07 January 2016 02:04:18

Date Time

Starting date:	<b>07/01/2016</b>	1:45:25 AM
Ending date:	<b>07/01/2016</b>	2:04:18 AM

Counters:

INSPECTED	<b>2000</b>		
ACCEPTED	<b>1606</b>	80.30	%
REJECTED	<b>382</b>	19.10	%
TECNICAL REJECT	<b>12</b>	0.60	%

Reject details:

ST0	AVAILABLE		ON
<i>Totals:</i>	305	15.25	%
Height	13	0.65	%
Shape	5	0.25	%
Black Spot	297	14.85	%
Colour Rings	0	0.00	%
ST1	AVAILABLE		ON
<i>Totals:</i>	12	0.60	%
Single Particles	11	0.55	%
Total Particles	12	0.60	%
	0	0.00	%
Fill Level	6	0.30	%
ST2	AVAILABLE		ON
<i>Totals:</i>	20	1.00	%
Single Particles	9	0.45	%
Total Particles	20	1.00	%
	0	0.00	%
	0	0.00	%
ST3	AVAILABLE		ON
<i>Totals:</i>	13	0.65	%
Still Particles	12	0.60	%
Tot. Still Particles	10	0.50	%
Charrings	0	0.00	%
Tot. Charrings	0	0.00	%

ST4	AVAILABLE		ON
<i>Totals:</i>	42	2.10	%
Single Particles	17	0.85	%
Total Particles	40	2.00	%
	0	0.00	%
	0	0.00	%
ST5	NOT AVAILABLE		OFF
<i>Totals:</i>	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%

LKD counters			
LKD1			ON
<i>Totals:</i>	2	0.10	%
LKD2			ON
<i>Totals:</i>	26	1.30	%

Approved by:

Verified by: