

Batch:

21490030

User:	1	robj
Product:	17	2ml update
Batch:	challe	nge 1803235
Notes:		

03 April 2018 13:37:05

Date Time

 Starting date:
 03/04/2018
 1:26:50 PM

 Ending date:
 03/04/2018
 1:37:05 PM

Counters:

INSPECTED	19		
ACCEPTED	5	26.32	%
REJECTED	14	73.68	%
TECNICAL REJECT	0	0.00	%

Reject details:

ST0	AVAILABLE		ON
Totals:	2	10.53	%
Height	1	5.26	%
Shape	2	10.53	%
Black Spot	2	10.53	%
Colour Rings	0	0.00	%
ST1	AVAILABLE		ON
Totals:	7	36.84	%
Single Particles	6	31.58	%
Total Particles	6	31.58	%
	0	0.00	%
Fill Level	1	5.26	%
ST2	AVAILABLE		ON
Totals:	9	47.37	%
Single Particles	9	47.37	%
Total Particles	9	47.37	%
	0	0.00	%
	0	0.00	%
ST3	AVAIL	AVAILABLE	
Totals:	1	5.26	%
Still Particles	1	5.26	%
Tot. Still Particles	1	5.26	%
Charrings	0	0.00	%
Tot. Charrings	0	0.00	%

ST4	AVAILABLE		ON
Totals:	9	47.37	%
Single Particles	8	42.11	%
Total Particles	9	47.37	%
	0	0.00	%
	0	0.00	%
ST5	NOT AVA	ILABLE	OFF
ST5 Totals:	NOT AVA	ILABLE 0.00	OFF %
	NOT AVA		
	0 0 0	0.00	%
	0 0 0 0	0.00	%

LK	D counters		
LKD1			ON
Totals:	2	10.53	%
LKD2			ON
Totals:	2	10.53	%

Approved by:	Verified by:	
--------------	--------------	--