

## Batch:

21490030

User:	1	robj
Product:	1	Production 1ml
Batch:	challe	enge 1802155
Notes:		
		10.5.1

19 February 2018 11:27:23

## Date Time

 Starting date:
 19/02/2018
 11:21:55 AM

 Ending date:
 19/02/2018
 11:27:23 AM

## Counters:

INSPECTED	18		
ACCEPTED	5	27.78	%
REJECTED	13	72.22	%
TECNICAL REJECT	0	0.00	%

## Reject details:

ST0	AVAILABLE		ON
Totals:	2	11.11	%
Height	1	5.56	%
Shape	0	0.00	%
Black Spot	2	11.11	%
Colour Rings	0	0.00	%
ST1	AVAILABLE		ON
Totals:	9	50.00	%
Single Particles	7	38.89	%
Total Particles	9	50.00	%
	0	0.00	%
Fill Level	0	0.00	%
ST2	AVAILABLE		ON
Totals:	10	55.56	%
Single Particles	10	55.56	%
Total Particles	10	55.56	%
	0	0.00	%
	0	0.00	%
ST3	AVAIL	ABLE	ON
Totals:	3	16.67	%
Still Particles	2	11.11	%
Tot. Still Particles	2	11.11	%
Charrings	0	0.00	%
Tot. Charrings	0	0.00	%

ST4	AVAILABLE		ON
Totals:	10	55.56	%
Single Particles	10	55.56	%
Total Particles	10	55.56	%
	0	0.00	%
	0	0.00	%
ST5	NOT AVA	ILABLE	OFF
ST5 Totals:	NOT AVA	ILABLE 0.00	OFF %
	NOT AVA		
	0 0 0	0.00	%
	0 0 0 0	0.00	% %

LKD counters			
LKD1			ON
Totals:	1	5.56	%
LKD2			ON
Totals:	0	0.00	%

Approved by:	Verified by:	
--------------	--------------	--