

User:	5	Supervisor
Product:	1	Ampoules. 1ml
Batch:	PQ1_TS01	
Notes:		

16 November 2015 14:26:40

### Date Time

Starting date:	<b>16/11/2015</b>	2:09:19 PM
Ending date:	<b>16/11/2015</b>	2:26:39 PM

### Counters:

INSPECTED	<b>2120</b>		
ACCEPTED	<b>1711</b>	80.71	%
REJECTED	<b>398</b>	18.77	%
TECNICAL REJECT	<b>11</b>	0.52	%

### Reject details:

ST0	AVAILABLE		ON
<i>Totals:</i>	91	4.29	%
Height	7	0.33	%
Shape	1	0.05	%
Black Spot	84	3.96	%
Colour Rings	0	0.00	%
ST1	AVAILABLE		ON
<i>Totals:</i>	52	2.45	%
Single Particles	21	0.99	%
Total Particles	52	2.45	%
	0	0.00	%
Fill Level	0	0.00	%
ST2	AVAILABLE		ON
<i>Totals:</i>	152	7.17	%
Single Particles	52	2.45	%
Total Particles	152	7.17	%
	0	0.00	%
	0	0.00	%
ST3	AVAILABLE		ON
<i>Totals:</i>	135	6.37	%
Still Particles	135	6.37	%
Tot. Still Particles	120	5.66	%
Charrings	0	0.00	%
Tot. Charrings	0	0.00	%

ST4	AVAILABLE		ON
<i>Totals:</i>	178	8.40	%
Single Particles	81	3.82	%
Total Particles	178	8.40	%
	0	0.00	%
	0	0.00	%
ST5	NOT AVAILABLE		OFF
<i>Totals:</i>	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%

LKD counters			
LKD1			ON
<i>Totals:</i>	0	0.00	%
LKD2			ON
<i>Totals:</i>	0	0.00	%

Approved by:

Verified by: