

User:	5	Supervisor
Product:	1	Ampoules. 1ml
Batch:	PQ1_TSO4	
Notes:		

09 January 2016 23:04:30

Date Time

Starting date:	09/01/2016	11:03:51 PM
Ending date:	09/01/2016	11:04:30 PM

Counters:

INSPECTED	10		
ACCEPTED	2	20.00	%
REJECTED	8	80.00	%
TECNICAL REJECT	0	0.00	%

Reject details:

ST0	AVAILABLE		ON
<i>Totals:</i>	4	40.00	%
Height	3	30.00	%
Shape	0	0.00	%
Black Spot	2	20.00	%
Colour Rings	0	0.00	%
ST1	AVAILABLE		ON
<i>Totals:</i>	3	30.00	%
Single Particles	3	30.00	%
Total Particles	3	30.00	%
	0	0.00	%
Fill Level	0	0.00	%
ST2	AVAILABLE		ON
<i>Totals:</i>	4	40.00	%
Single Particles	3	30.00	%
Total Particles	4	40.00	%
	0	0.00	%
	0	0.00	%
ST3	AVAILABLE		ON
<i>Totals:</i>	2	20.00	%
Still Particles	2	20.00	%
Tot. Still Particles	2	20.00	%
Charrings	0	0.00	%
Tot. Charrings	0	0.00	%

ST4	AVAILABLE		ON
<i>Totals:</i>	4	40.00	%
Single Particles	3	30.00	%
Total Particles	4	40.00	%
	0	0.00	%
	0	0.00	%
ST5	NOT AVAILABLE		OFF
<i>Totals:</i>	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%

LKD counters			
LKD1			ON
<i>Totals:</i>	1	10.00	%
LKD2			ON
<i>Totals:</i>	0	0.00	%

Approved by:

Verified by: