

User:	1	operator
Product:	2	Ampoules. 2ml
Batch:	PQ1_TS03	
Notes:	2ML PLAIN GLASS AMPOULE	

06 January 2016 03:29:55

Date Time

Starting date:	06/01/2016	3:22:31 AM
Ending date:	06/01/2016	3:29:54 AM

Counters:

INSPECTED	253		
ACCEPTED	25	9.88	%
REJECTED	219	86.56	%
TECNICAL REJECT	9	3.56	%

Reject details:

ST0	AVAILABLE		ON
<i>Totals:</i>	47	18.58	%
Height	0	0.00	%
Shape	0	0.00	%
Black Spot	47	18.58	%
Colour Rings	0	0.00	%
ST1	AVAILABLE		ON
<i>Totals:</i>	2	0.79	%
Single Particles	0	0.00	%
Total Particles	2	0.79	%
	0	0.00	%
Fill Level	0	0.00	%
ST2	AVAILABLE		ON
<i>Totals:</i>	2	0.79	%
Single Particles	1	0.40	%
Total Particles	2	0.79	%
	0	0.00	%
	0	0.00	%
ST3	AVAILABLE		ON
<i>Totals:</i>	7	2.77	%
Still Particles	7	2.77	%
Tot. Still Particles	5	1.98	%
Charrings	0	0.00	%
Tot. Charrings	0	0.00	%

ST4	AVAILABLE		ON
<i>Totals:</i>	5	1.98	%
Single Particles	4	1.58	%
Total Particles	5	1.98	%
	0	0.00	%
	0	0.00	%
ST5	NOT AVAILABLE		OFF
<i>Totals:</i>	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%

LKD counters			
LKD1			ON
<i>Totals:</i>	166	65.61	%
LKD2			ON
<i>Totals:</i>	19	7.51	%

Approved by:

Verified by: