

## Batch:

21490035

User:	1	operator
Product:	6	vials 50ml
Batch:	12345	67
Notes:		

09 June 2015 14:35:48

## Date Time

 Starting date:
 09/06/2015
 2:35:41 PM

 Ending date:
 09/06/2015
 2:35:48 PM

## Counters:

INSPECTED	40		
ACCEPTED	4	10.00	%
REJECTED	36	90.00	%
TECHNICAL REJECT	0	0.00	%

## Reject details:

ST0	NOT AVAILABLE		OFF
Totals:	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%
ST1	AVAILABLE		ON
Totals:	25	62.50	%
Min Still particles	19	47.50	%
Tot. Still Particles	25	62.50	%
	0	0.00	%
	0	0.00	%
ST2	AVAILABLE		ON
Totals:	1	2.50	%
Aluseal Area	1	2.50	%
Crimping	0	0.00	%
	0	0.00	%
	0	0.00	%
ST3	AVAILABLE		ON
Totals:	17	42.50	%
Single Particles	3	7.50	%
Total Particles	17	42.50	%
	0	0.00	%
Fill Level	0	0.00	%

ST4	AVAILABLE		ON
Totals:	28	70.00	%
Single Particles	3	7.50	%
Total Particles	28	70.00	%
	0	0.00	%
	0	0.00	%
ST5	AVAIL	ABLE	ON
ST5 Totals:	AVAIL/	ABLE 7.50	ON %
	AVAIL/ 3 1		
Totals:	3 1 3	7.50	%
<i>Totals:</i> Single Particles	3 1	7.50 2.50	% %

Approved by:	Verified by:	
--------------	--------------	--