

User:	1	robj
Product:	2	Production 2ml
Batch:	mag sulfate 2ml test	
Notes:		

14 November 2016 16:46:26

Date Time

Starting date:	14/11/2016	3:31:59 PM
Ending date:	14/11/2016	4:46:26 PM

Counters:

INSPECTED	80		
ACCEPTED	55	68.75	%
REJECTED	22	27.50	%
TECNICAL REJECT	3	3.75	%

Reject details:

ST0	AVAILABLE		ON
<i>Totals:</i>	11	13.75	%
Height	0	0.00	%
Shape	1	1.25	%
Black Spot	11	13.75	%
Colour Rings	0	0.00	%
ST1	AVAILABLE		ON
<i>Totals:</i>	6	7.50	%
Single Particles	0	0.00	%
Total Particles	0	0.00	%
	0	0.00	%
Fill Level	6	7.50	%
ST2	AVAILABLE		ON
<i>Totals:</i>	2	2.50	%
Single Particles	2	2.50	%
Total Particles	2	2.50	%
	0	0.00	%
	0	0.00	%
ST3	AVAILABLE		ON
<i>Totals:</i>	4	5.00	%
Still Particles	0	0.00	%
Tot. Still Particles	4	5.00	%
Charrings	0	0.00	%
Tot. Charrings	0	0.00	%

ST4	AVAILABLE		ON
<i>Totals:</i>	0	0.00	%
Single Particles	0	0.00	%
Total Particles	0	0.00	%
	0	0.00	%
	0	0.00	%
ST5	NOT AVAILABLE		OFF
<i>Totals:</i>	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%

LKD counters			
LKD1			ON
<i>Totals:</i>	0	0.00	%
LKD2			ON
<i>Totals:</i>	0	0.00	%

Approved by:

Verified by: