

User:	1	mitchf
Product:	4	Production 10ml
Batch:	challenge 1711146b (protocol)	
Notes:		

23 November 2017 12:25:47

Date Time

Starting date:	23/11/2017	12:23:08 PM
Ending date:	23/11/2017	12:25:46 PM

Counters:

INSPECTED	17		
ACCEPTED	5	29.41	%
REJECTED	12	70.59	%
TECNICAL REJECT	0	0.00	%

Reject details:

ST0	AVAILABLE		OFF
<i>Totals:</i>	0	0.00	%
Height	0	0.00	%
Shape	0	0.00	%
Black Spot	0	0.00	%
Colour Rings	0	0.00	%
ST1	AVAILABLE		ON
<i>Totals:</i>	10	58.82	%
Single Particles	10	58.82	%
Total Particles	10	58.82	%
	0	0.00	%
Fill Level	0	0.00	%
ST2	AVAILABLE		ON
<i>Totals:</i>	11	64.71	%
Single Particles	11	64.71	%
Total Particles	11	64.71	%
	0	0.00	%
	0	0.00	%
ST3	AVAILABLE		ON
<i>Totals:</i>	1	5.88	%
Still Particles	1	5.88	%
Tot. Still Particles	0	0.00	%
Charrings	0	0.00	%
Tot. Charrings	0	0.00	%

ST4	AVAILABLE		ON
<i>Totals:</i>	12	70.59	%
Single Particles	12	70.59	%
Total Particles	12	70.59	%
	0	0.00	%
	0	0.00	%
ST5	NOT AVAILABLE		OFF
<i>Totals:</i>	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%

LKD counters			
LKD1			OFF
<i>Totals:</i>	0	0.00	%
LKD2			OFF
<i>Totals:</i>	0	0.00	%

Approved by:

Verified by: