

User:	1	rojb
Product:	13	10mL particle test
Batch:	test sodium chloride	
Notes:		

08 May 2017 08:30:26

### Date Time

Starting date:	<b>08/05/2017</b>	8:26:54 AM
Ending date:	<b>08/05/2017</b>	8:30:26 AM

### Counters:

INSPECTED	<b>7</b>		
ACCEPTED	<b>0</b>	0.00	%
REJECTED	<b>6</b>	85.71	%
TECNICAL REJECT	<b>1</b>	14.29	%

### Reject details:

ST0	AVAILABLE		ON
<i>Totals:</i>	2	28.57	%
Height	0	0.00	%
Shape	0	0.00	%
Black Spot	2	28.57	%
Colour Rings	0	0.00	%
ST1	AVAILABLE		ON
<i>Totals:</i>	2	28.57	%
Single Particles	2	28.57	%
Total Particles	2	28.57	%
	0	0.00	%
Fill Level	0	0.00	%
ST2	AVAILABLE		ON
<i>Totals:</i>	2	28.57	%
Single Particles	2	28.57	%
Total Particles	2	28.57	%
	0	0.00	%
	0	0.00	%
ST3	AVAILABLE		ON
<i>Totals:</i>	0	0.00	%
Still Particles	0	0.00	%
Tot. Still Particles	0	0.00	%
Charrings	0	0.00	%
Tot. Charrings	0	0.00	%

ST4	AVAILABLE		ON
<i>Totals:</i>	2	28.57	%
Single Particles	2	28.57	%
Total Particles	2	28.57	%
	0	0.00	%
	0	0.00	%
ST5	NOT AVAILABLE		OFF
<i>Totals:</i>	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%

LKD counters			
LKD1			ON
<i>Totals:</i>	2	28.57	%
LKD2			ON
<i>Totals:</i>	0	0.00	%

Approved by:

Verified by: