

User:	5	Supervisor
Product:	3	Ampoules. 5ml
Batch:	111111	
Notes:		

29 September 2015 09:20:04

Date Time

Starting date:	29/09/2015	8:52:34 AM
Ending date:	29/09/2015	9:20:03 AM

Counters:

INSPECTED	1917		
ACCEPTED	1766	92.12	%
REJECTED	128	6.68	%
TECNICAL REJECT	23	1.20	%

Reject details:

ST0	AVAILABLE		ON
<i>Totals:</i>	0	0.00	%
Height	0	0.00	%
Shape	0	0.00	%
Black Spot	0	0.00	%
Colour Rings	0	0.00	%
ST1	AVAILABLE		ON
<i>Totals:</i>	6	0.31	%
Single Particles	0	0.00	%
Total Particles	6	0.31	%
	0	0.00	%
Fill Level	0	0.00	%
ST2	AVAILABLE		ON
<i>Totals:</i>	29	1.51	%
Single Particles	5	0.26	%
Total Particles	27	1.41	%
	0	0.00	%
	0	0.00	%
ST3	AVAILABLE		ON
<i>Totals:</i>	0	0.00	%
Still Particles	0	0.00	%
Tot. Still Particles	0	0.00	%
Charrings	0	0.00	%
Tot. Charrings	0	0.00	%

ST4	AVAILABLE		ON
<i>Totals:</i>	114	5.95	%
Single Particles	7	0.37	%
Total Particles	114	5.95	%
	0	0.00	%
	0	0.00	%
ST5	NOT AVAILABLE		OFF
<i>Totals:</i>	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%
	0	0.00	%

LKD counters			
LKD1			ON
<i>Totals:</i>	1	0.05	%
LKD2			ON
<i>Totals:</i>	2	0.10	%

Approved by:

Verified by: