

INFUSION PUMP END-OF-SUPPORT NOTICE

Manufacturer: Baxter International

Product Line: COLLEAGUE 3 Infusion Pumps

Notice Date: July 2024

Product Discontinuation Notice

Baxter International is announcing the discontinuation of technical support for COLLEAGUE 3 series infusion pumps. This notice outlines the support termination schedule and transition options.

End-of-Support Date: June 30, 2026

Effective June 30, 2026, Baxter will discontinue technical support and spare parts availability for all COLLEAGUE 3 infusion pump models.

Affected Models

Model Number	Configuration	Support Ends
COLLEAGUE 3 Basic	Single Channel, Standard Drug Library	06/30/2026
COLLEAGUE 3 Dual Channel	Two Infusion Channels, Enhanced Library	06/30/2026
COLLEAGUE 3 PCA	Patient-Controlled Analgesia Module	06/30/2026

Support Phase-Out Schedule

- July 2024 - June 2025:** Full support available. No further software updates planned.
- July 2025 - June 2026:** Limited support. Spare parts available on request but not guaranteed.
- July 2026 onward:** No support. Parts and service unavailable.

Spare Parts Availability During Support Period

Baxter will provide the following spare parts on a limited basis through June 30, 2026:

- Pump head and mechanism components
- Door assemblies and seals
- Power supply modules
- Display screens (if available)

Note: Battery modules and obsolete electronic components may not be available. Facilities should order critical spare parts by December 31, 2025.

Recommended Replacement Strategy

Phase 1: Assessment (Q3 2024)

- Inventory all COLLEAGUE 3 pumps in service
- Assess condition and expected useful life
- Identify critical clinical needs (ICU, oncology, PCA, etc.)

Phase 2: Planning (Q4 2024 - Q1 2025)

- Evaluate replacement platforms (Baxter INFUSOR 7000 or competitive products)
- Obtain pricing and lead times
- Budget for replacement hardware and staff training

Phase 3: Implementation (Q2 2025 - Q2 2026)

- Schedule phased replacement before June 30, 2026
- Conduct staff training on new systems (2-4 hours per user)
- Plan for overlap period where both systems run in parallel

Replacement Platform Options

Replacement Option	Key Features	Support Period
Baxter INFUSOR 7000	Wireless connectivity, advanced therapy library, electronic charting integration	10 years
Hospira Plum A+ Series	High-acuity infusion, barcode capability, cloud monitoring	7 years
B. Braun Space Station	Multi-channel platform, modular design, European regulatory compliance	8 years

Regulatory Compliance

Continuing to operate unsupported medical devices may result in:

- FDA warning letters or citations during inspections
- Inability to demonstrate compliance with medical device reporting requirements
- Increased liability in event of adverse patient outcome
- Patient safety concerns and potential loss of accreditation

Support Contact

For questions regarding COLLEAGUE 3 support or replacement:

- Baxter Clinical Support: 1-800-933-0303
- Equipment Services Manager: Available by appointment
- Email: equipment-eol@baxter.com

This bulletin contains synthetic product information created for training purposes.

Please contact manufacturers directly for actual support timelines and replacement guidance.