

Illumina DNA Prep

Reference Guide



This document and its contents are proprietary to Illumina, Inc. and its affiliates ("Illumina"), and are intended solely for the contractual use of its customer in connection with the use of the product(s) described herein and for no other purpose. This document and its contents shall not be used or distributed for any other purpose and/or otherwise communicated, disclosed, or reproduced in any way whatsoever without the prior written consent of Illumina. Illumina does not convey any license under its patent, trademark, copyright, or common-law rights nor similar rights of any third parties by this document.

The instructions in this document must be strictly and explicitly followed by qualified and properly trained personnel in order to ensure the proper and safe use of the product(s) described herein. All of the contents of this document must be fully read and understood prior to using such product(s).

FAILURE TO COMPLETELY READ AND EXPLICITLY FOLLOW ALL OF THE INSTRUCTIONS CONTAINED HEREIN MAY RESULT IN DAMAGE TO THE PRODUCT(S), INJURY TO PERSONS, INCLUDING TO USERS OR OTHERS, AND DAMAGE TO OTHER PROPERTY, AND WILL VOID ANY WARRANTY APPLICABLE TO THE PRODUCT(S).

ILLUMINA DOES NOT ASSUME ANY LIABILITY ARISING OUT OF THE IMPROPER USE OF THE PRODUCT(S) DESCRIBED HEREIN (INCLUDING PARTS THEREOF OR SOFTWARE).

© 2020 Illumina, Inc. All rights reserved.

All trademarks are the property of Illumina, Inc. or their respective owners. For specific trademark information, see www.illumina.com/company/legal.html.

Revision History

| Document | Date | Description of Change |
|------------------------------|------------|---|
| 1000000025416 v09 | June 2020 | Added dilution information for the NextSeq 2000 Sequencing System. Added information for IDT for Illumina Nextera Indexes. |
| 1000000025416 v08 | May 2020 | Changed product name from Nextera DNA Flex to Illumina DNA Prep. Renamed index and reagent kits. Updated catalog numbers for UD Index kits sets C and D. Removed obsolete UD Indexes sets A-D kit option. Updated the Additional Resources to remove document # 1000000070581, which applies only to the Illumina DNA Prep with Enrichment protocol. |
| 1000000025416 v07 | May 2019 | Added information on the IDT for Illumina Nextera Indexes Sets B, C, and D including kit information and preparation procedures. Revised Additional Resources section to provide more clarity on the resources available. Revised Prepare for Pooling to provide more clarity on the pooling information. Included TWB pipetting technique information in steps. Clarified language throughout document to provide consistency throughout Nextera guides. |
| Document # 1000000025416 v06 | March 2019 | Corrected PTC program thermal cycler temperature. |
| Document # 1000000025416 v05 | March 2019 | Corrected loading concentration values. Added information about separate workflow component requirements, DNA purity assessment, and about quantification and normalization of libraries. Added Local Run Manager Guide to additional resources. Added reagent overage information and PTC program settings to tagmentation step. Revised tagmentation amplification to include AMP program information and correct multichannel pipette volume. Revised workflow diagram to include RSB reagents. Revised cleanup step to include separate steps for small PCR amplicons and standard DNA input and to correct safe stopping storage days. Removed "Chapter 3 Sequencing." Added supplementary SPB step to Analytical Fragment Analyzer. Revised final loading concentrations in dilute libraries step. Added information about reagent storage temperature to ensure performance. Added new table with component and kit information and re-organized index kit information. Added Fragment Analyzer and Bioanalyzer to consumables table, added Fragment Analyzer and Agilent Technologies to equipment table, and added CD and UD acronyms. |

| Document | Date | Description of Change |
|---------------------------------|-----------------|--|
| Document # 1000000025416 v04 | October 2018 | Corrected average library size. |
| Document # 1000000025416 v03 | October 2018 | <p>Updated Index Adapter terminology.</p> <p>Updated to include IDT® for Illumina®-Nextera™ DNA UD Indexes Set A (96 Indexes, 96 Samples).</p> <p>Updated diluting to starting concentration information.</p> <p>Added clarification in regards to ordering index adapters.</p> <p>Added additional resource information for Unique Dual Indexes.</p> <p>Added catalog number information for IDT® for Illumina®-Nextera™ DNA UD Indexes Set A (96 Indexes, 96 Samples) and Axygen® 1.7 mL MaxyClear Snaplock Microcentrifuge Tubes.</p> <p>Updated storage information for Lysis Reagent Kit.</p> <p>Clarified PCR Amplicons information.</p> <p>Clarified instructions when safe stopping is an option.</p> <p>Moved recommended read lengths for each system to the support site.</p> <p>Moved blood and lysis consumables to their own table.</p> <p>Revised step-by-step instructions to be more succinct.</p> <p>Reorganized the following content to improve continuity:</p> <ul style="list-style-type: none"> • Rearranged DNA input recommendations. • Moved information on blood and saliva lysis preparation and procedures. |
| Document # 1000000025416 v02 | June 2018 | Added information about PCR Amplicons. |
| Document # 1000000025416 v01 | April 2018 | Replaced references to the <i>Nextera DNA Flex Pooling Guide</i> (document # 1000000031471) with the <i>Index Adapters Pooling Guide</i> (document # 1000000041074). Pooling information is consolidated into the <i>Index Adapters Pooling Guide</i> . |
| Document # 1000000025416 v00 | October 2017 | Initial release. |

Table of Contents

| | |
|--|-----------|
| Chapter 1 Overview | 1 |
| Introduction | 1 |
| DNA Input Recommendations | 1 |
| Blood and Saliva Input Recommendations | 2 |
| Sample Input Recommendations | 2 |
| PCR Amplicons | 3 |
| Additional Resources | 3 |
| Chapter 2 Protocol | 4 |
| Introduction | 4 |
| Tips and Techniques | 4 |
| Library Prep Workflow | 6 |
| Tagment Genomic DNA | 7 |
| Post Tagmentation Cleanup | 8 |
| Amplify Tagmented DNA | 9 |
| Clean Up Libraries | 11 |
| Pool Libraries | 13 |
| Check Library Quality (Optional) | 13 |
| Dilute Libraries to the Starting Concentration | 15 |
| Appendix A Supplemental Procedures | 17 |
| Introduction | 17 |
| Blood Lysis | 17 |
| Saliva Lysis | 19 |
| Appendix B Supporting Information | 22 |
| Introduction | 22 |
| How the Illumina DNA Prep Assay Works | 22 |
| Kit Contents | 24 |
| Consumables and Equipment | 27 |
| Acronyms | 29 |
| Technical Assistance | 30 |

Chapter 1 Overview

| | |
|--|---|
| Introduction | 1 |
| DNA Input Recommendations | 1 |
| Blood and Saliva Input Recommendations | 2 |
| Sample Input Recommendations | 2 |
| PCR Amplicons | 3 |
| Additional Resources | 3 |

Introduction

This protocol explains how to prepare up to 384 dual-indexed paired-end libraries from DNA using the Illumina DNA Prep workflow.

The Illumina DNA Prep protocol:

- ▶ Uses tagmentation, and enzymatic reaction, to fragment DNA and add adapter sequences in only 15 minutes.
- ▶ Innovates sample normalization at inputs ≥ 100 ng.
- ▶ Streamlines sample pooling and sequencing.
- ▶ Master mixed reagents reduce containers, pipetting, and hands-on time.
- ▶ Requires as little as 1 ng input.
- ▶ Can prepare libraries directly from whole blood or saliva samples when using an extraction protocol.

DNA Input Recommendations

The Illumina DNA Prep protocol is compatible with DNA inputs of 1–500 ng or higher. For human DNA samples and other large complex genomes, the recommended minimum DNA input is 100–500 ng. For small genomes (eg microbial), the DNA input amount can be reduced to as low as 1 ng (modifying the PCR cycling conditions accordingly).

Assess DNA purity to make sure that the initial DNA sample does not contain > 1 mM EDTA and is free of organic contaminants, such as phenol and ethanol. These substances can interfere with the tagmentation reaction and result in assay failure.

DNA Input 100–500 ng

For DNA inputs between 100–500 ng, quantifying and normalizing the initial DNA sample is not required.

DNA Input < 100 ng

This protocol does not normalize final library yields from < 100 ng DNA input. Therefore, quantification and normalization of libraries before sequencing is required.

If using < 100 ng DNA input, quantifying the initial DNA sample to determine the number of PCR cycles required is recommended. Use a fluorometric-based method to quantify double-stranded DNA input. Avoid methods that measure total nucleic acid, such as NanoDrop or other UV absorbance methods. For more information see [Sample Input Recommendations on page 2](#).

Assess DNA Purity

UV absorbance is a common method used for assessing the purity of a DNA sample. The ratio of absorbance at 260 nm to absorbance at 280 nm provides an indication of sample purity. This protocol is optimized for DNA with 260/280 absorbance ratio values of 1.8–2.0, which indicates a pure DNA sample.

For a secondary indication of sample purity, use the ratio of absorbance at 260 nm to absorbance at 230 nm. Target a 260/230 ratio of 2.0–2.2. Values outside this range indicate the presence of contaminants. For a complete list of contaminants, including sources, avoidance, and effects on the library preparation, see the *Nextera XT Troubleshooting Technical Note*.

Dilute the starting material in 10 mM Tris-HCl, pH 7.5–8.5. Incomplete tagmentation caused by contaminants can cause library preparation failure, poor clustering, or low quality sequencing results.

Blood and Saliva Input Recommendations

The Illumina DNA Prep protocol is compatible with fresh whole blood (requires the Flex Lysis Reagent Kit) and saliva sample inputs. For information about protocols specific to blood and saliva, see [Blood Lysis on page 17](#) or [Saliva Lysis on page 19](#).

When starting with 10 µl liquid whole blood in EDTA tubes or 30 µl saliva in Oragene tubes, expect normalization of libraries equal to that observed when using ≥ 100 ng DNA input. Blood and saliva are heterogeneous sample types, therefore the ability of Illumina DNA Prep to generate normalized libraries depends on the total amount of DNA obtained from the lysed sample. The following factors can adversely affect normalization of library independent of kit performance:

- ▶ Viscosity of the saliva samples
- ▶ Blood sample age
- ▶ Storage conditions
- ▶ Underlying medical conditions affecting white blood cell counts

Sample Input Recommendations

The Illumina DNA Prep workflow is compatible with blood and saliva samples when using the following protocols and reagent kits:

- ▶ Illumina Blood Lysis Protocol (blood) with Flex Lysis Reagent Kit
- ▶ Illumina Saliva Lysis Protocol (saliva)

The recommended PCR cycles for the BLT PCR program are adjusted based on sample input concentration and quality. For more information, see [Amplify Tagmented DNA on page 9](#).

Table 1 DNA Input Recommendations

| Total DNA Input (ng) | Quantification of Input DNA Recommended | Normalized Library Yield |
|----------------------|---|--------------------------|
| 1–9 | | |
| 10–24 | | |
| 25–49 | Yes | No |
| 50–99 | | |
| 100–500 | No | Yes |
| Blood/Saliva | No | Yes |

PCR Amplicons

When starting with PCR amplicons, the PCR amplicon must be > 150 bp. The standard clean up protocol depletes libraries < 500 bp. Therefore, Illumina recommends that amplicons < 500 bp undergo a 1.8 x sample purification bead volume ratio to supernatant during [Clean Up Libraries on page 11](#). Shorter amplicons can otherwise be lost during the library cleanup step.

Tagmentation cannot add an adapter directly to the distal end of a fragment, so a drop in sequencing coverage of ~50 bp from each distal end is expected. To ensure sufficient coverage of the amplicon target region, design primers to extend beyond the target region by 50 bp per end.

Additional Resources

The following resources provide instructions and guidelines for using the Illumina DNA Prep kit to prepare libraries. Visit the [Illumina DNA Prep support pages](#) for additional information:

- ▶ Compatible products and requirements for recording sample information, sequencing libraries, and analyzing data.
- ▶ Questions and answers about using the kit.
- ▶ Training videos about the kit and courses for related products and subjects.
- ▶ The latest versions of the kit documentation.

| Resource | Description |
|--|--|
| Custom Protocol Selector | A tool for generating end-to-end instructions tailored to your library prep method, run parameters, and analysis method, with options to refine the level of detail. |
| <i>Illumina DNA Prep Checklist</i> (document # 1000000033561) | Provides a checklist of the protocol steps intended for experienced users. |
| <i>Illumina DNA Prep Consumables and Equipment List</i> (document # 1000000033564) | Provides an interactive checklist of user-supplied consumables and equipment. |
| <i>Index Adapter Pooling Guide</i> (document # 1000000041074) | Provides pooling guidelines and dual-index strategies for using the 10 base pair IDT for Illumina DNA/RNA UD Indexes or 8 base pair Nextera CD Indexes with the Illumina DNA Prep kit. |
| <i>Illumina Adapter Sequences</i> (document # 1000000002694) | Provides the nucleotide sequences that comprise Illumina oligonucleotides used in Illumina sequencing technologies. |
| <i>Illumina Free Adapter Blocking Reagent</i> (document # 1000000047585) | Provides the protocol to block excess free adapter, minimize index hopping, and enhance data quality. |
| IDT for Illumina DNA/RNA UD Indexes support page | Provides information about IDT for Illumina DNA/RNA Unique Dual (UD) indexes. |

Chapter 2 Protocol

| | |
|--|----|
| Introduction | 4 |
| Tips and Techniques | 4 |
| Library Prep Workflow | 6 |
| Tagment Genomic DNA | 7 |
| Post Tagmentation Cleanup | 8 |
| Amplify Tagmented DNA | 9 |
| Clean Up Libraries | 11 |
| Pool Libraries | 13 |
| Check Library Quality (Optional) | 13 |
| Dilute Libraries to the Starting Concentration | 15 |

Introduction

This chapter describes the Illumina DNA Prep protocol.

- ▶ Review the planned complete sequencing workflow, from sample through analysis, to ensure compatibility of products and experiment parameters.
- ▶ Before proceeding, confirm kit contents and make sure that you have the required components, equipment, and consumables. This protocol requires library prep reagents and index adapters. Index adapters are sold separately. See [Supporting Information on page 22](#)
- ▶ Follow the protocols in the order shown, using the specified volumes and incubation parameters.

Prepare for Pooling

If you plan to pool libraries, record information about your samples before starting library prep using Illumina Experiment Manager (IEM), Local Run Manager, or BaseSpace Sequence Hub. For information on the tools compatible with your sequencing system, visit the [Illumina DNA Prep Product Compatibility page](#).

- ▶ For low-plexity pooling strategies (2-plex to 9-plex), see the *Index Adapters Pooling Guide (document # 1000000041074)*.
- ▶ For index adapter sequences and information about recording the sequences, see *Illumina Adapter Sequences (document # 1000000041074)*.

Tips and Techniques

Unless a safe stopping point is specified in the protocol, proceed immediately to the next step.

Avoiding Cross-Contamination

- ▶ When adding or transferring samples or reagent master mixes, change tips between **each sample**.
- ▶ When adding index adapters with a multichannel pipette, change tips between **each row** or **each column**. If using a single channel pipette, change tips between each sample.
- ▶ [Tubes] Open only one index adapter tube at a time to prevent misplacing caps. Remove unused index adapter tubes from the working area.

Sealing the Plate

- ▶ Always seal the 96-well plate with the adhesive seal using a rubber roller to cover the plate before the following steps in the protocol:
 - ▶ Shaking steps

- ▶ Thermal cycling steps
- ▶ Centrifuge steps
- ▶ Microseal 'B' adhesive seals are effective at -40°C to 110°C, and suitable for skirted or semiskirted PCR plates. Use microseal 'B' seals for thermal cycling or short-term storage.
- ▶ Microseal 'F' foil seals are effective at temperatures down to -70°C and are suitable for storing the 96-well plates containing the final libraries long-term.

Handling Bead-Linked Transposomes (BLT)

- ▶ Store the BLT stock tube upright in the refrigerator so that the beads are always submerged in the buffer.
- ▶ Vortex the BLT stock tube thoroughly until the beads are resuspended. To avoid resettling the beads, centrifugation before pipetting is not recommended.
- ▶ If beads are adhered to the side or top of a 96-well plate, centrifuge at $280 \times g$ for 3 seconds, and then pipette to resuspend.
- ▶ When washing beads:
 - ▶ Use the appropriate magnetic stand for the plate.
 - ▶ Keep the plate on the magnetic stand until the instructions specify to remove it.
 - ▶ Do not agitate the plate while it is on the magnetic stand.
 - ▶ Do not disturb the bead pellet.
 - ▶ If beads are aspirated into pipette tips, dispense back into the plate on the magnetic stand and wait until the liquid is clear (~2 minutes).
 - ▶ Dispense tagment wash buffer (TWB) directly onto the beads.
 - ▶ If liquid becomes adhered to the side or top of the tube or well, centrifuge at $280 \times g$ for 3 seconds to pull volume into liquid.

Handling Tagment Wash Buffer (TWB)

- ▶ Pipette slowly to minimize foaming.

Preparing Index Plates

- ▶ Illumina DNA Prep is compatible with IDT[®] for Illumina[®] DNA/RNA Unique Dual (UD), IDT for Illumina Nextera DNA Unique Dual (UD), and Nextera DNA Combinatorial Dual (CD) Indexes.
- ▶ Each index plate is for single use only.
- ▶ IDT[®] for Illumina[®] DNA/RNA UD Indexes use 10 base pair index codes that differ from Nextera DNA CD indexes, which use 8 base pair index codes. Confirm your sequencing system is configured for 10 base pair index codes.

Prepare index plates as follows.


- ▶ Centrifuge at $1000 \times g$ for 1 minute to settle liquid away from the seal.
- ▶ [< 96 samples] Pierce the foil seal on the index adapter plate using a new pipette tip for each well for only the number of samples being processed.
- ▶ [96 samples] Align a new Eppendorf 96-well PCR plate above the index adapter plate and press down to puncture the foil seal on all 96 wells. Press down slowly to avoid tipping the volume over.
- ▶ Discard the empty Eppendorf plate used to puncture the foil seal.

Library Prep Workflow

The following diagram illustrates the Illumina DNA Prep workflow. Safe stopping points are marked between steps.

Time estimates are based on preparing 16 samples using a multichannel pipette.

Figure 1 Illumina DNA Prep Workflow



Tagment Genomic DNA

This step uses the Bead-Linked Transposomes (BLT) to tagment DNA, which is a process that fragments and tags the DNA with adapter sequences.

Consumables

- ▶ Bead-Linked Transposomes (BLT)
- ▶ Tagmentation Buffer 1 (TB1)
- ▶ Nuclease-free water
- ▶ 96-well PCR plate
- ▶ Microseal 'B' adhesive seal
- ▶ 1.7 ml microcentrifuge tubes
- ▶ 8-tube strip
- ▶ Pipette tips
 - ▶ 20 µl multichannel pipettes
 - ▶ 200 µl multichannel pipettes

About Reagents

- ▶ BLT must be stored at temperatures above 2°C. Do not use BLT that has been stored below 2°C.

Preparation

- 1 Prepare the following consumables:

| Item | Storage | Instructions |
|------|----------------|---|
| BLT | 2°C to 8°C | Bring to room temperature. Vortex to mix. Do not centrifuge before pipetting. |
| TB1 | -25°C to -15°C | Bring to room temperature. Vortex to mix. |

- 2 Save the following TAG program on the thermal cycler:
 - ▶ Choose the preheat lid option and set to 100°C
 - ▶ Set the reaction volume to 50 µl
 - ▶ 55°C for 15 minutes
 - ▶ Hold at 10°C

Procedure

- 1 Add 2–30 µl DNA to each well of a 96-well PCR plate so that the total input amount is 100–500 ng.
- 2 If DNA volume < 30 µl, add nuclease-free water to the DNA samples to bring the total volume to 30 µl.
- 3 Vortex BLT (yellow cap) vigorously for 10 seconds to resuspend. Repeat as necessary.
- 4 Combine the following volumes to prepare the tagmentation master mix. Multiply each volume by the number of samples being processed.
 - ▶ BLT (11 µl)
 - ▶ TB1 (11 µl)
 Reagent overage is included in the volume to ensure accurate pipetting.

- 5 Vortex the tagmentation master mix thoroughly to resuspend.
- 6 Divide the tagmentation master mix volume equally into an 8-tube strip.
- 7 Using a 200 µl multichannel pipette, transfer 20 µl tagmentation master mix to each well of the plate containing a sample. Use fresh tips for each sample column.
- 8 Discard the 8-tube strip after the tagmentation master mix has been dispensed.
- 9 Pipette each sample 10 times to resuspend. Use fresh tips for each sample column.
- 10 Seal the plate with Microseal 'B', place on the preprogrammed thermal cycler, and run the TAG program.

Post Tagmentation Cleanup

This step washes the adapter-tagged DNA on the BLT before PCR amplification.

Consumables

- ▶ Tagment Stop Buffer (TSB)
- ▶ Tagment Wash Buffer (TWB)
- ▶ 96-well plate magnet
- ▶ Microseal 'B' adhesive seal
- ▶ Pipette tips
 - ▶ 20 µl multichannel pipettes
 - ▶ 200 µl multichannel pipettes

Preparation

- 1 Prepare the following consumables:

| Item | Storage | Instructions |
|------|--------------|---|
| TSB | 15°C to 30°C | If precipitates are observed, heat at 37°C for 10 minutes, and then vortex until precipitates are dissolved. Use at room temperature. |
| TWB | 15°C to 30°C | Use at room temperature. |

- 2 Save the following PTC program on the thermal cycler:
 - ▶ Choose the preheat lid option and set to 100°C
 - ▶ Set the reaction volume to 60 µl
 - ▶ 37°C for 15 minutes
 - ▶ Hold at 10°C

Procedure

- 1 Add 10 µl TSB to the tagmentation reaction.
- 2 Slowly pipette each well 10 times to resuspend the beads.
- 3 Seal the plate with Microseal 'B', place on the preprogrammed thermal cycler, and run the PTC program.
- 4 Place the plate on the magnetic stand and wait until liquid is clear (~3 minutes).
- 5 Using a multichannel pipette, remove and discard supernatant.

- 6 Wash two times as follows:
 - a Remove the sample plate from the magnetic stand and use a deliberately slow pipetting technique to add 100 μ l TWB directly onto the beads. A deliberately slow pipetting technique minimizes the potential of TWB foaming to avoid incorrect volume aspiration and incomplete mixing.
 - b Pipette slowly until beads are fully resuspended.
 - c Place the plate on the magnetic stand and wait until the liquid is clear (~3 minutes).
 - d Using a multichannel pipette, remove and discard supernatant.
- 7 Remove the plate from the magnetic stand and use a deliberately slow pipetting technique to add 100 μ l TWB directly onto the beads.
- 8 Pipette each well slowly to resuspend the beads.
- 9 Seal the plate and place on the magnetic stand until the liquid is clear (~3 minutes). Keep on the magnetic stand until step 4 of the *Procedure* section in *Amplify Tagmented DNA*.
The TWB remains in the wells to prevent overdrying of the beads.

Amplify Tagmented DNA

This step amplifies the tagmented DNA using a limited-cycle PCR program. The PCR step adds Index 1 (i7) adapters, Index 2 (i5) adapters, and sequences required for sequencing cluster generation. To confirm the indexes selected for low plexity pooling have the appropriate color balance, see the *Index Adapters Pooling Guide* (document # 1000000041074).

Index adapter tubes or plates are ordered separately from the library prep components. For a list of compatible index adapters for use with this protocol, see *Kit Contents* on page 24.

Consumables

- ▶ Enhanced PCR Mix (EPM)
- ▶ Index adapters (tubes or plate)
- ▶ Nuclease-free water
- ▶ Microseal 'B' adhesive seal
- ▶ 1.7 ml microcentrifuge tubes
- ▶ Pipette tips
 - ▶ 20 μ l multichannel pipettes
 - ▶ 200 μ l multichannel pipettes

About Reagents

- ▶ Index adapter plates
 - ▶ A well may contain >10 μ l of index adapters.
 - ▶ Do not add samples to the index adapter plate.
 - ▶ Each well of the index plate is single use only.
- ▶ Index adapter tubes
 - ▶ Open only one index adapter tube at a time to prevent misplacing caps. Alternatively, use fresh caps after opening each tube.

Preparation

- 1 Prepare the following consumables:

| Item | Storage | Instructions |
|----------------|----------------|---|
| EPM | -25°C to -15°C | Thaw on ice. Invert to mix, then briefly centrifuge. |
| Index Adapters | -25°C to -15°C | Thaw at room temperature. [Tubes] Vortex to mix, then centrifuge briefly. [Plates] Spin briefly before use. |

- 2 Save the following BLT PCR program on a thermal cycler using the appropriate number of PCR cycles, indicated in the table below.
 - ▶ Choose the preheat lid option and set to 100°C
 - ▶ 68°C for 3 minutes
 - ▶ 98°C for 3 minutes
 - ▶ (X) cycles of:
 - ▶ 98°C for 45 seconds
 - ▶ 62°C for 30 seconds
 - ▶ 68°C for 2 minutes
 - ▶ 68°C for 1 minutes
 - ▶ Hold at 10°C

| Total DNA Input (ng) | Number of PCR Cycles (X) |
|----------------------|--------------------------|
| 1–9 | 12 |
| 10–24 | 8 |
| 25–49 | 6 |
| 50–99 | 5 |
| 100–500 | 5 |
| Blood/Saliva | 5 |

Procedure

- 1 Combine the following volumes to prepare the PCR master mix. Multiply each volume by the number of samples being processed.
 - ▶ EPM (22 µl)
 - ▶ Nuclease-free water (22 µl)
 Reagent overage is included in the volume to ensure accurate pipetting.
- 2 Vortex, and then centrifuge the PCR master mix at 280 × g for 10 seconds.
- 3 With the plate on the magnetic stand, use a 200 µl multichannel pipette to remove and discard supernatant.

Foam that remains on the well walls does not adversely affect the library.
- 4 Remove from the magnet.
- 5 Immediately add 40 µl PCR master mix directly onto the beads in each sample well.

- 6 Immediately pipette to mix until the beads are fully resuspended. Alternatively, seal the plate and use a plate shaker at 1600 rpm for 1 minute.
- 7 Seal the sample plate and centrifuge at $280 \times g$ for 3 seconds.
- 8 Add the appropriate index adapters to each sample.

| Index Kit Type | Kit Configuration | Volume of Index Adapter per Sample |
|----------------------|-------------------|--|
| 24 plex (dual index) | Individual tubes | 5 μ l i7 adapter 5 μ l i5 adapter |
| 96 plex (dual index) | 96-well plate | 10 μ l pre-paired i7 and i5 index adapters |

- 9 Using a pipette set to 40 μ l, pipette 10 times to mix. Alternatively, seal the plate and use a plate shaker at 1600 rpm for 1 minute.
- 10 Seal the plate with Microseal 'B', and then centrifuge at $280 \times g$ for 30 seconds.
- 11 Place on the thermal cycler and run the BLT PCR program.

SAFE STOPPING POINT

If you are stopping, store at 2°C to 8°C for up to 3 days.

Clean Up Libraries

This step uses double-sided bead purification procedure to purify the amplified libraries.

Consumables

- ▶ Sample Purification Beads (SPB)
- ▶ Resuspension Buffer (RSB)
- ▶ Freshly prepared 80% ethanol (EtOH)
- ▶ 96-well 0.8 ml Polypropylene Deepwell Storage Plate (midi plate) (2)
- ▶ 96-well PCR plate
- ▶ Microseal 'B' adhesive seal
- ▶ Microseal 'F' foil seal
- ▶ 1.7 ml microcentrifuge tubes
- ▶ Nuclease-free water

About Reagents

- ▶ Sample Purification Beads
 - ▶ Must be at room temperature before use
 - ▶ Vortex before each use
 - ▶ Vortex frequently to make sure that beads are evenly distributed
 - ▶ Aspirate and dispense slowly due to the viscosity of the solution

Preparation

- 1 Prepare the following consumables:

| Item | Storage | Instructions |
|------|----------------|---|
| SPB | 2°C to 8°C | Let stand at room temperature for 30 minutes. Vortex and invert to mix. |
| RSB | -25°C to -15°C | Thaw and bring to room temperature. Vortex to mix. |

- 2 Prepare fresh 80% EtOH from absolute ethanol.

Procedure

- 1 Centrifuge at 280 × g for 1 minute to collect contents at the bottom of the well.
- 2 Place the plate on the magnetic stand and wait until the liquid is clear (~5 minutes).
- 3 Transfer 45 µl supernatant from each well of the PCR plate to the corresponding well of a new midi plate.
- 4 Vortex and invert SPB multiple times to resuspend.
- 5 For standard DNA input, perform the following steps.
 - a Add 40 µl nuclease-free water to each well containing supernatant.
 - b Add 45 µl SPB to each well containing supernatant.
 - c Pipette each well 10 times to mix. Alternatively, seal the plate and use a plate shaker at 1600 rpm for 1 minute.
 - d Seal the plate and incubate at room temperature for 5 minutes.
 - e Place on the magnetic stand and wait until the liquid is clear (~5 minutes).
 - f During incubation, thoroughly vortex the SPB (*undiluted* stock tube), and then add 15 µl to each well of a *new* midi plate.
 - g Transfer 125 µl supernatant from each well of the first plate into the corresponding well of the second plate (containing 15 µl *undiluted* SPB).
 - h Pipette each well in the second plate 10 times to mix. Alternatively, seal the plate and use a plate shaker at 1600 rpm for 1 minute.
 - i Discard the first plate.
- 6 For small PCR amplicon input, perform the following steps.
 - a Add 81 µl SPB to each midi plate well containing supernatant.
 - b Pipette each well 10 times to mix. Alternatively, seal the plate and use a plate shaker at 1600 rpm for 1 minute.
- 7 Incubate the sealed midi plate at room temperature for 5 minutes.
- 8 Place on the magnetic stand and wait until the liquid is clear (~5 minutes).
- 9 Without disturbing the beads, remove and discard supernatant.
- 10 Wash two times as follows.
 - a With the plate on the magnetic stand, add 200 µl fresh 80% EtOH without mixing.
 - b Incubate for 30 seconds.
 - c Without disturbing the beads, remove and discard supernatant.
- 11 Use a 20 µl pipette to remove and discard residual EtOH.
- 12 Air-dry on the magnetic stand for 5 minutes.
- 13 Remove from the magnetic stand.
- 14 Add 32 µl RSB to the beads.

- 15 Pipette to resuspend.
- 16 Incubate at room temperature for 2 minutes.
- 17 Place the plate on the magnetic stand and wait until the liquid is clear (~2 minutes).
- 18 Transfer 30 µl supernatant to a new 96-well PCR plate.

SAFE STOPPING POINT

If you are stopping, seal the plate with Microseal 'B' adhesive or Microseal 'F' foil seal, and store at -25°C to -15°C for up to 30 days.

Pool Libraries

When the DNA input is 100-500 ng, quantifying and normalizing individual libraries generated in the same experiment is not necessary. However, the final yield of libraries generated in separate experiments can vary slightly.

To achieve optimal cluster density, pool equal library volumes and quantify the pool before sequencing.

DNA Inputs of 100–500 ng

- 1 Combine 5 µl of each library (up to 384 libraries) in a 1.7 ml microcentrifuge tube.
- 2 Vortex to mix, and then centrifuge.
- 3 Quantify the library pool using a dsDNA fluorescent dye method, such as Qubit or PicoGreen.

For DNA Inputs of < 100 ng

- 1 Quantify each library individually using Qubit or PicoGreen.

Check Library Quality (Optional)

- 1 Run 1 µl library or pooled libraries on one of the following instruments:
 - ▶ Advanced Analytical Fragment Analyzer with the HS-NGS High Sensitivity 474 kit.
 - ▶ Add 1 µl RSB to the library to achieve the 2 µl volume required for Fragment Analyzer.
 - ▶ Agilent 2100 Bioanalyzer with a High Sensitivity DNA kit.

The following figures show typical library size profiles with an average fragment size of 600 bp when analyzed with a size range of 150–1500 bp.

Figure 2 Example Fragment Analyzer Trace





Figure 3 Example Bioanalyzer Trace



Dilute Libraries to the Starting Concentration

This step dilutes libraries to the starting concentration for your sequencing system and is the first step in a serial dilution. After diluting to the starting concentration, libraries are ready to be denatured and diluted to the final loading concentration.

For sequencing, Illumina recommends the read lengths indicated in the table below.

Table 2 Recommended Read Length on Illumina Systems

| Sequencing System | Read Length |
|--|-------------|
| NovaSeq 6000, HiSeq X*, HiSeq 3000 and HiSeq 4000, NextSeq 500 and NextSeq 550, MiSeq, MiniSeq, iSeq 100 | 2 x 151 |
| HiSeq 2000, HiSeq 2500 (high output) | 2 x 126 |
| HiSeq 2500 (rapid run) | 2 x 101** |

*Not compatible with IDT for Illumina DNA/RNA UD Indexes (10bp)

**Assumes the use of the 200 cycle kit

IDT for Illumina DNA/RNA UD Indexes uses 10 base pair index codes that differ from the Nextera DNA CD Indexes, and which use 8 base pair index codes. This change in base pair index codes can require adjustments to your sequencing run set up.

- Calculate the molarity value of the library or pooled libraries using the following formula.
 - For libraries qualified on a Bioanalyzer, use the average size obtained for the library.
 - For all other qualification methods, use 600 bp as the average library size.
- Using the molarity value, calculate the volumes of RSB and library needed to dilute libraries to the starting concentration for your system.

| Sequencing System | Starting Concentration (nM) | Final Loading Concentration (pM) |
|---|-----------------------------|---|
| HiSeq 2500 and HiSeq 2000 (high output modes) | 2 | 12 |
| HiSeq 2500 (rapid run mode) | 2 | 8.5 |
| HiSeq X, HiSeq 4000, and HiSeq 3000 | 2–3 | 200–300 |
| iSeq 100 | 2 | 200 |
| MiniSeq | 2 | 1.2–1.3 |
| MiSeq (v3 reagents) | 4 | 12 |
| NextSeq 550 and NextSeq 500 | 2 | 1.2–1.3 |
| NextSeq 2000 | 2 | 750 |
| NovaSeq 6000 | 2 | See document # 1000000019358 (NovaSeq 6000 System Guide) |

- Dilute libraries using RSB:
 - Libraries quantified as a multiplexed library pool**—Dilute the pool to the starting concentration for your system.
 - Libraries quantified individually**—Dilute each library to the starting concentration for your system. Add 10 µl each diluted library to a tube to create a multiplexed library pool.
- Follow the denature and dilute instructions for your system to dilute to the final loading concentration.

- ▶ For the iSeq 100 System, see the system guide for dilution instructions (libraries are automatically denatured).
- ▶ For the NovaSeq 6000 System, see the system guide for pool and denature instructions.
- ▶ For the HiSeq 4000 and HiSeq 3000 Systems, see the cBot 2 or cBot system guide for reagent preparation instructions.
- ▶ For all other systems, see the denature and dilute libraries guide.

The final loading concentrations are a starting point and general guideline. Optimize concentrations for your workflow and quantification method over subsequent sequencing runs or by flow cell titration.

Supplemental Procedures

| | |
|--------------------|----|
| Introduction | 17 |
| Blood Lysis | 17 |
| Saliva Lysis | 19 |

Introduction

This section provides instructions for optional procedures within the Illumina DNA Prep workflow.

Blood Lysis

Use this protocol when performing the Illumina DNA Prep workflow using blood sample inputs with the Flex Lysis Reagent Kit. This protocol has been validated using fresh whole blood collected in EDTA collection tubes. Store the blood at 4°C and process it within 3 days. The use of frozen blood has not been validated and therefore cannot be recommended.

This protocol is expected to generate > 100 ng of DNA output at the end of the blood lysis step.



CAUTION

Blood is a potential source of infectious diseases. Follow site-specific procedures to ensure the safe handling of blood samples. During the lysis protocol, make sure that the entire blood sample is fully lysed (brown in color following the heat incubation step) before proceeding to subsequent steps.

Consumables

- ▶ EDTA collection tubes (for blood sample collection)
- ▶ Blood Lysis Buffer (BLB)
- ▶ Proteinase K (PK1)
- ▶ Sample Purification Beads (SPB)
- ▶ Freshly prepared 80% ethanol (EtOH)
- ▶ Nuclease-free water
- ▶ 96-well PCR plate

About Reagents

- ▶ Sample Purification Beads
 - ▶ Must be at room temperature before use
 - ▶ Vortex before each use
 - ▶ Vortex frequently to make sure that beads are evenly distributed
 - ▶ Aspirate and dispense slowly due to the viscosity of the solution

Preparation

- 1 Prepare the following consumables.

| Item | Storage | Instructions |
|------|----------------|--|
| BLB | 15°C to 30°C* | If frozen, thaw at room temperature. If precipitates are observed, heat at 37°C for 10 minutes and vortex until resuspended. |
| PK1 | -25°C to -15°C | Place on ice until needed. |
| SPB | 2°C to 8°C** | Let stand at room temperature for 30 minutes. Vortex and invert to mix. |

*BLB is shipped -25°C to -15°C, but stored at 15°C to 30°C.

**SPB is included in the Illumina DNA Prep Kit.

- 2 Prepare fresh 80% EtOH from absolute ethanol.
- 3 Save the following BLP program on the thermal cycler:
 - ▶ Choose the preheat lid option and set to 100°C
 - ▶ 56°C for 10 minutes

Procedure

- 1 Combine the following volumes to prepare the lysis master mix. Multiply each volume by the number of samples being processed.
 - ▶ BLB (7 µl)
 - ▶ PK1 (2 µl)
 - ▶ Nuclease-free water (31 µl)
 Reagent overage is included in the volume to ensure accurate pipetting.
- 2 Vortex and centrifuge the lysis master mix.
- 3 Invert the EDTA tube 10 times to mix.
- 4 Transfer 10 µl blood from the tube to one well of a 96-well PCR plate.
- 5 Add 40 µl lysis master mix to each sample.
- 6 Vortex and invert SPB multiple times to resuspend.
- 7 Add 20 µl SPB to the well.
- 8 Using a pipette set to 50 µl, slowly pipette 10 times to mix.
- 9 Seal the plate, place on the preprogrammed thermal cycler, and run the BLP program.
- 10 Place on a magnetic stand and wait 5 minutes.

The liquid will not become clear due to the dark brown color of the blood from the lysis reaction. The beads migrate after 5 minutes.
- 11 Without disturbing the beads, remove and discard supernatant.
- 12 If beads are aspirated into pipette tips, dispense back to the plate on the magnetic stand, and wait until the liquid is clear (~2 minutes).
- 13 Add 150 µl fresh 80% EtOH to the well.
- 14 Incubate on the magnetic stand for 30 seconds.
- 15 Pipette to remove and discard the EtOH.

- 16 Use a 20 µl pipette to remove and discard all residual EtOH.
- 17 Remove the plate from the magnetic stand.
- 18 Add 30 µl nuclease-free water and pipette to resuspend.
- 19 Proceed immediately to step 3 of *Tagment Genomic DNA on page 7* or stop and store the sample bead mixture.

SAFE STOPPING POINT

If you are stopping before proceeding to *Tagment Genomic DNA on page 7*, seal the plate with a Microseal 'B' adhesive seal, and store the sample bead mixture at 2°C to 8°C for up to 3 days.

Saliva Lysis

Use this protocol when performing the Illumina DNA Prep workflow using saliva sample inputs. This protocol is validated for saliva collected only in Oragene DNA Saliva collection tubes. The saliva is mixed with the Oragene DX Solution contained in the collection tube, making it stable at room temperature.

This protocol is expected to generate > 100 ng of DNA output at the end of the saliva lysis step.



WARNING

Saliva is a potential source of infectious diseases. Follow site-specific procedures to ensure the safe handling of saliva samples.

Consumables

- ▶ Sample Purification Beads (SPB)
- ▶ Oragene DNA collection tubes (for saliva sample collection)
- ▶ Freshly prepared 80% ethanol (EtOH)
- ▶ Nuclease-free water
- ▶ 96-well PCR plate

About Reagents

- ▶ Sample Purification Beads
 - ▶ Must be at room temperature before use
 - ▶ Vortex before each use
 - ▶ Vortex frequently to make sure that beads are evenly distributed
 - ▶ Aspirate and dispense slowly due to the viscosity of the solution

Preparation

- 1 Prepare the following consumables.

| Item | Storage | Instructions |
|--|------------------|---|
| Saliva samples in Oragene DNA collection tubes | Room temperature | Any time after sample collection, incubate for a minimum of 1 hour at 50°C in a water bath or an air incubator (as recommended by DNA Genotek) to lyse the cells. Following heat treatment, store at room temperature. For information on long-term storage of Oragene/saliva samples at room temperature and guarantees, see the DNA Genotek website. |
| SPB | 2°C to 8°C* | Let stand at room temperature for 30 minutes. Vortex and invert to mix. |

*SPB is included in the Illumina DNA Prep Kit.

- 2 Prepare fresh 80% EtOH from absolute ethanol.

Procedure

- 1 For each sample, add 20 µl nuclease-free water to one well of a 96-well PCR plate.
- 2 Vortex the heat-treated Oragene DNA collection tube.
- 3 Transfer 30 µl saliva sample from the tube to the well containing water. Slowly pipette to mix. For viscous samples, use a wide-bore pipette tip for more accurate pipetting.
- 4 Vortex and invert SPB multiple times to resuspend.
- 5 Add 20 µl SPB to the well.
- 6 Using a pipette set to 50 µl, slowly pipette 10 times to mix.
- 7 Incubate at room temperature for 5 minutes.
- 8 Place on a magnetic stand and wait 5 minutes.
- 9 Without disturbing the beads, remove and discard supernatant.
- 10 If beads are aspirated into pipette tips, dispense back to the plate on the magnetic stand, and wait until the liquid is clear (~2 minutes).
- 11 Add 150 µl fresh 80% EtOH to the well.
- 12 Incubate on the magnetic stand for 30 seconds.
- 13 Pipette to remove and discard the EtOH.
- 14 Use a 20 µl pipette to remove and discard all residual EtOH.
- 15 Remove the plate from the magnetic stand.
- 16 Add 30 µl nuclease-free water and pipette to resuspend.
- 17 Proceed immediately to step 3 of *Tagment Genomic DNA* on page 7 or stop and store the sample bead mixture.

SAFE STOPPING POINT

If you are stopping before proceeding to *Tagment Genomic DNA on page 7*, seal the plate with a Microseal 'B' adhesive seal, and store the sample bead mixture at 2°C to 8°C for up to 3 days.

Supporting Information

| | |
|---|----|
| Introduction | 22 |
| How the Illumina DNA Prep Assay Works | 22 |
| Kit Contents | 24 |
| Consumables and Equipment | 27 |
| Acronyms | 29 |


Introduction

The protocol described in this guide assumes that you have reviewed the contents of this section, confirmed workflow contents, and obtained all required consumables and equipment.

How the Illumina DNA Prep Assay Works

The Illumina DNA Prep library prep kit uses a bead-based transposome complex to tagment genomic DNA, which is a process that fragments DNA and then tags the DNA with adapter sequences in one step. After it is saturated with input DNA, the bead-based transposome complex fragments a set number of DNA molecules. This fragmentation provides flexibility to use a wide DNA input range to generate normalized libraries of consistent tight fragment size distribution. Following tagmentation, a limited-cycle PCR adds Illumina DNA Prep-specific index adapter sequences to the ends of a DNA fragment. This step enables capability across all Illumina sequencing platforms. A subsequent Sample Purification Beads (SPB) cleanup step then purifies libraries for use on an Illumina sequencer. The double-stranded DNA library is denatured before hybridization of the biotin probe oligonucleotide pool.

Figure 4 Illumina DNA Prep Workflow



Kit Contents

Completing the Illumina DNA Prep protocol requires library prep reagents and index adapters. The number of index adapters required depends on the number of samples to be uniquely indexed for your experiment. Depending on the sample input type and sequencing requirements, the protocol might require additional, optional consumables.

| Component | Kit Options | Catalog # |
|--|--|-----------|
| Library prep reagents | Illumina DNA Prep – (M) Tagmentation (24 Samples) | 20018704 |
| | Illumina DNA Prep – (M) Tagmentation (96 Samples) | 20018705 |
| Index adapters ¹ | IDT for Illumina DNA/RNA UD Indexes Set A, Tagmentation (96 Indexes, 96 Samples) | 20027213 |
| | IDT for Illumina DNA/RNA UD Indexes Set B, Tagmentation (96 Indexes, 96 Samples) | 20027214 |
| | IDT for Illumina DNA/RNA UD Indexes Set C, Tagmentation (96 Indexes, 96 Samples) | 20042666 |
| | IDT for Illumina DNA/RNA UD Indexes Set D, Tagmentation (96 Indexes, 96 Samples) | 20042667 |
| | IDT for Illumina Nextera DNA UD Indexes Set C (96 indexes, 96 samples) | 20027215 |
| | IDT for Illumina Nextera DNA UD Indexes Set D (96 Indexes, 96 Samples) | 20027216 |
| | IDT for Illumina Nextera DNA UD Indexes Set A-D (384 Indexes, 384 Samples) | 20027217 |
| | Nextera DNA CD Indexes (24 Indexes, 24 Samples) | 20018707 |
| | Nextera DNA CD Indexes (96 Indexes, 96 Samples) | 20018708 |
| | | |
| [Optional] Blood Lysis ² | Flex Lysis Reagent Kit (96 samples) | 20018706 |
| [Optional] Additional Reagents | Illumina Adapter Blocking Reagents (12 reactions) | 20024144 |
| | Illumina Adapter Blocking Reagents (48 reactions) | 20024145 |
| [Optional] Additional Accessories for Nextera DNA CD Indexes (24 Samples) | Bag of 48 Index Adapter replacement caps, orange | 15026585 |
| | Bag of 32 Index Adapter replacement caps, white | 15026586 |

¹ Index kits are not interchangeable with Nextera XT index kits.

² Required when starting the protocol from fresh whole blood samples.

Illumina DNA Prep Contents

Illumina DNA Prep — Beads and Buffers

These reagents are shipped at 2°C to 8°C. Promptly store reagents at the indicated tube temperature to ensure proper performance.

| Tube Quantity | | Acronym | Reagent Name | Storage Temperature |
|---------------|------------|---------|---------------------------|---------------------|
| 24 Samples | 96 Samples | | | |
| 1 | 1 | SPB | Sample Purification Beads | 2°C to 8°C |
| 1 | 4 | TSB | Tagment Stop Buffer | Room temperature |
| 1 | 1 | TWB | Tagment Wash Buffer | Room temperature |

Illumina DNA Prep — PCR and Buffers, Store at -25°C to -15°C

| Tube Quantity | | Acronym | Reagent Name |
|---------------|------------|---------|-----------------------|
| 24 Samples | 96 Samples | | |
| 1 | 1 | RSB | Resuspension Buffer |
| 1 | 4 | TB1 | Tagmentation Buffer 1 |
| 1 | 4 | EPM | Enhancement PCR Mix |

Illumina DNA Prep — Tagmentation (M) Beads, Store at 2°C to 8°C

| Tube Quantity | | Acronym | Reagent Name |
|---------------|------------|---------|--------------------------|
| 24 Samples | 96 Samples | | |
| 1 | 4 | BLT | Bead-Linked Transposomes |

Index Kit Contents

For index adapter sequences, see *Illumina Adapter Sequences (document # 1000000002694)*.

Nextera DNA CD Indexes (24 Indexes, 24 Samples) Tubes, Store at -25°C to -15°C

| Quantity | Index Name | Description |
|----------|------------|-------------|
| 1 | H503 | DNA Adapter |
| 1 | H505 | DNA Adapter |
| 1 | H506 | DNA Adapter |
| 1 | H517 | DNA Adapter |
| 1 | H710 | DNA Adapter |
| 1 | H705 | DNA Adapter |
| 1 | H706 | DNA Adapter |
| 1 | H707 | DNA Adapter |
| 1 | H711 | DNA Adapter |
| 1 | H714 | DNA Adapter |

IDT for Illumina DNA/RNA UD Indexes, IDT for Illumina Nextera DNA UD Indexes, or Nextera DNA CD Indexes (96 Indexes, 96 Samples) Plates, Store at -25°C to -15°C

| Quantity | Description |
|----------|-----------------------------|
| 1 | 96 Dual Adapter Index Plate |

Flex Lysis Reagent Kit (Optional)







These reagents are shipped at -25°C to -15°C. Promptly store reagents at the indicated tube temperature to ensure proper performance.





| Quantity | Acronym | Description | Storage Temperature |
|----------|---------|--------------------|---------------------|
| 4 | BLB | Blood Lysis Buffer | 15°C to 30°C |
| 4 | PK1 | Proteinase K | -25°C to -15°C |

Sample Purification Beads are included in the Illumina DNA Prep 24 sample and 96 sample kits.

Symbol Descriptions

The following table describes the symbols on the shipment packaging, consumable, or consumable packaging.

| Symbol | Description |
|---|---|
|  | Indicates the direction to the top of the box. |
|  | Indicates that the contents are fragile and must be handled with care. |
|  | Storage temperature range in degrees Celsius. Store the consumable within the indicated range. ¹ |
|  | The date the consumable expires. For best results, use the consumable before this date. |
|  | Indicates the manufacturer (Illumina). |
|  | The intended use is Research Use Only (RUO). |

| Symbol | Description |
|---|---|
|  | Indicates the part number so that the consumable can be identified. ² |
|  | Indicates the batch code to identify the manufacturing batch or lot of the consumable. ¹ |
|  | Indicates that caution is necessary. |
|  | Indicates a health hazard. |

¹ Storage temperature can differ from shipping temperature.² REF identifies the individual component, while LOT identifies the lot or batch the component belongs to.

Consumables and Equipment

Confirm that all required user-supplied consumables and equipment are present and available before starting the protocol.

The protocol has been optimized and validated using the items listed. Comparable performance is not guaranteed when using alternate consumables and equipment.

Consumables

| Consumable | Supplier |
|--------------------------------|----------------------|
| 10 µl pipette tips | General lab supplier |
| 10 µl multichannel pipettes | General lab supplier |
| 10 µl single channel pipettes | General lab supplier |
| 20 µl pipette tips | General lab supplier |
| 20 µl multichannel pipettes | General lab supplier |
| 20 µl single channel pipettes | General lab supplier |
| 200 µl pipette tips | General lab supplier |
| 200 µl multichannel pipettes | General lab supplier |
| 200 µl single channel pipettes | General lab supplier |

| Consumable | Supplier |
|---|--|
| 1000 µl pipette tips | General lab supplier |
| 1000 µl single channel pipettes | General lab supplier |
| 96-well 0.8 ml Polypropylene Deepwell Storage (midi plate) | Thermo Fisher Scientific, catalog # AB-0859 |
| Hard-Shell 96-well PCR plates | Bio-Rad, catalog # HSP-9601 |
| 1.7 mL Microcentrifuge Tubes | General lab supplier |
| Microseal 'B' adhesive seals | Bio-Rad, catalog # MSB-1001 |
| Microseal 'F' foil seals | Bio-Rad, catalog # MSF-1001 |
| RNase/DNase-free multichannel reagent reservoirs, disposable | VWR, catalog # 89094-658 |
| Ethanol 200 proof (absolute) for molecular biology (500 ml) | Sigma-Aldrich, product # E7023 |
| Nuclease-free water | General lab supplier |
| Qubit dsDNA HS Assay Kit | One of the following, depending on quantification method: <ul style="list-style-type: none"> • ThermoFisher Scientific, catalog # Q32851 • ThermoFisher Scientific, catalog # Q32854 |
| Quant-iT™ PicoGreen® dsDNA Assay Kit | ThermoFisher Scientific, catalog # P11496 |
| Qubit Assay Tubes | ThermoFisher Scientific, catalog # Q32856 |
| One of the following kits, depending on quantification method: <ul style="list-style-type: none"> • [Fragment Analyzer] High Sensitivity NGS Fragment Analysis Kit • [Bioanalyzer] Agilent High Sensitivity DNA Kit | One of the following suppliers, depending on instrument: <ul style="list-style-type: none"> • Advanced Analytical, catalog # DNF-474 • Agilent, catalog # 5067-4626 |

Consumables for Blood and Saliva Input

| Consumable | Supplier |
|---|---------------------------------------|
| [Blood] Flex Lysis Reagent Kit | Illumina, catalog # 20015884 |
| [Blood] EDTA Blood Collection tubes | Becton Dickinson |
| [Saliva] Oragene DNA Collection Kit for Saliva | Genotek, catalog # OGR-500 or OGD-510 |

Equipment

| Equipment | Supplier |
|---|--|
| Magnetic Stand-96 | Thermo Fisher Scientific, catalog # AM10027 |
| Microplate centrifuge | General lab supplier |
| Microcentrifuge | General lab supplier |
| Vortexer | General lab supplier |
| Qubit® Fluorometer 3.0 | ThermoFisher Scientific, catalog # Q33216, Q33217 or Q33218 |
| One of the following analyzers: Advanced Analytical: • Fragment Analyzer™ Agilent Technologies: • 2100 Bioanalyzer Desktop System | Advanced Analytical: • See web product pages for catalog numbers Agilent Technologies: • Part # G2940CA |
| [Saliva] Water bath or air incubator reaching 50°C | As recommended by DNA Genotek, see product pages. |

Thermal Cyclers

The following table lists the recommended settings for the thermal cycler. If your lab has a thermal cycler that is not listed, validate the thermal cycler before performing the protocol.

| Thermal Cycler | Temp Mode | Lid Temp | Vessel Type |
|-------------------------------------|----------------------------|----------|--------------------------------|
| Bio-Rad C-1000 Touch thermal cycler | Calculated | Heated | Plate |
| Bio-Rad DNA Engine Tetrad 2 | Calculated | Heated | Polypropylene plates and tubes |
| MJ Research DNA Engine Tetrad | Calculated | Heated | Plate |
| Eppendorf Mastercycler Pro S | Gradient S, Simulated Tube | Heated | Plate |

Acronyms

| Acronym | Definition |
|---------|-----------------------------|
| BLB | Blood Lysis Buffer |
| BLT | Bead Linked Transposome |
| CD | Combinatorial Dual |
| EPM | Enhanced PCR Mix |
| EtOH | Ethanol |
| IEM | Illumina Experiment Manager |
| PK1 | Proteinase K |
| RSB | Resuspension Buffer |
| SPB | Sample Purification Beads |
| TB1 | Tagmentation Buffer 1 |
| TSB | Tagment Stop Buffer |
| TWB | Tagment Wash Buffer |
| UD | Unique Dual |

Technical Assistance

For technical assistance, contact Illumina Technical Support.

Website: www.illumina.com
Email: techsupport@illumina.com

Illumina Customer Support Telephone Numbers

| Region | Toll Free | Regional |
|-----------------|-----------------|----------------|
| North America | +1.800.809.4566 | |
| Australia | +1.800.775.688 | |
| Austria | +43 800006249 | +43 19286540 |
| Belgium | +32 80077160 | +32 34002973 |
| China | 400.066.5835 | |
| Denmark | +45 80820183 | +45 89871156 |
| Finland | +358 800918363 | +358 974790110 |
| France | +33 805102193 | +33 170770446 |
| Germany | +49 8001014940 | +49 8938035677 |
| Hong Kong | 800960230 | |
| Ireland | +353 1800936608 | +353 016950506 |
| Italy | +39 800985513 | +39 236003759 |
| Japan | 0800.111.5011 | |
| Netherlands | +31 8000222493 | +31 207132960 |
| New Zealand | 0800.451.650 | |
| Norway | +47 800 16836 | +47 21939693 |
| Singapore | +1.800.579.2745 | |
| South Korea | +82 80 234 5300 | |
| Spain | +34 911899417 | +34 800300143 |
| Sweden | +46 850619671 | +46 200883979 |
| Switzerland | +41 565800000 | +41 800200442 |
| Taiwan | 00806651752 | |
| United Kingdom | +44 8000126019 | +44 2073057197 |
| Other countries | +44.1799.534000 | |

Safety data sheets (SDSs)—Available on the Illumina website at support.illumina.com/sds.html.

Product documentation—Available for download in PDF from the Illumina website. Go to support.illumina.com, select a product, then select **Documentation & Literature**.



Illumina

5200 Illumina Way

San Diego, California 92122 U.S.A.

+1.800.809.ILMN (4566)

+1.858.202.4566 (outside North America)

techsupport@illumina.com

www.illumina.com

For Research Use Only. Not for use in diagnostic procedures.

© 2020 Illumina, Inc. All rights reserved.

illumina®