

WeeklyMT:2025.04.16

- Gas study short summary and schedule
- E80-CDC status

Yuto.Kimura

Gas study short summary

Ar-CO₂(90:10) and Ar-C₂H₆(50:50)
→ Ar-CO₂ and Ar-C₂H₆, omitted gas ratio

- what I did and what I know

Ar-C ₂ H ₆	⁵⁵ Fe	⁹⁰ Sr	Cosmic
test chamber	Gas Gain(Our standard)	ToT, Efficiency	Not yet
E15-CDC	None	None	ToT, Efficiency

- In Ar-CO₂ case, ToT of ⁹⁰Sr is relatively smaller than that of the Ar-C₂H₆ case, when comparing the same gas gain point.
- Checking Cosmic-ray data, the difference becomes even more serious.

Ar-CO ₂	⁵⁵ Fe	⁹⁰ Sr	Cosmic
test chamber	Gas Gain(Our standard)	ToT , Efficiency (small..)	Not yet
E15-CDC	None	None	ToT, Efficiency (very small..)

Gas study short summary

- what I did and what I know

Ar-C ₂ H ₆	⁵⁵ Fe	⁹⁰ Sr	Cosmic
test chamber	Gas Gain(Our standard)	ToT, Efficiency	Not yet
E15-CDC	None	None	ToT, Efficiency

- **The Ar-CO₂ gas mixture shows broader QDC and ToT spectra,**
and exhibits **more noise than Ar-C₂H₆,**
resulting in lower detection efficiency, **even under sufficiently high voltage.**

Ar-CO ₂	⁵⁵ Fe	⁹⁰ Sr	Cosmic
test chamber	Gas Gain(Our standard)	ToT , Efficiency (small..)	Not yet
E15-CDC	None	None	ToT, Efficiency (very small...)

Gas study short summary

- what I should
 - I want to complete the experiment only in the test chamber.
 - もしE15-CDC-Cosmicと同じような結果なら「そーゆーもんか」と理解する。
違う結果であれば、、、うーん

Ar-C2H6	55Fe	90Sr	Cosmic
test chamber	Gas Gain(Our standard)	ToT, Efficiency	Not yet
E15-CDC	None	None	ToT, Efficiency

Ar-CO2	55Fe	90Sr	Cosmic
test chamber	Gas Gain(Our standard)	ToT , Efficiency (small..)	Not yet
E15-CDC	None	None	ToT, Efficiency (very small...)

c.f.) Gas studyの辞書的なものの写し; 概要

Overview	内容1	内容2	Ar-CO2	Ar-C2H6	コメント1	コメント2	コメント3
Garfield	Electric Field		<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>			
	Drift Velocity		<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	sauli等にある過去データと形(Velocity vs 電場)が一致しない。		
	Diffusion		<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>			
	Gain		<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	使う関数によって絶対値が変わるが、相対比はほぼ一定。		
	Drift Time		<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>			
Test Chamber	波形 (55Fe)	analog (raw)	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	以下、線源位置はHV側。	Gainは波形 (55Fe)を積分することによって導出。	増幅率が等しいのは2200Vと2700V
		analog (after ASD)	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>			
	波形 (90Sr)	analog (raw)	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	55Feの場合と相対的には、大きな矛盾は無い。	宇宙線だと55Feと相対的に大きく異なる結果になるのか？	
		analog (after ASD)	<input type="checkbox"/>	<input checked="" type="checkbox"/>			
	QDC (55Fe)	analog (after ASD)	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	55Feの場合と相対的には、大きな矛盾は無い。	そりゃそう。	
	TDC (90Sr)	TOT	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	Analog sig (55Fe)の場合と相対的には、大きな矛盾は無い。	TOT(Ar-CO2) < TOT(Ar-C2H6) が期待されたが。。。	再現性確認のためデータ再取得したい。
		検出効率	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>			
	TDC (宇宙線)	TOT	<input type="checkbox"/>	<input type="checkbox"/>	データはあるが統計少ない。		再測定の必要あり。
E15-CDC		検出効率	<input type="checkbox"/>	<input type="checkbox"/>			
	TDC (宇宙線)	TOT	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	TOT(Ar-CO2) < TOT(Ar-C2H6)。Ar-CO2で検出効率が低い所同じ増幅率であるが、TOTにすると明らかにAr-CO2の方が小さい。		
		検出効率	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	TOT次第		検出効率が等しいのは2400Vと2800V (TOTとコンシステント)
		トラッキング効率	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	TOT次第		
E80-CDC		Residual	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	等しい検出効率 --> 等しいResidual。 (ArCO2 2400V == Arc2H6 2800V)		
	波形 (90Sr)	analog (raw)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Test Chamberに比べて波高がとて小さい。(何分の1?)		
		analog (after ASD)	<input type="checkbox"/>	<input type="checkbox"/>			
	TDC (宇宙線)	部分的ASD	<input checked="" type="checkbox"/>	<input type="checkbox"/>	お試し。ノイズだらけ。TOTのピークはかろうじて見える。		

c.f.) Gas studyの辞書的なものの写し; TOT, 90Sr, Test chamber

TOT比較

*増幅率が等しいのは2200Vと2700Vくらい? TOTピークが等しいのは2350Vと2800V、検出効率が等しいのは2400Vと2800V。

Test chamber (90Sr)

ArCO2 (90:10)

run# (nuc1:~/HUL_MHTDC/data/)

HV [V]

Gain

Field wire current [uA]

TOT peak [ch]

Wire Efficiency

32


2200

1.50E+04


0.48

50


tot_all_wires32



tot_all_wires33



tot_all_wires34



ArC2H6 (50:50)

run# (nuc1:~/HUL_MHTDC/data/)

HV [V]

Gain

Field wire current [uA]

TOT peak [ch]

Wire Efficiency

38


2600

9.18E+03


0.34

45


tot_all_wires38



tot_all_wires39



tot_all_wires40



宇山殿

41


2800

2.59E+04


0

105

tot_all_wires41



c.f.) Gas studyの辞書的なものの写し; TOT, cosmic, E15-CDC



Test chamber schedule in short term

- 2025.4.16(Wed),(Today), 12:00 :
 - Ar-C₂H₆ 50ml/min to replace 2.5 times
(~92%, maybe enough because there was no gas change)
- 2025.4.17(Thu), 18:00 :
 - Ar-C₂H₆ 25ml/min
 - data taking of 90Sr, 2600,2700,2800V (~1hour)
 - data taking of cosmic, 2600V (~24hours)
- 2025.4.18(Fri), 18:00 :
 - data taking of cosmic, 2700V (~24hours)
- 2025.4.19(Sat), 18:00 :
 - data taking of cosmic, 2800V (~24hours)
 - Ar-CO₂ 50ml/min to replace 5 times (~99.3%)
- 2025.4.22(Tue), 18:00 :
 - data taking of cosmic, 2200V (~24hours)
- 2025.4.23(Wed), 18:00 :
 - data taking of cosmic, 2300V (~24hours)
- 2025.4.24(Thu), 18:00 :
 - data taking of cosmic, 2400V (~24hours) (End)

残り~82 ml/mins*4days
-> (25 ml/minで)~12 days

A				B	C	D
				CDC [SCCM]		ArC ₂ H ₆ [MPa]
time				in	out	primary
2024/05/30 1:03				46.18	39.71	1.1
2024/05/30 8:45				31.2	37.5	1
2024/05/30 17:05				46	42	1
2024/05/31 1:23				91.4	85.3	1
2024/05/31 8:48				101	97.3	1
2024/05/31 16:36				123	95.5	0.7
2024/06/01 1:02				92	90	0.5
2024/06/01 9:02				75.1	69.3	0.5
2024/06/01 16:20				85.3	67	0.4
2024/06/02 12:54				93.65	81.59	0.2
2024/06/02 9:08				36.7	79.8	0
2024/06/02 17:00				38	32	0
2024/06/03 1:00				95.96	84.43	0
2024/06/03 9:03				70.8	66.2	0
2024/06/03 16:55				87.2	75.5	0
2024/06/04 2:18				94.91	83.97	0
2024/06/04 8:58				77.4	68.6	0
2024/06/04 16:54				84.5	60.4	0
2024/06/05 9:13				84.6	80.6	3.5
2024/06/05 16:30				37	24	3.2

Total 82 ml/min * 5days (E73_2024)

E80-CDC

- HV問題、未解決。
 - エイジング不足と見て、ひたすら2000Vかけていたが果たして効果はあるのか
- リークディテクター待ち（Gas in-outの非保存の確認）
- この後、一度2300Vくらいまであげてみるか？