Tide courier card: the courier scans every handover in the COC and records both sender and receiver identifiers.

Tide cold box card: a sealed temperature logger remains inside the box until the receiving desk opens it.

Tide delay card: an overdue routine sample is reviewed by the laboratory supervisor; the courier cannot waive the deadline.

Tide storm card: activation of the storm task requires the field coordinator signature before departure.

Tide bottle card: the backup bottle is not opened unless the laboratory supervisor requests a repeat analysis.

Tide chain card: a missing handover signature starts a chain exception even when the water temperature is acceptable.

Tide terminology: DUP means a duplicate field sample, which is distinct from FB and from a repeated laboratory reading.

Tide blank card: the field blank is prepared at the site using the approved blank water and the same handling routine.

Tide transport card: the vehicle route may change, but the collection timestamp is never reset.

Tide audit card: collection clocks are synchronized to UTC before the crew departs.

Tide debris card: large floating debris is recorded but not deliberately packed into routine dissolved-sample bottles.

Tide instrument card: failed conductivity calibration is repeated once with fresh standard; a second failure removes the instrument from service.

Tide recovery card: a lost bottle ID is recovered only by matching the event record and COC under two-person review.

Tide archive card: chain exception decisions are retained for 7 years.

Tide contact card: safety incidents go to the field coordinator through the team channel, not to a published personal number.

Tide cancellation card: an unsafe riverbank stops sampling at that site; a previous result is never copied as a new sample.

Tide reporting card: a rejected sample appears as rejected with a reason, not as a zero concentration result.

Tide scope card: operational handling limits alone do not establish drinking-water compliance.