## EUROPEAN ORGANISATION FOR NUCLEAR RESEARCH (CERN)



JINST 19 (2024) P02009

DOI: 10.1088/1748-0221/19/02/P02009



1st March 2024

# Electron and photon energy calibration with the ATLAS detector using LHC Run 2 data

## The ATLAS Collaboration

This paper presents the electron and photon energy calibration obtained with the ATLAS detector using 140 fb<sup>-1</sup> of LHC proton–proton collision data recorded at  $\sqrt{s}$  = 13 TeV between 2015 and 2018. Methods for the measurement of electron and photon energies are outlined, along with the current knowledge of the passive material in front of the ATLAS electromagnetic calorimeter. The energy calibration steps are discussed in detail, with emphasis on the improvements introduced in this paper. The absolute energy scale is set using a large sample of Z-boson decays into electron-positron pairs, and its residual dependence on the electron energy is used for the first time to further constrain systematic uncertainties. The achieved calibration uncertainties are typically 0.05% for electrons from resonant Z-boson decays, 0.4% at  $E_{\rm T} \sim 10$  GeV, and 0.3% at  $E_{\rm T} \sim 1$  TeV; for photons at  $E_{\rm T} \sim 60$  GeV, they are 0.2% on average. This is more than twice as precise as the previous calibration. The new energy calibration is validated using  $J/\psi \rightarrow ee$  and radiative Z-boson decays.

# **Contents**

| 1  | Introduction                                                                    | 3  |
|----|---------------------------------------------------------------------------------|----|
| 2  | Electron and photon reconstruction with the ATLAS detector                      | 4  |
|    | 2.1 The ATLAS detector                                                          | 4  |
|    | 2.2 Energy measurement, electron and photon reconstruction and identification   | 5  |
| 3  | Collision data and simulation                                                   | 6  |
|    | 3.1 Dataset                                                                     | 6  |
|    | 3.2 Simulation samples                                                          | 7  |
|    | 3.3 Passive material model                                                      | 7  |
| 4  | Overview of the calibration procedure                                           | 9  |
| 5  | Effects on the uniformity and stability of the energy response                  | 10 |
|    | 5.1 Uniformity                                                                  | 10 |
|    | 5.2 Stability                                                                   | 10 |
|    | 5.3 ADC non-linearity correction                                                | 11 |
|    | 5.4 Energy response in high and medium gain                                     | 12 |
| 6  | Intercalibration of the EM calorimeter layers                                   | 15 |
|    | 6.1 Presampler energy scale                                                     | 15 |
|    | 6.2 Intercalibration of the first and second calorimeter layers                 | 16 |
| 7  | Determination of the energy scale and resolution with $Z \rightarrow ee$ events | 20 |
| 8  | Photon-specific calibration                                                     | 24 |
|    | 8.1 Modelling of the photon reconstruction classification                       | 24 |
|    | 8.2 Out-of-cluster energy leakage mis-modelling                                 | 24 |
| 9  | Electron and photon energy scale uncertainties                                  | 25 |
| 10 | Energy linearity and constraints on the calibration uncertainties               | 26 |
|    | 10.1 Energy linearity measurement                                               | 26 |
|    | 10.2 Constraints on the calibration systematic uncertainties                    | 31 |
| 11 | Calibration cross-checks                                                        | 33 |
|    | 11.1 Checks using $J/\psi \to ee$ events                                        | 33 |
|    | 11.2 Checks using $Z \to \ell\ell\gamma$ events                                 | 34 |
| 12 | 2 Conclusion                                                                    | 37 |

## 1 Introduction

During the 2015–2018 data-taking period (Run 2) of the Large Hadron Collider at CERN, the ATLAS experiment accumulated a large sample of proton–proton collisions at  $\sqrt{s} = 13$  TeV, corresponding to an integrated luminosity of 140 fb<sup>-1</sup>. Such a sample provides significant opportunities for improvements in detector performance and calibration precision, further exploration of the Standard Model and searches for new physics. Optimal energy reconstruction and calibration of the electromagnetic calorimeter is necessary for all analyses involving electrons and photons, and especially for precise measurements of the masses and properties of the Higgs, W and Z bosons. The present paper describes the calorimeter energy calibration using the Run 2 data sample.

The calibration scheme comprises several steps: a simulation-based optimization of the energy measurement for electrons and photons, corrections for observed differences between data and simulation, calibration of the layers of the calorimeter and a final adjustment of the global energy scale using the abundant sample of electron–positron pairs from Z-boson decays. The resulting calibration corrections are validated using electrons from  $J/\psi$  decays and photons from radiative Z-boson decays.

The procedure applied to the full Run 2 dataset is similar to the one in Refs. [1, 2]. Compared to the previous publication, the methodology has been updated in order to reduce the impact of the dominant sources of uncertainty: the offline reconstruction of electrons and photons in the calorimeter moved from a clustering algorithm that produced fixed-size clusters to one producing variable-size 'superclusters' [3]; muons are now used for the presampler calibration, instead of electrons and photons; the layer intercalibration is now obtained by combining scales extracted using both electrons and muons; finally, dedicated data allowed further studies of the intercalibration of the high- and medium-gain electronics readouts. For the first time, the resulting calibration uncertainties are further constrained using a precise measurement of the energy dependence, or linearity, of the calorimeter response.

The calibration steps are discussed in the following, with special focus where improved methods have been utilized. Section 2.1 briefly describes the ATLAS detector and summarizes the electron and photon reconstruction algorithms applied in this analysis. Section 3 describes the data and simulated event samples used for the studies, as well as the present knowledge of the passive material in front of the calorimeter. Section 4 gives an overview of the calibration procedure and details the changes relative to the previous procedure. Section 5 enumerates corrections applied to the data to balance geometrical inhomogeneities, and to account for residual non-linearity in the electronics readout. This section also details an improved analysis of the transition between the high- and medium-gain electronics readouts. Section 6 describes the calibration of the calorimeter layers, namely constraints on the presampler energy scale and the relative responses of the first and second compartments. Section 7 combines the results of these studies to extract the final adjustment of the global energy scale. Calibration corrections specific to photons are described in Section 8. Calibration uncertainties at this stage are discussed in Section 9. The linearity of the electron energy scale is studied in Section 10. Finally, the validity of the energy calibration is established using independent samples of electron–positron pairs from  $J/\psi$  decays, and photons from radiative Z-boson decays, as discussed in Section 11.

## 2 Electron and photon reconstruction with the ATLAS detector

#### 2.1 The ATLAS detector

The ATLAS experiment [4] at the LHC is a multipurpose detector with cylindrical geometry covering almost  $4\pi$  in solid angle. Closest to the collision point, ATLAS is instrumented with an inner tracking detector (ID) covering the pseudorapidity range of  $|\eta| < 2.5$  and consisting of a silicon pixel detector, including the insertable B-layer installed as a new innermost layer before Run 2 [5, 6], followed by a silicon strip detector (SCT), and a transition radiation tracker (TRT) in the region  $|\eta| < 2.0$ . The ID is surrounded by a superconducting solenoid producing an axial magnetic field of 2 T such that the ensemble enables efficient reconstruction of the tracks and momenta of charged particles, measurement of primary and secondary vertices, discrimination between electrons and pions, and reconstruction of photon conversions in ID material at radii up to 800 mm.

The solenoid surrounding the ID is encompassed by an electromagnetic (EM) calorimeter, which consists of lead absorbers folded in an accordion geometry and immersed in a liquid-argon (LAr) bath. The EM calorimeter is divided into three regions, each contained in a separate cryostat: the barrel section (EMB) covering the central pseudorapidity region  $|\eta| < 1.475$  and two endcaps (EMEC) covering the acceptance regions  $1.375 < |\eta| < 3.2$ . The EMB and EMEC are segmented longitudinally into three (two) layers within  $|\eta| < 2.5$  (2.5 <  $|\eta| < 3.2$ ) with variable cell sizes, such that the direction of photon showers can be measured. The first layer (Layer 1) spans the regions  $|\eta| < 1.45$  and  $1.5 < |\eta| < 2.4$ , and has a thickness between three and five radiation lengths  $(X_0)$ , depending on  $\eta$ , and cells with a fine segmentation of  $0.003 \times 0.1$  in  $\Delta \eta \times \Delta \phi$  in the EMB, providing excellent discrimination between single photon showers and the showers of two nearly collinear photons from high-momentum pion decay. The second layer (Layer 2), with a cell granularity of  $0.025 \times 0.025$  in  $\Delta \eta \times \Delta \phi$ , has a thickness between 17  $X_0$  and 20  $X_0$  and collects most of the energy deposited in the calorimeter by electron and photon showers. A third layer (Layer 3) with a thickness of 2–10  $X_0$  and a coarser granularity of  $0.05 \times 0.025$  in  $\Delta \eta \times \Delta \phi$  is used to collect the energy tails of very energetic showers. A thin presampler (PS) layer, placed in front of the accordion layers and covering the region  $|\eta|$  < 1.8, is used to correct for energy losses upstream of the calorimeter. This detector consists of a 1 cm (0.5 cm) active LAr layer in the barrel (endcap) region with a coarse granularity of  $0.025 \times 0.1$  in  $\Delta \eta \times \Delta \phi$ . The typical structure of the EM barrel calorimeter is illustrated in Figure 1. Scintillators are placed between the barrel and endcap cryostats (1.37 <  $|\eta|$  < 1.52) to improve the energy measurement in this region.

The EM calorimeter is surrounded by an iron/scintillator hadron calorimeter in the region  $|\eta| < 1.7$ . In the endcap regions, copper/LAr calorimeters are used up to  $|\eta| = 3.2$ . Energy measurements at higher  $|\eta|$ , up to  $|\eta| = 4.9$ , are made using a combination of forward copper/LAr and tungsten/LAr modules placed inside the endcap cryostats with the EMEC. Muons are accurately measured and identified up to  $|\eta| = 2.7$  by a muon spectrometer located behind the calorimeters. It consists of three air-core superconducting toroids with eight coils each, precision tracking chambers, and fast chambers for triggering up to  $|\eta| = 2.4$ .

A two-level trigger system is used to select events. The first-level trigger is implemented in hardware and uses a subset of the detector information to accept events at a rate below 100 kHz. This is followed by a

<sup>&</sup>lt;sup>1</sup> ATLAS uses a right-handed coordinate system with its origin at the nominal interaction point (IP) in the centre of the detector and the z-axis along the beam pipe. The x-axis points from the IP to the centre of the LHC ring, and the y-axis points upward. Cylindrical coordinates  $(r, \phi)$  are used in the transverse plane,  $\phi$  being the azimuthal angle around the z-axis. The pseudorapidity is defined in terms of the polar angle  $\theta$  as  $\eta = -\ln \tan(\theta/2)$ . The transverse energy is defined as  $E_T = E/\cosh \eta$ .




Figure 1: Sketch of a barrel EM calorimeter module [4].

software-based high-level trigger that reduces the accepted event rate to 1 kHz on average depending on the data-taking conditions.

#### 2.2 Energy measurement, electron and photon reconstruction and identification

The current generated in an EM calorimeter cell by ionizing particles is collected, amplified, and shaped to reduce the impact of out-of time showers, especially in high instantaneous luminosity conditions [7]. The signal is sampled at 40 MHz and digitized by a 12-bit analogue-to-digital converter (ADC) in three different electronics readout gains, high, medium and low, in the front-end boards. The signal observed in the sample matching the trigger time defines which gain to use for the readout. Four digitized samples (two before the sample with highest energy, and one after) are sent to the back-end electronics. The energy deposited in the calorimeter cell is estimated through an optimal filtering procedure applied to the four samples after pedestal subtraction [8], corrected by factors describing the conversion from ADC count to current and from current to energy. The pedestal, ADC-to-current conversion and signal shape of all calorimeter cells are derived from specific electronics calibration data. The pile-up dependence of the resulting energy value is corrected for using the measured instantaneous luminosity for the considered bunch crossings.

Electrons and photons are reconstructed from energy deposits in the cells, using a dynamic, variable cluster-size algorithm to form superclusters [3, 9], allowing the recovery of energy from bremsstrahlung photons or from electrons from photon conversions. In this method an electron candidate is identified as a supercluster matching a track reconstructed in the ID. If a match is found, the track is re-fitted to account for

bremsstrahlung. Superclusters without a matching track in the ID define unconverted-photon candidates. Converted-photon candidates are defined as a cluster matching a track that originates from a conversion vertex. The fraction of photons that convert varies from 20% in the central region to 65% in the endcaps.

Selection criteria are applied after reconstruction to identify genuine electrons while rejecting a large fraction of fake electrons. Depending on the desired background rejection factor and the specific needs of each analysis, four operating points, called Very Loose, Loose, Medium and Tight, are optimized for electrons in bins of  $\eta$  and  $E_{\rm T}$ . The criteria were chosen by using a likelihood discriminant based on a number of track and cluster properties for which probability distributions are derived from electron and pion candidates in data. The identification efficiency of the selection is measured using electrons from Z-boson and  $J/\psi$  decays. Descriptions of the methods, the used samples and the obtained results are given in Refs. [3, 10, 11].

Prompt photons are identified using two selection criteria, Loose and Tight, which are based on the EM shower shapes. The working points are defined in bins of  $|\eta|$  and, in the case of the Tight selection, also in bins of  $E_{\rm T}$ . The Loose criterion is independent of the conversion status of the photon. The Tight selection uses shower shape information from the first calorimeter layer and it is optimized separately for the two cases, accounting for the opening angle of the  $e^+e^-$  pair in the magnetic field, which may impact the response for converted photons. The identification efficiency is measured using distinct data samples: inclusive photon production, photons from radiative *Z*-boson decays, and electrons from *Z*-boson decays after modifying their shower shapes to resemble those of photons. The corresponding analyses are described in detail in Refs. [3, 11, 12].

## 3 Collision data and simulation

#### 3.1 Dataset

The analyses described in this paper use the full pp collision dataset recorded by ATLAS between 2015 and 2018 with the LHC operating at a centre-of-mass energy of  $\sqrt{s} = 13$  TeV and 25 ns bunch spacing. The sample corresponds to an integrated luminosity of 140 fb<sup>-1</sup> after quality cuts [13, 14]; the typical mean number of interactions per bunch crossing,  $\langle \mu \rangle$ , was on average 13, 25 and 37 for the 2015, 2016 and 2017–2018 data, respectively. Special samples, called 'low-pile-up' samples in the following, were recorded in 2017 and 2018 at low instantaneous luminosity, with  $\langle \mu \rangle \sim 2$ ; after applying data-quality requirements, the corresponding integrated luminosity amounts to 340 pb<sup>-1</sup>.

The measurements of the electromagnetic energy response use a large sample of  $Z \rightarrow ee$  events selected with single-electron and dielectron triggers [15]. The dielectron high-level triggers use a transverse energy threshold ranging from 12 GeV (2015) to 17 or 24 GeV (2016–2018) and a Loose (2015) or Very Loose (2016–2018) identification criterion. The single-electron high-level trigger has a transverse energy threshold ranging from 24 GeV in 2015 and most of 2016 to 26 GeV at the end of 2016 and during 2017 and 2018; it applies Tight identification and loose track-based isolation criteria [10]. The offline selection for the energy calibration measurement requires two electrons satisfying Medium identification, loose isolation and  $E_T > 27$  GeV, resulting in  $\sim 57$  million  $Z \rightarrow ee$  candidate events.

A sample of  $J/\psi \to ee$  events with at least two electron candidates with  $E_T > 5$  GeV and  $|\eta| < 2.4$  is used to validate the electron energy scale at low  $E_T$ . It was collected using dedicated prescaled dielectron triggers with asymmetric  $E_T$  thresholds ranging from 4 to 14 GeV. The sample contains ~260000 events.

Samples of  $Z \to \ell\ell\gamma$  events ( $\ell = e, \mu$ ), used to validate the photon energy scale, were selected with the same triggers as for the  $Z \to ee$  sample for the electron channel and with single-muon or dimuon triggers [16] for the muon channel. The high-level dimuon (single-muon) trigger's transverse momentum threshold was 14 (26) GeV; a loose track-based isolation criterion was applied in the high-level single-muon trigger. The  $\mu\mu\gamma$  ( $ee\gamma$ ) samples, after requiring two muons (electrons) with Medium identification [17], transverse momentum  $p_T > 15$  GeV (18 GeV) and one tightly identified and loosely isolated photon with  $E_T > 15$  GeV, contain ~210000 (~100000) events.

#### 3.2 Simulation samples

Large Monte Carlo (MC) samples of  $Z \to \ell\ell$  and  $W \to \ell\nu$  events were simulated at next-to-leading order (NLO) in QCD using Powheg [18] and the Pythia 8 [19] parton shower model. The CT10 [20] parton distribution function (PDF) set was used in the matrix element calculation. The AZNLO set of tuned parameters [21] and the CTEQ6L1 [22] PDF set were used in the modelling of non-perturbative effects. Photos++ 3.52 [23] was used for QED emissions from electroweak vertices and charged leptons.

Both non-prompt (originating from b-hadron decays) and prompt (not originating from b-hadron decays)  $J/\psi \rightarrow ee$  samples were generated using PYTHIA 8. The A14 set of tuned parameters [24] was used together with the CTEQ6L1 PDF set, and EVTGEN 1.2.0 [25] was used to model b- and c-hadron decays.

Samples of  $Z \to \ell\ell\gamma$  events with photon  $E_T > 10\,\text{GeV}$  were generated with Sherpa 2.2.4 [26] using QCD leading-order matrix elements with up to three additional partons in the final state. The NNPDF3.0nnlo PDF set was used in conjunction with the dedicated parton shower tuning developed by the Sherpa authors.

The energy resolution of the new reconstruction algorithm was optimized using samples of 40 million single-electron and single-photon events simulated without pile-up. Their transverse energy distribution covers the range from 1 GeV to 3 TeV. Smaller samples with a flat  $\langle \mu \rangle$  spectrum between 0 and 60 were also simulated to assess the performance as a function of  $\langle \mu \rangle$ .

An extensive software suite [27] is used in data simulation, in the reconstruction and analysis of real and simulated data, in detector operations, and in the trigger and data acquisition systems of the experiment. The generated events were processed through the full ATLAS detector simulation [28] based on Geant4 [29]. The MC events were simulated with additional interactions in the same or neighbouring bunch crossings to match the pile-up conditions during LHC operations. The overlaid *pp* collisions were generated with the soft QCD processes of Pythia 8 using the A3 set of tuned parameters [30] and the NNPDF2.3Lo PDF [31]. Although this set of tuned parameters improves the modelling of minimum-bias data relative to the A2 set [32] used previously, it overestimates the hadronic activity measured using charged-particle tracks by roughly 3%. Simulated events were weighted to reproduce the distribution of the average number of interactions per bunch crossing in data, scaled down by a factor 1.03.

#### 3.3 Passive material model

Measurements of electron and photon energies are affected by the passive material in front of the EM calorimeter. The simulation-based energy calibration accounts for this effect, but any differences between the simulated detector model and the actual detector produce discrepancies between the energy responses. For electrons of  $E_{\rm T} \approx 40$  GeV, such discrepancies are absorbed through the Z-based energy calibration,




Figure 2: (a) Amount of material traversed by a particle, in units of radiation lengths  $X/X_0$ , as a function of  $|\eta|$  in the nominal simulation. (b) Measured difference between the data and the nominal simulation of the detector material up to the first layer of the EM calorimeter. The data points are obtained using the longitudinal shower profile of electrons from Z-boson decays in partial Run 2 data, and have negligible statistical uncertainties. The blue band summarizes the material uncertainties, relative to the nominal simulation, as determined in Run 1 calibration studies. The green band includes additional uncertainties related to the introduction of new material into the detector for Run 2, and covers potential mis-modelling of the insertable B-layer and a modified inner-detector patch panel (PP0). At high  $|\eta|$ , these uncertainties are dominated by a 50% uncertainty in the simulated PP0 material.

but biases remain that depend on the particle type (electron, unconverted or converted photon) and energy. The model used for the passive material and its associated uncertainties was derived from Run 1 and partial Run 2 data [1, 10], and is summarized briefly below.

The model of the ID is based on measurements performed during its construction [4], leading to a 5% uncertainty in the amount of material. In Run 2, a 10% uncertainty is assigned to the material of the insertable B-layer, and 25% to that of the inner-detector service patch panel [33], affecting the high- $|\eta|$  region.

The material between the ID and the LAr calorimeter was probed [1] using the longitudinal development of electron and photon showers. The ratio of the energies deposited in the first and second accordion layers was found to be sensitive to the total amount of material traversed by these particles before entering the calorimeter. After calibration of the layers' energy response, this ratio was measured with a relative precision of about 1% in the barrel, and about 2% in the endcaps, leading to a determination of the passive material with a typical precision of  $5\%X_0$  for  $|\eta| < 1.3$ ,  $10\%X_0$  for  $1.6 < |\eta| < 2.1$ , and up to  $20\%X_0$  for  $2.1 < |\eta| < 2.4$ . The impact of the passive material uncertainties on the energy measurement was parameterized using simulation as a function of the particle type,  $E_T$ , and  $|\eta|$ . The corresponding calibration uncertainties mostly affect electrons at low  $E_T$  and unconverted photons, and typically reach 1%-2%. No measurement was performed in the barrel–endcap transition regions, so these regions have larger uncertainties.

The resulting material model is presented in Figure 2(a), which summarizes the passive material between the interaction point and the calorimeter. The measured passive material is compared with the simulation and its uncertainties in Figure 2(b).

## 4 Overview of the calibration procedure

The different steps in the procedure used to calibrate the energy response for electrons and photons described in this paper are illustrated in Figure 3, and summarized below.




Figure 3: Schematic overview of the electron and photon energy calibration procedure in ATLAS.

The energy of an electron or photon candidate is built from the energy of a cluster of cells in the electromagnetic calorimeter. The measurement of electron and photon energies is optimized using a simulation-based boosted-decision-tree regression algorithm, combining energy deposits belonging to the reconstructed supercluster, in the presampler and in the calorimeter layers. The optimization is performed separately for electrons, converted photons and unconverted photons, taking into account the particle position. It is the same as the one used in Ref. [3] and the methodology is discussed in detail in Ref. [2].

Since the EM calorimeter is segmented in depth, the longitudinal layers should be calibrated separately to provide a correct description of the calorimeter response as a function of  $E_T$  (step 1). After these corrections, the simulation-based calibration is applied identically to the cluster energies reconstructed from collision data and simulated event samples (step 2).

A set of additional corrections is applied to data to account for response variations not included in the simulation in specific detector regions, e.g. regions with non-optimal high voltage, azimuthal non-uniformities, or biases associated with the liquid-argon calorimeter's electronics calibration (step 3). The stability of the calorimeter response as a function of azimuth, time and pile-up is also studied.

A final adjustment of the calorimeter response is derived from samples of  $Z \rightarrow ee$  events, so that the peak of the Z resonance reconstructed in data coincides with that in the simulation. The response corrections are applied to the data. Using the same event samples, it is found that the resolution in data is slightly worse than that in simulation, and appropriate corrections are derived and applied to the simulation to match the data (steps 4, 5). The passive material model and the intercalibration corrections carry uncertainties that affect the energy calibration with a specific dependence on the particle  $E_{\rm T}$ . Measuring the residual energy dependence of the energy scale thus allows further adjustments of the calibration model, and provides additional constraints on the associated uncertainties. High-precision measurements, such as

those measuring the masses of the Higgs and W bosons, will profit significantly from these uncertainty reductions.

The calibration factors extracted from  $Z \to ee$  events are assumed to reflect the intrinsic response of the calorimeter, and are thus applied identically to electrons and photons. Nevertheless, photon-specific corrections are needed to account for differences in the lateral development of electron and photon showers (step 6). Finally, the calibration chain is validated in data with low- $E_{\rm T}$  electron candidates from  $J/\psi \to ee$  decays, and with photon candidates from radiative Z-boson decays (step 7).

## 5 Effects on the uniformity and stability of the energy response

This section discusses the stability of the calorimeter energy response as a function of azimuthal angle, time, and pile-up. The dependence of the energy reconstruction on the readout electronics (ADC calibration and readout gain) is also discussed. Not all effects are modelled in the simulation. Corrections are defined for the ADC calibration and the azimuthal non-uniformity, while a systematic uncertainty is assigned for the observed dependence of the energy measurement on the readout gain.

#### **5.1** Uniformity

Gravity-induced mechanical deformations of the calorimeter cause variations in the size of the liquid-argon gaps between the absorbers as a function of azimuthal angle. The resulting energy response variations are at the level of 0.1%-0.2% in the barrel, and up to 1% in the endcaps. Energy corrections are derived from the modulations of the response as a function of  $\phi$ , separately in six intervals of absolute pseudorapidity (0-0.6, 0.6-1.0, 1.0-1.37, 1.37-1.55, 1.55-1.82, 1.82-2.47). In each  $|\eta|$  interval, the relative response is defined from profiles of  $E_T/\langle E_T \rangle$  in Z-boson decays, where  $E_T$  is the electron transverse energy at a given  $\phi$  value, and  $\langle E_T \rangle$  is the average over  $\phi$ .

The effect of this correction is illustrated in Figure 4. While the  $\phi$ -averaged energy response is unchanged by construction, the better uniformity is expected to yield a small improvement in the overall energy resolution. In practice, the resolution's constant term, discussed in Section 7, is reduced by 5%–10% in the endcaps, where the correction is most significant.

#### 5.2 Stability

The stability of the calorimeter response is studied using the reconstructed peak position of the dielectron mass distribution,  $m_{ee}/\langle m_{ee} \rangle$ , in  $Z \to ee$  candidate events. This is illustrated in Figure 5(a) for the data taken between 2015 and 2018. Stability at the level of 0.05% is observed over the full data-taking period.

Figure 5(b) shows  $m_{ee}/\langle m_{ee}\rangle$  as a function of the average number of interactions per bunch crossing for the data collected between 2015 and 2018. The bipolar shaping of the calorimeter signals [7] protects the energy measurement against pile-up fluctuations, and after correcting for bunch-to-bunch variations of the instantaneous luminosity, the residual dependence of the energy scale on  $\langle \mu \rangle$  is below 0.1%. The small increase in energy observed in data is consistent with the MC expectation over most of the  $\langle \mu \rangle$  range and is related to the new dynamical clustering used for the energy measurement [3]. The high- $\langle \mu \rangle$  region mostly




Figure 4: Electron energy response as a function of electron azimuth  $\phi$ , in the endcaps, before (red dots) and after (black dots) the  $\phi$ -uniformity energy correction. This correction was computed from 2015+2016 data and applied to the full Run 2 sample. The observed non-zero residuals have no impact on the final result

reflects data taken in 2017, and the discrepancies observed in this region justify the extraction of dedicated energy-scale corrections for each data-taking year. More details are given in Section 7.

The stability of the response as a function of the number of reconstructed collision vertices  $(N_{\text{vtx}})$  [34] is shown in Figure 5(c). Classifying events according to  $N_{\text{vtx}}$ , related to the number of interactions in the specific bunch crossing, biases the pile-up activity of colliding bunches relative to the average. In this case the compensation of the pile-up contributions to the reconstructed energy by the bipolar shaping becomes imperfect, giving rise to the observed slope. The description of this effect in the simulation is accurate to 0.1% for  $\langle \mu \rangle < 25$ , rising to 0.5% at high  $N_{\text{vtx}}$ . The larger discrepancy at high  $N_{\text{vtx}}$  reflects the effect observed for  $\langle \mu \rangle$ .

#### 5.3 ADC non-linearity correction

The energy reconstruction in a LAr calorimeter cell is based on a linear conversion from ADC counts to current, and an additional factor converting current into energy. In practice, the energy response of each cell is determined during dedicated electronics calibration runs, parameterizing the relation between the injected current and the measured ADC count using a linear function. The fits to the calibration data show non-zero residuals, caused by intrinsic non-linear behaviour of the electronics. This is illustrated in Figure 6 for an example cell and a calibration run performed in medium gain (MG). The residuals deviate from a linear fit by about 0.3%; comparable deviations are observed in high-gain (HG) calibration runs.

The implications of this non-linearity are twofold. First, the residuals of the ADC-to-current correction contribute to the non-linearity of measurements of electron and photon energies, and must therefore be accounted for in view of the linearity analysis in Section 10. A correction is implemented, based on fits of a fifth-order polynomial to the residuals, separately for each cell of the calorimeter, as exemplified in Figure 6. The parameterized residual is then added to the cell energy estimate. The impact on electron or photon cluster energies is estimated by repeating this procedure for all cells belonging to the cluster, and computing the modified cluster energy. The final correction is built such that it does not modify the cluster energies of electrons at  $E_T = 40$  GeV, where the global energy scale is set (see Section 7). Results of this procedure for electrons, unconverted photons and converted photons are illustrated in Figure 7. Cluster energies are increased by about 0.4% at low  $E_T$ , and decreased by about 0.2% at high  $E_T$ , with a




Figure 5: Relative variation of the peak position of the reconstructed dielectron mass distribution in  $Z \to ee$  events as a function of (a) time, (b) the average number of pile-up interactions,  $\langle \mu \rangle$ , and (c) the number of reconstructed collision vertices,  $N_{\text{vtx}}$ .

moderate dependence on particle type and pseudorapidity. A relative uncertainty of 30% is assigned to this correction.

In addition, non-linearities in the ADC-to-current conversion affect comparisons of the energy response in different readout gains. For a given energy, the measured ADC count and the residuals of the current conversion depend on the gain in which the cell is recorded, so that a direct comparison of the energy response in HG and MG is difficult to interpret. The ADC non-linearity correction described above removes this bias. The energy measurement in high and medium gains can then be compared directly, as described in the next section.

#### 5.4 Energy response in high and medium gain

Non-linearity in the cell energy measurement introduces a dependence of the energy response on the energy of the reconstructed particle. As discussed above, the linearity of the readout electronics is better than a few per mille in each of the three gains used to digitize the calorimeter signals. The relative calibration of the high, medium and low readout gains is assumed to be perfect in the simulation, but is less well




Figure 6: Output of a calibration run performed in medium gain, for an example cell in the second layer of the EM calorimeter. The relation between injected current, in DAC (digital-to-analogue converter) units, and ADC counts is assumed to be linear (top panel). The evolution of the residuals (in DAC units) between the measurements and the linear fit as a function of ADC counts (bottom panel) is parameterized with a fifth-order polynomial and used to correct the cell energy.




Figure 7: Relative cluster energy correction as a function of  $E_{\rm T}$ , for (a)  $|\eta| < 0.8$  and (b)  $1.8 < |\eta| < 2.4$  for electrons (black), unconverted photons (red) and converted photons (blue). For each particle type, the size of the envelope reflects the dependence of the correction on  $\eta$ , at a given  $E_{\rm T}$ .

understood in data [1, 2]. The HG and MG readouts are most relevant for the calibration studies presented in this paper, and discussed below.

In the standard configuration, the HG readout is used for the majority of cells in clusters from electrons in  $Z \rightarrow ee$  decays, especially in the barrel, where the transition to medium gain occurs for second-layer cell energies above 25 GeV, corresponding to electrons of  $E_T = 50$ –60 GeV. The relative calibration of the HG and MG readouts is studied using 0.3 fb<sup>-1</sup> of data recorded in 2017 and 2018 using special conditions. For these data, the threshold to switch from HG to MG readout for the cells in the second layer was lowered, typically by a factor of three, such that almost all electrons from Z-boson decays have at least the highest-energy cell in Layer 2 recorded in MG. The reconstructed dielectron invariant mass distribution in the special runs is compared with the distribution observed in 1.5 fb<sup>-1</sup> of data with the standard gain transition, and recorded around the same time. The procedure described in Section 5.3 is

used to calibrate the ADC-to-current conversion function in both gains. An example comparison is shown in Figure 8(a) for electrons with  $|\eta| < 0.8$ . The invariant mass distributions built in standard runs are parameterized with a sum of three Gaussian functions in (i, j) categories, where i and j label the five  $|\eta|$  regions<sup>2</sup> where the electrons of the pair are reconstructed. The same shapes are used to describe the corresponding distributions for the data taken in the special runs, but with the means and widths of the Gaussian functions modified by a multiplicative factor  $\sqrt{(1 + \alpha_{G,i})(1 + \alpha_{G,j})}$ , where  $\alpha_{G,i}$  is the energy scale difference between the two datasets in the  $|\eta|$  region i. These  $\alpha_{G,i}$  parameters are then extracted from a simultaneous fit of all (i, j) regions.

The measured values of  $\alpha_G$  are shown in Figure 8(b) as a function of  $|\eta|$ . The uncertainties shown are statistical only; systematic uncertainties are negligible for this study. For perfectly intercalibrated HG and MG readouts,  $\alpha_G = 0$ . Instead, small but significant differences are observed, especially for  $0.8 < |\eta| < 1.37$ . Compared to the previous analysis of this effect in Ref. [2],  $\alpha_G$  is found to be smaller by about a factor of two; this change is mostly driven by the improved ADC calibration using a residual correction discussed in Section 5.3.

The relative difference between the energy responses in HG and MG can be written as a function of the particle type,  $E_T$  and  $\eta$  as follows:

$$\frac{\Delta E}{E} = \alpha_{\rm G}(\eta) \cdot \frac{1}{\delta_Z(\eta)} \cdot \delta_{\rm MG}^{e,\gamma}(\eta, E_{\rm T}),\tag{1}$$

where

- $\delta_Z(\eta)$  quantifies the fractional change in energy for electrons from Z-boson decays between the data with lower and standard thresholds for a given change in the energy recorded in MG. This sensitivity factor is about 0.3 to 0.4 (0.2 to 0.25) in the barrel (endcap) calorimeter. It takes into account the fact that only a fraction of the electron energy is recorded in MG in the special-settings data taking, while in data recorded with normal settings some cells in the second layer can be read out in MG. This is particularly true for the endcaps, where the electron energies are larger.
- $\delta_{\mathrm{MG}}^{e,\gamma}(\eta,E_{\mathrm{T}})$  quantifies, for a given particle, the fractional change in total energy for a given change in the energy recorded in MG, for standard gain thresholds. It is estimated using simulated single-particle samples. It is close to zero up to  $E_{\mathrm{T}} = 50$  GeV for electrons and  $E_{\mathrm{T}} = 40$  GeV for photons, and rises to reach a maximum value of about 0.7 for  $E_{\mathrm{T}} \sim 400$  GeV in the barrel, and of about 0.5 for  $E_{\mathrm{T}} \sim 200$  GeV in the endcap. Above these values,  $\delta_{\mathrm{MG}}^{e,\gamma}(\eta,E_{\mathrm{T}})$  decreases as the low-gain readout starts contributing.

The gain dependence of the energy response,  $\Delta E/E$ , is considered as a systematic uncertainty in the energy measurement; the size of the uncertainty is defined as the full size of the observed dependence. The effect increases with energy, and typically reaches 0.1% in the barrel, and 0.4% in the endcaps, with low dependence on the particle type.

For completeness, a corresponding, independent uncertainty is defined for the transition between the MG and low-gain (LG) readouts. The uncertainty is defined based on Eq. (1), adapted using a dedicated parameterisation  $\delta_{LG}^{e,\gamma}(\eta,E_T)$ . It has a sub-dominant impact for the energy range considered in the present analysis.

<sup>&</sup>lt;sup>2</sup> The definition of these regions is motivated by the calorimeter geometry.




Figure 8: (a) Example dielectron invariant mass distributions, one from events collected in special runs and the other from standard runs, and (b) corresponding energy scale factors and their statistical uncertainties as a function of  $|\eta|$ .

## 6 Intercalibration of the EM calorimeter layers

#### **6.1 Presampler energy scale**

The presampler energy scale  $\alpha_{PS}$  is defined as the ratio of the presampler energies in data and simulation. The analysis presented in Refs. [1, 2] used samples of electrons and photons to measure  $\alpha_{PS}$ . However, the interplay between  $\alpha_{PS}$  and the amount of material in front of the presampler produces a strong correlation between the corresponding uncertainties. In this paper, the determination of  $\alpha_{PS}$  is performed using muon candidates selected in the low-pile-up data sample. The muon energy deposits are insensitive to the material in front of the presampler, and therefore provide a direct measurement of the presampler energies. A high-purity sample of  $W \to \mu \nu$  and  $Z \to \mu \mu$  events is selected by requiring one or two isolated muons, with additional criteria for the transverse energy and mass, and in the Z-boson case, the dimuon mass.

Although the signal-to-noise ratio for a typical energy deposit from a muon in a presampler cell is rather low, the near absence of pile-up in low-pile-up data ensures it is still significant if the sample is large enough. Figure 9 shows examples of muon energy deposits in data and MC samples. The noise distributions, as measured by the energy deposits in the neighbouring cells located at  $\Delta \eta = \pm 0.025$  from the crossed cells, are also shown for the simulation. The peak position in data and simulation matters for this study, while differences in the width, as observed in the endcaps, have no impact. The mean energy deposits measured in a cell are  $\sim 45,100$  and 75 MeV at  $\eta \sim 0,1.2$  and 1.6, respectively.

The mean values of these distributions are extracted within the interval [-1.6, 1.6] GeV for data and simulation, and  $\alpha_{PS}$  is determined from their ratio. The measurement is performed in nine pseudorapidity bins corresponding to the size of the presampler modules ( $\Delta \eta = 0.2$  up to  $|\eta| = 1.4$ ,  $\Delta \eta = 0.12$  in the last barrel module and  $\Delta \eta = 0.3$  in the endcaps).

Systematic uncertainties in the measurements of  $\alpha_{PS}$  are estimated as the envelope of the difference between the nominal  $\alpha_{PS}$  values and alternative measurements obtained by 1) varying the muon candidate selection, i.e. selecting candidates crossing a cell close to its centre; 2) varying the presampler energy definition, i.e. using a cluster of three cells in  $\eta$  instead of a single cell; 3) subtracting a pedestal energy estimated from




Figure 9: Distributions of muon energy deposits in presampler cells traversed by muon tracks, in two  $|\eta|$  regions: (a) barrel and (b) endcap. The corresponding distributions for the neighbouring cells in  $\eta$  are shown for the simulation by the dashed histograms.

neighbouring cells in  $\phi$ ; and 4) varying the interval used to determine the mean of the  $E_{PS}$  distributions. These changes contribute equally to the total systematic uncertainty. At a given  $|\eta|$ , the values of  $\alpha_{PS}$  are found to be compatible for positive and negative pseudorapidity, and are averaged. The results are shown in Figure 10. The statistical uncertainties are at the percent level, while the systematic uncertainties vary between 2% and 4% depending on  $|\eta|$ . The results obtained here using muons are compatible with the earlier electron- and photon-based measurements.




Figure 10: Measured presampler energy scale  $\alpha_{PS}$  as a function of  $|\eta|$ . The error bars represent the statistical uncertainty and the yellow band shows the total uncertainty.

#### 6.2 Intercalibration of the first and second calorimeter layers

The intercalibration of the first and second calorimeter layers is paramount in controlling the linearity of the electron and photon responses. Differently from previous iterations, both the muon and electron candidates from *Z*-boson decays are used here to measure the relative calibration of the two layers.

#### 6.2.1 Intercalibration using muons

The measurements using muon candidates from Z-boson decays closely follow the analysis in Ref. [2]. The intercalibration factor is defined as  $\alpha_{12} = (\langle E_1 \rangle / \langle E_2 \rangle)^{\text{data}} / (\langle E_1 \rangle / \langle E_2 \rangle)^{\text{MC}}$  where  $\langle E_i \rangle$  is the mean value of the distribution of the energy deposited in layer i. Two methods differing in the estimation of  $\langle E_i \rangle$  are considered: the most probable value method extracts it from a fit of a Landau distribution convolved with a noise template. The truncated mean (TM) method uses the mean of the distribution computed over a restricted window to minimize the sensitivity to the tails. In both methods, these quantities are measured in intervals of  $\langle \mu \rangle$  and the final  $\alpha_{12}$  measurement is obtained by extrapolating linearly to  $\langle \mu \rangle = 0$ . The extrapolation parameters are determined from the measurements in the interval  $\langle \mu \rangle \in [10, 40]$ . In the TM method, the extrapolation was validated by comparing the nominal results with the  $\alpha_{12}$  value estimated with the low-pile-up data sample, for which muon candidates from W-boson decays are considered in addition to candidates from  $Z \to \mu\mu$  decays. This is illustrated for  $0.3 < |\eta| < 0.4$  in Figure 11(a); a comparison of the measurements using the standard data sample and the low-pile-up one is shown in Figure 11(b). For both methods, calibration uncertainties are derived by varying the muon selection cuts and the fitting or averaging ranges, as described in Ref. [2]. Agreement between the two measurements is excellent and well within the total uncertainty of the nominal result. The total uncertainty in the muon measurements varies from about 0.7% in the barrel to about 2% in the endcaps.




Figure 11: (a) Evolution of the truncated mean of the muon energy deposit distribution for  $0.3 < |\eta| < 0.4$ , for the first (dots) and second (triangles) calorimeter layers, in data (open symbols) and in simulation (full symbols), as a function of the average number of pile-up interactions per bunch crossing  $\langle \mu \rangle$ . The values obtained from the low-pile-up samples are also shown, together with the values extracted from a MC sample without pile-up. The lines show the result of linear fits to the points for  $\langle \mu \rangle \in [10, 40]$  and the dotted lines show the extrapolation to lower and higher  $\langle \mu \rangle$ . (b) Comparison of the evolution with  $|\eta|$  of  $\alpha_{12}$  obtained from extrapolation at  $\mu=0$ , for the low-pile-up and standard data samples, for the TM method. The yellow band shows the total uncertainty of the nominal results, obtained from the standard dataset. The error bars on the data points are statistical only.

#### **6.2.2** Intercalibration using electrons

In the previous analyses [1, 2], electron probes were used only as a cross-check of the layer intercalibration performed with muons. From the few discrepancies between electron and muon correction factors observed in some pseudorapidity regions, systematic uncertainties were derived and applied to the muon result. The

method first described in Ref. [1] was revisited with the full Run 2 data sample in order to better constrain the layer intercalibration using electrons, and to combine the electron-based measurement with the result obtained from muons.

The dependence of the energy response on the depth of the EM shower allows a direct extraction of  $\alpha_{12}$ . This parameter is determined in 25 bins of  $|\eta| \in [0, 2.4]$ , using the ratio of the measured energy E and momentum p of electron candidates, E/p, and the distribution of the invariant mass of electron pairs,  $m_{ee}$ . The E/p distribution is affected by bremsstrahlung from interactions between the electrons and the material in front of the calorimeter, and is modelled using a Crystal Ball distribution [35] in the range 0.9 < E/p < 1.3. For this method, only the region  $|\eta| < 1.37$  is considered since the ID momentum resolution deteriorates rapidly at large  $|\eta|$ . The invariant mass distribution is fitted in each  $(|\eta|, E_1/E_2)$  bin<sup>3</sup>, in the range [80, 100] GeV, using the convolution of a Crystal Ball function and a Breit–Wigner distribution, while the small background in data is modelled by a second-order Chebyshev polynomial. In both the E/p and  $m_{ee}$  methods, the most probable value of the fitted Crystal Ball function is used as the estimator. Examples of fits to the dielectron invariant mass and the E/p distributions are illustrated in Figure 12.




Figure 12: Distributions of (a) the dielectron invariant mass and (b) electron E/p for electrons with  $0.1 < |\eta| < 0.2$  and  $0.1 < E_1/E_2 < 0.2$ . The fitted functions used to estimate the most probable value are superimposed.

In a given  $|\eta|$  bin, the dependence of the E/p and  $m_{ee}$  estimators on  $E_1/E_2$  is determined in data and in simulation. For a perfect layer intercalibration, the ratio of the estimator in data and the nominal simulation is expected to be constant, and any mis-calibration induces a slope. A constant data-to-simulation ratio is recovered by rescaling  $E_1$  in data and recomputing  $m_{ee}$  and E/p accordingly, adjusting  $\alpha_{12}$  to minimize the deviation of the ratio from a horizontal line. Figures 13(a) and 13(b) show the dependence of  $m_{ee}$  and E/p on  $E_1/E_2$  for data and simulation, for  $|\eta| \in [0.1, 0.2]$  ( $m_{ee}$ ) and  $|\eta| \in [1.0, 1.1]$  (E/p). Data-to-simulation ratios are shown in the bottom panels. For  $|\eta| \in [0.1, 0.2]$ , the uncorrected ratio is compatible with a constant, indicating a good layer intercalibration, while for  $|\eta| \in [1.0, 1.1]$ , a constant ratio is recovered after rescaling  $E_1$  by 0.97. The measurements of  $\alpha_{12}$  using  $m_{ee}$  or E/p are compatible in the whole  $|\eta|$  range.

The following sources of systematic uncertainty are considered:

<sup>&</sup>lt;sup>3</sup> Each  $m_{ee}$  value is considered in the two relevant ( $|\eta|$ ,  $E_1/E_2$ ) bins, where  $E_i$  is the cluster energy contained in Layer i.




Figure 13: Evolution of (a)  $m_{ee}$  and (b) E/p estimators as a function of  $E_1/E_2$  for the data (full dots), the data with a scaling of  $E_1$  (red squares), and the nominal simulation (open squares). The  $m_{ee}$  and E/p figures show values for electrons in the pseudorapidity bin  $0.1 < |\eta| < 0.2$  and  $1.0 < |\eta| < 1.1$  respectively. The bottom panels show the ratio to the nominal simulation. In addition, the evolution for a simulation with additional material is also shown. For this sample, the range  $|\eta| < 1.35$  is used.

- The effect of uncertainties in the amount of material is estimated from simulations in which the passive material in the barrel cryostat, the SCT and the endcaps of the TRT is increased according to the uncertainties described in Section 3.3. The distorted MC samples are used as pseudo-data, and the deviation of  $\alpha_{12}$  from 1 defines the associated uncertainty. As can be seen in ratio panels of Figure 13, the method is rather insensitive to passive material variations (the constant ratio of the distorted MC to nominal MC values in the range  $|\eta| < 1.35$  is also observed in finer  $|\eta|$  bins). The uncertainty ranges from  $\sim 0.5\%$  at low  $|\eta|$  to  $\sim 5\%$  for  $|\eta| \in [1.4, 1.8]$ .
- The intervals used to fit the energy response are modified to [70, 100] GeV or [84, 98] GeV for the  $m_{ee}$  method, and to [0.95, 1.5] for the E/p method. The difference between the  $\alpha_{12}$  values measured in the nominal and alternative intervals is used as an uncertainty.
- The uncertainty in the presampler energy scale also impacts the measurement. The corresponding uncertainty in  $\alpha_{12}$  can be as large as 1% at the end of the barrel for the E/p method.
- Variations of the electron identification working point, the residual bias of the method, and the limited size of the Z → ee sample contribute with much smaller uncertainties.

The precision of the  $\alpha_{12}$  measurement with electrons ranges from 0.7% to 2% in the barrel, and from 1.5% to 6.2% in the endcap calorimeters. In the barrel–endcap transition regions and in the first half of the endcap calorimeters, the distributions in data and simulation differ significantly, regardless of  $\alpha_{12}$ , leading to poor convergence of the minimization procedure and increased uncertainties.

#### 6.2.3 Combination

The two muon measurements and the two electron measurements are combined using the BLUE prescription [36]. In each channel, the statistical correlation between the two methods is ignored, as systematic uncertainties dominate; the latter are found to be largely uncorrelated as a function of  $|\eta|$ . First, the muon

and electron measurements are combined separately. The results are shown as the blue and red open dots in Figure 14(a). The uncertainty in the combined muon measurement includes the systematic uncertainty induced by the extrapolation of this calibration to electron and photon showers [1]. This uncertainty accounts for the uncertainties in the simulation of the ionization current induced by muons, and how these uncertainties affect EM showers. These two combined results are then combined to provide the final measurement, illustrated in Figure 14(a). When the  $\chi^2$  of the combination is larger than one, the combined uncertainties are rescaled by a factor  $\sqrt{\chi^2}$ . The total uncertainty is presented in Figure 14(b), and varies between 0.6% in the central part of the barrel to 3% at  $|\eta| \sim 2.4$ . The inclusion of the electron measurement allows the uncertainty to be reduced by a factor of  $\sim$ 1.8 in the first half of the barrel. In the endcaps, the combined uncertainties are dominated by the differences between the electron and muon results.




Figure 14: (a) Relative calibration scale factor  $\alpha_{12}$  of the first and second EM calorimeter layers as a function of  $|\eta|$ . Open blue squares correspond to the results obtained from the study of muon energy deposits in  $Z \to \mu\mu$  events, combining the *truncated mean* and *most probable value* methods. Open red circles show the extracted combined values obtained from the study of the dependence of the dielectron invariant mass  $m_{ee}$  and the E/p ratio as a function of  $E_1/E_2$  in  $Z \to ee$  events. The final scale factors, combining electron and muon results, are shown as the black solid lines. (b) Corresponding total uncertainties, taking into account rescaling in the case of large  $\chi^2$ .

## 7 Determination of the energy scale and resolution with $Z \rightarrow ee$ events

Electron energy scale and resolution corrections are determined using electron pairs from Z-boson decays. For electrons reconstructed in  $\eta$  regions labelled i and j, the difference between the positions of the resonance in data and simulation is used to determine invariant mass scale corrections  $\alpha_{ij}$ , defined by  $m_{ij}^{\text{corr}} = m_{ij}/(1 + \alpha_{ij})$ . Similarly, a correction to the mass resolution is parameterized as  $(\sigma_m/m)_{ij}^{\text{corr}} = (\sigma_m/m)_{ij} \oplus c_{ij}$ . The values of  $\alpha_{ij}$  and  $c_{ij}$  are those which give the best agreement between the invariant mass distributions in data and simulation, separately for each (i, j) category.

The analysis in Section 6.1 measures an absolute energy scale for the presampler, but the calibration procedure for the first and second accordion layers presented in Section 6.2 only determines their relative response. Since the contribution of the third accordion layer to the energy measurement is negligible for electrons from Z-boson decays, the only remaining degree of freedom is the overall energy scale

of the accordion calorimeter. Accordion energy<sup>4</sup> scale corrections for single electrons, defined by  $E_i^{\text{acc},\text{corr}} = E_i^{\text{acc}}/(1 + \alpha_i^{\text{acc}})$ , are at first order related to the invariant mass scale factors  $\alpha_{ij}$  following

$$\alpha_{ij} = \frac{f_i^{\rm acc}\alpha_i^{\rm acc} + f_j^{\rm acc}\alpha_j^{\rm acc}}{2},$$

where  $f_i^{\rm acc}$ , shown in Figure 15, is determined from simulation. It represents the sensitivity of the total calibrated electron energy to the energy measured in the accordion calorimeter, for electrons in  $\eta$  bin i. It is expected to be smaller than one since part of the electron energy is deposited in the presampler (for  $|\eta| < 1.8$ ) and in the scintillators in the transition regions between the barrel and endcap calorimeters  $(1.4 < |\eta| < 1.6)$ . The accordion energy scale corrections are applied only to the data.




Figure 15: Sensitivity of the calibrated electron energy to the energy measured in the accordion calorimeter,  $f^{acc}$ , as a function of  $|\eta|$  for electrons from Z-boson decays.

The invariant mass resolution correction  $c_{ij}$  can be expressed in terms of single-electron energy resolution corrections,  $c_i$ , as  $c_{ij} = (c_i \oplus c_j)/2$ . The resolution correction is applied to the reconstructed energy in simulation.

The invariant mass window considered for the determination of  $\alpha_{ij}$  and  $c_{ij}$  is  $80 < m_{ee} < 100$  GeV, and the  $\alpha_i^{\rm acc}$  and  $c_i$  parameters are extracted from simultaneous fits of all categories. Possible biases from imperfect modelling of Z-boson production are removed by reweighting the simulated two-dimensional Z-boson transverse momentum and rapidity distribution to the data. Figure 16 shows the results for (a)  $\alpha_i^{\rm acc}$  and (b)  $c_i$  derived in 68 and 24  $\eta$  intervals, respectively, separately for the 2015, 2016, 2017 and 2018 data samples. Changes in the  $\alpha_i^{\rm acc}$  values between successive years are mainly due to variations of the LAr temperature and the instantaneous luminosity. The temperature variations induce changes in the charge/energy collection, affecting the energy response by about -2%/K [37]. Increases in luminosity during Run 2 imply that more energy is deposited in the liquid-argon gap, which creates a higher current in the high-voltage lines, effectively reducing the high voltage applied to the gap and changing the response by up to 0.1% in the endcap regions. Given the small size of the observed dependence, dedicated energy scale corrections for each data-taking year provide adequate stability for the energy measurement.

The energy scale correction is applied as an overall correction to the energy measured in the three accordion layers, and the Z-based calibration fit is repeated to verify that the procedure converges. The residual energy scale factors  $\alpha_i^{\text{closure}}$  obtained from this iteration are also shown in Figure 16(a). Their values are

<sup>&</sup>lt;sup>4</sup> The accordion energy is the sum of the energies in the three layers of the accordion calorimeter:  $E^{acc} = \sum_{layer=1}^{3} E_{layer}$ 




Figure 16: (a) Energy scale calibration factors  $\alpha_i^{\rm acc}$  and  $\alpha_i^{\rm closure}$ , and (b) the additional constant term  $c_i$ , as a function of  $\eta$ . The shaded areas correspond to the statistical uncertainties. The bottom panels show the differences between (a)  $\alpha_i^{\rm acc}$  and (b)  $c_i$  measured in a given data-taking period and the measurements using 2018 data.

below  $10^{-4}$  everywhere except in the transition regions,  $1.37 < |\eta| < 1.52$ , where scintillators installed between the barrel and endcap cryostats contribute to the energy measurement but are not the subject of a specific calibration. The residual non-closure is applied as a final correction to the reconstructed energy.

For the constant term resolution corrections  $c_i$ , a dependence on the pile-up level is observed through the different values obtained from the 2015–2018 data. The dependence of the  $c_i$  values on the amount of pile-up is explained by the larger pile-up noise predicted by the simulation, compared with that observed in the data, for a given value of  $\mu$  [3]. A weighted average of the  $c_i$  values for the different years is applied in the analyses of the complete dataset. The additional constant term in the energy resolution is typically less than 1% in most of the barrel and between 1% and 2% in the endcaps.

Systematic uncertainties in the determination of  $\alpha_i^{\text{closure}}$  and  $c_i$  are assessed using variations of the event selections and fitting range. The event selection variations are a change of the electron identification criterion from Medium to Tight, removal of the isolation criterion, and addition of a cut on the momentum fraction lost by bremsstrahlung inside the ID. The mass window used for the fit is varied from  $80 < m_{ee} < 100 \text{ GeV}$  to  $87 < m_{ee} < 94.5 \text{ GeV}$ . The combined effect of these variations is a systematic uncertainty of about 0.05% in  $\alpha_i^{\text{closure}}$  over most of the calorimeter acceptance, but reaching 0.5% in the transition regions between the barrel and endcap calorimeters and at the edge of the calorimeter ( $|\eta| > 2.3$ ). The systematic uncertainty in the  $c_i$  corrections ranges between 0.1% and 0.2% across the detector acceptance. The dominant contribution to the systematic uncertainty comes from the mass window variation.

Figure 17 shows the invariant mass distribution of  $Z \to ee$  candidates in data and in simulation after applying the energy scale correction to the data and the resolution correction to the simulation. Background contamination is not taken into account in this comparison, but it is expected to be no more than 1% in the shown mass range. The uncertainty band corresponds to the propagation of the uncertainties in the  $\alpha_i^{\text{closure}}$  and  $c_i$  factors. Within these uncertainties, the data and simulation are in fair agreement.




Figure 17: Comparison of the invariant mass distributions of the electron pair in the selected  $Z \rightarrow ee$  candidates in data and simulation, after the calibration and resolution corrections are applied. The total number of events in the simulation is normalized to that in data. The ratio between data and simulation is shown in the bottom panel, where the uncertainty band represents the impact of the uncertainties in the calibration and resolution correction factors.

## 8 Photon-specific calibration

#### 8.1 Modelling of the photon reconstruction classification

The simulation-based energy reconstruction procedure, discussed in Section 4, is optimized and applied separately for unconverted photons and converted photons. Therefore, a difference between data and simulation in the rates of classifying photons as converted or unconverted generates a bias in the photon energy scale. Misclassifications arise from inefficiencies in the conversion-finding algorithm and from the incorrect classification of genuine unconverted photons as converted photons by matching the cluster to pile-up-induced tracks. The rates of correct and incorrect classification are measured using a sample of photons selected from radiative Z events. As described in Ref. [12], these rates are evaluated, in both data and simulation, using the ratio of the energies deposited in the first and the second layers of the calorimeter to discriminate between genuine converted and unconverted photons. The uncertainty in the energy scale is evaluated, as a function of  $|\eta|$  and  $E_T$ , by reweighting the conversion fractions in a sample of simulated single photons according to the values obtained from the radiative Z sample in simulation and data, respectively, and it is taken to be the relative difference of the energy responses.

For photons with  $E_{\rm T}=60$  GeV, the uncertainty in the energy scale for unconverted photon candidates is about 0.02% in the barrel and 0.02%-0.13% in the endcaps. For converted photon candidates, it is about 0.12% in the barrel and smaller than 0.01% in the endcaps. For photons with lower energy the uncertainty increases significantly: for  $E_{\rm T}=15$  GeV, it amounts to 0.18% in the barrel and 0.08%-0.67% in the endcaps for unconverted photons. For converted photons, it becomes 0.69%-1.31% in the barrel and 0.01%-0.1% in the endcaps. This systematic uncertainty is considered as a single source, correlated between converted and unconverted photons.

#### 8.2 Out-of-cluster energy leakage mis-modelling

Electrons and photons deposit about 1% to 6% of their energy outside of the cluster used in the reconstruction, depending on  $E_T$ ,  $\eta$  and the particle type. This effect is corrected for by the MC-based energy response calibration. However, a bias in the reconstructed energy could appear in data if this lateral leakage is mis-modelled by the simulation. For electrons, the global energy scale correction (Section 7) absorbs any potential discrepancy at  $\langle E_T \rangle \approx 40$  GeV.<sup>5</sup> To take into account possible differences between electron and photon showers related to the different probabilities for interaction with the material in front of the calorimeter, the lateral energy leakage in the calorimeter outside the area of the cluster is studied directly in data and simulation.

The  $Z \to \ell\ell\gamma$  and  $Z \to ee$  event samples are used to estimate this difference. The lateral energy leakage, l, is defined by comparing the energy collected in the second-layer cells belonging to the supercluster,  $E_{\text{nom}}^{\text{L2}}$ , with the energy deposited in second-layer cells in a larger rectangular window of size  $7 \times 11$  in  $\eta \times \phi$  around it,  $E_{7 \times 11}^{\text{L2}}$ :

$$l = \frac{E_{7 \times 11}^{L2}}{E_{\text{nom}}^{L2}} - 1.$$

The difference between data and simulation is presented in Figure 18 as a function of  $E_T$  and in two pseudorapidity regions for electrons, unconverted photons and converted photons.

<sup>&</sup>lt;sup>5</sup> The energy dependence of the difference between data and simulation for electron lateral leakage was studied and found to be small, below 0.1%, and is accounted for as a  $(E_T, |\eta|)$ -dependent systematic uncertainty.




Figure 18: The difference between the leakage fractions in data and simulation, in (a) the barrel and (b) the endcaps as a function of  $E_{\rm T}$  and the particle type. A variable bin size was chosen to make optimal use of the available samples. The last bin covers the range  $E_{\rm T} > 40$  GeV, and the corresponding markers are set near the  $E_{\rm T}$ -average in that bin.

The double difference between electrons and photons in data and simulation,  $\alpha_l$ , defined as:

$$\alpha_l = (l_e - l_\gamma)^{\text{data}} - (l_e - l_\gamma)^{\text{MC}},$$

is estimated separately for converted and unconverted photons and is used to correct the photon energy scale. The reconstruction of photon conversions and the classification into converted photon and unconverted photon categories have an impact on  $\alpha_l$  that is estimated by reweighting the simulation to match the data (as detailed in Section 8.1), and the difference between the nominal  $\alpha_l$  values and the values obtained without reweighting is taken as a systematic uncertainty. The  $\alpha_l$  values vary from -(0.3-0.2)% at low  $E_T$  to -(0.1-0.05)% at high  $E_T$ , with an absolute uncertainty ranging from 0.01% to 0.07% depending on the kinematic bins and the particle type.

Discrepancies between leakage modelling and data were observed in previous publications [1, 2], and the full size of the discrepancies was considered as a systematic uncertainty in the photon energy calibration. The statistical power of the Run 2 data allows an improved mapping of the  $|\eta|$  and  $E_T$  dependence of this effect, and a correction is derived. The systematic variations, defined as above, reduce the residual uncertainty and allow a decrease in the corresponding calibration uncertainty by a factor of about two.

# 9 Electron and photon energy scale uncertainties

The complete systematic uncertainty model contains 64 and 67 independent uncertainty variations for the electron and photon energy scales, respectively.

Uncertainties in the material upstream of the calorimeter are derived in Refs. [1–3] from a combination of detector construction information and *in situ* measurements, and are evaluated for up to nine  $|\eta|$  regions.

The uncertainty model for the cell readout non-linearity has been revised with respect to Ref. [2]. The present analysis considers separate sources of uncertainty for the transitions between High and Medium gain, and between Medium and Low gain in Layer 2. In the previous analyses, a single uncertainty was

used to cover both gain transitions. In addition, because of the improved strategy for the  $\alpha_{12}$  measurement (now also using an electron sample), the size of the uncertainty associated with the transition between High and Medium gain in Layer 1 was re-estimated for the current analysis. The corresponding sources of uncertainty are considered fully correlated in pseudorapidity.

The presampler calibration and the intercalibration of the first and second accordion layers are estimated in nine and seven  $|\eta|$  regions, respectively, with corresponding sources of uncertainty. Uncertainties corresponding to the Z scale calibration, photon reconstruction and classification, pile-up modelling, and lateral shower shape development are all fully correlated in pseudorapidity. For the last category, an exception is the energy scale dependence on the shower width in  $\eta$ , where the region  $1.52 < |\eta| < 1.82$  is considered to be independent of the rest of the pseudorapidity interval.

The full list of systematic uncertainty sources is summarized in Table 1. The impact of the most significant ones is illustrated as a function of the electron or photon  $E_T$  in Figure 19, for two pseudorapidity values.

The uncertainty in the energy resolution is shown in Figure 20 as a function of  $E_{\rm T}$ , for electrons and unconverted photons, at  $|\eta|=0.3$ . Only the uncertainty related to the resolution correction (Section 7) was updated in this analysis with respect to the descriptions provided in Refs. [1–3]. At high  $E_{\rm T}$ , the resolution uncertainties are larger than those presented in Ref. [2], where fixed-size clusters were used for electron and photon reconstruction. The present clustering algorithm is more sensitive to pile-up fluctuations. This effect is larger for electrons than for photons.

## 10 Energy linearity and constraints on the calibration uncertainties

#### 10.1 Energy linearity measurement

In order to test the  $E_T$  dependence of the energy scale corrections  $\alpha_i$ , the procedure in Section 7 is repeated in bins of  $|\eta|$  and  $E_T$ . Specifically, an extended definition of the calibrated energy is introduced as

$$E^{\text{data,corr}} = E^{\text{data}}/[(1 + \alpha_i)(1 + \alpha'_j)],$$

where the energy scale factors  $\alpha_i$  are left at their values determined above, and  $\alpha'_j$  quantifies the energy dependence of the energy scale (in the absence of energy dependence,  $\alpha' = 0$ ). The index j labels a two-dimensional mapping of the electron phase space, with the following bin boundaries:

- 0, 0.6, 1.0, 1.37, 1.55, 1.82, 2.47 in  $|\eta|$ ;
- 27, 33, 38, 44, 50, 62, 100,  $\sqrt{s}/2$  in  $E_T$  [GeV].

Events are assigned to a bin j if either one of the two decay electrons falls in this bin, and  $\alpha'_j$  is varied to obtain the best agreement between the invariant mass distributions in data and simulation, separately for each bin j. Since the second electron is distributed randomly, iterations are required and the procedure is repeated until convergence is reached.

The  $\alpha'$  coefficients are shown in Figure 21 as a function of  $|\eta|$  and  $E_{\rm T}$ . The observed energy dependence is significant, and is measured with a typical precision of 0.03%, including the effect of systematic variations as described in Section 7. In the regions  $1.37 < |\eta| < 1.55$  and  $1.55 < |\eta| < 1.82$ , the precision is about 0.3%. The measurement precision is better than the calibration uncertainty derived in the previous sections, as can be seen from the outer uncertainty band in Figure 21. The energy linearity measurement can thus be used to further constrain the uncertainty model, as described in the following.

Table 1: List of the independent systematic uncertainties affecting the energy calibration and their division into a number of  $|\eta|$  regions, between which the uncertainties are not correlated.

| Source of uncertainty                                 | Methodology                                             | Description | Number of $ \eta $ regions |
|-------------------------------------------------------|---------------------------------------------------------|-------------|----------------------------|
| ID material                                           | Run 1 detector construction                             | [1, 3]      | 4                          |
|                                                       | Pixel services description                              | [2, 3]      | 2                          |
| Material presampler to calorimeter                    | Run 1 measurement with unconv. photon                   | [1, 3]      | 10                         |
| $( \eta  < 1.8)$                                      | Simulation of long. shower shape unconv. photon         | [1]         | 2                          |
| Material ID to presampler                             | Run 1 measurement with electrons                        | [1, 3]      | 9                          |
| $( \eta  < 1.8)$                                      | Simulation of long. shower shape electrons              | [1]         | 2                          |
| Material ID to calorimeter                            | Run 1 measurement with electrons                        | [1, 3]      | 3                          |
| $( \eta  > 1.8)$                                      | Simulation of long. shower shape electrons              | [1]         | 1                          |
| All material ID to calorimeter                        | Variations of Geant4 physics list                       | [1]         | 1                          |
| Cell readout non-linearity                            | ADC non-linearity                                       | Section 5.3 | 1                          |
| ·                                                     | Medium gain/High gain Layer 2                           | Section 5.4 | 1                          |
|                                                       | Low gain/Medium gain Layer 2                            | Section 5.4 | 1                          |
|                                                       | Medium gain/High gain Layer 1                           | [2]         | 1                          |
|                                                       | Pile-up shift                                           | [2]         | 1                          |
| Presampler calibration                                | $\alpha_{\rm PS}$ measurement                           | Section 6.1 | 9                          |
| Layer 1/Layer 2 calibration                           | $\alpha_{12}$ measurement                               | Section 6.2 | 7                          |
| Barrel–endcap gap scintillator $(1.4 <  \eta  < 1.6)$ | Scintillator calibration                                | [2]         | 3                          |
| $Z \rightarrow ee$ calibration                        | Statistical uncertainty                                 | Section 7   | 1                          |
|                                                       | Systematic uncertainty                                  | Section 7   | 1                          |
| Conversion reconstruction                             | Classification (efficiency and fake rate)               | Section 8.1 | 1                          |
|                                                       | Radius dependence of conversion reconstruction          | [1]         | 1                          |
| Lateral shower shape modelling                        | Dependence on shower $\eta$ width                       | [2]         | 2                          |
| •                                                     | Lateral leakage for electrons                           | Section 8.2 | 1                          |
|                                                       | Lateral leakage for unconv. photons                     | Section 8.2 | 1                          |
|                                                       | Lateral leakage for conv. photons                       | Section 8.2 | 1                          |
| Pile-up modelling                                     | Mis-modelling of pile-up noise vs $\langle \mu \rangle$ | [3]         | 1                          |




Figure 19: Relative energy scale calibration uncertainty for (a, b) electrons, (c, d) unconverted photons and (e, f) converted photons, as a function of  $E_T$  for (a, c, e)  $|\eta| = 0.3$  and (b, d, f)  $|\eta| = 2.1$ . For converted photons at  $|\eta| = 0.3$ , the relative uncertainty reaches 2.1% at  $E_T = 10$  GeV. The total uncertainty is shown along with the main contributions, which are represented by the signed impact of a one-sided variation of the corresponding uncertainty. Only a one-sided variation for each uncertainty source is shown for clarity, except for the uncertainty related to the *in situ* global energy scale determination with  $Z \rightarrow ee$  candidate events.




Figure 20: Relative uncertainty in the energy resolution,  $\delta_{\sigma}/\sigma$  as a function of  $E_{\rm T}$  for (a) electrons and (b) unconverted photons at  $|\eta|=0.3$ . The total uncertainty is shown along with the breakdown into the different contributions.




Figure 21: Comparison of the measured values of  $\alpha'$  with the pre-fit and post-fit linearity models. The full lines represent the nominal pre- and post-fit models, and the bands represent the corresponding uncertainties. The measurements in the  $|\eta|$  range corresponding to the barrel–endcap transition regions,  $1.37 \le |\eta| < 1.55$ , and from the  $J/\psi \to ee$  analysis are not included in the energy linearity fit. The analysis described in Section 11.1 was repeated with the present  $\eta$  binning to obtain the open dots in these figures.

#### 10.2 Constraints on the calibration systematic uncertainties

Accounting for the energy scale calibration using  $Z \rightarrow ee$  events, varying a given source of uncertainty by one standard deviation affects the reconstructed energy of an electron as follows:

$$\delta_{\text{rel}} E_k(E_{\text{T}}, \eta) = \Delta_{\text{rel}} E_k(E_{\text{T}}, \eta) - \Delta_{\text{rel}} E_k(\langle E_{\text{T}} \rangle, \eta),$$

where k labels the source of uncertainty (see Table 1),  $\Delta_{\rm rel}E_k(E_{\rm T},\eta)$  represents its fractional impact on the reconstructed energy, and  $\langle E_{\rm T} \rangle \approx 40$  GeV is the average transverse energy for electrons produced in Z-boson decays. The quantities  $\Delta_{\rm rel}E_k(E_{\rm T},\eta)$  are estimated by varying the corresponding sources of uncertainty in the reconstruction or detector simulation. The Z-based calibration absorbs the systematic uncertainty effect for electrons with  $E_{\rm T}=\langle E_{\rm T} \rangle$  and leaves the residual effect  $\delta_{\rm rel}E_k(E_{\rm T},\eta)$ . In this model, the total effect of all systematic variations on the linearity can be parameterized as

$$\alpha'_{\text{mod},j} = \sum_{k} \delta_{\text{rel}} E_{jk} \theta_k,$$

where  $\delta_{\text{rel}}E_{jk}$  is the average of  $\delta_{\text{rel}}E_k(E_T, \eta)$  over  $E_T$  and  $\eta$  in bin j, and  $\theta_k$  is the normally distributed nuisance parameter (NP) associated with the source k. The calibration model is fitted to the data by adjusting the nuisance parameters  $\theta$ , minimizing the following  $\chi^2$ :

$$\chi^{2} = \sum_{j_{1}, j_{2}} \left[ \alpha'_{j_{1}} - \alpha'_{\text{mod}, j_{1}}(\theta) \right] C_{j_{1}, j_{2}}^{-1} \left[ \alpha'_{j_{2}} - \alpha'_{\text{mod}, j_{2}}(\theta) \right] + \sum_{k} \theta_{k}^{2}, \tag{2}$$

where C is the covariance of the  $\alpha'$  measurements, calculated from the statistical and systematic uncertainties in the measurement procedure. The model contains 46 nuisance parameters<sup>6</sup> with the corresponding constraint terms, and  $7\times5$  measured values of  $\alpha'$ . The transition regions between the barrel and endcap calorimeters are not included in the fit. The nuisance parameters represent uncertainties in the passive material model, electromagnetic shower shape development, readout electronics calibration and layer intercalibration; each of these uncertainty classes may be subdivided into up to 12 pseudorapidity regions.

The  $\chi^2$  minimization is performed analytically. Denoting the fitted values of the nuisance parameters by  $\hat{\theta}_k$  and their post-fit covariance by V, the post-fit linearity and its uncertainty are defined as:

$$\hat{\alpha}'_{\text{mod},j} = \sum_{k} \delta_{\text{rel}} E_{jk} \, \hat{\theta}_{k} \quad , \quad \delta \hat{\alpha}'_{\text{mod},j} = \left[ \sum_{k_1, k_2} \delta_{\text{rel}} E_{jk_1} \, \delta_{\text{rel}} E_{jk_2} \, V_{k_1 k_2} \right]^{1/2} .$$

For comparison, the pre-fit expressions are  $\alpha'_{\mathrm{mod},j} = 0$  and  $\delta\alpha'_{\mathrm{mod},j} = \left[\sum_k \delta_{\mathrm{rel}} E_{jk}^2\right]^{1/2}$ . The minimization result is sensitive to the details of the covariance matrix C assigned to the linearity measurements. A global goodness-of-fit of  $\chi^2/N = 90/35$  is obtained, assuming full correlation of the  $\alpha'_j$  systematic uncertainties across  $E_{\mathrm{T}}$  bins within each  $|\eta|$  bin, but ignoring correlations across  $|\eta|$  bins. The MC statistical uncertainty is accounted for in the evaluation of the systematic uncertainties. For  $|\eta|$  bins with a partial  $\chi^2$  per degree of freedom  $\chi^2_{\mathrm{bin}}/N_{\mathrm{bin}}$  greater than one, the  $\alpha'$  measurement uncertainties are rescaled by a factor  $\sqrt{\chi^2_{\mathrm{bin}}/N_{\mathrm{bin}}}$ . This scaling typically increases the fit uncertainties by 5%. The final goodness-of-fit is  $\chi^2/N = 41/35$ , corresponding to a p-value of 0.22.

<sup>&</sup>lt;sup>6</sup> The full model contains 64 nuisance parameters relevant for the electron energy scale (Section 9). Two of them related to the final energy scale determination (Section 7) are not part of the  $\theta_k$  and 16 are removed from the fit by a pruning procedure.

The results of the analysis are illustrated in Figures 21 and 22. With few exceptions, the measured values of  $\alpha'_j$  are well within the initial calibration uncertainties. The fit  $\chi^2$  in Eq. (2) captures the measured non-linearities, and reduces the uncertainties by up to a factor of two for  $E_T < 50$  GeV, and up to three for  $E_T \sim 150$  GeV. As can be seen in Figure 22, the reduction in uncertainty is mostly driven by the nuisance parameters associated with the cell-level non-linearity (ADC corrections, and HG/MG transition) and the shower development (lateral leakage and shower width). Most other nuisance parameters are typically constrained by 5%–10%. All nuisance parameter shifts are within the initially assigned uncertainties, with the exception of a 1.5 $\sigma$  effect observed for the presampler calibration in one specific  $|\eta|$  region.




Figure 22: Shifts and constraints on the nuisance parameters of the systematic uncertainty model from the energy linearity fit. A digit after the NP name represents a given  $\eta$  range (a priori different for different NP sources).

The impact of the linearity fit on the electron energy scale uncertainty is illustrated in Figure 23(a) as a function of  $E_{\rm T}$  and  $|\eta|$ . The precision for electrons with  $E_{\rm T}\sim 40$  GeV is mostly unchanged, since these particles are typical of on-shell Z-boson decays and their calibration is essentially determined by the energy scale analysis in Section 7. Energy scale uncertainties for electrons with  $E_{\rm T}=10$  GeV or  $E_{\rm T}=1$  TeV are typically reduced by 30%–50%, and vary from 0.2%–0.3% for  $|\eta|<1$  and  $|\eta|>1.8$  to between 0.5% and 1% for  $1<|\eta|<1.8$ .

The impact of the present analysis on photon calibration uncertainties is shown in Figure 23(b) for converted and unconverted photons, and for  $E_T = 60$  GeV, which is typical for photons from Higgs boson decays. Uncertainties for converted photons, which are experimentally close to electrons, are only moderately reduced for this energy. For unconverted photons, the energy calibration uncertainty is typically reduced by 30% in the barrel, and by up to a factor of two in the endcaps.




Figure 23: Total relative systematic uncertainty in the energy scale as a function of  $|\eta|$  for (a) electrons with  $E_T = 10$  GeV, 40 GeV or 1 TeV and (b) photons with  $E_T = 60$  GeV, after the constraints from the linearity fit. The bottom panels show the ratio of the post-fit to pre-fit uncertainties.

#### 11 Calibration cross-checks

#### 11.1 Checks using $J/\psi \rightarrow ee$ events

The known mass of the  $J/\psi$  resonance provides a completely independent check of the energy calibration for electrons with transverse energy in the range from 5 to 30 GeV. For this, the full calibration procedure discussed in the previous sections, including the energy scale derived from  $Z \to ee$  events, is applied. The difference between data and simulation for  $J/\psi \to ee$  events is then quantified using residual energy scale factors extracted from the peak positions of the reconstructed invariant mass. If the energy calibration is correct, the residual energy scale factors should be consistent with zero within the combined uncertainties of the  $J/\psi \to ee$  measurement and the systematic uncertainty of the energy calibration.

The event selection requires two tightly identified, loosely isolated, opposite-sign electron candidates with  $E_T > 5$  GeV and  $|\eta| < 2.4$ . The primary vertex must be located within |z| = 150 mm of the IP, and the dielectron invariant mass must be in the range [1, 5] GeV.

The residual energy scale factors are denoted by  $\alpha_i$ , where i labels the kinematic bin, and determined as follows:

•  $J/\psi$  particles can be produced promptly or in b-hadron decays. The hadronic activity surrounding the decay electrons differs in both cases, biasing the energy-scale determination if the relative event fractions are not modelled accurately. The prompt fraction is extracted from a fit to the proper decay-time distribution of the data and is found to be between 76% and 82% depending on the leading electron's  $E_{\rm T}$ , with uncertainties of up to 4%. The simulated prompt and non-prompt samples are then combined using the measured fractions.

- The data are divided into categories depending on the  $\eta$  values of the two selected electrons. The  $\eta$  bin boundaries are -2.4, -1.52, -1.37, -1.0, -0.8, -0.4, 0, 0.4, 0.8, 1.0, 1.37, 1.52, and 2.4. The region  $1.37 < |\eta| < 1.52$  is not considered for the nominal results. The measurement can also be performed as a function of  $E_{\rm T}$  and integrated in  $\eta$ . In this case, the  $E_{\rm T}$  bin boundaries are 5, 10, 15, 20, and 30 GeV.
- Comparison of the dielectron invariant mass distributions in data and simulation requires the background contributions to be subtracted from the data sample. The subtraction is performed by fitting a signal+background distribution to the data, separately in each category. The total distribution is expressed as the sum of two double-sided Crystal Ball functions to describe the  $J/\psi$  and  $\psi(2S)$  resonances, and a second-order Chebyshev polynomial to represent the continuum background. The parameters describing the resonance distributions are fixed to values determined from simulation, with the exception of the peak positions, which are parameterized using energy scales  $\alpha_i$  and  $\alpha_j$ . An example of an invariant mass fit is shown in Figure 24.




Figure 24: An example dielectron invariant mass distribution with the fitted signal, background and  $\psi(2S)$  contributions. Both electrons are required to have  $|\eta| < 0.4$ .

• A simultaneous fit to all categories is performed to extract the residual energy scale factors. The considered systematic uncertainties are related to the modelling of signal and background, the fitted mass range, the uncertainties in the prompt fraction, and the modelling of the pseudorapidity distribution of electrons in simulation. The fit is repeated, varying each of the above sources of uncertainty in turn, and the deviations from the nominal  $\alpha$  value are added in quadrature to obtain the final uncertainty in the measurement.

The results are given in Figure 25, where the evolution of  $\alpha$  is shown as a function of (a)  $\eta$  and (b)  $E_{\rm T}$ , before and after including the constraints from the linearity fit. The residual post-fit scale factors are below 0.5% and are compatible with zero within the total calibration uncertainty.

#### 11.2 Checks using $Z \to \ell\ell\gamma$ events

Radiative Z-boson decay events provide a way to investigate the validity of the final photon energy scale. The selection requires a Z-boson candidate decaying into two opposite-sign electrons or muons and a photon from final-state radiation. Electrons (muons) must meet medium identification criteria with




Figure 25: Variation of the residual energy scale as a function of (a)  $\eta$  and (b)  $E_{\rm T}$ , as measured with  $J/\psi \to ee$  events. The data points and uncertainty bands are shown for both the pre- and post-linearity-fit energy scale models. The uncertainty bands correspond to the energy calibration uncertainty for the energy range of the  $J/\psi \to ee$  decays.

 $E_{\rm T} > 18$  (15) GeV. Tightly identified photons with  $E_{\rm T} > 10$  GeV are selected. Electrons and photons in the barrel–endcap transition regions are not considered. Loose isolation requirements are applied to all objects. The invariant mass of the dilepton+photon system is required to be in the range  $80 < m_{\ell\ell\gamma} < 100$  GeV and the dilepton invariant mass must be in the range  $40 < m_{\ell\ell} < 80$  GeV. Figure 26 shows the  $m_{\ell\ell\gamma}$  distributions for the (a) dielectron and (b) dimuon channels. The inclusive residual photon energy scale factors are measured to be  $(3.3 \pm 2.0) \times 10^{-3}$  and  $(1.4 \pm 1.1) \times 10^{-3}$  in the  $Z \to ee\gamma$  and  $Z \to \mu\mu\gamma$  samples, respectively. This is consistent with the larger offset observed in Figure 26(a) than in Figure 26(b). The residual scales measured in the two channels agree within one standard deviation. The uncertainties combine the statistical uncertainty and the systematic uncertainty originating from the uncertainty in the lepton-energy calibration (see Ref. [38] for a description of the muon momentum scale determination).

To extract the residual photon energy scale, MC templates are compared with the data distribution through a  $\chi^2$  test, for the electron and muon channels independently. These templates are built by modifying the photon  $E_T$  by a factor  $(1 + \alpha)$ , where  $\alpha$  is varied from -0.0200 to 0.0200 in steps of 0.0004, and the overall energy scale residual is given by the template providing the lowest  $\chi^2$ . The two channels are compatible and are statistically combined.

The dependence of  $\alpha$  on the energy and pseudorapidity of the photon is illustrated in Figure 27, separately for converted photons and unconverted photons. The residual photon energy scales are compared with the total energy calibration uncertainty for photons from  $Z \to \ell\ell\gamma$  decays, before and after including the constraints from the linearity fit. The error bars assigned to the data are typically 0.15%, and are dominated by the uncertainty in the lepton-energy calibration. As in the  $J/\psi$  analysis, the linearity-constrained calibration tends to reduce the values of the residual calibration factors, which are found to be within the corresponding calibration uncertainties. Mild tension is observed for  $|\eta| > 1.8$ , for unconverted photons; this effect is driven by low-energy photons, and disappears at high  $E_T$  as indicated by Figure 27(b).




Figure 26: Comparison of the (a)  $ee\gamma$  and (b)  $\mu\mu\gamma$  invariant mass distributions in data and simulation.




Figure 27: Variation of the residual energy scale for (a, b) unconverted and (c, d) converted photons as a function of (a, c)  $|\eta|$  and (b, d)  $E_T$ , as measured with  $Z \to \ell\ell\gamma$  events. The data points and uncertainty bands are shown for both the pre- and post-linearity-fit energy scale models. The uncertainty bands correspond to the energy calibration uncertainty for photons from  $Z \to \ell\ell\gamma$  decays.

### 12 Conclusion

This paper presents the energy calibration for electrons and photons reconstructed in 140 fb<sup>-1</sup> of 13 TeV proton–proton collision data recorded by ATLAS during Run 2 of the LHC. All of the major sources of uncertainty have been re-evaluated since the previous publications, and new methods are introduced to further reduce their impact. Improved methods to calibrate energies in the calorimeter cells and layers, and an improved measurement of lateral energy leakage from reconstructed electron and photon energy clusters, reduce the *a priori* calibration uncertainty by about 30%. In addition, a precise measurement of the energy linearity, using electrons from *Z*-boson decays, provides a further reduction by about a factor of two. The overall calibration uncertainty is reduced by a factor of 2–3, depending on the particle type, pseudorapidity and energy. The achieved calibration uncertainties are typically 0.05% for electrons from *Z*-boson decays, 0.4% at  $E_T \sim 10$  GeV, and 0.3% at  $E_T \sim 1$  TeV; for photons at  $E_T \sim 60$  GeV, they are 0.2% on average. These improvements are validated using independent samples of  $J/\psi \rightarrow ee$  decays and radiative *Z*-boson decays. The achieved precision is adequate for high-precision measurements of fundamental parameters such as the masses and properties of the Higgs, *W* and *Z* bosons, and is expected to improve the sensitivity of searches and measurements at the weak scale.

# Acknowledgements

We thank CERN for the very successful operation of the LHC and its injectors, as well as the support staff at CERN and at our institutions worldwide without whom ATLAS could not be operated efficiently.

The crucial computing support from all WLCG partners is acknowledged gratefully, in particular from CERN, the ATLAS Tier-1 facilities at TRIUMF/SFU (Canada), NDGF (Denmark, Norway, Sweden), CC-IN2P3 (France), KIT/GridKA (Germany), INFN-CNAF (Italy), NL-T1 (Netherlands), PIC (Spain), RAL (UK) and BNL (USA), the Tier-2 facilities worldwide and large non-WLCG resource providers. Major contributors of computing resources are listed in Ref. [39].

We acknowledge the support of ANPCyT, Argentina; YerPhI, Armenia; ARC, Australia; BMWFW and FWF, Austria; ANAS, Azerbaijan; CNPq and FAPESP, Brazil; NSERC, NRC and CFI, Canada; CERN; ANID, Chile; CAS, MOST and NSFC, China; Minciencias, Colombia; MEYS CR, Czech Republic; DNRF and DNSRC, Denmark; IN2P3-CNRS and CEA-DRF/IRFU, France; SRNSFG, Georgia; BMBF, HGF and MPG, Germany; GSRI, Greece; RGC and Hong Kong SAR, China; ISF and Benoziyo Center, Israel; INFN, Italy; MEXT and JSPS, Japan; CNRST, Morocco; NWO, Netherlands; RCN, Norway; MEiN, Poland; FCT, Portugal; MNE/IFA, Romania; MESTD, Serbia; MSSR, Slovakia; ARRS and MIZŠ, Slovenia; DSI/NRF, South Africa; MICINN, Spain; SRC and Wallenberg Foundation, Sweden; SERI, SNSF and Cantons of Bern and Geneva, Switzerland; MOST, Taipei; TENMAK, Türkiye; STFC, United Kingdom; DOE and NSF, United States of America.

In addition, individual groups and members have received support from BCKDF, CANARIE, CRC and DRAC, Canada; PRIMUS 21/SCI/017 and UNCE SCI/013, Czech Republic; COST, ERC, ERDF, Horizon 2020, ICSC-NextGenerationEU and Marie Skłodowska-Curie Actions, European Union; Investissements d'Avenir Labex, Investissements d'Avenir Idex and ANR, France; DFG and AvH Foundation, Germany; Herakleitos, Thales and Aristeia programmes co-financed by EU-ESF and the Greek NSRF, Greece; BSF-NSF and MINERVA, Israel; Norwegian Financial Mechanism 2014-2021, Norway; NCN and NAWA, Poland; La Caixa Banking Foundation, CERCA Programme Generalitat de Catalunya and PROMETEO

and GenT Programmes Generalitat Valenciana, Spain; Göran Gustafssons Stiftelse, Sweden; The Royal Society and Leverhulme Trust, United Kingdom.

Individual members wish to acknowledge support from Chile: Agencia Nacional de Investigación y Desarrollo (FONDECYT 1190886, FONDECYT 1210400, FONDECYT 1230987); China: National Natural Science Foundation of China (NSFC - 12175119, NSFC 12275265); European Union: European Research Council (ERC - 948254), Horizon 2020 Framework Programme (MUCCA - CHIST-ERA-19-XAI-00), Italian Center for High Performance Computing, Big Data and Quantum Computing (ICSC, NextGenerationEU), Marie Sklodowska-Curie Actions (EU H2020 MSC IF GRANT NO 101033496); France: Agence Nationale de la Recherche (ANR-20-CE31-0013, ANR-21-CE31-0022), Investissements d'Avenir Idex (ANR-11-LABX-0012), Investissements d'Avenir Labex (ANR-11-LABX-0012); Germany: Baden-Württemberg Stiftung (BW Stiftung-Postdoc Eliteprogramme), Deutsche Forschungsgemeinschaft (DFG - CR 312/5-1); Italy: Istituto Nazionale di Fisica Nucleare (FELLINI G.A. n. 754496, ICSC, NextGenerationEU); Japan: Japan Society for the Promotion of Science (JSPS KAKENHI 22H01227, JSPS KAKENHI JP21H05085, JSPS KAKENHI JP22H04944); Netherlands: Netherlands Organisation for Scientific Research (NWO Veni 2020 - VI.Veni.202.179); Norway: Research Council of Norway (RCN-314472); Poland: Polish National Agency for Academic Exchange (PPN/PPO/2020/1/00002/U/00001), Polish National Science Centre (NCN 2021/42/E/ST2/00350, NCN OPUS nr 2022/47/B/ST2/03059, NCN UMO-2019/34/E/ST2/00393, UMO-2020/37/B/ST2/01043); Slovenia: Slovenian Research Agency (ARIS grant J1-3010); Spain: BBVA Foundation (LEO22-1-603), Generalitat Valenciana (Artemisa, FEDER, IDIFEDER/2018/048), La Caixa Banking Foundation (LCF/BQ/PI20/11760025), Ministry of Science and Innovation (RYC2019-028510-I, RYC2020-030254-I), PROMETEO and GenT Programmes Generalitat Valenciana (CIDEGENT/2019/023, CIDEGENT/2019/027); Sweden: Swedish Research Council (VR 2022-03845); Switzerland: Swiss National Science Foundation (SNSF - PCEFP2\_194658); United Kingdom: Leverhulme Trust (Leverhulme Trust RPG-2020-004); United States of America: Neubauer Family Foundation.

#### References

- [1] ATLAS Collaboration, *Electron and photon energy calibration with the ATLAS detector using LHC Run 1 data*, Eur. Phys. J. C **74** (2014) 3071, arXiv: 1407.5063 [hep-ex].
- [2] ATLAS Collaboration, *Electron and photon energy calibration with the ATLAS detector using* 2015–2016 LHC proton–proton collision data, JINST **14** (2019) P03017, arXiv: 1812.03848 [hep-ex].
- [3] ATLAS Collaboration, Electron and photon performance measurements with the ATLAS detector using the 2015–2017 LHC proton–proton collision data, JINST 14 (2019) P12006, arXiv: 1908.00005 [hep-ex].
- [4] ATLAS Collaboration, *The ATLAS Experiment at the CERN Large Hadron Collider*, JINST **3** (2008) S08003.
- [5] ATLAS Collaboration, ATLAS Insertable B-Layer: Technical Design Report, ATLAS-TDR-19; CERN-LHCC-2010-013, 2010, url: https://cds.cern.ch/record/1291633, Addendum: ATLAS-TDR-19-ADD-1; CERN-LHCC-2012-009, 2012, url: https://cds.cern.ch/record/1451888.

- [6] B. Abbott et al., *Production and integration of the ATLAS Insertable B-Layer*, JINST **13** (2018) T05008, arXiv: 1803.00844 [physics.ins-det].
- [7] H. Abreu et al., *Performance of the electronic readout of the ATLAS liquid argon calorimeters*, JINST **5** (2010) P09003.
- [8] W. E. Cleland and E. G. Stern,

  Signal processing considerations for liquid ionization calorimeters in a high rate environment,

  Nucl. Instrum. Meth. A 338 (1994) 467.
- [9] ATLAS Collaboration,

  Topological cell clustering in the ATLAS calorimeters and its performance in LHC Run 1,

  Eur. Phys. J. C 77 (2017) 490, arXiv: 1603.02934 [hep-ex].
- [10] ATLAS Collaboration, *Electron reconstruction and identification in the ATLAS experiment using* the 2015 and 2016 LHC proton–proton collision data at  $\sqrt{s} = 13$  TeV, Eur. Phys. J. C **79** (2019) 639, arXiv: 1902.04655 [hep-ex].
- [11] ATLAS Collaboration, *Electron and photon efficiencies in LHC Run 2 with the ATLAS experiment*, (2023), arXiv: 2308.13362 [hep-ex].
- [12] ATLAS Collaboration, *Measurement of the photon identification efficiencies with the ATLAS detector using LHC Run 2 data collected in 2015 and 2016*, Eur. Phys. J. C **79** (2019) 205, arXiv: 1810.05087 [hep-ex].
- [13] ATLAS Collaboration, *ATLAS data quality operations and performance for 2015–2018 data-taking*, JINST **15** (2020) P04003, arXiv: 1911.04632 [physics.ins-det].
- [14] ATLAS Collaboration,

  Luminosity determination in pp collisions at  $\sqrt{s} = 13$  TeV using the ATLAS detector at the LHC, (2022), arXiv: 2212.09379 [hep-ex].
- [15] ATLAS Collaboration, *Performance of electron and photon triggers in ATLAS during LHC Run 2*, Eur. Phys. J. C **80** (2020) 47, arXiv: 1909.00761 [hep-ex].
- [16] ATLAS Collaboration, *Performance of the ATLAS muon triggers in Run 2*, JINST **15** (2020) P09015, arXiv: 2004.13447 [hep-ex].
- [17] ATLAS Collaboration, *Muon reconstruction performance of the ATLAS detector in proton–proton collision data at*  $\sqrt{s} = 13$  *TeV*, Eur. Phys. J. C **76** (2016) 292, arXiv: 1603.05598 [hep-ex].
- [18] S. Alioli, P. Nason, C. Oleari and E. Re, NLO vector-boson production matched with shower in POWHEG, JHEP **07** (2008) 060, arXiv: **0805.4802** [hep-ph].
- [19] T. Sjöstrand, S. Mrenna and P. Skands, *A brief introduction to PYTHIA 8.1*, Comput. Phys. Commun. **178** (2008) 852, arXiv: **0710.3820** [hep-ph].
- [20] H.-L. Lai et al., *New parton distributions for collider physics*, Phys. Rev. D **82** (2010) 074024, arXiv: 1007.2241 [hep-ph].
- [21] ATLAS Collaboration, Measurement of the  $Z/\gamma^*$  boson transverse momentum distribution in pp collisions at  $\sqrt{s} = 7$  TeV with the ATLAS detector, JHEP **09** (2014) 145, arXiv: 1406.3660 [hep-ex].
- [22] J. Pumplin et al.,

  New Generation of Parton Distributions with Uncertainties from Global QCD Analysis,

  JHEP 07 (2002) 012, arXiv: hep-ph/0201195.

- [23] N. Davidson, T. Przedzinski and Z. Was, *PHOTOS Interface in C++: Technical and physics documentation*, Comput. Phys. Commun. **199** (2016) 86, arXiv: 1011.0937 [hep-ph].
- [24] ATLAS Collaboration, *ATLAS Pythia 8 tunes to 7 TeV data*, ATL-PHYS-PUB-2014-021, 2014, url: https://cds.cern.ch/record/1966419.
- [25] D. J. Lange, *The EvtGen particle decay simulation package*, Nucl. Instrum. Meth. A **462** (2001) 152.
- [26] E. Bothmann et al., Event generation with Sherpa 2.2, SciPost Phys. 7 (2019) 034, arXiv: 1905.09127 [hep-ph].
- [27] ATLAS Collaboration, *The ATLAS Collaboration Software and Firmware*, ATL-SOFT-PUB-2021-001, 2021, URL: https://cds.cern.ch/record/2767187.
- [28] ATLAS Collaboration, *The ATLAS Simulation Infrastructure*, Eur. Phys. J. C **70** (2010) 823, arXiv: 1005.4568 [physics.ins-det].
- [29] S. Agostinelli et al., Geant 4 a simulation toolkit, Nucl. Instrum. Meth. A 506 (2003) 250.
- [30] ATLAS Collaboration, *The Pythia 8 A3 tune description of ATLAS minimum bias and inelastic measurements incorporating the Donnachie–Landshoff diffractive model*, ATL-PHYS-PUB-2016-017, 2016, URL: https://cds.cern.ch/record/2206965.
- [31] NNPDF Collaboration, R. D. Ball et al., *Parton distributions with LHC data*, Nucl. Phys. B **867** (2013) 244, arXiv: 1207.1303 [hep-ph].
- [32] ATLAS Collaboration, Summary of ATLAS Pythia 8 tunes, ATL-PHYS-PUB-2012-003, 2012, URL: https://cds.cern.ch/record/1474107.
- [33] ATLAS Collaboration, *Study of the material of the ATLAS inner detector for Run 2 of the LHC*, JINST **12** (2017) P12009, arXiv: 1707.02826 [hep-ex].
- [34] ATLAS Collaboration, *Reconstruction of primary vertices at the ATLAS experiment in Run 1 proton–proton collisions at the LHC*, Eur. Phys. J. C **77** (2017) 332, arXiv: 1611.10235 [hep-ex].
- [35] M. Oreglia, A Study of the Reactions  $\psi' \to \gamma \gamma \psi$ , (1980), URL: https://www-public.slac.stanford.edu/sciDoc/docMeta.aspx?slacPubNumber=slac-r-236.
- [36] A. Valassi, Combining correlated measurements of several different physical quantities, Nucl. Instrum. Meth. A **500** (2003) 391.
- [37] C. de La Taille and L. Serin, Temperature dependance of the ATLAS electromagnetic calorimeter signal. Preliminary drift time measurement, ATL-LARG-95-029, 1995, URL: https://cds.cern.ch/record/686091.
- [38] ATLAS Collaboration, Studies of the muon momentum calibration and performance of the ATLAS detector with pp collisions at  $\sqrt{s} = 13$  TeV, (2022), arXiv: 2212.07338 [hep-ex].
- [39] ATLAS Collaboration, *ATLAS Computing Acknowledgements*, ATL-SOFT-PUB-2023-001, 2023, URL: https://cds.cern.ch/record/2869272.

## The ATLAS Collaboration

```
G. Aad 10102, B. Abbott 10120, K. Abeling 1055, N.J. Abicht 1049, S.H. Abidi 1029, A. Aboulhorma 1035e,
H. Abramowicz 151, H. Abreu 150, Y. Abulaiti 117, B.S. Acharya 169a,69b,m, C. Adam Bourdarios 4,
L. Adamczyk 686a, S.V. Addepalli 626, M.J. Addison 6101, J. Adelman 6115, A. Adiguzel 621c,
T. Adye 134, A.A. Affolder 136, Y. Afik 36, M.N. Agaras 13, J. Agarwala 73a,73b,
A. Aggarwal ©<sup>100</sup>, C. Agheorghiesei ©<sup>27c</sup>, A. Ahmad ©<sup>36</sup>, F. Ahmadov ©<sup>38,y</sup>, W.S. Ahmed ©<sup>104</sup>,
S. Ahuja (1995), X. Ai (1962a), G. Aielli (1976a,76b), A. Aikot (1916a), M. Ait Tamlihat (1935e), B. Aitbenchikh (1935a),
I. Aizenberg 169, M. Akbiyik 100, T.P.A. Åkesson 98, A.V. Akimov 37, D. Akiyama 168,
N.N. Akolkar (D<sup>24</sup>, K. Al Khoury (D<sup>41</sup>, G.L. Alberghi (D<sup>23b</sup>, J. Albert (D<sup>165</sup>, P. Albicocco (D<sup>53</sup>),
G.L. Albouy 60, S. Alderweireldt 52, M. Aleksa 536, I.N. Aleksandrov 538, C. Alexa 5276,
T. Alexopoulos (10), F. Alfonsi (12), M. Algren (15), M. Alhroob (120), B. Ali (13), H.M.J. Ali (19),
S. Ali 148, S.W. Alibocus 192, M. Aliev 145, G. Alimonti 171a, W. Alkakhi 155, C. Allaire 166,
B.M.M. Allbrooke <sup>146</sup>, J.F. Allen <sup>52</sup>, C.A. Allendes Flores <sup>137f</sup>, P.P. Allport <sup>20</sup>, A. Aloisio <sup>72a,72b</sup>,
F. Alonso <sup>690</sup>, C. Alpigiani <sup>6138</sup>, M. Alvarez Estevez <sup>699</sup>, A. Alvarez Fernandez <sup>6100</sup>,
M. Alves Cardoso 656, M.G. Alviggi 672a,72b, M. Aly 6101, Y. Amaral Coutinho 83b, A. Ambler 6104,
C. Amelung<sup>36</sup>, M. Amerl <sup>1010</sup>, C.G. Ames <sup>1010</sup>, D. Amidei <sup>1010</sup>, S.P. Amor Dos Santos <sup>10130</sup>a,
K.R. Amos © 163, V. Ananiev © 125, C. Anastopoulos © 139, T. Andeen © 11, J.K. Anders © 36,
S.Y. Andrean (D<sup>47a,47b</sup>, A. Andreazza (D<sup>71a,71b</sup>, S. Angelidakis (D<sup>9</sup>, A. Angerami (D<sup>41,ab</sup>,
A.V. Anisenkov (1937), A. Annovi (1974a), C. Antel (1956), M.T. Anthony (19139), E. Antipov (19145).
M. Antonelli ©<sup>53</sup>, F. Anulli ©<sup>75a</sup>, M. Aoki ©<sup>84</sup>, T. Aoki ©<sup>153</sup>, J.A. Aparisi Pozo ©<sup>163</sup>, M.A. Aparo ©<sup>146</sup>,
L. Aperio Bella <sup>648</sup>, C. Appelt <sup>618</sup>, A. Apyan <sup>626</sup>, N. Aranzabal <sup>636</sup>, S.J. Arbiol Val <sup>687</sup>,
C. Arcangeletti \mathbb{D}^{53}, A.T.H. Arce \mathbb{D}^{51}, E. Arena \mathbb{D}^{92}, J-F. Arguin \mathbb{D}^{108}, S. Argyropoulos \mathbb{D}^{54},
J.-H. Arling (D<sup>48</sup>, O. Arnaez (D<sup>4</sup>, H. Arnold (D<sup>114</sup>, G. Artoni (D<sup>75a,75b</sup>, H. Asada (D<sup>111</sup>, K. Asai (D<sup>118</sup>),
S. Asai 10153, N.A. Asbah 1061, K. Assamagan 1029, R. Astalos 1028a, S. Atashi 10160, R.J. Atkin 1033a,
M. Atkinson<sup>162</sup>, H. Atmani<sup>35f</sup>, P.A. Atmasiddha <sup>106</sup>, K. Augsten <sup>132</sup>, S. Auricchio <sup>72a,72b</sup>,
A.D. Auriol 620, V.A. Austrup 6101, G. Avolio 636, K. Axiotis 656, G. Azuelos 6108, D. Babal 6286,
H. Bachacou 135, K. Bachas 152,p, A. Bachiu 34, F. Backman 47a,47b, A. Badea 61, T.M. Baer 106,
P. Bagnaia (D75a,75b), M. Bahmani (D18), A.J. Bailey (D163), V.R. Bailey (D162), J.T. Baines (D134), L. Baines (D94),
O.K. Baker 172, E. Bakos 15, D. Bakshi Gupta 8, V. Balakrishnan 120, R. Balasubramanian 114,
E.M. Baldin ^{\odot 37}, P. Balek ^{\odot 86a}, E. Ballabene ^{\odot 23b,23a}, F. Balli ^{\odot 135}, L.M. Baltes ^{\odot 63a}, W.K. Balunas ^{\odot 32},
J. Balz 100, E. Banas 187, M. Bandieramonte 129, A. Bandyopadhyay 124, S. Bansal 124,
L. Barak 151, M. Barakat 48, E.L. Barberio 105, D. Barberis 576,57a, M. Barbero 102,
M.Z. Barel © 114, K.N. Barends © 33a, T. Barillari © 110, M-S. Barisits © 36, T. Barklow © 143, P. Baron © 122,
D.A. Baron Moreno (D101), A. Baroncelli (D62a), G. Barone (D29), A.J. Barr (D126), J.D. Barr (D96),
L. Barranco Navarro (D<sup>47a,47b</sup>, F. Barreiro (D<sup>99</sup>, J. Barreiro Guimarães da Costa (D<sup>14a</sup>, U. Barron (D<sup>151</sup>),
M.G. Barros Teixeira 130a, S. Barsov 137, F. Bartels 163a, R. Bartoldus 143, A.E. Barton 1991,
P. Bartos (D<sup>28a</sup>, A. Basan (D<sup>100</sup>, M. Baselga (D<sup>49</sup>, A. Bassalat (D<sup>66,b</sup>, M.J. Basso (D<sup>156a</sup>, C.R. Basson (D<sup>101</sup>),
R.L. Bates \bigcirc^{59}, S. Batlamous ^{35e}, J.R. Batley \bigcirc^{32}, B. Batool \bigcirc^{141}, M. Battaglia \bigcirc^{136}, D. Battulga \bigcirc^{18},
M. Bauce \bigcirc^{75a,75b}, M. Bauer \bigcirc^{36}, P. Bauer \bigcirc^{24}, L.T. Bazzano Hurrell \bigcirc^{30}, J.B. Beacham \bigcirc^{51},
T. Beau © 127, J.Y. Beaucamp © 90, P.H. Beauchemin © 158, F. Becherer © 54, P. Bechtle © 24,
H.P. Beck (19,0), K. Becker (167), A.J. Beddall (182), V.A. Bednyakov (1838), C.P. Bee (1845),
L.J. Beemster 15, T.A. Beermann 36, M. Begalli 83d, M. Begel 29, A. Behera 145, J.K. Behr 48,
J.F. Beirer 055, F. Beisiegel 024, M. Belfkir 0159, G. Bella 0151, L. Bellagamba 023b, A. Bellerive 034,
P. Bellos (D<sup>20</sup>, K. Beloborodov (D<sup>37</sup>, D. Benchekroun (D<sup>35a</sup>, F. Bendebba (D<sup>35a</sup>, Y. Benhammou (D<sup>151</sup>),
M. Benoit <sup>©29</sup>, J.R. Bensinger <sup>©26</sup>, S. Bentvelsen <sup>©114</sup>, L. Beresford <sup>©48</sup>, M. Beretta <sup>©53</sup>,
```

```
E. Bergeaas Kuutmann 161, N. Berger 4, B. Bergmann 132, J. Beringer 17a, G. Bernardi 5,
C. Bernius 143, F.U. Bernlochner 24, F. Bernon 36,102, A. Berrocal Guardia 13, T. Berry 55,
P. Berta (133), A. Berthold (150), I.A. Bertram (191), S. Bethke (110), A. Betti (175a,75b), A.J. Bevan (194),
N.K. Bhalla <sup>654</sup>, M. Bhamjee <sup>633c</sup>, S. Bhatta <sup>6145</sup>, D.S. Bhattacharya <sup>6166</sup>, P. Bhattarai <sup>6143</sup>,
V.S. Bhopatkar (D<sup>121</sup>, R. Bi<sup>29,ai</sup>, R.M. Bianchi (D<sup>129</sup>, G. Bianco (D<sup>23b,23a</sup>, O. Biebel (D<sup>109</sup>, R. Bielski (D<sup>123</sup>,
M. Biglietti \bigcirc^{77a}, M. Bindi \bigcirc^{55}, A. Bingul \bigcirc^{21b}, C. Bini \bigcirc^{75a,75b}, A. Biondini \bigcirc^{92}, C.J. Birch-sykes \bigcirc^{101},
G.A. Bird (D<sup>20</sup>,134), M. Birman (D<sup>169</sup>), M. Biros (D<sup>133</sup>), S. Biryukov (D<sup>146</sup>), T. Bisanz (D<sup>49</sup>),
E. Bisceglie (D43b,43a, J.P. Biswal (D134, D. Biswas (D141, A. Bitadze (D101, K. Bjørke (D125, I. Bloch (D48,
C. Blocker (D<sup>26</sup>, A. Blue (D<sup>59</sup>, U. Blumenschein (D<sup>94</sup>, J. Blumenthal (D<sup>100</sup>, G.J. Bobbink (D<sup>114</sup>,
V.S. Bobrovnikov \bigcirc^{37}, M. Boehler \bigcirc^{54}, B. Boehm \bigcirc^{166}, D. Bogavac \bigcirc^{36}, A.G. Bogdanchikov \bigcirc^{37},
C. Bohm (D<sup>47</sup>a), V. Boisvert (D<sup>95</sup>), P. Bokan (D<sup>48</sup>), T. Bold (D<sup>86</sup>a), M. Bomben (D<sup>5</sup>), M. Bona (D<sup>94</sup>),
M. Boonekamp <sup>135</sup>, C.D. Booth <sup>95</sup>, A.G. Borbély <sup>59</sup>, I.S. Bordulev <sup>37</sup>, H.M. Borecka-Bielska <sup>108</sup>,
G. Borissov (D<sup>91</sup>, D. Bortoletto (D<sup>126</sup>, D. Boscherini (D<sup>23b</sup>, M. Bosman (D<sup>13</sup>, J.D. Bossio Sola (D<sup>36</sup>),
K. Bouaouda (D35a), N. Bouchhar (D163), J. Boudreau (D129), E.V. Bouhova-Thacker (D91), D. Boumediene (D40),
R. Bouquet 165, A. Boveia 119, J. Boyd 36, D. Boye 29, I.R. Boyko 38, J. Bracinik 20,
N. Brahimi 62d, G. Brandt 171, O. Brandt 32, F. Braren 48, B. Brau 103, J.E. Brau 123,
R. Brener 169, L. Brenner 114, R. Brenner 161, S. Bressler 169, D. Britton 159, D. Britzger 110,
I. Brock ^{\odot}2^4, G. Brooijmans ^{\odot}4^1, W.K. Brooks ^{\odot}1^{37f}, E. Brost ^{\odot}2^9, L.M. Brown ^{\odot}1^{65}, L.E. Bruce ^{\odot}6^1, T.L. Bruckler ^{\odot}1^{26}, P.A. Bruckman de Renstrom ^{\odot}8^7, B. Brüers ^{\odot}4^8, A. Bruni ^{\odot}2^{3b}, G. Bruni ^{\odot}2^{3b},
M. Bruschi (D<sup>23b</sup>, N. Bruscino (D<sup>75a,75b</sup>, T. Buanes (D<sup>16</sup>, Q. Buat (D<sup>138</sup>, D. Buchin (D<sup>110</sup>, A.G. Buckley (D<sup>59</sup>),
O. Bulekov \bigcirc^{37}, B.A. Bullard \bigcirc^{143}, S. Burdin \bigcirc^{92}, C.D. Burgard \bigcirc^{49}, A.M. Burger \bigcirc^{40},
B. Burghgrave <sup>108</sup>, O. Burlayenko <sup>1054</sup>, J.T.P. Burr <sup>1032</sup>, C.D. Burton <sup>1011</sup>, J.C. Burzynski <sup>10142</sup>,
E.L. Busch (D<sup>41</sup>, V. Büscher (D<sup>100</sup>, P.J. Bussey (D<sup>59</sup>, J.M. Butler (D<sup>25</sup>, C.M. Buttar (D<sup>59</sup>),
J.M. Butterworth <sup>6</sup>96, W. Buttinger <sup>6</sup>134, C.J. Buxo Vazquez <sup>6</sup>107, A.R. Buzykaev <sup>37</sup>7,
S. Cabrera Urbán (163), L. Cadamuro (166), D. Caforio (158), H. Cai (129), Y. Cai (144,14e), Y. Cai (144,14e), Y. Cai (154,14e), Y. Cai (1
V.M.M. Cairo (1036), O. Cakir (103a), N. Calace (1036), P. Calafiura (1017a), G. Calderini (10127), P. Calfayan (1068),
G. Callea 59, L.P. Caloba 83b, D. Calvet 640, S. Calvet 640, T.P. Calvet 10102, M. Calvetti 74a,74b,
R. Camacho Toro 127, S. Camarda 36, D. Camarero Munoz 26, P. Camarri 76a,76b,
M.T. Camerlingo (D<sup>72a,72b</sup>, D. Cameron (D<sup>36</sup>, C. Camincher (D<sup>165</sup>, M. Campanelli (D<sup>96</sup>, A. Camplani (D<sup>42</sup>),
V. Canale (D<sup>72a,72b</sup>, A. Canesse (D<sup>104</sup>, J. Cantero (D<sup>163</sup>, Y. Cao (D<sup>162</sup>, F. Capocasa (D<sup>26</sup>, M. Capua (D<sup>43b,43a</sup>,
A. Carbone \bigcirc^{71a,71b}, R. Cardarelli \bigcirc^{76a}, J.C.J. Cardenas \bigcirc^{8}, F. Cardillo \bigcirc^{163}, G. Carducci \bigcirc^{43b,43a},
T. Carli <sup>10</sup> <sup>36</sup>, G. Carlino <sup>17</sup> <sup>2a</sup>, J.I. Carlotto <sup>13</sup>, B.T. Carlson <sup>129,q</sup>, E.M. Carlson <sup>165,156a</sup>,
L. Carminati (D<sup>71a,71b</sup>, A. Carnelli (D<sup>135</sup>, M. Carnesale (D<sup>75a,75b</sup>, S. Caron (D<sup>113</sup>, E. Carquin (D<sup>137f</sup>,
S. Carrá (D<sup>71a</sup>, G. Carratta (D<sup>23b,23a</sup>, F. Carrio Argos (D<sup>33g</sup>, J.W.S. Carter (D<sup>155</sup>, T.M. Carter (D<sup>52</sup>),
M.P. Casado (D<sup>13,i</sup>, M. Caspar (D<sup>48</sup>, F.L. Castillo (D<sup>4</sup>, L. Castillo Garcia (D<sup>13</sup>, V. Castillo Gimenez (D<sup>163</sup>,
N.F. Castro (D130a,130e), A. Catinaccio (D36), J.R. Catmore (D125), V. Cavaliere (D29), N. Cavalli (D23b,23a),
V. Cavasinni (D<sup>74a,74b</sup>, Y.C. Cekmecelioglu (D<sup>48</sup>, E. Celebi (D<sup>21a</sup>, F. Celli (D<sup>126</sup>, M.S. Centonze (D<sup>70a,70b</sup>,
V. Cepaitis 656, K. Cerny 122, A.S. Cerqueira 83a, A. Cerri 146, L. Cerrito 76a,76b, F. Cerutti 17a,
B. Cervato 141, A. Cervelli 23b, G. Cesarini 53, S.A. Cetin 82, Z. Chadi 35a, D. Chakraborty 115,
J. Chan 6 170, W.Y. Chan 6 153, J.D. Chapman 6 22, E. Chapon 6 135, B. Chargeishvili 6 149b,
D.G. Charlton (D<sup>20</sup>, T.P. Charman (D<sup>94</sup>, M. Chatterjee (D<sup>19</sup>, C. Chauhan (D<sup>133</sup>, S. Chekanov (D<sup>6</sup>),
S.V. Chekulaev 156a, G.A. Chelkov 188,a, A. Chen 1610, B. Chen 1615, B. Chen 1615, H. Chen 1614c,
H. Chen (D<sup>29</sup>, J. Chen (D<sup>62c</sup>, J. Chen (D<sup>142</sup>, M. Chen (D<sup>126</sup>, S. Chen (D<sup>153</sup>, S.J. Chen (D<sup>14c</sup>, X. Chen (D<sup>62c</sup>, 135),
X. Chen (D<sup>14b</sup>,ae), Y. Chen (D<sup>62a</sup>, C.L. Cheng (D<sup>170</sup>), H.C. Cheng (D<sup>64a</sup>, S. Cheong (D<sup>143</sup>), A. Cheplakov (D<sup>38</sup>),
E. Cheremushkina <sup>648</sup>, E. Cherepanova <sup>6114</sup>, R. Cherkaoui El Moursli <sup>635e</sup>, E. Cheu <sup>67</sup>, K. Cheung <sup>65</sup>,
L. Chevalier 135, V. Chiarella 53, G. Chiarelli 74a, N. Chiedde 102, G. Chiodini 70a,
A.S. Chisholm <sup>©20</sup>, A. Chitan <sup>©27b</sup>, M. Chitishvili <sup>©163</sup>, M.V. Chizhov <sup>©38</sup>, K. Choi <sup>©11</sup>,
```

```
A.R. Chomont (D<sup>75a,75b</sup>, Y. Chou (D<sup>103</sup>, E.Y.S. Chow (D<sup>113</sup>, T. Chowdhury (D<sup>33g</sup>, K.L. Chu (D<sup>169</sup>),
M.C. Chu (10<sup>64a</sup>, X. Chu (10<sup>14a,14e</sup>, J. Chudoba (10<sup>131</sup>, J.J. Chwastowski (10<sup>87</sup>, D. Cieri (10<sup>110</sup>, K.M. Ciesla (10<sup>86a</sup>,
V. Cindro (D<sup>93</sup>, A. Ciocio (D<sup>17a</sup>, F. Cirotto (D<sup>72a,72b</sup>, Z.H. Citron (D<sup>169,k</sup>, M. Citterio (D<sup>71a</sup>, D.A. Ciubotaru<sup>27b</sup>,
B.M. Ciungu <sup>155</sup>, A. Clark <sup>56</sup>, P.J. Clark <sup>52</sup>, C. Clarry <sup>155</sup>, J.M. Clavijo Columbie <sup>48</sup>,
S.E. Clawson (D48), C. Clement (D47a,47b), J. Clercx (D48), Y. Coadou (D102), M. Cobal (D69a,69c),
A. Coccaro <sup>57b</sup>, R.F. Coelho Barrue <sup>130a</sup>, R. Coelho Lopes De Sa <sup>103</sup>, S. Coelli <sup>71a</sup>, H. Cohen <sup>151</sup>,
A.E.C. Coimbra (D<sup>71</sup>a,71b), B. Cole (D<sup>41</sup>, J. Collot (D<sup>60</sup>), P. Conde Muiño (D<sup>130a,130g</sup>, M.P. Connell (D<sup>33c</sup>),
S.H. Connell (D<sup>33c</sup>, I.A. Connelly (D<sup>59</sup>, E.I. Conroy (D<sup>126</sup>, F. Conventi (D<sup>72a,ag</sup>, H.G. Cooke (D<sup>20</sup>),
A.M. Cooper-Sarkar 10126, A. Cordeiro Oudot Choi 10127, L.D. Corpe 1040, M. Corradi 1075a,75b,
F. Corriveau 10104,w, A. Cortes-Gonzalez 1018, M.J. Costa 10163, F. Costanza 104, D. Costanzo 10139,
B.M. Cote 119, G. Cowan 95, K. Cranmer 170, D. Cremonini 23b,23a, S. Crépé-Renaudin 60,
F. Crescioli 127, M. Cristinziani 141, M. Cristoforetti 78a,78b, V. Croft 14, J.E. Crosby 121,
G. Crosetti (1043b,43a), A. Cueto (1099), T. Cuhadar Donszelmann (10160), H. Cui (1014a,14e), Z. Cui (107),
W.R. Cunningham <sup>59</sup>, F. Curcio <sup>43b,43a</sup>, P. Czodrowski <sup>36</sup>, M.M. Czurylo <sup>63b</sup>,
M.J. Da Cunha Sargedas De Sousa 657b,57a, J.V. Da Fonseca Pinto 683b, C. Da Via 6101,
W. Dabrowski ^{686a}, T. Dado ^{649}, S. Dahbi ^{633g}, T. Dai ^{6106}, D. Dal Santo ^{619}, C. Dallapiccola ^{6103},
M. Dam (10, J.R. Dandoy (10, J.R. Dandoy (10, M.F. Daneri (10, M.F. Daneri
M. Danninger (D<sup>142</sup>, V. Dao (D<sup>36</sup>, G. Darbo (D<sup>57b</sup>, S. Darmora (D<sup>6</sup>, S.J. Das (D<sup>29</sup>,ai), S. D'Auria (D<sup>71a,71b</sup>,
C. David (D<sup>156b</sup>, T. Davidek (D<sup>133</sup>, B. Davis-Purcell (D<sup>34</sup>, I. Dawson (D<sup>94</sup>, H.A. Day-hall (D<sup>132</sup>, K. De (D<sup>8</sup>),
R. De Asmundis <sup>12a</sup>, N. De Biase <sup>48</sup>, S. De Castro <sup>23b,23a</sup>, N. De Groot <sup>113</sup>, P. de Jong <sup>114</sup>,
H. De la Torre 1115, A. De Maria 114c, A. De Salvo 175a, U. De Sanctis 176a,76b, A. De Santo 1146,
J.B. De Vivie De Regie 60, D.V. Dedovich38, J. Degens 114, A.M. Deiana 44, F. Del Corso 23b,23a,
J. Del Peso (1099), F. Del Rio (1063a), F. Deliot (10135), C.M. Delitzsch (1049), M. Della Pietra (1072a,72b),
D. Della Volpe 656, A. Dell'Acqua 636, L. Dell'Asta 671a,71b, M. Delmastro 64, P.A. Delsart 660,
S. Demers \bigcirc^{172}, M. Demichev \bigcirc^{38}, S.P. Denisov \bigcirc^{37}, L. D'Eramo \bigcirc^{40}, D. Derendarz \bigcirc^{87}, F. Derue \bigcirc^{127},
P. Dervan (5)24, K. Desch (5)24, C. Deutsch (5)24, F.A. Di Bello (5)57b,57a, A. Di Ciaccio (5)76a,76b,
L. Di Ciaccio <sup>64</sup>, A. Di Domenico <sup>675a,75b</sup>, C. Di Donato <sup>672a,72b</sup>, A. Di Girolamo <sup>636</sup>,
G. Di Gregorio (1036), A. Di Luca (1078a,78b), B. Di Micco (1077a,77b), R. Di Nardo (1077a,77b), C. Diaconu (10102),
M. Diamantopoulou <sup>134</sup>, F.A. Dias <sup>114</sup>, T. Dias Do Vale <sup>142</sup>, M.A. Diaz <sup>137a,137b</sup>,
F.G. Diaz Capriles (D<sup>24</sup>, M. Didenko (D<sup>163</sup>, E.B. Diehl (D<sup>106</sup>, L. Diehl (D<sup>54</sup>, S. Díez Cornell (D<sup>48</sup>),
C. Diez Pardos 141, C. Dimitriadi 161,24,161, A. Dimitrievska 17a, J. Dingfelder 17a, J. 
S.J. Dittmeier (163b), F. Dittus (1536), F. Djama (1512), T. Djobava (15149b), J.I. Djuvsland (1514),
C. Doglioni (D<sup>101,98</sup>, A. Dohnalova (D<sup>28a</sup>, J. Dolejsi (D<sup>133</sup>, Z. Dolezal (D<sup>133</sup>, K.M. Dona (D<sup>39</sup>)
M. Donadelli 683c, B. Dong 6107, J. Donini 640, A. D'Onofrio 677a,77b, M. D'Onofrio 692,
J. Dopke <sup>134</sup>, A. Doria <sup>72a</sup>, N. Dos Santos Fernandes <sup>130a</sup>, P. Dougan <sup>101</sup>, M.T. Dova <sup>90</sup>,
A.T. Doyle <sup>59</sup>, M.A. Draguet <sup>126</sup>, E. Dreyer <sup>169</sup>, I. Drivas-koulouris <sup>10</sup>, M. Drnevich <sup>117</sup>,
A.S. Drobac 158, M. Drozdova 56, D. Du 62a, T.A. du Pree 114, F. Dubinin 57, M. Dubovsky 28a,
E. Duchovni (D<sup>169</sup>, G. Duckeck (D<sup>109</sup>, O.A. Ducu (D<sup>27b</sup>, D. Duda (D<sup>52</sup>, A. Dudarev (D<sup>36</sup>, E.R. Duden (D<sup>26</sup>,
M. D'uffizi 10101, L. Duflot 1066, M. Dührssen 1036, C. Dülsen 10171, A.E. Dumitriu 1027b, M. Dunford 1063a,
S. Dungs (D<sup>49</sup>, K. Dunne (D<sup>47a,47b</sup>, A. Duperrin (D<sup>102</sup>, H. Duran Yildiz (D<sup>3a</sup>, M. Düren (D<sup>58</sup>),
A. Durglishvili ©149b, B.L. Dwyer ©115, G.I. Dyckes ©17a, M. Dyndal ©86a, B.S. Dziedzic ©87,
Z.O. Earnshaw 146, G.H. Eberwein 126, B. Eckerova 28a, S. Eggebrecht 55,
E. Egidio Purcino De Souza 127, L.F. Ehrke 56, G. Eigen 16, K. Einsweiler 17a, T. Ekelof 161,
P.A. Ekman <sup>1</sup>0<sup>98</sup>, S. El Farkh <sup>1</sup>0<sup>35b</sup>, Y. El Ghazali <sup>1</sup>0<sup>35b</sup>, H. El Jarrari <sup>1</sup>0<sup>35e,148</sup>, A. El Moussaouy <sup>1</sup>0<sup>108</sup>,
V. Ellajosyula 161, M. Ellert 161, F. Ellinghaus 171, N. Ellis 1636, J. Elmsheuser 1629, M. Elsing 1636,
D. Emeliyanov \bigcirc^{134}, Y. Enari \bigcirc^{153}, I. Ene \bigcirc^{17a}, S. Epari \bigcirc^{13}, J. Erdmann \bigcirc^{49}, P.A. Erland \bigcirc^{87},
M. Errenst \bigcirc^{171}, M. Escalier \bigcirc^{66}, C. Escobar \bigcirc^{163}, E. Etzion \bigcirc^{151}, G. Evans \bigcirc^{130a}, H. Evans \bigcirc^{68},
```

```
L.S. Evans (D<sup>95</sup>, M.O. Evans (D<sup>146</sup>, A. Ezhilov (D<sup>37</sup>, S. Ezzargtouni (D<sup>35a</sup>, F. Fabbri (D<sup>59</sup>, L. Fabbri (D<sup>23b,23a</sup>,
G. Facini © 6, V. Fadeyev © 136, R.M. Fakhrutdinov © 37, S. Falciano © 75a, L.F. Falda Ulhoa Coelho © 36,
P.J. Falke (D<sup>24</sup>, J. Faltova (D<sup>133</sup>, C. Fan (D<sup>162</sup>, Y. Fan (D<sup>14a</sup>, Y. Fang (D<sup>14a,14e</sup>, M. Fanti (D<sup>71a,71b</sup>,
M. Faraj 69a,69b, Z. Farazpay 97, A. Farbin 8, A. Farilla 77a, T. Farooque 107, S.M. Farrington 52,
F. Fassi <sup>©35e</sup>, D. Fassouliotis <sup>©9</sup>, M. Faucci Giannelli <sup>©76a,76b</sup>, W.J. Fawcett <sup>©32</sup>, L. Fayard <sup>©66</sup>,
P. Federic 133, P. Federicova 131, O.L. Fedin 37,a, G. Fedotov 37, M. Feickert 170,
L. Feligioni (D)102, D.E. Fellers (D)123, C. Feng (D)62b, M. Feng (D)14b, Z. Feng (D)14b, M.J. Fenton (D)160,
A.B. Fenyuk<sup>37</sup>, L. Ferencz <sup>648</sup>, R.A.M. Ferguson <sup>691</sup>, S.I. Fernandez Luengo <sup>6137f</sup>,
P. Fernandez Martinez 13, M.J.V. Fernoux 10, J. Ferrando 14, A. Ferrari 16, P. Ferrari 114,113,
R. Ferrari 1073a, D. Ferrere 1056, C. Ferretti 10106, F. Fiedler 10100, P. Fiedler 10132, A. Filipčič 1093,
E.K. Filmer <sup>1</sup>, F. Filthaut <sup>113</sup>, M.C.N. Fiolhais <sup>130a,130c,c</sup>, L. Fiorini <sup>163</sup>, W.C. Fisher <sup>107</sup>,
T. Fitschen 1010, P.M. Fitzhugh 135, I. Fleck 10141, P. Fleischmann 10106, T. Flick 10171, M. Flores 1033d,ac,
L.R. Flores Castillo 64a, L. Flores Sanz De Acedo 536, F.M. Follega 78a,78b, N. Fomin 16,
J.H. Foo © 155, B.C. Forland 68, A. Formica © 135, A.C. Forti © 101, E. Fortin © 36, A.W. Fortman © 61,
M.G. Foti \bigcirc^{17a}, L. Fountas \bigcirc^{9,j}, D. Fournier \bigcirc^{66}, H. Fox \bigcirc^{91}, P. Francavilla \bigcirc^{74a,74b}, S. Francescato \bigcirc^{61},
S. Franchellucci 656, M. Franchini 623b,23a, S. Franchino 63a, D. Francis 6, L. Franco 6113,
V. Franco Lima 636, L. Franconi 648, M. Franklin 661, G. Frattari 626, A.C. Freegard 694,
W.S. Freund <sup>683b</sup>, Y.Y. Frid <sup>6151</sup>, J. Friend <sup>659</sup>, N. Fritzsche <sup>650</sup>, A. Froch <sup>654</sup>, D. Froidevaux <sup>636</sup>,
J.A. Frost <sup>126</sup>, Y. Fu <sup>62a</sup>, S. Fuenzalida Garrido <sup>137f</sup>, M. Fujimoto <sup>102</sup>, E. Fullana Torregrosa <sup>163,*</sup>,
K.Y. Fung \bigcirc^{64a}, E. Furtado De Simas Filho \bigcirc^{83b}, M. Furukawa \bigcirc^{153}, J. Fuster \bigcirc^{163}, A. Gabrielli \bigcirc^{23b,23a},
A. Gabrielli 155, P. Gadow 56, G. Gagliardi 576, L.G. Gagnon 17a, E.J. Gallas 126,
B.J. Gallop 6134, K.K. Gan 6119, S. Ganguly 6153, Y. Gao 652, F.M. Garay Walls 6137a,137b, B. Garcia<sup>29</sup>,
C. García 163, A. Garcia Alonso 114, A.G. Garcia Caffaro 172, J.E. García Navarro 163,
M. Garcia-Sciveres © 17a, G.L. Gardner © 128, R.W. Gardner © 39, N. Garelli © 158, D. Garg © 80,
R.B. Garg (143,n), J.M. Gargan<sup>52</sup>, C.A. Garner<sup>155</sup>, C.M. Garvey (133a), P. Gaspar (183b), V.K. Gassmann<sup>158</sup>,
G. Gaudio <sup>13</sup>, V. Gautam<sup>13</sup>, P. Gauzzi <sup>5</sup><sup>5a,75b</sup>, I.L. Gavrilenko <sup>37</sup>, A. Gavrilyuk <sup>37</sup>, C. Gay <sup>164</sup>,
G. Gaycken (D<sup>48</sup>, E.N. Gazis (D<sup>10</sup>, A.A. Geanta (D<sup>27b</sup>, C.M. Gee (D<sup>136</sup>, A. Gekow<sup>119</sup>, C. Gemme (D<sup>57b</sup>),
M.H. Genest \bigcirc^{60}, S. Gentile \bigcirc^{75a,75b}, A.D. Gentry \bigcirc^{112}, S. George \bigcirc^{95}, W.F. George \bigcirc^{20}, T. Geralis \bigcirc^{46},
P. Gessinger-Befurt 636, M.E. Geyik 6171, M. Ghani 6167, M. Ghneimat 6141, K. Ghorbanian 694,
A. Ghosal (D<sup>141</sup>, A. Ghosh (D<sup>160</sup>, A. Ghosh (D<sup>7</sup>, B. Giacobbe (D<sup>23b</sup>, S. Giagu (D<sup>75a,75b</sup>, T. Giani (D<sup>114</sup>,
P. Giannetti 10<sup>74a</sup>, A. Giannini 10<sup>62a</sup>, S.M. Gibson 10<sup>95</sup>, M. Gignac 10<sup>136</sup>, D.T. Gil 10<sup>86b</sup>, A.K. Gilbert 10<sup>86a</sup>,
B.J. Gilbert <sup>1</sup>0<sup>41</sup>, D. Gillberg <sup>1</sup>0<sup>34</sup>, G. Gilles <sup>1</sup>1<sup>14</sup>, N.E.K. Gillwald <sup>1</sup>0<sup>48</sup>, L. Ginabat <sup>1</sup>27,
D.M. Gingrich (D<sup>2</sup>,af), M.P. Giordani (D<sup>69a,69c</sup>), P.F. Giraud (D<sup>135</sup>), G. Giugliarelli (D<sup>69a,69c</sup>), D. Giugni (D<sup>71a</sup>),
F. Giuli <sup>©36</sup>, I. Gkialas <sup>©9,j</sup>, L.K. Gladilin <sup>©37</sup>, C. Glasman <sup>©99</sup>, G.R. Gledhill <sup>©123</sup>, G. Glemža <sup>©48</sup>,
M. Glisic<sup>123</sup>, I. Gnesi (D<sup>43b,f</sup>, Y. Go (D<sup>29,ai</sup>, M. Goblirsch-Kolb (D<sup>36</sup>, B. Gocke (D<sup>49</sup>, D. Godin<sup>108</sup>,
B. Gokturk <sup>©21a</sup>, S. Goldfarb <sup>©105</sup>, T. Golling <sup>©56</sup>, M.G.D. Gololo <sup>©33g</sup>, D. Golubkov <sup>©37</sup>,
J.P. Gombas © 107, A. Gomes © 130a, 130b, G. Gomes Da Silva © 141, A.J. Gomez Delegido © 163,
R. Gonçalo (D<sup>130a,130c</sup>, G. Gonella (D<sup>123</sup>, L. Gonella (D<sup>20</sup>, A. Gongadze (D<sup>149c</sup>, F. Gonnella (D<sup>20</sup>),
J.L. Gonski <sup>1</sup><sup>041</sup>, R.Y. González Andana <sup>1</sup><sup>52</sup>, S. González de la Hoz <sup>1</sup><sup>163</sup>, S. Gonzalez Fernandez <sup>1</sup><sup>13</sup>,
R. Gonzalez Lopez <sup>692</sup>, C. Gonzalez Renteria <sup>617a</sup>, M.V. Gonzalez Rodrigues <sup>648</sup>,
R. Gonzalez Suarez 161, S. Gonzalez-Sevilla 56, G.R. Gonzalvo Rodriguez 163, L. Goossens 56,
B. Gorini 636, E. Gorini 670a,70b, A. Gorišek 693, T.C. Gosart 6128, A.T. Goshaw 651, M.I. Gostkin 638,
S. Goswami <sup>121</sup>, C.A. Gottardo <sup>36</sup>, S.A. Gotz <sup>109</sup>, M. Gouighri <sup>35b</sup>, V. Goumarre <sup>48</sup>,
A.G. Goussiou 138, N. Govender 33c, I. Grabowska-Bold 86a, K. Graham 34, E. Gramstad 125,
S. Grancagnolo  <sup>1070a,70b</sup>, M. Grandi  <sup>146</sup>, C.M. Grant<sup>1,135</sup>, P.M. Gravila  <sup>1027f</sup>, F.G. Gravili  <sup>1070a,70b</sup>,
H.M. Gray 17a, M. Greco 70a,70b, C. Grefe 24, I.M. Gregor 48, P. Grenier 143, S.G. Grewe 10,
C. Grieco 13, A.A. Grillo 136, K. Grimm 31, S. Grinstein 13, J.-F. Grivaz 66, E. Gross 169,
```

```
J. Grosse-Knetter <sup>55</sup>, C. Grud <sup>106</sup>, J.C. Grundy <sup>126</sup>, L. Guan <sup>106</sup>, W. Guan <sup>29</sup>, C. Gubbels <sup>164</sup>,
J.G.R. Guerrero Rojas (D<sup>163</sup>, G. Guerrieri (D<sup>69a,69c</sup>, F. Guescini (D<sup>110</sup>, R. Gugel (D<sup>100</sup>, J.A.M. Guhit (D<sup>106</sup>,
A. Guida 18, T. Guillemin 4, E. Guilloton 167,134, S. Guindon 36, F. Guo 14a,14e, J. Guo 162c,
L. Guo (D48), Y. Guo (D106), R. Gupta (D48), R. Gupta (D129), S. Gurbuz (D24), S.S. Gurdasani (D54),
G. Gustavino 636, M. Guth 556, P. Gutierrez 5120, L.F. Gutierrez Zagazeta 128, M. Gutsche 550,
C. Gutschow 696, C. Gwenlan 6126, C.B. Gwilliam 692, E.S. Haaland 6125, A. Haas 6117,
M. Habedank <sup>©48</sup>, C. Haber <sup>©17a</sup>, H.K. Hadavand <sup>©8</sup>, A. Hadef <sup>©100</sup>, S. Hadzic <sup>©110</sup>, A.I. Hagan <sup>©91</sup>,
J.J. Hahn (D<sup>141</sup>, E.H. Haines (D<sup>96</sup>, M. Haleem (D<sup>166</sup>, J. Haley (D<sup>121</sup>, J.J. Hall (D<sup>139</sup>, G.D. Hallewell (D<sup>102</sup>),
L. Halser 19, K. Hamano 165, M. Hamer 124, G.N. Hamity 152, E.J. Hampshire 195, J. Ham 162b,
K. Han 62a, L. Han 614c, L. Han 62a, S. Han 617a, Y.F. Han 6155, K. Hanagaki 684, M. Hance 6136,
D.A. Hangal (D<sup>41</sup>, ab), H. Hanif (D<sup>142</sup>, M.D. Hank (D<sup>128</sup>, R. Hankache (D<sup>101</sup>, J.B. Hansen (D<sup>42</sup>),
J.D. Hansen ^{\odot}4^2, P.H. Hansen ^{\odot}4^2, K. Hara ^{\odot}15^7, D. Harada ^{\odot}5^6, T. Harenberg ^{\odot}17^1, S. Harkusha ^{\odot}3^7,
M.L. Harris © 103, Y.T. Harris © 126, J. Harrison © 13, N.M. Harrison © 119, P.F. Harrison 167,
N.M. Hartman (10) N.M. Hartmann (10), Y. Hasegawa (140), R. Hauser (10), C.M. Hawkes (120),
R.J. Hawkings 636, Y. Hayashi 6153, S. Hayashida 6111, D. Hayden 6107, C. Hayes 6106,
R.L. Hayes 114, C.P. Hays 126, J.M. Hays 194, H.S. Hayward 192, F. He 162a, M. He 14a,14e,
Y. He 154, Y. He 48, N.B. Heatley 94, V. Hedberg 98, A.L. Heggelund 125, N.D. Hehir 94,*,
C. Heidegger <sup>54</sup>, K.K. Heidegger <sup>54</sup>, W.D. Heidorn <sup>81</sup>, J. Heilman <sup>34</sup>, S. Heim <sup>48</sup>, T. Heim <sup>17a</sup>, J.G. Heinlein <sup>128</sup>, J.J. Heinrich <sup>123</sup>, L. Heinrich <sup>110</sup>, J. Heiplal <sup>131</sup>, L. Helary <sup>48</sup>, A. Held <sup>170</sup>,
S. Hellesund <sup>16</sup>, C.M. Helling <sup>164</sup>, S. Hellman <sup>164</sup>, R.C.W. Henderson <sup>91</sup>, L. Henkelmann <sup>163</sup>,
A.M. Henriques Correia<sup>36</sup>, H. Herde <sup>©98</sup>, Y. Hernández Jiménez <sup>©145</sup>, L.M. Herrmann <sup>©24</sup>,
T. Herrmann \bigcirc^{50}, G. Herten \bigcirc^{54}, R. Hertenberger \bigcirc^{109}, L. Hervas \bigcirc^{36}, M.E. Hesping \bigcirc^{100},
N.P. Hessey 156a, H. Hibi 185, E. Hill 155, S.J. Hillier 120, J.R. Hinds 107, F. Hinterkeuser 124,
M. Hirose 124, S. Hirose 157, D. Hirschbuehl 171, T.G. Hitchings 101, B. Hiti 193, J. Hobbs 145,
R. Hobincu 127e, N. Hod 169, M.C. Hodgkinson 139, B.H. Hodkinson 32, A. Hoecker 36,
D.D. Hofer 106, J. Hofer 1048, T. Holm 1024, M. Holzbock 10110, L.B.A.H. Hommels 1032,
B.P. Honan (101), J. Hong (102), T.M. Hong (129), B.H. Hooberman (101), W.H. Hopkins (101), Y. Horii (101),
S. Hou D<sup>148</sup>, A.S. Howard D<sup>93</sup>, J. Howarth D<sup>59</sup>, J. Hoya D<sup>6</sup>, M. Hrabovsky D<sup>122</sup>, A. Hrynevich D<sup>48</sup>, T. Hryn'ova D<sup>4</sup>, P.J. Hsu D<sup>65</sup>, S.-C. Hsu D<sup>138</sup>, Q. Hu D<sup>62a</sup>, Y.F. Hu D<sup>14a,14e</sup>, S. Huang D<sup>64b</sup>,
X. Huang 14c, X. Huang 14a, 14e, Y. Huang 139, Y. Huang 14a, Z. Huang 1011, Z. Hubacek 132,
M. Huebner \mathbb{D}^{24}, F. Huegging \mathbb{D}^{24}, T.B. Huffman \mathbb{D}^{126}, C.A. Hugli \mathbb{D}^{48}, M. Huhtinen \mathbb{D}^{36},
S.K. Huiberts \bigcirc^{16}, R. Hulsken \bigcirc^{104}, N. Huseynov \bigcirc^{12}, J. Huston \bigcirc^{107}, J. Huth \bigcirc^{61}, R. Hyneman \bigcirc^{143},
G. Iacobucci \bigcirc^{56}, G. Iakovidis \bigcirc^{29}, I. Ibragimov \bigcirc^{141}, L. Iconomidou-Fayard \bigcirc^{66}, P. Iengo \bigcirc^{72a,72b},
R. Iguchi 153, T. Iizawa 126, Y. Ikegami 84, N. Ilic 155, H. Imam 135a, M. Ince Lezki 56,
T. Ingebretsen Carlson (D<sup>47a,47b</sup>, G. Introzzi (D<sup>73a,73b</sup>, M. Iodice (D<sup>77a</sup>, V. Ippolito (D<sup>75a,75b</sup>, R.K. Irwin (D<sup>92</sup>),
M. Ishino 153, W. Islam 170, C. Issever 18,48, S. Istin 21a,ak, H. Ito 168, J.M. Iturbe Ponce 64a,
R. Iuppa (D<sup>78a,78b</sup>), A. Ivina (D<sup>169</sup>), J.M. Izen (D<sup>45</sup>), V. Izzo (D<sup>72a</sup>), P. Jacka (D<sup>131,132</sup>), P. Jackson (D<sup>1</sup>),
R.M. Jacobs • B.P. Jaeger • C.S. Jagfeld • G. Jain • Jain 
L. Jeanty 123, J. Jejelava 149a,z, P. Jenni 54,g, C.E. Jessiman 34, S. Jézéquel 4, C. Jia62b, J. Jia 145,
X. Jia 61, X. Jia 14a,14e, Z. Jia 14c, S. Jiggins 48, J. Jimenez Pena 15, S. Jin 14c, A. Jinaru 15,
O. Jinnouchi 154, P. Johansson 139, K.A. Johns 7, J.W. Johnson 136, D.M. Jones 32, E. Jones 48,
P. Jones (D32), R.W.L. Jones (D91), T.J. Jones (D92), H.L. Joos (D55,36), R. Joshi (D119), J. Jovicevic (D15),
X. Ju 17a, J.J. Junggeburth 17a, T. Junkermann 16a, A. Juste Rozas 13a, M.K. Juzek 87,
S. Kabana (137e), A. Kaczmarska (87, M. Kado (110), H. Kagan (119), M. Kagan (143), A. Kahn<sup>41</sup>,
A. Kahn \mathbb{D}^{128}, C. Kahra \mathbb{D}^{100}, T. Kaji \mathbb{D}^{153}, E. Kajomovitz \mathbb{D}^{150}, N. Kakati \mathbb{D}^{169}, I. Kalaitzidou \mathbb{D}^{54},
C.W. Kalderon <sup>©29</sup>, A. Kamenshchikov <sup>©155</sup>, N.J. Kang <sup>©136</sup>, D. Kar <sup>©33g</sup>, K. Karava <sup>©126</sup>,
```

```
M.J. Kareem (D156b), E. Karentzos (D54), I. Karkanias (D152), O. Karkout (D114), S.N. Karpov (D38),
Z.M. Karpova (D<sup>38</sup>, V. Kartvelishvili (D<sup>91</sup>, A.N. Karyukhin (D<sup>37</sup>, E. Kasimi (D<sup>152</sup>, J. Katzy (D<sup>48</sup>, S. Kaur (D<sup>34</sup>,
K. Kawade (140, M.P. Kawale (120, C. Kawamoto (188, T. Kawamoto (135, E.F. Kay (136, F.I. Kaya (158,
S. Kazakos <sup>107</sup>, V.F. Kazanin <sup>37</sup>, Y. Ke <sup>145</sup>, J.M. Keaveney <sup>33a</sup>, R. Keeler <sup>165</sup>, G.V. Kehris <sup>61</sup>,
J.S. Keller ^{\odot 34}, A.S. Kelly ^{96}, J.J. Kempster ^{\odot 146}, K.E. Kennedy ^{\odot 41}, P.D. Kennedy ^{\odot 100}, O. Kepka ^{\odot 131},
B.P. Kerridge <sup>167</sup>, S. Kersten <sup>171</sup>, B.P. Kerševan <sup>93</sup>, S. Keshri <sup>66</sup>, L. Keszeghova <sup>28a</sup>,
S. Ketabchi Haghighat 155, R.A. Khan 129, M. Khandoga 127, A. Khanov 121, A.G. Kharlamov 137,
T. Kharlamova 637, E.E. Khoda 6138, M. Kholodenko 637, T.J. Khoo 618, G. Khoriauli 6166,
J. Khubua (D<sup>149b</sup>, Y.A.R. Khwaira (D<sup>66</sup>, A. Kilgallon (D<sup>123</sup>, D.W. Kim (D<sup>47a,47b</sup>, Y.K. Kim (D<sup>39</sup>),
N. Kimura 696, M.K. Kingston 655, A. Kirchhoff 55, C. Kirfel 624, F. Kirfel 624, J. Kirk 6134,
A.E. Kiryunin \bigcirc^{110}, C. Kitsaki \bigcirc^{10}, O. Kivernyk \bigcirc^{24}, M. Klassen \bigcirc^{63a}, C. Klein \bigcirc^{34}, L. Klein \bigcirc^{166},
M.H. Klein (10) 106, M. Klein (10) 29, S.B. Klein (10) 56, U. Klein (10) 29, P. Klimek (10) 36, A. Klimentov (10) 29,
T. Klioutchnikova (D<sup>36</sup>, P. Kluit (D<sup>114</sup>, S. Kluth (D<sup>110</sup>, E. Kneringer (D<sup>79</sup>, T.M. Knight (D<sup>155</sup>, A. Knue (D<sup>49</sup>),
R. Kobayashi <sup>©88</sup>, D. Kobylianskii <sup>©169</sup>, S.F. Koch <sup>©126</sup>, M. Kocian <sup>©143</sup>, P. Kodyš <sup>©133</sup>,
D.M. Koeck (123), P.T. Koenig (124), T. Koffas (134), O. Kolay (150), M. Kolb (135), I. Koletsou (144),
T. Komarek 122, K. Köneke 154, A.X.Y. Kong 11, T. Kono 118, N. Konstantinidis 156,
P. Kontaxakis <sup>56</sup>, B. Konya <sup>98</sup>, R. Kopeliansky <sup>68</sup>, S. Koperny <sup>86a</sup>, K. Korcyl <sup>87</sup>, K. Kordas <sup>152</sup>, e,
G. Koren (D<sup>151</sup>, A. Korn (D<sup>96</sup>, S. Korn (D<sup>55</sup>, I. Korolkov (D<sup>13</sup>, N. Korotkova (D<sup>37</sup>, B. Kortman (D<sup>114</sup>,
O. Kortner 110, S. Kortner 110, W.H. Kostecka 1115, V.V. Kostyukhin 1141, A. Kotsokechagia 115,
A. Kotwal <sup>1</sup>0<sup>51</sup>, A. Koulouris <sup>1</sup>0<sup>36</sup>, A. Kourkoumeli-Charalampidi <sup>1</sup>0<sup>73a,73b</sup>, C. Kourkoumelis <sup>1</sup>0<sup>9</sup>.
E. Kourlitis 110,ad, O. Kovanda 146, R. Kowalewski 165, W. Kozanecki 153, A.S. Kozhin 157,
V.A. Kramarenko (D<sup>37</sup>, G. Kramberger (D<sup>93</sup>, P. Kramer (D<sup>100</sup>, M.W. Krasny (D<sup>127</sup>, A. Krasznahorkay (D<sup>36</sup>,
J.W. Kraus (D<sup>171</sup>, J.A. Kremer (D<sup>48</sup>, T. Kresse (D<sup>50</sup>, J. Kretzschmar (D<sup>92</sup>, K. Kreul (D<sup>18</sup>, P. Krieger (D<sup>155</sup>, S. Krishnamurthy (D<sup>103</sup>, M. Krivos (D<sup>133</sup>, K. Krizka (D<sup>20</sup>, K. Kroeninger (D<sup>49</sup>, H. Kroha (D<sup>110</sup>, J. Kroll (D<sup>131</sup>,
J. Kroll <sup>128</sup>, K.S. Krowpman <sup>107</sup>, U. Kruchonak <sup>38</sup>, H. Krüger <sup>24</sup>, N. Krumnack M.C. Kruse <sup>51</sup>,
J.A. Krzysiak <sup>1087</sup>, O. Kuchinskaia <sup>1037</sup>, S. Kuday <sup>103a</sup>, S. Kuehn <sup>1036</sup>, R. Kuesters <sup>1054</sup>, T. Kuhl <sup>1048</sup>,
V. Kukhtin <sup>©38</sup>, Y. Kulchitsky <sup>©37,a</sup>, S. Kuleshov <sup>©137d,137b</sup>, M. Kumar <sup>©33g</sup>, N. Kumari <sup>©48</sup>,
A. Kupco 131, T. Kupfer, A. Kupich 37, O. Kuprash 54, H. Kurashige 58, L.L. Kurchaninov 156a,
O. Kurdysh 66, Y.A. Kurochkin 637, A. Kurova 637, M. Kuze 6154, A.K. Kvam 6103, J. Kvita 6122,
T. Kwan 10104, N.G. Kyriacou 10106, L.A.O. Laatu 10102, C. Lacasta 10163, F. Lacava 1075a,75b,
H. Lacker 18, D. Lacour 127, N.N. Lad 196, E. Ladygin 138, B. Laforge 127, T. Lagouri 137e,
F.Z. Lahbabi (D35a, S. Lai (D55), I.K. Lakomiec (D86a, N. Lalloue (D60), J.E. Lambert (D165), S. Lammers (D68).
W. Lampl <sup>107</sup>, C. Lampoudis <sup>152</sup>, A.N. Lancaster <sup>115</sup>, E. Lançon <sup>129</sup>, U. Landgraf <sup>154</sup>,
M.P.J. Landon 694, V.S. Lang 554, R.J. Langenberg 6103, O.K.B. Langrekken 6125, A.J. Lankford 6160,
F. Lanni 🗗 36, K. Lantzsch 🕩 24, A. Lanza 🕩 73a, A. Lapertosa 🕩 57b,57a, J.F. Laporte 🕩 135, T. Lari 🕩 71a,
F. Lasagni Manghi (D<sup>23b</sup>, M. Lassnig (D<sup>36</sup>, V. Latonova (D<sup>131</sup>, A. Laudrain (D<sup>100</sup>, A. Laurier (D<sup>150</sup>),
S.D. Lawlor (139), Z. Lawrence (101), R. Lazaridou 167, M. Lazzaroni (171a,71b), B. Le<sup>101</sup>,
E.M. Le Boulicaut <sup>51</sup>, B. Leban <sup>93</sup>, A. Lebedev <sup>81</sup>, M. LeBlanc <sup>101</sup>, F. Ledroit-Guillon <sup>60</sup>,
A.C.A. Lee<sup>96</sup>, S.C. Lee <sup>148</sup>, S. Lee <sup>47a,47b</sup>, T.F. Lee <sup>92</sup>, L.L. Leeuw <sup>33c</sup>, H.P. Lefebvre <sup>95</sup>,
M. Lefebvre <sup>165</sup>, C. Leggett <sup>17a</sup>, G. Lehmann Miotto <sup>36</sup>, M. Leigh <sup>56</sup>, W.A. Leight <sup>103</sup>,
W. Leinonen 113, A. Leisos 152, M.A.L. Leite 83c, C.E. Leitgeb 48, R. Leitner 133,
K.J.C. Leney \mathbb{D}^{44}, T. Lenz \mathbb{D}^{24}, S. Leone \mathbb{D}^{74a}, C. Leonidopoulos \mathbb{D}^{52}, A. Leopold \mathbb{D}^{144}, C. Leroy \mathbb{D}^{108},
R. Les © 107, C.G. Lester © 32, M. Levchenko © 37, J. Levêque © 4, D. Levin © 106, L.J. Levinson © 169,
M.P. Lewicki 627, D.J. Lewis 64, A. Li 55, B. Li 62b, C. Li<sup>62a</sup>, C-Q. Li 62c, H. Li 62a, H. Li 62b,
H. Li 14c, H. Li 14a, 14e, Q.Y. Li 162c, K. Li 1538, L. Li 162c, M. Li 164a, 14e, Q.Y. Li 162a,
S. Li 14a,14e, S. Li 62d,62c,d, T. Li 5, X. Li 14a,14e,
S. Liang <sup>14a,14e</sup>, Z. Liang <sup>14a</sup>, M. Liberatore <sup>135</sup>, B. Liberti <sup>76a</sup>, K. Lie <sup>64c</sup>, J. Lieber Marin <sup>83b</sup>,
```

```
H. Lien \bigcirc^{68}, K. Lin \bigcirc^{107}, R.E. Lindley \bigcirc^{7}, J.H. Lindon \bigcirc^{2}, E. Lipeles \bigcirc^{128}, A. Lipniacka \bigcirc^{16},
A. Lister 164, J.D. Little 4, B. Liu 14a, B.X. Liu 142, D. Liu 162d, 52c, J.B. Liu 162a, J.K.K. Liu 1532,
K. Liu 62d,62c, M. Liu 62a, M.Y. Liu 62a, P. Liu 614a, Q. Liu 62d,138,62c, X. Liu 62a, Y. Liu 614d,14e,
Y.L. Liu 662b, Y.W. Liu 662a, J. Llorente Merino 142, S.L. Lloyd 694, E.M. Lobodzinska 48,
P. Loch <sup>1</sup>0, T. Lohse <sup>18</sup>, K. Lohwasser <sup>139</sup>, E. Loiacono <sup>48</sup>, M. Lokajicek <sup>131,*</sup>, J.D. Lomas <sup>20</sup>,
J.D. Long 162, I. Longarini 160, L. Longo 170a,70b, R. Longo 162, I. Lopez Paz 167,
A. Lopez Solis <sup>648</sup>, J. Lorenz <sup>6109</sup>, N. Lorenzo Martinez <sup>64</sup>, A.M. Lory <sup>6109</sup>,
G. Löschcke Centeno (D)146, O. Loseva (D)37, X. Lou (D)47a,47b, X. Lou (D)14a,14e, A. Lounis (D)66, J. Love (D)6,
P.A. Love (1991), G. Lu (1914a,14e), M. Lu (1980), S. Lu (19128), Y.J. Lu (1965), H.J. Lubatti (19138), C. Luci (1975a,75b),
F.L. Lucio Alves 14c, A. Lucotte 66, F. Luehring 68, I. Luise 145, O. Lukianchuk 66,
O. Lundberg (144, B. Lund-Jensen (144, N.A. Luongo (123, M.S. Lutz (151, A.B. Lux (125, D. Lynn (129,
H. Lyons<sup>92</sup>, R. Lysak <sup>131</sup>, E. Lytken <sup>98</sup>, V. Lyubushkin <sup>38</sup>, T. Lyubushkina <sup>38</sup>, M.M. Lyukova <sup>145</sup>,
H. Ma 629, K. Ma<sup>62a</sup>, L.L. Ma 62b, W. Ma 62a, Y. Ma 121, D.M. Mac Donell 165,
G. Maccarrone <sup>53</sup>, J.C. MacDonald <sup>100</sup>, P.C. Machado De Abreu Farias <sup>83b</sup>, R. Madar <sup>40</sup>,
W.F. Mader (D<sup>50</sup>, T. Madula (D<sup>96</sup>, J. Maeda (D<sup>85</sup>, T. Maeno (D<sup>29</sup>, H. Maguire (D<sup>139</sup>, V. Maiboroda (D<sup>135</sup>),
A. Maio (D130a,130b,130d), K. Maj (D86a), O. Majersky (D48), S. Majewski (D123), N. Makovec (D66),
V. Maksimovic 15, B. Malaescu 127, Pa. Malecki 87, V.P. Maleev 37, F. Malek 66, M. Mali 93,
D. Malito <sup>10</sup>95, U. Mallik <sup>10</sup>80, S. Maltezos <sup>10</sup>, S. Malyukov <sup>38</sup>, J. Mamuzic <sup>13</sup>, G. Mancini <sup>15</sup>3,
G. Manco (573a,73b), J.P. Mandalia (594), I. Mandić (593), L. Manhaes de Andrade Filho (583a),
I.M. Maniatis (D<sup>169</sup>, J. Manjarres Ramos (D<sup>102</sup>, aa, D.C. Mankad (D<sup>169</sup>, A. Mann (D<sup>109</sup>, B. Mansoulie (D<sup>135</sup>),
S. Manzoni 👨 36, L. Mao 📵 62c, X. Mapekula 📵 33c, A. Marantis 📵 152, r, G. Marchiori 📵 5,
M. Marcisovsky (D<sup>131</sup>, C. Marcon (D<sup>71a</sup>, M. Marinescu (D<sup>20</sup>, M. Marjanovic (D<sup>120</sup>, E.J. Marshall (D<sup>91</sup>)
Z. Marshall 17a, S. Marti-Garcia 16a, T.A. Martin 16a, V.J. Martin 15a, B. Martin dit Latour 16a,
L. Martinelli ©75a,75b, M. Martinez ©13,8, P. Martinez Agullo ©163, V.I. Martinez Outschoorn ©103,
P. Martinez Suarez 13, S. Martin-Haugh 134, V.S. Martoiu 27b, A.C. Martyniuk 96, A. Marzin 36,
D. Mascione (D<sup>78a,78b</sup>, L. Masetti (D<sup>100</sup>, T. Mashimo (D<sup>153</sup>, J. Masik (D<sup>101</sup>, A.L. Maslennikov (D<sup>37</sup>,
L. Massa (D<sup>23b</sup>, P. Massarotti (D<sup>72a,72b</sup>, P. Mastrandrea (D<sup>74a,74b</sup>, A. Mastroberardino (D<sup>43b,43a</sup>,
T. Masubuchi \mathbb{D}^{153}, T. Mathisen \mathbb{D}^{161}, J. Matousek \mathbb{D}^{133}, N. Matsuzawa \mathbb{D}^{153}, J. Maurer \mathbb{D}^{27b}, B. Maček \mathbb{D}^{93},
D.A. Maximov (D37, R. Mazini (D148, I. Maznas (D152, M. Mazza (D107, S.M. Mazza (D136,
E. Mazzeo (1071a,71b), C. Mc Ginn (1029), J.P. Mc Gowan (10104), S.P. Mc Kee (10106), E.F. McDonald (10105),
A.E. McDougall 114, J.A. Mcfayden 1146, R.P. McGovern 1128, G. Mchedlidze 1149b,
R.P. Mckenzie (D<sup>33g</sup>, T.C. Mclachlan (D<sup>48</sup>, D.J. Mclaughlin (D<sup>96</sup>, S.J. McMahon (D<sup>134</sup>,
C.M. Mcpartland <sup>192</sup>, R.A. McPherson <sup>165,w</sup>, S. Mehlhase <sup>109</sup>, A. Mehta <sup>192</sup>, D. Melini <sup>150</sup>,
B.R. Mellado Garcia <sup>©33g</sup>, A.H. Melo <sup>©55</sup>, F. Meloni <sup>©48</sup>, A.M. Mendes Jacques Da Costa <sup>©101</sup>,
H.Y. Meng (D<sup>155</sup>, L. Meng (D<sup>91</sup>, S. Menke (D<sup>110</sup>, M. Mentink (D<sup>36</sup>, E. Meoni (D<sup>43b,43a</sup>, G. Mercado (D<sup>115</sup>),
C. Merlassino © 126, L. Merola © 72a,72b, C. Meroni © 71a,71b, G. Merz 106, O. Meshkov © 37, J. Metcalfe © 6,
A.S. Mete 66, C. Meyer 68, J-P. Meyer 135, R.P. Middleton 134, L. Mijović 52, G. Mikenberg 169,
M. Mikestikova 131, M. Mikuž 193, H. Mildner 1010, A. Milic 136, C.D. Milke 144, D.W. Miller 139,
L.S. Miller <sup>1034</sup>, A. Milov <sup>169</sup>, D.A. Milstead <sup>47a,47b</sup>, T. Min<sup>14c</sup>, A.A. Minaenko <sup>1037</sup>,
I.A. Minashvili (^{149b}, L. Mince (^{159}), A.I. Mincer (^{117}, B. Mindur (^{186a}, M. Mineev (^{188}), Y. Mino (^{188}), L.M. Mir (^{189}), M. Miralles Lopez (^{163}), M. Mironova (^{17a}), A. Mishima (^{153}), M.C. Missio (^{113}),
A. Mitra 10167, V.A. Mitsou 10163, Y. Mitsumori 10111, O. Miu 10155, P.S. Miyagawa 1094,
T. Mkrtchyan 663a, M. Mlinarevic 696, T. Mlinarevic 696, M. Mlynarikova 636, S. Mobius 619,
P. Moder <sup>1048</sup>, P. Mogg <sup>1099</sup>, A.F. Mohammed <sup>14a,14e</sup>, S. Mohapatra <sup>1041</sup>, G. Mokgatitswane <sup>1033g</sup>,
L. Moleri (10) 169, B. Mondal (11) 131, S. Mondal (11) 132, K. Mönig (14), E. Monnier (11) 102,
L. Monsonis Romero<sup>163</sup>, J. Montejo Berlingen <sup>13</sup>, M. Montella <sup>119</sup>, F. Montereali <sup>77a,77b</sup>,
F. Monticelli (1990), S. Monzani (1969a,69c), N. Morange (1966), A.L. Moreira De Carvalho (19130a),
```

```
M. Moreno Llácer 163, C. Moreno Martinez 56, P. Morettini 57b, S. Morgenstern 36, M. Morii 61,
M. Morinaga 153, A.K. Morley 36, F. Morodei 75a,75b, L. Morvaj 36, P. Moschovakos 36,
B. Moser 36, M. Mosidze 149b, T. Moskalets 54, P. Moskvitina 113, J. Moss 31,1,
E.J.W. Moyse 103, O. Mtintsilana 133g, S. Muanza 1012, J. Mueller 129, D. Muenstermann 1091,
R. Müller 19, G.A. Mullier 161, A.J. Mullin 32, J.J. Mullin 128, D.P. Mungo 155, D. Munoz Perez 163,
F.J. Munoz Sanchez (101), M. Murin (101), W.J. Murray (107), A. Murrone (107), M. Muškinja (1017),
C. Mwewa (D<sup>29</sup>), A.G. Myagkov (D<sup>37</sup>,a, A.J. Myers (D<sup>8</sup>), G. Myers (D<sup>68</sup>), M. Myska (D<sup>132</sup>), B.P. Nachman (D<sup>17a</sup>),
O. Nackenhorst ( A. Nag ( A. Nag ( A. Nagai ( Nagai ( Nagano ( Nagano ( Nagano ( Nagai ( Nagano ( Nagai ( Nagai ( Nagai ( Nagano ( Nagai ( Nag
A.M. Nairz 636, Y. Nakahama 684, K. Nakamura 684, K. Nakkalil 65, H. Nanjo 6124, R. Narayan 644,
E.A. Narayanan 112, I. Naryshkin 37, M. Naseri 34, S. Nasri 159, C. Nass 24, G. Navarro 22a,
J. Navarro-Gonzalez <sup>163</sup>, R. Nayak <sup>151</sup>, A. Nayaz <sup>18</sup>, P.Y. Nechaeva <sup>37</sup>, F. Nechansky <sup>48</sup>,
L. Nedic (126), T.J. Neep (120), A. Negri (120), M. Negrini (123), C. Nellist (131), C. Nelson (134),
K. Nelson © 106, S. Nemecek © 131, M. Nessi © 36,h, M.S. Neubauer © 162, F. Neuhaus © 100,
J. Neundorf <sup>648</sup>, R. Newhouse <sup>6164</sup>, P.R. Newman <sup>620</sup>, C.W. Ng <sup>6129</sup>, Y.W.Y. Ng <sup>648</sup>, B. Ngair <sup>635e</sup>,
H.D.N. Nguyen 10108, R.B. Nickerson 10126, R. Nicolaidou 10135, J. Nielsen 10136, M. Niemeyer 1055,
J. Niermann (D<sup>55,36</sup>, N. Nikiforou (D<sup>36</sup>, V. Nikolaenko (D<sup>37,a</sup>, I. Nikolic-Audit (D<sup>127</sup>, K. Nikolopoulos (D<sup>20</sup>),
P. Nilsson (D<sup>29</sup>, I. Ninca (D<sup>48</sup>, H.R. Nindhito (D<sup>56</sup>, G. Ninio (D<sup>151</sup>, A. Nisati (D<sup>75a</sup>, N. Nishu (D<sup>2</sup>),
R. Nisius 110, J-E. Nitschke 50, E.K. Nkadimeng 33g, T. Nobe 153, D.L. Noel 32,
T. Nommensen 147, M.B. Norfolk 139, R.R.B. Norisam 696, B.J. Norman 34, J. Novak 93,
T. Novak (D48, L. Novotny (D132, R. Novotny (D112, L. Nozka (D122, K. Ntekas (D160,
N.M.J. Nunes De Moura Junior 683b, E. Nurse, J. Ocariz 6127, A. Ochi 685, I. Ochoa 6130a,
S. Oerdek (10<sup>48,t</sup>, J.T. Offermann (10<sup>39</sup>), A. Ogrodnik (10<sup>133</sup>), A. Oh (10<sup>101</sup>), C.C. Ohm (10<sup>144</sup>), H. Oide (10<sup>84</sup>),
R. Oishi 153, M.L. Ojeda 148, M.W. O'Keefe92, Y. Okumura 153, L.F. Oleiro Seabra 130a,
S.A. Olivares Pino 137d, D. Oliveira Damazio 29, D. Oliveira Goncalves 83a, J.L. Oliver 160,
Ö.O. Öncel 654, A.P. O'Neill 619, A. Onofre 6130a,130e, P.U.E. Onyisi 611, M.J. Oreglia 639,
G.E. Orellana (D)0, D. Orestano (D)77a,77b, N. Orlando (D)13, R.S. Orr (D)155, V. O'Shea (D)59,
L.M. Osojnak 128, R. Ospanov 62a, G. Otero y Garzon 30, H. Otono 89, P.S. Ott 63a,
G.J. Ottino 17a, M. Ouchrif 35d, J. Ouellette 29, F. Ould-Saada 125, M. Owen 59, R.E. Owen 134,
K.Y. Oyulmaz (D<sup>21a</sup>, V.E. Ozcan (D<sup>21a</sup>, F. Ozturk (D<sup>87</sup>, N. Ozturk (D<sup>8</sup>), S. Ozturk (D<sup>82</sup>, H.A. Pacey (D<sup>126</sup>),
A. Pacheco Pages ©13, C. Padilla Aranda ©13, G. Padovano ©75a,75b, S. Pagan Griso ©17a,
G. Palacino 68, A. Palazzo 70a,70b, S. Palestini 636, J. Pan 6172, T. Pan 64a, D.K. Panchal 611,
C.E. Pandini 114, J.G. Panduro Vazquez 55, H.D. Pandya 11, H. Pang 14b, P. Pani 548,
G. Panizzo 69a,69c, L. Paolozzi 56, C. Papadatos 108, S. Parajuli 44, A. Paramonov 6,
C. Paraskevopoulos <sup>10</sup>, D. Paredes Hernandez <sup>64b</sup>, K.R. Park <sup>41</sup>, T.H. Park <sup>155</sup>, M.A. Parker <sup>32</sup>,
F. Parodi D<sup>57b,57a</sup>, E.W. Parrish D<sup>115</sup>, V.A. Parrish D<sup>52</sup>, J.A. Parsons D<sup>41</sup>, U. Parzefall D<sup>54</sup>,
B. Pascual Dias 108, L. Pascual Dominguez 151, E. Pasqualucci 157a, S. Passaggio 157b, F. Pastore 1595,
P. Pasuwan (D<sup>47a,47b</sup>, P. Patel (D<sup>87</sup>, U.M. Patel (D<sup>51</sup>, J.R. Pater (D<sup>101</sup>, T. Pauly (D<sup>36</sup>, J. Pearkes (D<sup>143</sup>),
M. Pedersen \bigcirc^{125}, R. Pedro \bigcirc^{130a}, S.V. Peleganchuk \bigcirc^{37}, O. Penc \bigcirc^{36}, E.A. Pender \bigcirc^{52},
K.E. Penski 10109, M. Penzin 1037, B.S. Peralva 1083d, A.P. Pereira Peixoto 1060, L. Pereira Sanchez 1047a,47b,
D.V. Perepelitsa (D<sup>29</sup>,ai), E. Perez Codina (D<sup>156a</sup>, M. Perganti (D<sup>10</sup>, L. Perini (D<sup>71a,71b,*</sup>, H. Pernegger (D<sup>36</sup>,
O. Perrin ^{\bullet}<sup>0</sup><sup>40</sup>, K. Peters ^{\bullet}<sup>48</sup>, R.F.Y. Peters ^{\bullet}<sup>101</sup>, B.A. Petersen ^{\bullet}<sup>36</sup>, T.C. Petersen ^{\bullet}<sup>42</sup>, E. Petit ^{\bullet}<sup>102</sup>,
V. Petousis 132, C. Petridou 152,e, A. Petrukhin 141, M. Pettee 17a, N.E. Pettersson 36,
A. Petukhov  <sup>37</sup>, K. Petukhova <sup>133</sup>, R. Pezoa <sup>137</sup>f, L. Pezzotti <sup>36</sup>, G. Pezzullo <sup>172</sup>, T.M. Pham <sup>170</sup>,
T. Pham 105, P.W. Phillips 134, G. Piacquadio 145, E. Pianori 17a, F. Piazza 123, R. Piegaia 130,
D. Pietreanu <sup>©27b</sup>, A.D. Pilkington <sup>©101</sup>, M. Pinamonti <sup>©69a,69c</sup>, J.L. Pinfold <sup>©2</sup>,
B.C. Pinheiro Pereira 130a, A.E. Pinto Pinoargote 1100,135, L. Pintucci 69a,69c, K.M. Piper 146,
A. Pirttikoski <sup>©</sup><sup>56</sup>, D.A. Pizzi <sup>©</sup><sup>34</sup>, L. Pizzimento <sup>©</sup><sup>64b</sup>, A. Pizzini <sup>©</sup><sup>114</sup>, M.-A. Pleier <sup>©</sup><sup>29</sup>, V. Plesanovs<sup>54</sup>,
```

```
V. Pleskot 133, E. Plotnikova R. Poddar 4, R. Poettgen 98, L. Poggioli 127, I. Pokharel 55,
S. Polacek 133, G. Polesello 173a, A. Poley 142,156a, R. Polifka 132, A. Polini 23b, C.S. Pollard 167,
Z.B. Pollock (D<sup>119</sup>, V. Polychronakos (D<sup>29</sup>, E. Pompa Pacchi (D<sup>75a,75b</sup>, D. Ponomarenko (D<sup>113</sup>,
L. Pontecorvo (D<sup>36</sup>, S. Popa (D<sup>27a</sup>, G.A. Popeneciu (D<sup>27d</sup>, A. Poreba (D<sup>36</sup>, D.M. Portillo Quintero (D<sup>156a</sup>,
S. Pospisil 132, M.A. Postill 139, P. Postolache 27c, K. Potamianos 167, P.A. Potepa 86a,
I.N. Potrap (D<sup>38</sup>, C.J. Potter (D<sup>32</sup>, H. Potti (D<sup>1</sup>, T. Poulsen (D<sup>48</sup>, J. Poveda (D<sup>163</sup>, M.E. Pozo Astigarraga (D<sup>36</sup>,
A. Prades Ibanez 163, J. Pretel 54, D. Price 101, M. Primavera 70a, M.A. Principe Martin 199,
R. Privara 122, T. Procter 59, M.L. Proffitt 138, N. Proklova 128, K. Prokofiev 64c, G. Proto 110,
S. Protopopescu 29, J. Proudfoot 6, M. Przybycien 86a, W.W. Przygoda 86b, J.E. Puddefoot 139,
D. Pudzha 1037, D. Pyatiizbyantseva 1037, J. Qian 10106, D. Qichen 10101, Y. Qin 10101, T. Qiu 1052,
A. Quadt <sup>055</sup>, M. Queitsch-Maitland <sup>0101</sup>, G. Quetant <sup>056</sup>, R.P. Quinn <sup>0164</sup>, G. Rabanal Bolanos <sup>061</sup>,
D. Rafanoharana <sup>©54</sup>, F. Ragusa <sup>©71a,71b</sup>, J.L. Rainbolt <sup>©39</sup>, J.A. Raine <sup>©56</sup>, S. Rajagopalan <sup>©29</sup>,
E. Ramakoti 1037, I.A. Ramirez-Berend 1034, K. Ran 1048,14e, N.P. Rapheeha 1033g, H. Rasheed 1027b,
V. Raskina ©<sup>127</sup>, D.F. Rassloff ©<sup>63a</sup>, S. Rave ©<sup>100</sup>, B. Ravina ©<sup>55</sup>, I. Ravinovich ©<sup>169</sup>, M. Raymond ©<sup>36</sup>,
A.L. Read © 125, N.P. Readioff © 139, D.M. Rebuzzi © 73a,73b, G. Redlinger © 29, A.S. Reed © 110,
K. Reeves \bigcirc^{26}, J.A. Reidelsturz \bigcirc^{171}, D. Reikher \bigcirc^{151}, A. Rej \bigcirc^{49}, C. Rembser \bigcirc^{36}, A. Renardi \bigcirc^{48},
M. Renda ©27b, M.B. Rendel<sup>110</sup>, F. Renner ©48, A.G. Rennie ©160, A.L. Rescia ©48, S. Resconi ©71a,
M. Ressegotti \bigcirc <sup>57b,57a</sup>, S. Rettie \bigcirc <sup>36</sup>, J.G. Reyes Rivera \bigcirc <sup>107</sup>, E. Reynolds \bigcirc <sup>17a</sup>, O.L. Rezanova \bigcirc <sup>37</sup>,
P. Reznicek (D133), N. Ribaric (D91), E. Ricci (D78a,78b), R. Richter (D110), S. Richter (D47a,47b),
E. Richter-Was <sup>©86b</sup>, M. Ridel <sup>©127</sup>, S. Ridouani <sup>©35d</sup>, P. Rieck <sup>©117</sup>, P. Riedler <sup>©36</sup>, E.M. Riefel <sup>©47a,47b</sup>,
J.O. Rieger 114, M. Rijssenbeek 145, A. Rimoldi 73a,73b, M. Rimoldi 36, L. Rinaldi 23b,23a,
T.T. Rinn \mathbb{D}^{29}, M.P. Rinnagel \mathbb{D}^{109}, G. Ripellino \mathbb{D}^{161}, I. Riu \mathbb{D}^{13}, P. Rivadeneira \mathbb{D}^{48},
J.C. Rivera Vergara 6, F. Rizatdinova 6, E. Rizvi 6, B.A. Roberts 6, B.R. Roberts 6, B.R. Roberts 6, B.R. Roberts
S.H. Robertson © 104,w, D. Robinson © 32, C.M. Robles Gajardo 137f, M. Robles Manzano © 100,
A. Robson <sup>©59</sup>, A. Rocchi <sup>©76a,76b</sup>, C. Roda <sup>©74a,74b</sup>, S. Rodriguez Bosca <sup>©63a</sup>, Y. Rodriguez Garcia <sup>©22a</sup>,
A. Rodriguez Rodriguez <sup>654</sup>, A.M. Rodríguez Vera <sup>6156b</sup>, S. Roe<sup>36</sup>, J.T. Roemer <sup>6160</sup>,
A.R. Roepe-Gier (136), J. Roggel (171), O. Røhne (125), R.A. Rojas (103), C.P.A. Roland (127),
J. Roloff (D<sup>29</sup>), A. Romaniouk (D<sup>37</sup>), E. Romano (D<sup>73a,73b</sup>), M. Romano (D<sup>23b</sup>), A.C. Romero Hernandez (D<sup>162</sup>),
N. Rompotis (D<sup>92</sup>, L. Roos (D<sup>127</sup>, S. Rosati (D<sup>75a</sup>, B.J. Rosser (D<sup>39</sup>, E. Rossi (D<sup>126</sup>, E. Rossi (D<sup>72a,72b</sup>,
L.P. Rossi © 57b, L. Rossini © 54, R. Rosten © 119, M. Rotaru © 27b, B. Rottler © 54, C. Rougier © 102,aa,
D. Rousseau 666, D. Rousso 532, A. Roy 5162, S. Roy-Garand 5155, A. Rozanov 5102,
Z.M.A. Rozario <sup>59</sup>, Y. Rozen <sup>150</sup>, X. Ruan <sup>33g</sup>, A. Rubio Jimenez <sup>163</sup>, A.J. Ruby <sup>92</sup>,
V.H. Ruelas Rivera 1018, T.A. Ruggeri 101, A. Ruggiero 10126, A. Ruiz-Martinez 10163, A. Rummler 1036,
Z. Rurikova 654, N.A. Rusakovich 638, H.L. Russell 6165, G. Russo 675a,75b, J.P. Rutherfoord 67,
S. Rutherford Colmenares <sup>32</sup>, K. Rybacki<sup>91</sup>, M. Rybar <sup>133</sup>, E.B. Rye <sup>125</sup>, A. Ryzhov <sup>44</sup>,
J.A. Sabater Iglesias <sup>156</sup>, P. Sabatini <sup>163</sup>, L. Sabetta <sup>75a,75b</sup>, H.F-W. Sadrozinski <sup>136</sup>,
F. Safai Tehrani ©<sup>75a</sup>, B. Safarzadeh Samani ©<sup>134</sup>, M. Safdari ©<sup>143</sup>, S. Saha ©<sup>165</sup>, M. Sahinsoy ©<sup>110</sup>,
M. Saimpert (135), M. Saito (153), T. Saito (153), D. Salamani (136), A. Salnikov (143), J. Salt (154),
A. Salvador Salas 151, D. Salvatore 43b,43a, F. Salvatore 146, A. Salzburger 36, D. Sammel 54,
D. Sampsonidis 6152,e, D. Sampsonidou 6123, J. Sánchez 6163, A. Sanchez Pineda 64,
V. Sanchez Sebastian <sup>163</sup>, H. Sandaker <sup>125</sup>, C.O. Sander <sup>48</sup>, J.A. Sandesara <sup>103</sup>, M. Sandhoff <sup>171</sup>,
C. Sandoval (D<sup>22b</sup>, D.P.C. Sankey (D<sup>134</sup>, T. Sano (D<sup>88</sup>, A. Sansoni (D<sup>53</sup>, L. Santi (D<sup>75a,75b</sup>, C. Santoni (D<sup>40</sup>),
H. Santos © 130a,130b, S.N. Santpur © 17a, A. Santra © 169, K.A. Saoucha © 116b, J.G. Saraiva © 130a,130d,
J. Sardain <sup>107</sup>, O. Sasaki <sup>1084</sup>, K. Sato <sup>10157</sup>, C. Sauer <sup>63b</sup>, F. Sauerburger <sup>1054</sup>, E. Sauvan <sup>104</sup>,
P. Savard 155,af, R. Sawada 153, C. Sawyer 134, L. Sawyer 197, I. Sayago Galvan 163, C. Sbarra 153,
A. Sbrizzi (D<sup>23b,23a</sup>, T. Scanlon (D<sup>96</sup>, J. Schaarschmidt (D<sup>138</sup>, P. Schacht (D<sup>110</sup>, U. Schäfer (D<sup>100</sup>),
A.C. Schaffer 66,44, D. Schaile 10, R.D. Schamberger 145, C. Scharf 18, M.M. Schefer 19,
```

```
V.A. Schegelsky \bigcirc^{37}, D. Scheirich \bigcirc^{133}, F. Schenck \bigcirc^{18}, M. Schernau \bigcirc^{160}, C. Scheulen \bigcirc^{55},
C. Schiavi ©57b,57a, E.J. Schioppa ©70a,70b, M. Schioppa ©43b,43a, B. Schlag ©143,n, K.E. Schleicher ©54,
S. Schlenker <sup>1036</sup>, J. Schmeing <sup>171</sup>, M.A. Schmidt <sup>171</sup>, K. Schmieden <sup>100</sup>, C. Schmitt <sup>100</sup>,
N. Schmitt \bigcirc^{100}, S. Schmitt \bigcirc^{48}, L. Schoeffel \bigcirc^{135}, A. Schoening \bigcirc^{63b}, P.G. Scholer \bigcirc^{54}, E. Schopf \bigcirc^{126}, M. Schott \bigcirc^{100}, J. Schovancova \bigcirc^{36}, S. Schramm \bigcirc^{56}, F. Schroeder \bigcirc^{171}, T. Schroer \bigcirc^{56},
H-C. Schultz-Coulon 63a, M. Schumacher 54, B.A. Schumm 136, Ph. Schune 135, A.J. Schuy 138,
H.R. Schwartz <sup>136</sup>, A. Schwartzman <sup>143</sup>, T.A. Schwarz <sup>106</sup>, Ph. Schwemling <sup>135</sup>,
R. Schwienhorst 10<sup>107</sup>, A. Sciandra 10<sup>136</sup>, G. Sciolla 10<sup>26</sup>, F. Scuri 10<sup>74a</sup>, C.D. Sebastiani 10<sup>92</sup>,
K. Sedlaczek 115, P. Seema 18, S.C. Seidel 112, A. Seiden 136, B.D. Seidlitz 141, C. Seitz 148,
J.M. Seixas 683b, G. Sekhniaidze 672a, S.J. Sekula 644, L. Selem 660, N. Semprini-Cesari 623b,23a,
D. Sengupta 656, V. Senthilkumar 6163, L. Serin 666, L. Serkin 69a,69b, M. Sessa 676a,76b,
H. Severini ©<sup>120</sup>, F. Sforza ©<sup>57b,57a</sup>, A. Sfyrla ©<sup>56</sup>, E. Shabalina ©<sup>55</sup>, R. Shaheen ©<sup>144</sup>,
J.D. Shahinian <sup>128</sup>, D. Shaked Renous <sup>169</sup>, L.Y. Shan <sup>14a</sup>, M. Shapiro <sup>17a</sup>, A. Sharma <sup>36</sup>,
A.S. Sharma 164, P. Sharma 180, S. Sharma 184, P.B. Shatalov 187, K. Shaw 184, S.M. Shaw 1811,
A. Shcherbakova <sup>©37</sup>, Q. Shen <sup>©62c,5</sup>, P. Sherwood <sup>©96</sup>, L. Shi <sup>©96</sup>, X. Shi <sup>©14a</sup>, C.O. Shimmin <sup>©172</sup>,
J.D. Shinner <sup>1</sup>
<sup>95</sup>, I.P.J. Shipsey <sup>126</sup>, S. Shirabe <sup>56,h</sup>, M. Shiyakova <sup>38,u</sup>, J. Shlomi <sup>169</sup>,
M.J. Shochet (539, J. Shojaii (5105, D.R. Shope (5125, B. Shrestha (5120, S. Shrestha (5119,aj),
E.M. Shrif (533g), M.J. Shroff (5165), P. Sicho (5131), A.M. Sickles (5162), E. Sideras Haddad (5133g),
A. Sidoti <sup>©23b</sup>, F. Siegert <sup>©50</sup>, Dj. Sijacki <sup>©15</sup>, R. Sikora <sup>©86a</sup>, F. Sili <sup>©90</sup>, J.M. Silva <sup>©20</sup>,
M.V. Silva Oliveira (D<sup>29</sup>, S.B. Silverstein (D<sup>47a</sup>, S. Simion<sup>66</sup>, R. Simoniello (D<sup>36</sup>, E.L. Simpson (D<sup>59</sup>),
H. Simpson (106), L.R. Simpson (106), N.D. Simpson (182), S. Simsek (182), S. Sindhu (185), P. Sinervo (185),
S. Singh 155, S. Sinha 48, S. Sinha 10101, M. Sioli 1023b,23a, I. Siral 1036, E. Sitnikova 1048,
S.Yu. Sivoklokov 0^{37,*}, J. Sjölin 0^{47a,47b}, A. Skaf 0^{55}, E. Skorda 0^{20}, P. Skubic 0^{120}, M. Slawinska 0^{87},
V. Smakhtin<sup>169</sup>, B.H. Smart 134, J. Smiesko 36, S.Yu. Smirnov 37, Y. Smirnov 37,
L.N. Smirnova (D<sup>37</sup>,a, O. Smirnova (D<sup>98</sup>, A.C. Smith (D<sup>41</sup>, E.A. Smith (D<sup>39</sup>, H.A. Smith (D<sup>126</sup>,
J.L. Smith \mathbb{D}^{92}, R. Smith ^{143}, M. Smizanska \mathbb{D}^{91}, K. Smolek \mathbb{D}^{132}, A.A. Snesarev \mathbb{D}^{37}, S.R. Snider \mathbb{D}^{155},
H.L. Snoek 114, S. Snyder 29, R. Sobie 165, A. Soffer 151, C.A. Solans Sanchez 36,
E.Yu. Soldatov \bigcirc^{37}, U. Soldevila \bigcirc^{163}, A.A. Solodkov \bigcirc^{37}, S. Solomon \bigcirc^{26}, A. Soloshenko \bigcirc^{38},
K. Solovieva D<sup>54</sup>, O.V. Solovyanov D<sup>40</sup>, V. Solovyev D<sup>37</sup>, P. Sommer D<sup>36</sup>, A. Sonay D<sup>13</sup>,
W.Y. Song © 156b, J.M. Sonneveld © 114, A. Sopczak © 132, A.L. Sopio © 96, F. Sopkova © 28b,
I.R. Sotarriva Alvarez 154, V. Sothilingam<sup>63a</sup>, O.J. Soto Sandoval 137c,137b, S. Sottocornola 68,
R. Soualah 116b, Z. Soumaimi 35e, D. South 48, N. Soybelman 169, S. Spagnolo 70a,70b,
M. Spalla <sup>110</sup>, D. Sperlich <sup>54</sup>, G. Spigo <sup>36</sup>, S. Spinali <sup>91</sup>, D.P. Spiteri <sup>59</sup>, M. Spousta <sup>133</sup>,
E.J. Staats \bigcirc^{34}, A. Stabile \bigcirc^{71a,71b}, R. Stamen \bigcirc^{63a}, A. Stampekis \bigcirc^{20}, M. Standke \bigcirc^{24}, E. Stanecka \bigcirc^{87},
M.V. Stange <sup>©50</sup>, B. Stanislaus <sup>©17a</sup>, M.M. Stanitzki <sup>©48</sup>, B. Stapf <sup>©48</sup>, E.A. Starchenko <sup>©37</sup>,
G.H. Stark 136, J. Stark 102,aa, D.M. Starko 156b, P. Staroba 131, P. Starovoitov 163a, S. Stärz 104,
R. Staszewski 687, G. Stavropoulos 646, J. Steentoft 6161, P. Steinberg 629, B. Stelzer 6142,156a,
H.J. Stelzer (D<sup>129</sup>, O. Stelzer-Chilton (D<sup>156a</sup>, H. Stenzel (D<sup>58</sup>, T.J. Stevenson (D<sup>146</sup>, G.A. Stewart (D<sup>36</sup>),
J.R. Stewart 121, M.C. Stockton 36, G. Stoicea 27b, M. Stolarski 130a, S. Stonjek 110,
A. Straessner <sup>50</sup>, J. Strandberg <sup>144</sup>, S. Strandberg <sup>47a,47b</sup>, M. Stratmann <sup>171</sup>, M. Strauss <sup>120</sup>,
T. Strebler 102, P. Strizenec 28b, R. Ströhmer 166, D.M. Strom 123, L.R. Strom 48,
R. Stroynowski \mathbb{D}^{44}, A. Strubig \mathbb{D}^{47a,47b}, S.A. Stucci \mathbb{D}^{29}, B. Stugu \mathbb{D}^{16}, J. Stupak \mathbb{D}^{120}, N.A. Styles \mathbb{D}^{48},
D. Su 143, S. Su 162a, W. Su 162d, X. Su 162a, 66, K. Sugizaki 153, V.V. Sulin 1637, M.J. Sullivan 1692,
D.M.S. Sultan (D<sup>78a,78b</sup>, L. Sultanaliyeva (D<sup>37</sup>, S. Sultansoy (D<sup>3b</sup>, T. Sumida (D<sup>88</sup>, S. Sun (D<sup>106</sup>, S. Sun (D<sup>170</sup>),
O. Sunneborn Gudnadottir \mathbb{D}^{161}, N. Sur \mathbb{D}^{102}, M.R. Sutton \mathbb{D}^{146}, H. Suzuki \mathbb{D}^{157}, M. Svatos \mathbb{D}^{131},
M. Swiatlowski <sup>156a</sup>, T. Swirski <sup>166</sup>, I. Sykora <sup>28a</sup>, M. Sykora <sup>133</sup>, T. Sykora <sup>133</sup>, D. Ta <sup>100</sup>,
K. Tackmann \mathbb{D}^{48,t}, A. Taffard \mathbb{D}^{160}, R. Tafirout \mathbb{D}^{156a}, J.S. Tafoya Vargas \mathbb{D}^{66}, E.P. Takeva \mathbb{D}^{52},
```

```
Y. Takubo (1084), M. Talby (1012), A.A. Talyshev (1037), K.C. Tam (1046), N.M. Tamir (1511), A. Tanaka (10153),
J. Tanaka 10153, R. Tanaka 1066, M. Tanasini 1057b,57a, Z. Tao 10164, S. Tapia Araya 10137f,
S. Tapprogge 100, A. Tarek Abouelfadl Mohamed 107, S. Tarem 150, K. Tariq 14a, G. Tarna 102,27b,
G.F. Tartarelli (D<sup>71a</sup>, P. Tas (D<sup>133</sup>, M. Tasevsky (D<sup>131</sup>, E. Tassi (D<sup>43b,43a</sup>, A.C. Tate (D<sup>162</sup>, G. Tateno (D<sup>153</sup>,
Y. Tayalati (D<sup>35e,v</sup>, G.N. Taylor (D<sup>105</sup>, W. Taylor (D<sup>156b</sup>, A.S. Tee (D<sup>170</sup>, R. Teixeira De Lima (D<sup>143</sup>,
P. Teixeira-Dias (1995), J.J. Teoh (1915), K. Terashi (1915), J. Terron (1999), S. Terzo (1913), M. Testa (1953),
R.J. Teuscher (155,w), A. Thaler (179, O. Theiner (156, N. Themistokleous (152, T. Theveneaux-Pelzer (150, N. Themistokleous (152, T. Theveneaux-Pelzer (152, T. Theveneaux-Pelze
O. Thielmann \bigcirc^{171}, D.W. Thomas \bigcirc^{95}, J.P. Thomas \bigcirc^{20}, E.A. Thompson \bigcirc^{17a}, P.D. Thompson \bigcirc^{20}.
E. Thomson 128, Y. Tian 55, V. Tikhomirov 57,a, Yu.A. Tikhonov 57, S. Timoshenko<sup>37</sup>,
D. Timoshyn (D<sup>133</sup>, E.X.L. Ting (D<sup>1</sup>, P. Tipton (D<sup>172</sup>, S.H. Tlou (D<sup>33g</sup>, A. Tnourji (D<sup>40</sup>, K. Todome (D<sup>154</sup>,
S. Todorova-Nova 133, S. Todt<sup>50</sup>, M. Togawa 184, J. Tojo 189, S. Tokár 128a, K. Tokushuku 184,
O. Toldaiev 668, R. Tombs 32, M. Tomoto 84,111, L. Tompkins 143,n, K.W. Topolnicki 866,
E. Torrence 123, H. Torres 1010, aa, E. Torró Pastor 163, M. Toscani 150, C. Tosciri 1539, M. Tost 1511,
D.R. Tovey ^{\text{\tiny D}139}, A. Traeet^{16}, I.S. Trandafir ^{\text{\tiny D}27b}, T. Trefzger ^{\text{\tiny D}166}, A. Tricoli ^{\text{\tiny D}29}, I.M. Trigger ^{\text{\tiny D}156a},
S. Trincaz-Duvoid <sup>127</sup>, D.A. Trischuk <sup>26</sup>, B. Trocmé <sup>60</sup>, C. Troncon <sup>71a</sup>, L. Truong <sup>33c</sup>,
M. Trzebinski 687, A. Trzupek 687, F. Tsai 6145, M. Tsai 6106, A. Tsiamis 6152,e, P.V. Tsiareshka 37,
S. Tsigaridas 156a, A. Tsirigotis 152,r, V. Tsiskaridze 155, E.G. Tskhadadze 149a,
M. Tsopoulou \bigcirc^{152,e}, Y. Tsujikawa \bigcirc^{88}, I.I. Tsukerman \bigcirc^{37}, V. Tsulaia \bigcirc^{17a}, S. Tsuno \bigcirc^{84}, K. Tsuri \bigcirc^{118},
D. Tsybychev 145, Y. Tu 64b, A. Tudorache 27b, V. Tudorache 27b, A.N. Tuna 36,
S. Turchikhin <sup>©57b,57a</sup>, I. Turk Cakir <sup>©3a</sup>, R. Turra <sup>©71a</sup>, T. Turtuvshin <sup>©38,x</sup>, P.M. Tuts <sup>©41</sup>,
S. Tzamarias © 152,e, P. Tzanis © 10, E. Tzovara © 100, F. Ukegawa © 157, P.A. Ulloa Poblete © 137c,137b,
E.N. Umaka 629, G. Unal 636, M. Unal 611, A. Undrus 629, G. Unel 6160, J. Urban 628b,
P. Urquijo (105, P. Urrejola (137a), G. Usai (18, R. Ushioda (154, M. Usman (108, Z. Uysal (121b),
V. Vacek © 132, B. Vachon © 104, K.O.H. Vadla © 125, T. Vafeiadis © 36, A. Vaitkus © 96, C. Valderanis © 109,
E. Valdes Santurio (D47a,47b), M. Valente (D156a), S. Valentinetti (D23b,23a), A. Valero (D163),
E. Valiente Moreno (163), A. Vallier (102,aa), J.A. Valls Ferrer (163), D.R. Van Arneman (114),
T.R. Van Daalen 138, A. Van Der Graaf 49, P. Van Gemmeren 6, M. Van Rijnbach 125,36,
S. Van Stroud 6, I. Van Vulpen 6, M. Vanadia 76a,76b, W. Vandelli 6, M. Vandenbroucke 6, M. Vandenbroucke
E.R. Vandewall 121, D. Vannicola 151, L. Vannoli 57b,57a, R. Vari 575a, E.W. Varnes 77,
C. Varni 1017b, T. Varol 10148, D. Varouchas 1066, L. Varriale 10163, K.E. Varvell 10147, M.E. Vasile 1027b,
L. Vaslin<sup>84</sup>, G.A. Vasquez 165, A. Vasyukov 1838, F. Vazeille 1840, T. Vazquez Schroeder 1856,
J. Veatch (1031), V. Vecchio (1011), M.J. Veen (1010), I. Veliscek (10126), L.M. Veloce (10155), F. Veloso (10130a,130c),
S. Veneziano ^{\bullet 75a}, A. Ventura ^{\bullet 70a,70b}, S. Ventura Gonzalez ^{\bullet 135}, A. Verbytskyi ^{\bullet 110},
M. Verducci (D<sup>74a,74b</sup>, C. Vergis (D<sup>24</sup>, M. Verissimo De Araujo (D<sup>83b</sup>, W. Verkerke (D<sup>114</sup>,
J.C. Vermeulen 114, C. Vernieri 143, M. Vessella 1013, M.C. Vetterli 142, af, A. Vgenopoulos 152, e,
N. Viaux Maira 137f, T. Vickey 139, O.E. Vickey Boeriu 139, G.H.A. Viehhauser 126, L. Vigani 163b,
M. Villa (D<sup>23b,23a</sup>, M. Villaplana Perez (D<sup>163</sup>, E.M. Villhauer<sup>52</sup>, E. Vilucchi (D<sup>53</sup>, M.G. Vincter (D<sup>34</sup>,
G.S. Virdee \mathbb{D}^{20}, A. Vishwakarma \mathbb{D}^{52}, A. Visibile \mathbb{D}^{14}, C. Vittori \mathbb{D}^{36}, I. Vivarelli \mathbb{D}^{146},
E. Voevodina 110, F. Vogel 1019, J.C. Voigt 150, P. Vokac 132, Yu. Volkotrub 186a, J. Von Ahnen 1848,
E. Von Toerne • B. Vormwald • V. Vorobel • V. Vorobev • M. Vorobev • M. Vos • M. Vos • M. Vos • M. Voss • M. Voss • M. Voss • M. Vossebeld • M. Vozak • M.
M. Vreeswijk 114, R. Vuillermet 36, O. Vujinovic 100, I. Vukotic 39, S. Wada 157, C. Wagner 3,
J.M. Wagner 171, W. Wagner 171, S. Wahdan 171, H. Wahlberg 190, M. Wakida 111, J. Walder 1134,
R. Walker (10) 109, W. Walkowiak (11) 141, A. Wall (10) 128, T. Wamorkar (10) 6, A.Z. Wang (10) 136, C. Wang (10) 100,
C. Wang 62c, H. Wang 17a, J. Wang 64a, R.-J. Wang 100, R. Wang 61, R. Wang 66,
S.M. Wang 148, S. Wang 62b, T. Wang 62a, W.T. Wang 80, W. Wang 14a, X. Wang 14c,
X. Wang 6162, X. Wang 662c, Y. Wang 662d, Y. Wang 614c, Z. Wang 6106, Z. Wang 662d,51,62c,
```

```
Z. Wang 6006, A. Warburton 60104, R.J. Ward 6020, N. Warrack 6059, A.T. Watson 6020, H. Watson 6059,
M.F. Watson (D<sup>20</sup>), E. Watton (D<sup>59</sup>,134), G. Watts (D<sup>138</sup>), B.M. Waugh (D<sup>96</sup>), C. Weber (D<sup>29</sup>), H.A. Weber (D<sup>18</sup>),
M.S. Weber (19, S.M. Weber (1063a), C. Wei (1062a), Y. Wei (10126), A.R. Weidberg (10126), E.J. Weik (10117),
J. Weingarten <sup>1049</sup>, M. Weirich <sup>1000</sup>, C. Weiser <sup>1054</sup>, C.J. Wells <sup>1048</sup>, T. Wenaus <sup>1029</sup>, B. Wendland <sup>1049</sup>,
T. Wengler <sup>10</sup>36, N.S. Wenke <sup>110</sup>, N. Wermes <sup>24</sup>, M. Wessels <sup>63</sup>3a, A.M. Wharton <sup>91</sup>, A.S. White <sup>61</sup>,
A. White 6, M.J. White 1, D. Whiteson 6, L. Wickremasinghe 6, W. Wiedenmann 6, W. Wiedenmann 6, W. Wiedenmann 1, W. Wiedenman
C. Wiel 50, M. Wielers 134, C. Wiglesworth 42, D.J. Wilbern 120, H.G. Wilkens 36,
D.M. Williams (D<sup>41</sup>, H.H. Williams (128, S. Williams (D<sup>32</sup>, S. Willocq (D<sup>103</sup>, B.J. Wilson (D<sup>101</sup>),
P.J. Windischhofer <sup>©39</sup>, F.I. Winkel <sup>©30</sup>, F. Winklmeier <sup>©123</sup>, B.T. Winter <sup>©54</sup>, J.K. Winter <sup>©101</sup>,
M. Wittgen<sup>143</sup>, M. Wobisch <sup>197</sup>, Z. Wolffs <sup>114</sup>, J. Wollrath <sup>160</sup>, M.W. Wolter <sup>187</sup>, H. Wolters <sup>130a,130c</sup>,
A.F. Wongel <sup>648</sup>, E.L. Woodward <sup>641</sup>, S.D. Worm <sup>648</sup>, B.K. Wosiek <sup>687</sup>, K.W. Woźniak <sup>687</sup>,
S. Wozniewski <sup>55</sup>, K. Wraight <sup>59</sup>, C. Wu <sup>520</sup>, J. Wu <sup>514a,14e</sup>, M. Wu <sup>54a</sup>, M. Wu <sup>5113</sup>, S.L. Wu <sup>5170</sup>,
X. Wu 1056, Y. Wu 1062a, Z. Wu 10135, J. Wuerzinger 10110,ad, T.R. Wyatt 10101, B.M. Wynne 1052,
S. Xella (D42, L. Xia (D14c, M. Xia (D14b), J. Xiang (D64c, M. Xie (D62a, X. Xie (D62a, S. Xin (D14a, 14e),
A. Xiong 123, J. Xiong 17a, D. Xu 14a, H. Xu 62a, L. Xu 62a, R. Xu 128, T. Xu 16b, Y. Xu 16b,
Z. Xu D<sup>52</sup>, Z. Xu<sup>14c</sup>, B. Yabsley D<sup>147</sup>, S. Yacoob D<sup>33a</sup>, Y. Yamaguchi D<sup>154</sup>, E. Yamashita D<sup>153</sup>,
H. Yamauchi 157, T. Yamazaki 17a, Y. Yamazaki 185, J. Yan 12c, S. Yan 12c, Z. Yan 1525,
H.J. Yang 62c,62d, H.T. Yang 62a, S. Yang 62a, T. Yang 64c, X. Yang 636, X. Yang 614a, Y. Yang 644,
Y. Yang<sup>62a</sup>, Z. Yang <sup>62a</sup>, W-M. Yao <sup>17a</sup>, Y.C. Yap <sup>48</sup>, H. Ye <sup>14c</sup>, H. Ye <sup>55</sup>, J. Ye <sup>14a</sup>, S. Ye <sup>29</sup>,
X. Ye 6<sup>22</sup>a, Y. Yeh 6<sup>96</sup>, I. Yeletskikh 6<sup>38</sup>, B.K. Yeo 6<sup>17</sup>b, M.R. Yexley 6<sup>96</sup>, P. Yin 6<sup>41</sup>, K. Yorita 6<sup>168</sup>,
S. Younas (D<sup>27b</sup>, C.J.S. Young (D<sup>36</sup>, C. Young (D<sup>143</sup>, C. Yu (D<sup>14a,14e,ah</sup>, Y. Yu (D<sup>62a</sup>, M. Yuan (D<sup>106</sup>),
R. Yuan 6<sup>62b</sup>, L. Yue 6<sup>96</sup>, M. Zaazoua 6<sup>62a</sup>, B. Zabinski 6<sup>87</sup>, E. Zaid<sup>52</sup>, Z.K. Zak 6<sup>87</sup>,
T. Zakareishvili 149b, N. Zakharchuk 34, S. Zambito 55, J.A. Zamora Saa 137d,137b, J. Zang 153,
D. Zanzi 654, O. Zaplatilek 6132, C. Zeitnitz 6171, H. Zeng 614a, J.C. Zeng 6162, D.T. Zenger Jr 626,
O. Zenin 637, T. Ženiš 628a, S. Zenz 694, S. Zerradi 635a, D. Zerwas 666, M. Zhai 614a,14e,
B. Zhang 614c, D.F. Zhang 6139, J. Zhang 62b, J. Zhang 66, K. Zhang 614a,14e, L. Zhang 614c,
P. Zhang (D<sup>14a,14e</sup>, R. Zhang (D<sup>170</sup>, S. Zhang (D<sup>106</sup>, S. Zhang (D<sup>44</sup>, T. Zhang (D<sup>153</sup>, X. Zhang (D<sup>62c</sup>),
X. Zhang (10<sup>62b</sup>, Y. Zhang (10<sup>62c,5</sup>, Y. Zhang (10<sup>96</sup>, Y. Zhang (10<sup>14c</sup>, Z. Zhang (10<sup>17a</sup>, Z. Zhang (10<sup>66</sup>,
H. Zhao 138, P. Zhao 51, T. Zhao 62b, Y. Zhao 136, Z. Zhao 62a, A. Zhemchugov 38,
J. Zheng 14c, K. Zheng 162, X. Zheng 162a, Z. Zheng 143, D. Zhong 162, B. Zhou 166, H. Zhou 7,
N. Zhou 62c, Y. Zhou, C.G. Zhu 62b, J. Zhu 6106, Y. Zhu 66c, Y. Zhu 66a, X. Zhuang 614a,
K. Zhukov \bigcirc^{37}, V. Zhulanov \bigcirc^{37}, N.I. Zimine \bigcirc^{38}, J. Zinsser \bigcirc^{63b}, M. Ziolkowski \bigcirc^{141}, L. Živković \bigcirc^{15},
A. Zoccoli (23b,23a), K. Zoch (51, T.G. Zorbas (5139), O. Zormpa (546), W. Zou (541), L. Zwalinski (536).
```

<sup>&</sup>lt;sup>1</sup>Department of Physics, University of Adelaide, Adelaide; Australia.

<sup>&</sup>lt;sup>2</sup>Department of Physics, University of Alberta, Edmonton AB; Canada.

<sup>&</sup>lt;sup>3(a)</sup>Department of Physics, Ankara University, Ankara; <sup>(b)</sup>Division of Physics, TOBB University of Economics and Technology, Ankara; Türkiye.

<sup>&</sup>lt;sup>4</sup>LAPP, Université Savoie Mont Blanc, CNRS/IN2P3, Annecy; France.

<sup>&</sup>lt;sup>5</sup>APC, Université Paris Cité, CNRS/IN2P3, Paris; France.

<sup>&</sup>lt;sup>6</sup>High Energy Physics Division, Argonne National Laboratory, Argonne IL; United States of America.

<sup>&</sup>lt;sup>7</sup>Department of Physics, University of Arizona, Tucson AZ; United States of America.

<sup>&</sup>lt;sup>8</sup>Department of Physics, University of Texas at Arlington, Arlington TX; United States of America.

<sup>&</sup>lt;sup>9</sup>Physics Department, National and Kapodistrian University of Athens, Athens; Greece.

<sup>&</sup>lt;sup>10</sup>Physics Department, National Technical University of Athens, Zografou; Greece.

<sup>&</sup>lt;sup>11</sup>Department of Physics, University of Texas at Austin, Austin TX; United States of America.

<sup>&</sup>lt;sup>12</sup>Institute of Physics, Azerbaijan Academy of Sciences, Baku; Azerbaijan.

- <sup>13</sup>Institut de Física d'Altes Energies (IFAE), Barcelona Institute of Science and Technology, Barcelona; Spain.
- <sup>14(a)</sup>Institute of High Energy Physics, Chinese Academy of Sciences, Beijing; <sup>(b)</sup>Physics Department, Tsinghua University, Beijing; <sup>(c)</sup>Department of Physics, Nanjing University, Nanjing; <sup>(d)</sup>School of Science, Shenzhen Campus of Sun Yat-sen University; <sup>(e)</sup>University of Chinese Academy of Science (UCAS), Beijing; China.
- <sup>15</sup>Institute of Physics, University of Belgrade, Belgrade; Serbia.
- <sup>16</sup>Department for Physics and Technology, University of Bergen, Bergen; Norway.
- $^{17(a)}$ Physics Division, Lawrence Berkeley National Laboratory, Berkeley CA;  $^{(b)}$ University of California, Berkeley CA; United States of America.
- <sup>18</sup>Institut für Physik, Humboldt Universität zu Berlin, Berlin; Germany.
- <sup>19</sup>Albert Einstein Center for Fundamental Physics and Laboratory for High Energy Physics, University of Bern, Bern; Switzerland.
- <sup>20</sup>School of Physics and Astronomy, University of Birmingham, Birmingham; United Kingdom.
- <sup>21(a)</sup>Department of Physics, Bogazici University, Istanbul; <sup>(b)</sup>Department of Physics Engineering,
- Gaziantep University, Gaziantep; (c) Department of Physics, Istanbul University, Istanbul; Türkiye.
- <sup>22(a)</sup>Facultad de Ciencias y Centro de Investigaciónes, Universidad Antonio Nariño,
- Bogotá;(b) Departamento de Física, Universidad Nacional de Colombia, Bogotá; Colombia.
- <sup>23(a)</sup>Dipartimento di Fisica e Astronomia A. Righi, Università di Bologna, Bologna; <sup>(b)</sup>INFN Sezione di Bologna; Italy.
- <sup>24</sup>Physikalisches Institut, Universität Bonn, Bonn; Germany.
- <sup>25</sup>Department of Physics, Boston University, Boston MA; United States of America.
- <sup>26</sup>Department of Physics, Brandeis University, Waltham MA; United States of America.
- <sup>27(a)</sup>Transilvania University of Brasov, Brasov; <sup>(b)</sup>Horia Hulubei National Institute of Physics and Nuclear Engineering, Bucharest; <sup>(c)</sup>Department of Physics, Alexandru Ioan Cuza University of Iasi, Iasi; <sup>(d)</sup>National Institute for Research and Development of Isotopic and Molecular Technologies, Physics Department, Cluj-Napoca; <sup>(e)</sup>National University of Science and Technology Politechnica, Bucharest; <sup>(f)</sup>West University in Timisoara, Timisoara; <sup>(g)</sup>Faculty of Physics, University of Bucharest, Bucharest; Romania. <sup>28(a)</sup>Faculty of Mathematics, Physics and Informatics, Comenius University, Bratislava; <sup>(b)</sup>Department of Subnuclear Physics, Institute of Experimental Physics of the Slovak Academy of Sciences, Kosice; Slovak Republic.
- <sup>29</sup>Physics Department, Brookhaven National Laboratory, Upton NY; United States of America.
- <sup>30</sup>Universidad de Buenos Aires, Facultad de Ciencias Exactas y Naturales, Departamento de Física, y CONICET, Instituto de Física de Buenos Aires (IFIBA), Buenos Aires; Argentina.
- <sup>31</sup>California State University, CA; United States of America.
- <sup>32</sup>Cavendish Laboratory, University of Cambridge, Cambridge; United Kingdom.
- <sup>33(a)</sup>Department of Physics, University of Cape Town, Cape Town; <sup>(b)</sup> iThemba Labs, Western
- Cape; (c) Department of Mechanical Engineering Science, University of Johannesburg,
- Johannesburg; (d) National Institute of Physics, University of the Philippines Diliman
- (Philippines); (e) University of South Africa, Department of Physics, Pretoria; (f) University of Zululand,
- KwaDlangezwa; (g) School of Physics, University of the Witwatersrand, Johannesburg; South Africa.
- <sup>34</sup>Department of Physics, Carleton University, Ottawa ON; Canada.
- <sup>35(a)</sup> Faculté des Sciences Ain Chock, Réseau Universitaire de Physique des Hautes Energies Université Hassan II, Casablanca; <sup>(b)</sup> Faculté des Sciences, Université Ibn-Tofail, Kénitra; <sup>(c)</sup> Faculté des Sciences Semlalia, Université Cadi Ayyad, LPHEA-Marrakech; <sup>(d)</sup> LPMR, Faculté des Sciences, Université Mohammed Premier, Oujda; <sup>(e)</sup> Faculté des sciences, Université Mohammed V, Rabat; <sup>(f)</sup> Institute of Applied Physics, Mohammed VI Polytechnic University, Ben Guerir; Morocco.

- <sup>36</sup>CERN, Geneva; Switzerland.
- <sup>37</sup>Affiliated with an institute covered by a cooperation agreement with CERN.
- <sup>38</sup> Affiliated with an international laboratory covered by a cooperation agreement with CERN.
- <sup>39</sup>Enrico Fermi Institute, University of Chicago, Chicago IL; United States of America.
- <sup>40</sup>LPC, Université Clermont Auvergne, CNRS/IN2P3, Clermont-Ferrand; France.
- <sup>41</sup>Nevis Laboratory, Columbia University, Irvington NY; United States of America.
- <sup>42</sup>Niels Bohr Institute, University of Copenhagen, Copenhagen; Denmark.
- <sup>43(a)</sup>Dipartimento di Fisica, Università della Calabria, Rende; <sup>(b)</sup>INFN Gruppo Collegato di Cosenza, Laboratori Nazionali di Frascati; Italy.
- <sup>44</sup>Physics Department, Southern Methodist University, Dallas TX; United States of America.
- <sup>45</sup>Physics Department, University of Texas at Dallas, Richardson TX; United States of America.
- <sup>46</sup>National Centre for Scientific Research "Demokritos", Agia Paraskevi; Greece.
- <sup>47(a)</sup>Department of Physics, Stockholm University; <sup>(b)</sup>Oskar Klein Centre, Stockholm; Sweden.
- <sup>48</sup>Deutsches Elektronen-Synchrotron DESY, Hamburg and Zeuthen; Germany.
- <sup>49</sup>Fakultät Physik, Technische Universität Dortmund, Dortmund; Germany.
- <sup>50</sup>Institut für Kern- und Teilchenphysik, Technische Universität Dresden, Dresden; Germany.
- <sup>51</sup>Department of Physics, Duke University, Durham NC; United States of America.
- <sup>52</sup>SUPA School of Physics and Astronomy, University of Edinburgh, Edinburgh; United Kingdom.
- <sup>53</sup>INFN e Laboratori Nazionali di Frascati, Frascati; Italy.
- <sup>54</sup>Physikalisches Institut, Albert-Ludwigs-Universität Freiburg, Freiburg; Germany.
- <sup>55</sup>II. Physikalisches Institut, Georg-August-Universität Göttingen, Göttingen; Germany.
- <sup>56</sup>Département de Physique Nucléaire et Corpusculaire, Université de Genève, Genève; Switzerland.
- <sup>57(a)</sup>Dipartimento di Fisica, Università di Genova, Genova; (b) INFN Sezione di Genova; Italy.
- <sup>58</sup>II. Physikalisches Institut, Justus-Liebig-Universität Giessen, Giessen; Germany.
- <sup>59</sup>SUPA School of Physics and Astronomy, University of Glasgow, Glasgow; United Kingdom.
- <sup>60</sup>LPSC, Université Grenoble Alpes, CNRS/IN2P3, Grenoble INP, Grenoble; France.
- <sup>61</sup>Laboratory for Particle Physics and Cosmology, Harvard University, Cambridge MA; United States of America.
- <sup>62</sup>(a) Department of Modern Physics and State Key Laboratory of Particle Detection and Electronics, University of Science and Technology of China, Hefei; (b) Institute of Frontier and Interdisciplinary Science and Key Laboratory of Particle Physics and Particle Irradiation (MOE), Shandong University, Qingdao; (c) School of Physics and Astronomy, Shanghai Jiao Tong University, Key Laboratory for Particle Astrophysics and Cosmology (MOE), SKLPPC, Shanghai; (d) Tsung-Dao Lee Institute, Shanghai; China. (d) Kirchhoff-Institut für Physik, Ruprecht-Karls-Universität Heidelberg, Heidelberg; (b) Physikalisches Institut, Ruprecht-Karls-Universität Heidelberg, Heidelberg; Germany.
- $^{64(a)}$ Department of Physics, Chinese University of Hong Kong, Shatin, N.T., Hong Kong;  $^{(b)}$ Department of Physics, University of Hong Kong, Hong Kong;  $^{(c)}$ Department of Physics and Institute for Advanced Study, Hong Kong University of Science and Technology, Clear Water Bay, Kowloon, Hong Kong; China.
- <sup>65</sup>Department of Physics, National Tsing Hua University, Hsinchu; Taiwan.
- <sup>66</sup>IJCLab, Université Paris-Saclay, CNRS/IN2P3, 91405, Orsay; France.
- <sup>67</sup>Centro Nacional de Microelectrónica (IMB-CNM-CSIC), Barcelona; Spain.
- <sup>68</sup>Department of Physics, Indiana University, Bloomington IN; United States of America.
- $^{69(a)}$ INFN Gruppo Collegato di Udine, Sezione di Trieste, Udine; $^{(b)}$ ICTP, Trieste; $^{(c)}$ Dipartimento Politecnico di Ingegneria e Architettura, Università di Udine, Udine; Italy.
- <sup>70(a)</sup>INFN Sezione di Lecce; <sup>(b)</sup>Dipartimento di Matematica e Fisica, Università del Salento, Lecce; Italy.
- <sup>71</sup>(a) INFN Sezione di Milano; <sup>(b)</sup> Dipartimento di Fisica, Università di Milano, Milano; Italy.
- 72(a)INFN Sezione di Napoli; (b)Dipartimento di Fisica, Università di Napoli, Napoli; Italy.

- $^{73(a)}$ INFN Sezione di Pavia; $^{(b)}$ Dipartimento di Fisica, Università di Pavia, Pavia; Italy.
- $^{74(a)}$ INFN Sezione di Pisa;  $^{(b)}$ Dipartimento di Fisica E. Fermi, Università di Pisa, Pisa; Italy.
- <sup>75(a)</sup>INFN Sezione di Roma; <sup>(b)</sup>Dipartimento di Fisica, Sapienza Università di Roma, Roma; Italy.
- $^{76(a)}$ INFN Sezione di Roma Tor Vergata;  $^{(b)}$ Dipartimento di Fisica, Università di Roma Tor Vergata, Roma; Italy.
- $^{77(a)}$ INFN Sezione di Roma Tre; $^{(b)}$ Dipartimento di Matematica e Fisica, Università Roma Tre, Roma; Italy.
- <sup>78</sup>(a) INFN-TIFPA; (b) Università degli Studi di Trento, Trento; Italy.
- <sup>79</sup>Universität Innsbruck, Department of Astro and Particle Physics, Innsbruck; Austria.
- <sup>80</sup>University of Iowa, Iowa City IA; United States of America.
- <sup>81</sup>Department of Physics and Astronomy, Iowa State University, Ames IA; United States of America.
- <sup>82</sup>Istinye University, Sariyer, Istanbul; Türkiye.
- <sup>83(a)</sup>Departamento de Engenharia Elétrica, Universidade Federal de Juiz de Fora (UFJF), Juiz de Fora; <sup>(b)</sup>Universidade Federal do Rio De Janeiro COPPE/EE/IF, Rio de Janeiro; <sup>(c)</sup>Instituto de Física, Universidade de São Paulo, São Paulo; <sup>(d)</sup>Rio de Janeiro State University, Rio de Janeiro; Brazil.
- <sup>84</sup>KEK, High Energy Accelerator Research Organization, Tsukuba; Japan.
- <sup>85</sup>Graduate School of Science, Kobe University, Kobe; Japan.
- $^{86(a)}$ AGH University of Krakow, Faculty of Physics and Applied Computer Science, Krakow;  $^{(b)}$ Marian Smoluchowski Institute of Physics, Jagiellonian University, Krakow; Poland.
- <sup>87</sup>Institute of Nuclear Physics Polish Academy of Sciences, Krakow; Poland.
- <sup>88</sup>Faculty of Science, Kyoto University, Kyoto; Japan.
- <sup>89</sup>Research Center for Advanced Particle Physics and Department of Physics, Kyushu University, Fukuoka ; Japan.
- <sup>90</sup>Instituto de Física La Plata, Universidad Nacional de La Plata and CONICET, La Plata; Argentina.
- <sup>91</sup>Physics Department, Lancaster University, Lancaster; United Kingdom.
- <sup>92</sup>Oliver Lodge Laboratory, University of Liverpool, Liverpool; United Kingdom.
- <sup>93</sup>Department of Experimental Particle Physics, Jožef Stefan Institute and Department of Physics, University of Ljubljana, Ljubljana; Slovenia.
- <sup>94</sup>School of Physics and Astronomy, Queen Mary University of London, London; United Kingdom.
- <sup>95</sup>Department of Physics, Royal Holloway University of London, Egham; United Kingdom.
- <sup>96</sup>Department of Physics and Astronomy, University College London, London; United Kingdom.
- <sup>97</sup>Louisiana Tech University, Ruston LA; United States of America.
- <sup>98</sup>Fysiska institutionen, Lunds universitet, Lund; Sweden.
- <sup>99</sup>Departamento de Física Teorica C-15 and CIAFF, Universidad Autónoma de Madrid, Madrid; Spain.
- <sup>100</sup>Institut für Physik, Universität Mainz, Mainz; Germany.
- <sup>101</sup>School of Physics and Astronomy, University of Manchester, Manchester; United Kingdom.
- <sup>102</sup>CPPM, Aix-Marseille Université, CNRS/IN2P3, Marseille; France.
- <sup>103</sup>Department of Physics, University of Massachusetts, Amherst MA; United States of America.
- <sup>104</sup>Department of Physics, McGill University, Montreal QC; Canada.
- <sup>105</sup>School of Physics, University of Melbourne, Victoria; Australia.
- <sup>106</sup>Department of Physics, University of Michigan, Ann Arbor MI; United States of America.
- <sup>107</sup>Department of Physics and Astronomy, Michigan State University, East Lansing MI; United States of America.
- <sup>108</sup>Group of Particle Physics, University of Montreal, Montreal QC; Canada.
- <sup>109</sup>Fakultät für Physik, Ludwig-Maximilians-Universität München, München; Germany.
- <sup>110</sup>Max-Planck-Institut für Physik (Werner-Heisenberg-Institut), München; Germany.
- <sup>111</sup>Graduate School of Science and Kobayashi-Maskawa Institute, Nagoya University, Nagoya; Japan.

- <sup>112</sup>Department of Physics and Astronomy, University of New Mexico, Albuquerque NM; United States of America.
- <sup>113</sup>Institute for Mathematics, Astrophysics and Particle Physics, Radboud University/Nikhef, Nijmegen; Netherlands.
- <sup>114</sup>Nikhef National Institute for Subatomic Physics and University of Amsterdam, Amsterdam; Netherlands.
- $^{115}$ Department of Physics, Northern Illinois University, DeKalb IL; United States of America.  $^{116(a)}$ New York University Abu Dhabi, Abu Dhabi;  $^{(b)}$ University of Sharjah, Sharjah; United Arab Emirates.
- <sup>117</sup>Department of Physics, New York University, New York NY; United States of America.
- <sup>118</sup>Ochanomizu University, Otsuka, Bunkyo-ku, Tokyo; Japan.
- <sup>119</sup>Ohio State University, Columbus OH; United States of America.
- <sup>120</sup>Homer L. Dodge Department of Physics and Astronomy, University of Oklahoma, Norman OK; United States of America.
- <sup>121</sup>Department of Physics, Oklahoma State University, Stillwater OK; United States of America.
- <sup>122</sup>Palacký University, Joint Laboratory of Optics, Olomouc; Czech Republic.
- <sup>123</sup>Institute for Fundamental Science, University of Oregon, Eugene, OR; United States of America.
- <sup>124</sup>Graduate School of Science, Osaka University, Osaka; Japan.
- <sup>125</sup>Department of Physics, University of Oslo, Oslo; Norway.
- <sup>126</sup>Department of Physics, Oxford University, Oxford; United Kingdom.
- <sup>127</sup>LPNHE, Sorbonne Université, Université Paris Cité, CNRS/IN2P3, Paris; France.
- <sup>128</sup>Department of Physics, University of Pennsylvania, Philadelphia PA; United States of America.
- <sup>129</sup>Department of Physics and Astronomy, University of Pittsburgh, Pittsburgh PA; United States of America.
- <sup>130</sup>(a) Laboratório de Instrumentação e Física Experimental de Partículas LIP, Lisboa; (b) Departamento de Física, Faculdade de Ciências, Universidade de Lisboa, Lisboa; (c) Departamento de Física, Universidade de Coimbra, Coimbra; (d) Centro de Física Nuclear da Universidade de Lisboa, Lisboa; (e) Departamento de Física, Universidade do Minho, Braga; (f) Departamento de Física Teórica y del Cosmos, Universidad de Granada, Granada (Spain); (g) Departamento de Física, Instituto Superior Técnico, Universidade de Lisboa, Lisboa; Portugal.
- <sup>131</sup>Institute of Physics of the Czech Academy of Sciences, Prague; Czech Republic.
- <sup>132</sup>Czech Technical University in Prague, Prague; Czech Republic.
- <sup>133</sup>Charles University, Faculty of Mathematics and Physics, Prague; Czech Republic.
- <sup>134</sup>Particle Physics Department, Rutherford Appleton Laboratory, Didcot; United Kingdom.
- <sup>135</sup>IRFU, CEA, Université Paris-Saclay, Gif-sur-Yvette; France.
- <sup>136</sup>Santa Cruz Institute for Particle Physics, University of California Santa Cruz, Santa Cruz CA; United States of America.
- $^{137(a)}$ Departamento de Física, Pontificia Universidad Católica de Chile, Santiago;  $^{(b)}$ Millennium Institute for Subatomic physics at high energy frontier (SAPHIR), Santiago;  $^{(c)}$ Instituto de Investigación Multidisciplinario en Ciencia y Tecnología, y Departamento de Física, Universidad de La Serena;  $^{(d)}$ Universidad Andres Bello, Department of Physics, Santiago;  $^{(e)}$ Instituto de Alta Investigación, Universidad de Tarapacá, Arica;  $^{(f)}$ Departamento de Física, Universidad Técnica Federico Santa María, Valparaíso; Chile.
- <sup>138</sup>Department of Physics, University of Washington, Seattle WA; United States of America.
- <sup>139</sup>Department of Physics and Astronomy, University of Sheffield, Sheffield; United Kingdom.
- <sup>140</sup>Department of Physics, Shinshu University, Nagano; Japan.
- <sup>141</sup>Department Physik, Universität Siegen, Siegen; Germany.

- <sup>142</sup>Department of Physics, Simon Fraser University, Burnaby BC; Canada.
- <sup>143</sup>SLAC National Accelerator Laboratory, Stanford CA; United States of America.
- <sup>144</sup>Department of Physics, Royal Institute of Technology, Stockholm; Sweden.
- <sup>145</sup>Departments of Physics and Astronomy, Stony Brook University, Stony Brook NY; United States of America.
- <sup>146</sup>Department of Physics and Astronomy, University of Sussex, Brighton; United Kingdom.
- <sup>147</sup>School of Physics, University of Sydney, Sydney; Australia.
- <sup>148</sup>Institute of Physics, Academia Sinica, Taipei; Taiwan.
- <sup>149</sup>(*a*)E. Andronikashvili Institute of Physics, Iv. Javakhishvili Tbilisi State University, Tbilisi; (*b*) High Energy Physics Institute, Tbilisi State University, Tbilisi; (*c*) University of Georgia, Tbilisi; Georgia.
- <sup>150</sup>Department of Physics, Technion, Israel Institute of Technology, Haifa; Israel.
- <sup>151</sup>Raymond and Beverly Sackler School of Physics and Astronomy, Tel Aviv University, Tel Aviv; Israel.
- <sup>152</sup>Department of Physics, Aristotle University of Thessaloniki, Thessaloniki; Greece.
- <sup>153</sup>International Center for Elementary Particle Physics and Department of Physics, University of Tokyo, Tokyo; Japan.
- <sup>154</sup>Department of Physics, Tokyo Institute of Technology, Tokyo; Japan.
- <sup>155</sup>Department of Physics, University of Toronto, Toronto ON; Canada.
- <sup>156</sup>(a) TRIUMF, Vancouver BC; (b) Department of Physics and Astronomy, York University, Toronto ON; Canada.
- <sup>157</sup>Division of Physics and Tomonaga Center for the History of the Universe, Faculty of Pure and Applied Sciences, University of Tsukuba, Tsukuba; Japan.
- <sup>158</sup>Department of Physics and Astronomy, Tufts University, Medford MA; United States of America.
- <sup>159</sup>United Arab Emirates University, Al Ain; United Arab Emirates.
- <sup>160</sup>Department of Physics and Astronomy, University of California Irvine, Irvine CA; United States of America.
- <sup>161</sup>Department of Physics and Astronomy, University of Uppsala, Uppsala; Sweden.
- <sup>162</sup>Department of Physics, University of Illinois, Urbana IL; United States of America.
- <sup>163</sup>Instituto de Física Corpuscular (IFIC), Centro Mixto Universidad de Valencia CSIC, Valencia; Spain.
- <sup>164</sup>Department of Physics, University of British Columbia, Vancouver BC; Canada.
- <sup>165</sup>Department of Physics and Astronomy, University of Victoria, Victoria BC; Canada.
- <sup>166</sup>Fakultät für Physik und Astronomie, Julius-Maximilians-Universität Würzburg, Würzburg; Germany.
- <sup>167</sup>Department of Physics, University of Warwick, Coventry; United Kingdom.
- <sup>168</sup>Waseda University, Tokyo; Japan.
- <sup>169</sup>Department of Particle Physics and Astrophysics, Weizmann Institute of Science, Rehovot; Israel.
- <sup>170</sup>Department of Physics, University of Wisconsin, Madison WI; United States of America.
- <sup>171</sup>Fakultät für Mathematik und Naturwissenschaften, Fachgruppe Physik, Bergische Universität Wuppertal, Wuppertal; Germany.
- <sup>172</sup>Department of Physics, Yale University, New Haven CT; United States of America.
- <sup>a</sup> Also Affiliated with an institute covered by a cooperation agreement with CERN.
- <sup>b</sup> Also at An-Najah National University, Nablus; Palestine.
- <sup>c</sup> Also at Borough of Manhattan Community College, City University of New York, New York NY; United States of America.
- <sup>d</sup> Also at Center for High Energy Physics, Peking University; China.
- <sup>e</sup> Also at Center for Interdisciplinary Research and Innovation (CIRI-AUTH), Thessaloniki; Greece.
- <sup>f</sup> Also at Centro Studi e Ricerche Enrico Fermi; Italy.
- <sup>g</sup> Also at CERN, Geneva; Switzerland.
- <sup>h</sup> Also at Département de Physique Nucléaire et Corpusculaire, Université de Genève, Genève;

#### Switzerland.

- <sup>i</sup> Also at Departament de Fisica de la Universitat Autonoma de Barcelona, Barcelona; Spain.
- <sup>j</sup> Also at Department of Financial and Management Engineering, University of the Aegean, Chios; Greece.
- <sup>k</sup> Also at Department of Physics, Ben Gurion University of the Negev, Beer Sheva; Israel.
- <sup>1</sup> Also at Department of Physics, California State University, Sacramento; United States of America.
- <sup>m</sup> Also at Department of Physics, King's College London, London; United Kingdom.
- <sup>n</sup> Also at Department of Physics, Stanford University, Stanford CA; United States of America.
- <sup>o</sup> Also at Department of Physics, University of Fribourg, Fribourg; Switzerland.
- <sup>p</sup> Also at Department of Physics, University of Thessaly; Greece.
- <sup>q</sup> Also at Department of Physics, Westmont College, Santa Barbara; United States of America.
- <sup>r</sup> Also at Hellenic Open University, Patras; Greece.
- <sup>s</sup> Also at Institucio Catalana de Recerca i Estudis Avancats, ICREA, Barcelona; Spain.
- <sup>t</sup> Also at Institut für Experimentalphysik, Universität Hamburg, Hamburg; Germany.
- <sup>u</sup> Also at Institute for Nuclear Research and Nuclear Energy (INRNE) of the Bulgarian Academy of Sciences, Sofia; Bulgaria.
- <sup>v</sup> Also at Institute of Applied Physics, Mohammed VI Polytechnic University, Ben Guerir; Morocco.
- <sup>w</sup> Also at Institute of Particle Physics (IPP); Canada.
- <sup>x</sup> Also at Institute of Physics and Technology, Mongolian Academy of Sciences, Ulaanbaatar; Mongolia.
- <sup>y</sup> Also at Institute of Physics, Azerbaijan Academy of Sciences, Baku; Azerbaijan.
- <sup>z</sup> Also at Institute of Theoretical Physics, Ilia State University, Tbilisi; Georgia.
- aa Also at L2IT, Université de Toulouse, CNRS/IN2P3, UPS, Toulouse; France.
- ab Also at Lawrence Livermore National Laboratory, Livermore; United States of America.
- ac Also at National Institute of Physics, University of the Philippines Diliman (Philippines); Philippines.
- ad Also at Technical University of Munich, Munich; Germany.
- ae Also at The Collaborative Innovation Center of Quantum Matter (CICQM), Beijing; China.
- af Also at TRIUMF, Vancouver BC; Canada.
- ag Also at Università di Napoli Parthenope, Napoli; Italy.
- ah Also at University of Chinese Academy of Sciences (UCAS), Beijing; China.
- ai Also at University of Colorado Boulder, Department of Physics, Colorado; United States of America.
- aj Also at Washington College, Chestertown, MD; United States of America.
- ak Also at Yeditepe University, Physics Department, Istanbul; Türkiye.
- \* Deceased