Signal Evaluation & CAPA Record

Product code **LLZ** — System, Image Processing, Radiological · LLZ report spike: evaluate for corrective action

APPROVED FOR CAPA INITIATION by Demo Reviewer — open items in §4 remain with the CAPA owner

Document control

Document ID	**SER-LLZ-20260831-001**	**Revision**	1.0
Status	Approved for CAPA initiation (CAPA open)	**Generated (UTC)**	2026-10-28 01:55:46
Effective date	2026-10-28	**CAPA ID**	CAPA-DRAFT-0CCA1700
Signal as-of	2026-08-31	**Source data**	openFDA MAUDE device events · demo.db (latest openFDA date_received 20260831)

Device

Product code / classification	LLZ — System, Image Processing, Radiological (21 CFR 892.2050, Class 2, per openFDA)
Device name(s) in reports	IntelliSpace Cardiovascular
510(k) number(s) in reports	K153022 (openFDA `pma_pmn_number`)
Manufacturer(s)	Philips Medical Systems Nederland B.V., Best, NL

Approvals (signature manifestation modeled on 21 CFR 11.50: printed name, date/time, meaning)

Role	Printed name	Date / time (UTC)	Meaning	Audit reference
Prepared by	`copilot:muse-spark`	2026-10-28 01:55:05	Authorship (software draft)	capa.draft entry in Appendix
QA/RA reviewer	**Demo Reviewer**	2026-10-28 01:55:46	**Approved**	`seq 2 · a78ab36853d0`
Reviewer comment	Demo approval: open CAPA and assign owner.			
CAPA owner (acceptance)	_____	_____	Responsibility	—

Not a 21 CFR Part 11 signature. The reviewer name is recorded by the Actions page without authenticated identity (no 11.100/11.200/11.300 controls, no 11.70 signature-to-record linking, system not validated per 11.10(a)). The block is *modeled on* 11.50 so a human-readable record shows who decided, when, and what the decision meant; a manufacturer must re-execute approval in its validated eQMS.

Number integrity. 16 distinct figures printed in this record were recomputed from stored openFDA records by the claim-level gate at generation: 16 matched, 0 did not. Each figure carries its claim in the source HTML (`data-claim`) for independent re-verification.

Contents

1. Document control & approvals 1
2. Signal evaluation 2
3. Evidence — supporting MAUDE reports 3
4. CAPA record (ISO 13485:2016 §8.5.2 / §8.5.3) 5
5. MDR assessment worksheet (Form 3500A map) 7
6. Narrative summary 9
Appendix A. Audit trail 10

2. Signal evaluation

2.1 Method

Deterministic detector `maudesignal.monitor.deterioration` over stored openFDA MAUDE reports, anchored on FDA *date_received*. Current window: the 7 days ending on the as-of date (2026-08-25 to 2026-08-31). Baseline: the 4 weeks before it (2026-07-28 to 2026-08-24). Rules (several may fire; precedence new_signal > severity_escalation > spike):

spike	current-window count ≥ 2× the baseline weekly mean **and** ≥ 3 reports
severity_escalation	Injury + Death share of the current window exceeds the baseline share by ≥ 15 percentage points (baseline non-empty)
new_signal	first-ever reports for the product code, or an event type never seen before the current window

2.2 Result: SPIKE (rules fired: spike)

Metric	Current window	Baseline (4 weeks)
Reports received	**33**	**35**
Weekly mean (baseline ÷ 4)	—	**8.75**
Spike threshold (2 × weekly mean)	**17.50**	—
Ratio current ÷ weekly mean	**3.77×**	—
Injury + Death reports	**0**	**1**
Injury + Death share	**0.0%**	**2.9%**
Event type “Malfunction” (current window)	**33**	—

2.3 Weekly report counts




Figure 1 — Reports per week by FDA date_received (B1–B4 = baseline weeks). Deterministic chart drawn from the stored records; not an AI-generated image.

> A spike is a statistical signal for human evaluation. It does not establish a device nonconformity, a root cause, or MDR reportability.

3. Evidence — supporting MAUDE reports

33 reports cited by CAPA-DRAFT-0CCA1700. Narrative excerpts are verbatim from openFDA (first 300 characters, whitespace collapsed). Report IDs link to the FDA MAUDE record.

#	MAUDE report	Mfr report no.	Received	Event type	Brand	Narrative excerpt (first 300 chars)
1	50426770	3003768277-2026-101324	2026-08-25	Malfunction	IntelliSpace Cardiovascular	Philips received a complaint regarding IntelliSpace Cardiovascular (ISCV), indicating that they have a case that needs to be corrected. The device was in clinical use at the time of the event, and no adverse event or patient harm was reported (PR (b)(4)).Philips has started an investigation of this…
2	50426773	3003768277-2026-101323	2026-08-25	Malfunction	IntelliSpace Cardiovascular	Philips received a complaint regarding IntelliSpace Cardiovascular (ISCV), indicating that unable to finalize report on Echo Module and encounter an error 500. The device was in clinical use at the time of the event, and no adverse event or patient harm was reported (PR (b)(4)).Philips has started…
3	50448213	3003768277-2026-101328	2026-08-26	Malfunction	IntelliSpace Cardiovascular	Philips received a complaint regarding IntelliSpace Cardiovascular (ISCV), indicating that an error 500 occurs for one study, preventing the doctor from processing the report. The device was in clinical use at the time of the event, and no adverse event or patient harm was reported (PR (b)(4)).Phil…
4	50448215	3003768277-2026-101327	2026-08-26	Malfunction	IntelliSpace Cardiovascular	Philips received a complaint on the IntelliSpace Cardiovascular (ISCV) indicating that customer reports intermittent HTTP 500 Internal Server Errors affecting all users and all study types. The device was in clinical use at the time of the event, and no adverse events or harm were reported in the c…
5	50448216	3003768277-2026-101325	2026-08-26	Malfunction	IntelliSpace Cardiovascular	Philips received a complaint regarding IntelliSpace Cardiovascular (ISCV), indicating that user experiencing connectivity error with the workstation. The device was in clinical use at the time of the event, and no adverse event or patient harm was reported (PR (b)(4)). Philips has started an invest…
6	50448217	3003768277-2026-101326	2026-08-26	Malfunction	IntelliSpace Cardiovascular	Philips received a complaint regarding IntelliSpace Cardiovascular (ISCV), indicating that failure to transfer the report to MedArchiver. The device was in clinical use at the time of the event, and no adverse event or patient harm was reported (PR (b)(4)).Philips has started an investigation of th…
7	50451032	3003768277-2026-101329	2026-08-26	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
8	50453041	3003768277-2026-101330	2026-08-26	Malfunction	IntelliSpace Cardiovascular	Philips received a complaint on the IntelliSpace Cardiovascular (ISCV) indicating that customer facing issues with the Copy study feature. The device was in clinical use at the time of the event, and no adverse events or harm were reported in the complaint ((b)(4)).Philips has started an investigat…
9	50502511	3003768277-2026-101334	2026-08-28	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
10	50502514	3003768277-2026-101335	2026-08-28	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
11	50502888	3003768277-2026-101337	2026-08-28	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
12	50503657	3003768277-2026-101341	2026-08-28	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
13	50503658	3003768277-2026-101339	2026-08-28	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)

#	MAUDE report	Mfr report no.	Received	Event type	Brand	Narrative excerpt (first 300 chars)
14	50503662	3003768277-2026-101340	2026-08-28	Malfunction	IntelliSpace Cardiovascular	Philips identified a software defect in ISCV (IntelliSpace Cardiovascular) system wherein the user was unable to finalize the report. The device was in clinical use at the time of event.Based on the new information that Philips became aware, this has been determined to be a reportable malfunction….
15	50504539	3003768277-2026-101343	2026-08-28	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
16	50504540	3003768277-2026-101342	2026-08-28	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
17	50504541	3003768277-2026-101344	2026-08-28	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
18	50504545	3003768277-2026-101345	2026-08-28	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
19	50513025	3003768277-2026-101346	2026-08-30	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
20	50513122	3003768277-2026-101336	2026-08-30	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
21	50515017	3003768277-2026-101347	2026-08-30	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
22	50515470	3003768277-2026-101349	2026-08-31	Malfunction	IntelliSpace Cardiovascular	Philips identified a software defect in ISCV (IntelliSpace Cardiovascular) system wherein unable to fetch images for some records. The device was in clinical use at the time of event.Based on the new information that Philips became aware, this has been determined to be a reportable malfunction. Und…
23	50516576	3003768277-2026-101352	2026-08-31	Malfunction	IntelliSpace Cardiovascular	Philips identified a software defect in ISCV (Intelli Space Cardiovascular) system wherein the user could not log into system. The device was in clinical use at the time of event. Based on the new information that Philips became aware, this has been determined to be a reportable malfunction. Under…
24	50516578	3003768277-2026-101351	2026-08-31	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
25	50516613	3003768277-2026-101350	2026-08-31	Malfunction	IntelliSpace Cardiovascular	Philips identified a software defect in ISCV (IntelliSpace Cardiovascular) system wherein the user was unable to retrieve images from Vue Picture Archiving and Communication System (PACS). The device was in clinical use at the time of event.Based on the new information that Philips became aware, th…
26	50517192	3003768277-2026-101353	2026-08-31	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
27	50517935	3003768277-2026-101354	2026-08-31	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
28	50519907	3003768277-2026-101357	2026-08-31	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
29	50519908	3003768277-2026-101355	2026-08-31	Malfunction	IntelliSpace CardioVascular	(no narrative text in the openFDA record)
30	50519912	3003768277-2026-101356	2026-08-31	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
31	50519913	3003768277-2026-101358	2026-08-31	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
32	50531298	3003768277-2026-101360	2026-08-31	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)
33	50531301	3003768277-2026-101361	2026-08-31	Malfunction	IntelliSpace Cardiovascular	(no narrative text in the openFDA record)

4. CAPA record (ISO 13485:2016 §8.5.2 / §8.5.3, incorporated by 21 CFR Part 820 — QMSR)

CAPA ID	CAPA-DRAFT-0CCA1700	**Status**	APPROVED
Title	LLZ report spike: evaluate for corrective action		
Opened (UTC)	2026-10-28 01:55:05	**Opened by**	`copilot:muse-spark`
Source	Postmarket surveillance signal SER-LLZ-20260831-001 (§2) · complaint/MDR data (§3) · MDR worksheet (§5)		

Under the QMSR (effective 2026-02-02) CAPA requirements are those of ISO 13485:2016 §8.5.2 and §8.5.3; the former 21 CFR 820.100 is superseded and is not cited.

4.1 (a) Review of nonconformities, including complaints

Problem statement as drafted (copilot:muse-spark, recorded verbatim from the CAPA draft):

> The overnight monitor classified product code LLZ as 'spike' for the week ending 2026-08-31. The cited MAUDE reports from that week are attached for complaint review.

Complaint/MDR input reviewed: **33** MAUDE reports (§3); signal classification **spike** (§2).

4.2 (b) Determination of the cause(s) of nonconformity

MaudeSignal does not determine root cause.

To be completed by CAPA owner

_____

_____

_____

_____

Name: _____ Date (UTC): _____

4.3 (c) Evaluation of the need for action to ensure nonconformities do not recur

Input: deterministic rule fired — spike. Need-for-action decision and risk assessment rationale:

To be completed by CAPA owner

_____

_____

_____

Name: _____ Date (UTC): _____

4.4 (d) Planning, documenting and implementing action, including updating documentation

To be completed by CAPA owner

_____

_____

_____

Name: _____ Date (UTC): _____

4.5 (e) Verification that the action does not adversely affect the ability to meet regulatory requirements or the safety and performance of the device

To be completed by CAPA owner

Name: _____ Date (UTC): _____

4.6 (f) Review of the effectiveness of the corrective action

Effectiveness criteria, monitoring period and result:

To be completed by CAPA owner

Name: _____ Date (UTC): _____

4.7 Records of results of investigation and action (§8.5.2, final paragraph)

This document, its revision history and the audit trail in Appendix A. Reporting to management review: required — schedule with management review.

4.8 Preventive action (§8.5.3) — potential nonconformities in similar devices / product codes

To be completed by CAPA owner

Name: _____ Date (UTC): _____

4.9 Complaint handling link (§8.2.2; 21 CFR 820.35(a))

Complaint records for events reportable under Part 803 must hold device name, date received, UDI/identifiers, complainant contact, nature and details, correction/corrective action and any reply. MaudeSignal supplies the public MAUDE view only; link the manufacturer's complaint file numbers here:

To be completed by CAPA owner

5. MDR assessment worksheet (FDA Form 3500A map)

DRAFT WORKSHEET — NOT FOR SUBMISSION

Reportability is a human regulatory decision.

5.1 Reporting clock (advisory — 21 CFR 803.50 / 803.53)

Awareness date (803.3 / 803.50)	**2026-10-28** (assumed = CAPA draft creation date (demo); confirm)
30-calendar-day report due (803.50(a))	2026-11-27
5-work-day report due, if triggered (803.53(a) remedial action / (b) FDA request)	2026-11-04

Dates computed by `copilot.tools.deadline_check` (same code as the copilot). A clock only — it does not decide that any event is reportable.

5.2 Reportability determination (human)

☐	803.50(a)(1): information reasonably suggests the device may have caused or contributed to a death or serious injury
☐	803.50(a)(2): the device malfunctioned and would be likely to cause or contribute to a death or serious injury if the malfunction recurred
☐	803.53(a): remedial action needed to prevent an unreasonable risk of substantial harm (5-day report) · ☐ 803.53(b): FDA written request
☐	Not reportable — rationale documented in the MDR event file (803.18(b)(1)(i))

To be completed by CAPA owner

_____

_____

Name: _____ Date (UTC): _____

5.3 Form 3500A fields — pre-filled from openFDA where available

One 3500A is filed per reportable event. Pre-fills are the values common to the cited reports, from openFDA; blank cells require manufacturer information (21 CFR 803.52: “if known or reasonably known”).

Sec.	Item	Field	803.52	Pre-fill	Source
A	A1–A5	Patient identifier, age, sex, weight, race/ethnicity	803.52(a)	Not pre-filled — MaudeSignal handles no patient-identifying information	—
B	B1	Type of report (adverse event / product problem …)	803.52(b)	_____	—
B	B2	Outcomes attributed to adverse event	803.52(b)	_____	—
B	B3	Date of event	803.52(b)	Per report — see Evidence table	—
B	B5	Describe event or problem	803.52(b)	Per report — MAUDE narrative excerpts in Evidence table	mdr_text
B	B6–B7	Relevant tests/laboratory data; other relevant history	803.52(b)	_____	—
D	D1	Brand name	803.52(c)	IntelliSpace Cardiovascular	openFDA device event: device.brand_name
D	D2a	Product code (procode)	803.52(c)	LLZ	openFDA device event: device.device_report_product_code

DRAFT WORKSHEET NOT FOR SUBMISSION

Sec.	Item	Field	803.52	Pre-fill	Source
D	D2b	Common device name	803.52(c)	Cardiovascular Image and Information Management System	openFDA device event: device.generic_name
D	D3	Manufacturer name, city, state, country	803.52(c)	Philips Medical Systems Nederland B.V., Best, NL	openFDA device event: device.manufacturer_d_*
D	D4	Model #, lot #, catalog #, serial #, UDI	803.52(c)	Model: IntelliSpace CardioVascular	openFDA device event: device.model_number
D	D5–D10	Operator, implant/explant dates, reprocessing, 3rd-party service, availability for evaluation, concomitant products	803.52(c)	_____	—
E	E1–E4	Initial reporter (name, address, health professional, occupation)	803.52(d)	_____	—
F	F1–F13	User facility / importer block	803.52(f)(11)	Not applicable to a manufacturer report unless correcting UF/importer data	—
G	G1	Contact office and manufacturing site	803.52(e)	_____	—
G	G2	Report source	803.52(e)	_____	—
G	G3	Date received by manufacturer (awareness)	803.52(e)	See awareness date above (human entry)	user input
G	G4	PMA / 510(k) number	803.52(e)	K153022	openFDA device event: pma_pmn_number
G	G6	Type of report (initial / follow-up / 5-day …)	803.52(e)	_____	—
G	G7	Adverse event term(s)	803.52(e)	_____	—
G	G8	Manufacturer report number	803.52(e)	_____	—
H	H1	Type of reportable event (death / serious injury / malfunction)	803.52(f)	openFDA event_type observed: Malfunction	openFDA device event: event_type (reference only)
H	H3	Device evaluated by manufacturer? (summary or reason not)	803.52(f)	_____	—
H	H6	Adverse event problem codes (device, patient, evaluation)	803.52(f)	_____	—
H	H7	Remedial action initiated (type)	803.52(f)	_____	—
H	H9	Correction/removal report number (21 USC 360i(g))	803.52(f)	_____	—
H	H10	Additional manufacturer narrative	803.52(f)	_____	—
—	FDA	Date received by FDA (MAUDE)	—	2026-08-25 to 2026-08-31	openFDA device event: date_received

MDR procedures and event files: 21 CFR 803.17 (written procedures) and 803.18 (event files incl. the reportability deliberation; retain the longer of 2 years or the expected device life).

DRAFT WORKSHEET NOT FOR SUBMISSION

6. Narrative summary

> AI-drafted (recorded run, 2026-10-28T01:55:10Z). Drafted by muse-spark-1.3 (skill dossier-narrative v1.0.0); every sentence was re-checked by the deterministic claim-level gate against stored openFDA records. Withheld sentences are shown struck through with the gate's reason. Shown: 1 · withheld: 0.

VERIFIED

Appendix A. Audit trail

Entries of the append-only, SHA-256 hash-chained `action_audit` table for CAPA-DRAFT-0CCA1700 and SER-LLZ-20260831-001. Time stamps are computer-generated (UTC).

Seq	Time (UTC)	Actor	Action	Entity	Entry hash (prefix)
1	2026-10-28 01:55:05	`copilot:muse-spark`	`capa.draft`	`CAPA-DRAFT-0CCA1700`	`723dde8e8667`
2	2026-10-28 01:55:46	`Demo Reviewer`	`capa.approved`	`CAPA-DRAFT-0CCA1700`	`a78ab36853d0`

> **Chain verification** (`Database.verify_audit_chain()`, all entries recomputed): INTACT.

> Modeled on 21 CFR 11.10(e) (secure, computer-generated, time-stamped audit trails). Not claimed as Part 11 compliant: the system is not validated, has no user authentication, and a database administrator could rewrite the whole chain consistently; the hash chain detects edits to individual entries only.

Generated by MaudeSignal 0.1.0 (maudesignal.dossier). Regulatory sources: docs/dossier_sources.md.