



Certificate of Analysis

REPORTED September 24, 2026

CLIENT

VERIFY peptidetestcanada.ca/verify

EMAIL info@peptidetestcanada.ca

SCAN TO VERIFY



[PTC-JSU7DRL9](#)

Lab ID: **BLD - 295**

END CUSTOMER

NAME Aura Peptides

§ 01 Sample Information

FOR RESEARCH PURPOSES ONLY

Sample ID:	BLD-295	Received:	09/11/2026
Compound:	Wolverine BPC-157 / TB-500 Blend	Analyzed:	09/24/2026
Report No.:	COA-2026-SC-01666	Components:	2 peptides
Lot Number:	BB20-2608-003	Total Claim:	20.00 mg
Total Mass:	81.0 mg	Total Net:	24.46 mg
Appearance:	Lyophilized powder		

SAMPLE PHOTOGRAPH



§ 02 Analytical Results

Qualitative and Quantitative chemical analysis by Liquid Chromatography Tandem Mass Spectrometry (HPLC-MS/MS)
Instrument · Agilent 1290 Infinity II with Agilent 6495 QQQ

PEPTIDE	LABEL CLAIM	NET CONTENT	CHROM. PURITY	FILL ACC.	STATUS
BPC-157	10.00 mg	12.44 mg	99.64 %	124.4 %	CONFIRMED
TB-500	10.00 mg	12.03 mg	99.61 %	120.3 %	CONFIRMED
TOTAL NET CONTENT	20.00 mg	24.46 mg	—	122.3 %	

BLEND COMPOSITION share of total net peptide content




PRECURSOR [M+2H]²⁺

710.40 Da

Quantifier transition ● MRM

710.40 > 70.10 CE 22 V



ROLE	TRANSITION	CE
● Quantifier	710.4 > 70.1	22 V


Agilent Poroshell 120 EC-C18 · 2.1×100 mm · 2.7 μm · Solvent A: 0.1 % formic acid in H₂O · Solvent B: 0.1 % formic acid in Acetonitrile

PRECURSOR [M+2H]²⁺

445.30 Da

Quantifier transition ● MRM

445.30 > 241.20 CE 12 V



ROLE	TRANSITION	CE
● Quantifier	445.3 > 241.2	12 V

Agilent Poroshell 120 EC-C18 · 2.1×100 mm · 2.7 μm · Solvent A: 0.1 % formic acid in H₂O · Solvent B: 0.1 % formic acid in Acetonitrile

DISCLAIMER · RESEARCH USE ONLY
 Results relate **only to the sample as received** by Side-Chain Analytics Inc. and the tests performed, as and when received — not to any other unit, vial, lot, or batch, and not to any other product bearing the same name, label, or description. This report is not a verification of any lot or batch. **For research use only:** this report is not a certificate of quality, safety, efficacy, or regulatory compliance; the material tested is not authorized by Health Canada for human or veterinary use; and this report is not for human or veterinary administration, clinical or diagnostic use, or label claims. Side-Chain Analytics Inc. does not manufacture, source, or sell the material tested and makes **no warranty, express or implied**; its liability shall not exceed the testing fee. This report may not be reproduced except in full, and may not be reproduced, displayed, quoted, or referenced — in whole or in part, including by any badge, seal, or link — in connection with the advertising, offer for sale, sale, or distribution of any product. This certificate is valid only as currently shown at sidechainanalytics.com/verify and may be withdrawn.